


Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG061020\
 Data File : BG045686.D
 Acq On : 10 Jun 2020 11:14
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG051520.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jun 01 15:07:00 2020



AutoFind: Scans 2501, 2502, 2503; Background Corrected with Scan 2494

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	28.0	16173	PASS
68	69	0.00	2	0.6	120	PASS
69	198	0.00	100	35.1	20232	PASS
70	69	0.00	2	0.3	63	PASS
127	198	10	80	43.1	24856	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	57712	PASS
199	198	5	9	6.3	3656	PASS
275	198	10	60	31.5	18204	PASS
365	198	1	100	5.8	3345	PASS
441	198	0.01	100	16.6	9573	PASS
442	198	50	100	99.2	57237	PASS
443	442	15	24	19.1	10940	PASS