

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG061215\
 Data File : BG017663.D
 Acq On : 13 Jun 2015 7:48
 Operator : TP/IZ
 Sample : G2522-02MS
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 MW-19MS

Quant Time: Jun 15 04:30:40 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG061215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jun 13 02:22:44 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.52	152	19615	20.00	ng	0.00
21) Naphthalene-d8	10.32	136	75074	20.00	ng	0.00
38) Acenaphthene-d10	14.21	164	61328	20.00	ng	0.00
63) Phenanthrene-d10	16.96	188	172777	20.00	ng	0.00
75) Chrysene-d12	21.16	240	253941	20.00	ng	0.00
86) Perylene-d12	23.36	264	243575	20.00	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
5) 2-Fluorophenol	5.12	112	76126	75.16	ng	0.00
7) Phenol-d6	6.74	99	69748	47.00	ng	0.00
23) Nitrobenzene-d5	8.69	82	181431	101.44	ng	0.00
41) 2,4,6-Tribromophenol	15.71	330	173365	163.77	ng	0.00
44) 2-Fluorobiphenyl	12.82	172	449690	101.41	ng	0.00
78) Terphenyl-d14	19.61	244	1153853	94.79	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	2.95	88	8299	16.45	ng	# 71
3) Pyridine	3.35	79	18713	14.33	ng	# 92
4) n-Nitrosodimethylamine	3.28	42	16942	22.56	ng	# 85
6) Aniline	6.86	93	33056	16.73	ng	# 85
8) 2-Chlorophenol	7.10	128	49620	40.36	ng	93
9) Benzaldehyde	6.67	77	2057	2.05	ng	92
10) Phenol	6.76	94	25592	16.58	ng	96
11) bis(2-Chloroethyl)ether	6.96	93	52897	46.22	ng	96
12) 1,3-Dichlorobenzene	7.41	146	55383	38.57	ng	97
13) 1,4-Dichlorobenzene	7.56	146	59252	39.17	ng	92
14) 1,2-Dichlorobenzene	7.88	146	59408	42.25	ng	97
15) Benzyl Alcohol	7.78	79	52435	32.06	ng	96
16) 2,2'-oxybis(1-Chloropropan	8.08	45	23139	50.47	ng	82
17) 2-Methylphenol	8.01	107	41608	37.20	ng	89
18) Hexachloroethane	8.59	117	25267	40.22	ng	89
19) n-Nitroso-di-n-propylamine	8.34	70	58577	44.00	ng	# 87
20) 3+4-Methylphenols	8.34	107	50678	32.92	ng	# 78
22) Acetophenone	8.35	105	103576	52.35	ng	96
24) Nitrobenzene	8.74	77	112716	58.18	ng	# 95
25) Isophorone	9.26	82	163646	48.72	ng	97
26) 2-Nitrophenol	9.44	139	36773	49.39	ng	# 89
27) 2,4-Dimethylphenol	9.53	122	54624	45.60	ng	# 85
28) bis(2-Chloroethoxy)methane	9.75	93	74322	51.08	ng	98
29) 2,4-Dichlorophenol	9.99	162	70535	55.23	ng	96
30) 1,2,4-Trichlorobenzene	10.19	180	80760	48.92	ng	# 89
31) Naphthalene	10.37	128	188697	46.57	ng	# 92
33) 4-Chloroaniline	10.50	127	31512	19.00	ng	97
34) Hexachlorobutadiene	10.66	225	61423	43.37	ng	96
35) Caprolactam	11.28	113	2028	3.74	ng	# 78
36) 4-Chloro-3-methylphenol	11.66	107	79243	48.52	ng	90
37) 2-Methylnaphthalene	12.00	142	157373	48.58	ng	# 89
39) 1,2,4,5-Tetrachlorobenzene	12.38	216	126042	51.05	ng	97
40) Hexachlorocyclopentadiene	12.35	237	88027	81.41	ng	91
42) 2,4,6-Trichlorophenol	12.64	196	76045	49.66	ng	97


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG061215\
 Data File : BG017663.D
 Acq On : 13 Jun 2015 7:48
 Operator : TP/IZ
 Sample : G2522-02MS
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 MW-19MS

Quant Time: Jun 15 04:30:40 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG061215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jun 13 02:22:44 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) 2,4,5-Trichlorophenol	12.73	196	77554	48.42	ng	95
45) 1,1'-Biphenyl	13.03	154	232792	51.76	ng	97
46) 2-Chloronaphthalene	13.07	162	181674	48.69	ng	97
47) 2-Nitroaniline	13.30	65	67587	51.29	ng	97
48) Acenaphthylene	13.93	152	286239	45.13	ng	98
49) Dimethylphthalate	13.68	163	271172	51.64	ng	98
50) 2,6-Dinitrotoluene	13.80	165	58851	51.11	ng	97
51) Acenaphthene	14.27	154	181624	48.22	ng	87
52) 3-Nitroaniline	14.14	138	16553	16.09	ng	# 82
53) 2,4-Dinitrophenol	14.35	184	77409	98.61	ng	94
54) Dibenzofuran	14.61	168	313704	50.44	ng	99
55) 4-Nitrophenol	14.49	139	26608	41.08	ng	96
56) 2,4-Dinitrotoluene	14.60	165	89361	53.47	ng	# 98
57) Fluorene	15.26	166	267647	50.16	ng	95
58) 2,3,4,6-Tetrachlorophenol	14.85	232	87018	49.90	ng	97
59) Diethylphthalate	15.05	149	293226	50.32	ng	99
60) 4-Chlorophenyl-phenylether	15.26	204	176747	52.90	ng	92
61) 4-Nitroaniline	15.30	138	26989	24.61	ng	93
62) Azobenzene	15.55	77	318901	50.42	ng	96
64) 4,6-Dinitro-2-methylphenol	15.36	198	63981	46.59	ng	92
65) n-Nitrosodiphenylamine	15.48	169	249012	49.46	ng	98
66) 4-Bromophenyl-phenylether	16.16	248	119726	52.10	ng	96
67) Hexachlorobenzene	16.27	284	126162	50.45	ng	# 89
68) Atrazine	16.44	200	113974	45.91	ng	97
69) Pentachlorophenol	16.63	266	124257	105.65	ng	96
70) Phenanthrene	17.00	178	461303	47.84	ng	97
71) Anthracene	17.10	178	472136	49.31	ng	97
72) Carbazole	17.38	167	441234	50.26	ng	100
73) Di-n-butylphthalate	17.94	149	564659	52.56	ng	99
74) Fluoranthene	19.04	202	697206	50.54	ng	99
76) Benzidine	19.23	184	51481	8.67	ng	99
77) Pyrene	19.40	202	728380	50.80	ng	98
79) Butylbenzylphthalate	20.30	149	266180	51.43	ng	96
80) Benzo(a)anthracene	21.15	228	764210	48.78	ng	100
81) 3,3'-Dichlorobenzidine	21.09	252	87238	15.56	ng	# 96
82) Chrysene	21.20	228	709735	50.23	ng	98
83) Bis(2-ethylhexyl)phthalate	21.09	149	388427	51.08	ng	95
84) Di-n-octyl phthalate	21.94	149	646336	51.02	ng	100
85) Indeno(1,2,3-cd)pyrene	25.60	276	879880	47.88	ng	99
87) Benzo(b)fluoranthene	22.71	252	777011	49.49	ng	99
88) Benzo(k)fluoranthene	22.75	252	739834	48.98	ng	99
89) Benzo(a)pyrene	23.27	252	730689	50.39	ng	99
90) Dibenzo(a,h)anthracene	25.60	278	748259	49.87	ng	# 98
91) Benzo(g,h,i)perylene	26.28	276	701255	49.14	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed







#1

1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.52 min Scan# 822
Delta R.T. -0.01 min
Lab File: BG017663.D
Acq: 13 Jun 2015 7:48

Instrument :
BNA_G
ClientSampleId :
MW-19MS




Tgt Ion:152 Resp: 19615
Ion Ratio Lower Upper
152 100
150 126.6 119.4 179.0
115 65.4 58.2 87.2




#2

1,4-Dioxane
Concen: 16.45 ng
RT: 2.95 min Scan# 44
Delta R.T. -0.00 min
Lab File: BG017663.D
Acq: 13 Jun 2015 7:48

Tgt Ion: 88 Resp: 8299
Ion Ratio Lower Upper
88 100
58 50.0 29.6 44.4#
43 42.2 18.3 27.5#





#3

Pyridine

Concen: 14.33 ng

RT: 3.35 min Scan# 113

Delta R.T. -0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48

Instrument :

BNA_G

ClientSampleId :

MW-19MS


Tgt Ion: 79 Resp: 18713

Ion Ratio Lower Upper

79 100

52 29.3 24.7 37.1




51 35.9 23.3 34.9#



Abundance Ion 79.00 (78.70 to 79.70): BGC

Ion 52.00 (51.70 to 52.70): BGC

Ion 51.00 (50.70 to 51.70): BGC



#4

n-Nitrosodimethylamine

Concen: 22.56 ng

RT: 3.28 min Scan# 100

Delta R.T. 0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48


Tgt Ion: 42 Resp: 16942

Ion Ratio Lower Upper

42 100

74 78.5 74.9 112.3



44 2.3 1.5 2.3#




Abundance Ion 42.00 (41.70 to 42.70): BGC

Ion 74.00 (73.70 to 74.70): BGC

Ion 44.00 (43.70 to 44.70): BGC





#5

2-Fluorophenol

Concen: 75.16 ng

RT: 5.12 min Scan# 413

Delta R.T. 0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48

Instrument :

BNA_G

ClientSampleId :

MW-19MS


Tgt Ion: 112 Resp: 76126

Ion Ratio Lower Upper

112 100

64 57.9 47.0 70.6




63 47.6 38.6 58.0



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BGC

Ion 63.00 (62.70 to 63.70): BGC



#6

Aniline

Concen: 16.73 ng

RT: 6.86 min Scan# 709

Delta R.T. -0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48


Tgt Ion: 93 Resp: 33056

Ion Ratio Lower Upper

93 100

66 45.7 30.1 45.1#



65 29.1 17.0 25.4#




Abundance Ion 93.00 (92.70 to 93.70): BGC

Ion 66.00 (65.70 to 66.70): BGC

Ion 65.00 (64.70 to 65.70): BGC





#7

Phenol-d6

Concen: 47.00 ng

RT: 6.74 min Scan# 689

Delta R.T. 0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48

Instrument :

BNA_G

ClientSampleId :

MW-19MS





Tgt Ion: 99 Resp: 69748

Ion Ratio Lower Upper

99 100

42 19.8 17.4 26.2

71 60.6 47.4 71.0



#8

2-Chlorophenol

Concen: 40.36 ng

RT: 7.10 min Scan# 750

Delta R.T. -0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48




Tgt Ion: 128 Resp: 49620


Ion Ratio Lower Upper

128 100

130 36.3 8.8 48.8

64 53.6 31.1 71.1





#9

Benzaldehyde

Concen: 2.05 ng

RT: 6.67 min Scan# 678

Delta R.T. 0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48

Instrument :

BNA_G

ClientSampleId :

MW-19MS


Tgt Ion: 77 Resp: 2057

Ion Ratio Lower Upper

77 100

105 79.2 62.3 102.3




106 69.9 60.7 100.7



Abundance Ion 77.00 (76.70 to 77.70): BGC

Ion 105.00 (104.70 to 105.70): BGC

Ion 106.00 (105.70 to 106.70): BGC



#10

Phenol

Concen: 16.58 ng

RT: 6.76 min Scan# 693

Delta R.T. 0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48


Tgt Ion: 94 Resp: 25592

Ion Ratio Lower Upper

94 100

65 40.8 16.5 56.5



66 62.4 40.6 80.6




Abundance Ion 94.00 (93.70 to 94.70): BGC

Ion 65.00 (64.70 to 65.70): BGC

Ion 66.00 (65.70 to 66.70): BGC









#11
bis(2-Chloroethyl)ether
Concen: 46.22 ng
RT: 6.96 min Scan# 727
Delta R.T. 0.00 min
Lab File: BG017663.D
Acq: 13 Jun 2015 7:48




Instrument :
BNA_G
ClientSampleId :
MW-19MS


Tgt Ion	Ratio	Lower	Upper
93	100		
63	57.1	40.5	80.5
95	32.5	11.0	51.0



#12
1,3-Dichlorobenzene
Concen: 38.57 ng
RT: 7.41 min Scan# 804
Delta R.T. 0.00 min
Lab File: BG017663.D
Acq: 13 Jun 2015 7:48

Tgt Ion	Ratio	Lower	Upper
146	100		
148	61.4	48.6	73.0
75	52.4	45.7	68.5









#13
 1,4-Dichlorobenzene
 Concen: 39.17 ng
 RT: 7.56 min Scan# 829
 Delta R.T. -0.00 min
 Lab File: BG017663.D
 Acq: 13 Jun 2015 7:48




Instrument :
 BNA_G
 ClientSampleId :
 MW-19MS


Tgt Ion:146 Resp: 59252
 Ion Ratio Lower Upper
 146 100
 148 62.6 46.3 69.5
 111 41.1 37.9 56.9



#14
 1,2-Dichlorobenzene
 Concen: 42.25 ng
 RT: 7.88 min Scan# 883
 Delta R.T. 0.00 min
 Lab File: BG017663.D
 Acq: 13 Jun 2015 7:48

Tgt Ion:146 Resp: 59408
 Ion Ratio Lower Upper
 146 100
 148 56.1 43.8 65.8
 111 48.2 36.3 54.5





#15

Benzyl Alcohol

Concen: 32.06 ng

RT: 7.78 min Scan# 866

Delta R.T. -0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48

Instrument :

BNA_G

ClientSampleId :

MW-19MS


Tgt Ion: 79 Resp: 52435

Ion Ratio Lower Upper

79 100

108 48.9 44.6 67.0




77 66.3 53.1 79.7



Abundance Ion 79.00 (78.70 to 79.70): BGC

Ion 108.00 (107.70 to 108.70): BGC

Ion 77.00 (76.70 to 77.70): BGC



#16

2,2'-oxybis(1-Chloropropane)

Concen: 50.47 ng

RT: 8.08 min Scan# 917

Delta R.T. 0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48


Tgt Ion: 45 Resp: 23139

Ion Ratio Lower Upper

45 100

77 38.3 33.8 73.8



79 37.1 26.2 66.2




Abundance Ion 45.00 (44.70 to 45.70): BGC

Ion 77.00 (76.70 to 77.70): BGC

Ion 79.00 (78.70 to 79.70): BGC









#17
2-Methylphenol
Concen: 37.20 ng
RT: 8.01 min Scan# 905
Delta R.T. 0.00 min
Lab File: BG017663.D
Acq: 13 Jun 2015 7:48




Instrument :
BNA_G
ClientSampleId :
MW-19MS


Tgt Ion:107 Resp: 41608
Ion Ratio Lower Upper
107 100
108 103.2 91.6 137.4
77 72.8 63.2 94.8
79 70.5 67.8 101.8



#18
Hexachloroethane
Concen: 40.22 ng
RT: 8.59 min Scan# 1005
Delta R.T. 0.00 min
Lab File: BG017663.D
Acq: 13 Jun 2015 7:48

Tgt Ion:117 Resp: 25267
Ion Ratio Lower Upper
117 100
119 102.3 72.8 109.2
201 124.8 91.0 136.6









#19
n-Nitroso-di-n-propylamine
Concen: 44.00 ng
RT: 8.34 min Scan# 962
Delta R.T. -0.00 min
Lab File: BG017663.D
Acq: 13 Jun 2015 7:48




Instrument :
BNA_G
ClientSampleId :
MW-19MS


Tgt Ion: 70 Resp: 58577
Ion Ratio Lower Upper
70 100
42 50.8 31.6 47.4#
101 14.2 6.6 9.8#
130 26.2 19.8 29.6



#20
3+4-Methylphenols
Concen: 32.92 ng
RT: 8.34 min Scan# 961
Delta R.T. 0.00 min
Lab File: BG017663.D
Acq: 13 Jun 2015 7:48

Tgt Ion: 107 Resp: 50678
Ion Ratio Lower Upper
107 100
108 91.9 66.0 106.0
77 97.6 39.6 79.6#
79 48.6 21.3 61.3









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.32 min Scan# 1299
Delta R.T. 0.00 min
Lab File: BG017663.D
Acq: 13 Jun 2015 7:48




Instrument :
BNA_G
ClientSampleId :
MW-19MS


Tgt Ion:	136	Resp:	75074
Ion	Ratio	Lower	Upper
136	100		
137	11.2	8.7	13.1
54	4.6	4.7	7.1#
68	4.4	2.4	3.6#



#22
Acetophenone
Concen: 52.35 ng
RT: 8.35 min Scan# 964
Delta R.T. -0.00 min
Lab File: BG017663.D
Acq: 13 Jun 2015 7:48

Tgt Ion:	105	Resp:	103576
Ion	Ratio	Lower	Upper
105	100		
71	3.7	2.8	4.2
51	29.8	21.5	32.3
120	25.8	22.2	33.4









#23
 Nitrobenzene-d5
 Concen: 101.44 ng
 RT: 8.69 min Scan# 1022
 Delta R.T. 0.00 min
 Lab File: BG017663.D
 Acq: 13 Jun 2015 7:48




Instrument :
 BNA_G
 ClientSampleId :
 MW-19MS


Tgt Ion	Ratio	Lower	Upper
82	100		
128	34.4	27.3	40.9
54	45.4	37.0	55.4



#24
 Nitrobenzene
 Concen: 58.18 ng
 RT: 8.74 min Scan# 1029
 Delta R.T. -0.00 min
 Lab File: BG017663.D
 Acq: 13 Jun 2015 7:48

Tgt Ion	Ratio	Lower	Upper
77	100		
123	33.2	28.0	42.0
65	17.4	10.5	15.7









#25
Isophorone
Concen: 48.72 ng
RT: 9.26 min Scan# 1118
Delta R.T. 0.00 min
Lab File: BG017663.D
Acq: 13 Jun 2015 7:48




Instrument :
BNA_G
ClientSampleId :
MW-19MS


Tgt Ion: 82 Resp: 163646
Ion Ratio Lower Upper
82 100
95 11.8 9.0 13.6
138 18.7 13.8 20.6



#26
2-Nitrophenol
Concen: 49.39 ng
RT: 9.44 min Scan# 1149
Delta R.T. -0.00 min
Lab File: BG017663.D
Acq: 13 Jun 2015 7:48

Tgt Ion: 139 Resp: 36773
Ion Ratio Lower Upper
139 100
109 54.8 57.3 85.9#
65 72.3 58.6 88.0









#27
 2,4-Dimethylphenol
 Concen: 45.60 ng
 RT: 9.53 min Scan# 1164
 Delta R.T. 0.00 min
 Lab File: BG017663.D
 Acq: 13 Jun 2015 7:48




Instrument :
 BNA_G
 ClientSampleId :
 MW-19MS


Tgt Ion: 122 Resp: 54624
 Ion Ratio Lower Upper
 122 100
 107 156.6 112.3 168.5
 121 64.8 41.0 61.4#



#28
 bis(2-Chloroethoxy)methane
 Concen: 51.08 ng
 RT: 9.75 min Scan# 1201
 Delta R.T. -0.00 min
 Lab File: BG017663.D
 Acq: 13 Jun 2015 7:48

Tgt Ion: 93 Resp: 74322
 Ion Ratio Lower Upper
 93 100
 95 32.3 26.2 39.4
 123 14.2 9.8 14.6









#29
 2,4-Dichlorophenol
 Concen: 55.23 ng
 RT: 9.99 min Scan# 1243
 Delta R.T. -0.00 min
 Lab File: BG017663.D
 Acq: 13 Jun 2015 7:48




Instrument :
 BNA_G
 ClientSampleId :
 MW-19MS


Tgt Ion	Ratio	Lower	Upper
162	100		
164	61.3	45.4	85.4
98	37.3	16.0	56.0



#30
 1,2,4-Trichlorobenzene
 Concen: 48.92 ng
 RT: 10.19 min Scan# 1276
 Delta R.T. -0.00 min
 Lab File: BG017663.D
 Acq: 13 Jun 2015 7:48

Tgt Ion	Ratio	Lower	Upper
180	100		
182	103.4	75.0	112.4
145	33.0	21.0	31.6









#31
Naphthalene
Concen: 46.57 ng
RT: 10.37 min Scan# 1307
Delta R.T. -0.00 min
Lab File: BG017663.D
Acq: 13 Jun 2015 7:48




Instrument :
BNA_G
ClientSampleId :
MW-19MS


Tgt Ion	Ratio	Lower	Upper
128	100		
129	11.5	7.0	10.4#
127	15.8	10.2	15.2#



#33
4-Chloroaniline
Concen: 19.00 ng
RT: 10.50 min Scan# 1329
Delta R.T. 0.00 min
Lab File: BG017663.D
Acq: 13 Jun 2015 7:48

Tgt Ion	Ratio	Lower	Upper
127	100		
129	30.7	27.4	41.2
65	43.2	33.8	50.8
92	27.4	21.6	32.4





#34

Hexachlorobutadiene

Concen: 43.37 ng

RT: 10.66 min Scan# 1356

Delta R.T. 0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48

Instrument :

BNA_G

ClientSampled :

MW-19MS


Tgt Ion:225 Resp: 61423

Ion Ratio Lower Upper

225 100

223 67.9 54.2 81.2




227 58.2 52.0 78.0



Abundance Ion 225.00 (224.70 to 225.70): E

Ion 223.00 (222.70 to 223.70): E

Ion 227.00 (226.70 to 227.70): E



#35

Caprolactam

Concen: 3.74 ng

RT: 11.28 min Scan# 1462

Delta R.T. -0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48


Tgt Ion:113 Resp: 2028

Ion Ratio Lower Upper

113 100

55 94.5 49.3 89.3#



56 54.7 45.2 85.2




Abundance Ion 113.00 (112.70 to 113.70): E

Ion 55.00 (54.70 to 55.70): BG0

Ion 56.00 (55.70 to 56.70): BG0









#36
 4-Chloro-3-methylphenol
 Concen: 48.52 ng
 RT: 11.66 min Scan# 1527
 Delta R.T. 0.00 min
 Lab File: BG017663.D
 Acq: 13 Jun 2015 7:48




Instrument :
 BNA_G
 ClientSampleId :
 MW-19MS


Tgt Ion:107 Resp: 79243
 Ion Ratio Lower Upper
 107 100
 144 22.6 17.3 25.9
 142 75.4 52.6 79.0



#37
 2-Methylnaphthalene
 Concen: 48.58 ng
 RT: 12.00 min Scan# 1584
 Delta R.T. -0.00 min
 Lab File: BG017663.D
 Acq: 13 Jun 2015 7:48

Tgt Ion:142 Resp: 157373
 Ion Ratio Lower Upper
 142 100
 141 97.0 71.8 107.8
 115 62.2 41.0 61.4#









#38
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.21 min Scan# 1960
Delta R.T. -0.00 min
Lab File: BG017663.D
Acq: 13 Jun 2015 7:48




Instrument :
BNA_G
ClientSampleId :
MW-19MS


Tgt Ion:164 Resp: 61328
Ion Ratio Lower Upper
164 100
162 89.9 75.0 112.4
160 46.5 30.3 45.5#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 51.05 ng
RT: 12.38 min Scan# 1649
Delta R.T. -0.00 min
Lab File: BG017663.D
Acq: 13 Jun 2015 7:48

Tgt Ion:216 Resp: 126042
Ion Ratio Lower Upper
216 100
214 75.6 63.2 94.8
179 18.4 15.8 23.8
108 19.6 15.8 23.8









#40
Hexachlorocyclopentadiene
Concen: 81.41 ng
RT: 12.35 min Scan# 1644
Delta R.T. -0.00 min
Lab File: BG017663.D
Acq: 13 Jun 2015 7:48

Instrument :
BNA_G
ClientSampleId :
MW-19MS

Tgt Ion: 237 Resp: 88027
Ion Ratio Lower Upper
237 100
235 58.7 46.6 86.6
272 11.9 0.0 32.3




Abundance Ion 236.80 (236.50 to 237.50): E
Ion 234.90 (234.60 to 235.60): E
Ion 271.75 (271.45 to 272.45): E








#41
2,4,6-Tribromophenol
Concen: 163.77 ng
RT: 15.71 min Scan# 2216
Delta R.T. -0.00 min
Lab File: BG017663.D
Acq: 13 Jun 2015 7:48

Tgt Ion: 330 Resp: 173365
Ion Ratio Lower Upper
330 100
332 99.1 76.6 114.8
141 43.1 34.6 51.8



Abundance Ion 329.80 (329.50 to 330.50): E
Ion 331.80 (331.50 to 332.50): E
Ion 141.00 (140.70 to 141.70): E









#42
2,4,6-Trichlorophenol
Concen: 49.66 ng
RT: 12.64 min Scan# 1693
Delta R.T. 0.00 min
Lab File: BG017663.D
Acq: 13 Jun 2015 7:48


Instrument :
BNA_G
ClientSampleId :
MW-19MS


Tgt Ion:196 Resp: 76045
Ion Ratio Lower Upper
196 100
198 104.1 85.3 127.9
200 31.6 28.2 42.2



#43
2,4,5-Trichlorophenol
Concen: 48.42 ng
RT: 12.73 min Scan# 1708
Delta R.T. 0.00 min
Lab File: BG017663.D
Acq: 13 Jun 2015 7:48

Tgt Ion:196 Resp: 77554
Ion Ratio Lower Upper
196 100
198 94.6 70.9 106.3
97 48.7 40.2 60.4
132 22.2 16.3 24.5









#44
 2-Fluorobiphenyl
 Concen: 101.41 ng
 RT: 12.82 min Scan# 1724
 Delta R.T. -0.00 min
 Lab File: BG017663.D
 Acq: 13 Jun 2015 7:48




Instrument :
 BNA_G
 ClientSampleId :
 MW-19MS


Tgt Ion	Ratio	Lower	Upper
172	100		
171	35.3	28.2	42.2
170	22.6	17.4	26.2



#45
 1,1'-Biphenyl
 Concen: 51.76 ng
 RT: 13.03 min Scan# 1760
 Delta R.T. -0.00 min
 Lab File: BG017663.D
 Acq: 13 Jun 2015 7:48

Tgt Ion	Ratio	Lower	Upper
154	100		
153	41.2	22.9	62.9
76	12.9	0.0	31.8









#46
2-Chloronaphthalene
Concen: 48.69 ng
RT: 13.07 min Scan# 1767
Delta R.T. -0.00 min
Lab File: BG017663.D
Acq: 13 Jun 2015 7:48




Instrument :
BNA_G
ClientSampleId :
MW-19MS


Tgt Ion	Ratio	Lower	Upper
162	100		
127	41.9	31.6	47.4
164	32.3	25.0	37.6



#47
2-Nitroaniline
Concen: 51.29 ng
RT: 13.30 min Scan# 1805
Delta R.T. 0.00 min
Lab File: BG017663.D
Acq: 13 Jun 2015 7:48

Tgt Ion	Ratio	Lower	Upper
65	100		
92	58.3	48.4	72.6
138	78.6	64.8	97.2





#48

Acenaphthylene

Concen: 45.13 ng

RT: 13.93 min Scan# 1913

Delta R.T. 0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48





Instrument :

BNA_G

ClientSampleId :

MW-19MS

Tgt Ion:	152	Resp:	286239
Ion	Ratio	Lower	Upper
152	100		
151	21.9	17.0	25.4
153	14.8	10.5	15.7



#49

Dimethylphthalate

Concen: 51.64 ng




RT: 13.68 min Scan# 1870


Delta R.T. -0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48

Tgt Ion:	163	Resp:	271172
Ion	Ratio	Lower	Upper
163	100		
194	7.4	5.7	8.5
164	10.2	7.1	10.7









#50
2,6-Dinitrotoluene
Concen: 51.11 ng
RT: 13.80 min Scan# 1891
Delta R.T. 0.00 min
Lab File: BG017663.D
Acq: 13 Jun 2015 7:48




Instrument :
BNA_G
ClientSampleId :
MW-19MS


Tgt Ion:165 Resp: 58851
Ion Ratio Lower Upper
165 100
63 88.5 70.2 105.4
89 91.7 69.8 104.8



#51
Acenaphthene
Concen: 48.22 ng
RT: 14.27 min Scan# 1971
Delta R.T. -0.00 min
Lab File: BG017663.D
Acq: 13 Jun 2015 7:48

Tgt Ion:154 Resp: 181624
Ion Ratio Lower Upper
154 100
153 126.0 87.2 130.8
152 55.8 41.0 61.4





#52

3-Nitroaniline

Concen: 16.09 ng

RT: 14.14 min Scan# 1948

Delta R.T. 0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48

Instrument :

BNA_G

ClientSampleId :

MW-19MS





Tgt Ion:138 Resp: 16553

Ion Ratio Lower Upper

138 100

108 9.0 11.4 17.2#

92 158.2 109.0 163.6



#53

2,4-Dinitrophenol

Concen: 98.61 ng

RT: 14.35 min Scan# 1985

Delta R.T. 0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48




Tgt Ion:184 Resp: 77409


Ion Ratio Lower Upper

184 100

63 85.0 66.5 99.7

154 70.8 50.5 75.7





#54

Dibenzofuran

Concen: 50.44 ng

RT: 14.61 min Scan# 2029

Delta R.T. 0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48

Instrument :

BNA_G

ClientSampleId :

MW-19MS


Tgt Ion:168 Resp: 313704

Ion Ratio Lower Upper

168 100

139 41.4 32.8 49.2




169 13.0 10.3 15.5



Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 169.00 (168.70 to 169.70): E



#55

4-Nitrophenol

Concen: 41.08 ng

RT: 14.49 min Scan# 2009

Delta R.T. 0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48


Tgt Ion:139 Resp: 26608

Ion Ratio Lower Upper

139 100

109 155.8 138.9 178.9



65 141.1 115.1 155.1




Abundance Ion 139.00 (138.70 to 139.70): E

Ion 109.00 (108.70 to 109.70): E

Ion 65.00 (64.70 to 65.70): BGC









#56
2,4-Dinitrotoluene
Concen: 53.47 ng
RT: 14.60 min Scan# 2027
Delta R.T. 0.00 min
Lab File: BG017663.D
Acq: 13 Jun 2015 7:48




Instrument :
BNA_G
ClientSampleId :
MW-19MS


Tgt Ion:165 Resp: 89361
Ion Ratio Lower Upper
165 100
63 77.4 61.9 92.9
89 107.9 88.9 133.3
182 6.6 4.0 6.0#



#57
Fluorene
Concen: 50.16 ng
RT: 15.26 min Scan# 2140
Delta R.T. 0.00 min
Lab File: BG017663.D
Acq: 13 Jun 2015 7:48

Tgt Ion:166 Resp: 267647
Ion Ratio Lower Upper
166 100
165 101.2 76.2 114.4
167 15.3 12.2 18.2









#58
 2,3,4,6-Tetrachlorophenol
 Concen: 49.90 ng
 RT: 14.85 min Scan# 2070
 Delta R.T. 0.00 min
 Lab File: BG017663.D
 Acq: 13 Jun 2015 7:48




Instrument :
 BNA_G
 ClientSampleId :
 MW-19MS


Tgt Ion:232 Resp: 87018
 Ion Ratio Lower Upper
 232 100
 131 41.1 31.3 46.9
 130 3.2 2.8 4.2
 166 23.1 17.3 25.9



#59
 Diethylphthalate
 Concen: 50.32 ng
 RT: 15.05 min Scan# 2103
 Delta R.T. -0.00 min
 Lab File: BG017663.D
 Acq: 13 Jun 2015 7:48

Tgt Ion:149 Resp: 293226
 Ion Ratio Lower Upper
 149 100
 177 24.3 19.3 28.9
 150 13.6 10.2 15.2









#60
 4-Chlorophenyl-phenylether
 Concen: 52.90 ng
 RT: 15.26 min Scan# 2140
 Delta R.T. 0.00 min
 Lab File: BG017663.D
 Acq: 13 Jun 2015 7:48




Instrument :
 BNA_G
 ClientSampleId :
 MW-19MS


Tgt Ion:204 Resp: 176747
 Ion Ratio Lower Upper
 204 100
 206 31.5 28.9 43.3
 141 50.4 45.3 67.9



#61
 4-Nitroaniline
 Concen: 24.61 ng
 RT: 15.30 min Scan# 2147
 Delta R.T. -0.00 min
 Lab File: BG017663.D
 Acq: 13 Jun 2015 7:48

Tgt Ion:138 Resp: 26989
 Ion Ratio Lower Upper
 138 100
 92 67.0 46.1 86.1
 108 191.5 184.4 224.4





#62

Azobenzene

Concen: 50.42 ng

RT: 15.55 min Scan# 2189

Delta R.T. -0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48

Instrument :

BNA_G

ClientSampleId :

MW-19MS

Tgt Ion: 77 Resp: 318901


Ion Ratio Lower Upper

77 100

182 25.5 8.2 48.2

105 12.9 0.0 33.0

51 31.9 14.7 54.7






Abundance Ion 77.00 (76.70 to 77.70): BGC

Ion 182.00 (181.70 to 182.70): BGC

Ion 105.00 (104.70 to 105.70): BGC

Ion 51.00 (50.70 to 51.70): BGC



#63

Phenanthrene-d10

Concen: 20.00 ng

RT: 16.96 min Scan# 2429

Delta R.T. -0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48


Tgt Ion: 188 Resp: 172777

Ion Ratio Lower Upper

188 100

94 5.2 3.9 5.9



80 6.8 5.0 7.6




Abundance Ion 188.00 (187.70 to 188.70): BGC

Ion 94.00 (93.70 to 94.70): BGC

Ion 80.00 (79.70 to 80.70): BGC









#64
 4,6-Dinitro-2-methylphenol
 Concen: 46.59 ng
 RT: 15.36 min Scan# 2157
 Delta R.T. -0.00 min
 Lab File: BG017663.D
 Acq: 13 Jun 2015 7:48




Instrument :
 BNA_G
 ClientSampleId :
 MW-19MS


Tgt Ion:198 Resp: 63981
 Ion Ratio Lower Upper
 198 100
 51 30.3 16.2 56.2
 105 32.2 15.7 55.7



#65
 n-Nitrosodiphenylamine
 Concen: 49.46 ng
 RT: 15.48 min Scan# 2177
 Delta R.T. -0.00 min
 Lab File: BG017663.D
 Acq: 13 Jun 2015 7:48

Tgt Ion:169 Resp: 249012
 Ion Ratio Lower Upper
 169 100
 168 67.1 52.3 78.5
 167 37.0 28.5 42.7









#66
4-Bromophenyl-phenylether
Concen: 52.10 ng
RT: 16.16 min Scan# 2292
Delta R.T. 0.00 min
Lab File: BG017663.D
Acq: 13 Jun 2015 7:48




Instrument :
BNA_G
ClientSampleId :
MW-19MS


Tgt Ion:248 Resp: 119726
Ion Ratio Lower Upper
248 100
250 95.2 79.0 118.6
141 70.7 59.0 88.6



#67
Hexachlorobenzene
Concen: 50.45 ng
RT: 16.27 min Scan# 2311
Delta R.T. -0.00 min
Lab File: BG017663.D
Acq: 13 Jun 2015 7:48

Tgt Ion:284 Resp: 126162
Ion Ratio Lower Upper
284 100
142 42.4 27.8 41.6#
249 37.2 25.7 38.5









#68
 Atrazine
 Concen: 45.91 ng
 RT: 16.44 min Scan# 2341
 Delta R.T. 0.00 min
 Lab File: BG017663.D
 Acq: 13 Jun 2015 7:48




Instrument :
 BNA_G
 ClientSampleId :
 MW-19MS


Tgt Ion:200 Resp: 113974
 Ion Ratio Lower Upper
 200 100
 173 22.9 0.8 40.8
 215 60.4 38.4 78.4



#69
 Pentachlorophenol
 Concen: 105.65 ng
 RT: 16.63 min Scan# 2372
 Delta R.T. -0.00 min
 Lab File: BG017663.D
 Acq: 13 Jun 2015 7:48

Tgt Ion:266 Resp: 124257
 Ion Ratio Lower Upper
 266 100
 268 64.6 52.7 79.1
 264 60.7 52.2 78.4





#70

Phenanthrene

Concen: 47.84 ng

RT: 17.00 min Scan# 2436

Delta R.T. -0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48

Instrument :

BNA_G

ClientSampleId :

MW-19MS


Tgt Ion:178 Resp: 461303

Ion Ratio Lower Upper

178 100

176 21.0 16.2 24.4




179 17.7 12.3 18.5



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#71

Anthracene

Concen: 49.31 ng

RT: 17.10 min Scan# 2452

Delta R.T. 0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48


Tgt Ion:178 Resp: 472136

Ion Ratio Lower Upper

178 100

176 20.1 14.8 22.2



179 15.4 11.9 17.9




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E









#72
 Carbazole
 Concen: 50.26 ng
 RT: 17.38 min Scan# 2500
 Delta R.T. 0.00 min
 Lab File: BG017663.D
 Acq: 13 Jun 2015 7:48




Instrument :
 BNA_G
 ClientSampleId :
 MW-19MS


Tgt Ion	Ratio	Lower	Upper
167	100		
166	20.5	16.3	24.5
139	12.9	10.6	15.8



#73
 Di-n-butylphthalate
 Concen: 52.56 ng
 RT: 17.94 min Scan# 2595
 Delta R.T. -0.00 min
 Lab File: BG017663.D
 Acq: 13 Jun 2015 7:48

Tgt Ion	Ratio	Lower	Upper
149	100		
150	9.1	7.2	10.8
104	9.8	7.4	11.0





#74

Fluoranthene

Concen: 50.54 ng

RT: 19.04 min Scan# 2782

Delta R.T. 0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48

Instrument :

BNA_G

ClientSampleId :

MW-19MS


Tgt Ion:202 Resp: 697206

Ion Ratio Lower Upper

202 100

101 5.0 0.0 25.1




203 17.8 0.0 38.2



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#75

Chrysene-d12

Concen: 20.00 ng

RT: 21.16 min Scan# 3144

Delta R.T. -0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48


Tgt Ion:240 Resp: 253941

Ion Ratio Lower Upper

240 100

120 4.6 3.5 5.3



236 24.8 19.5 29.3




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E





#76

Benzidine

Concen: 8.67 ng

RT: 19.23 min Scan# 2815

Delta R.T. 0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48

Instrument :

BNA_G

ClientSampleId :

MW-19MS


Tgt Ion:184 Resp: 51481

Ion Ratio Lower Upper

184 100

185 14.8 12.2 18.4




183 13.6 11.2 16.8



Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E



#77

Pyrene

Concen: 50.80 ng

RT: 19.40 min Scan# 2844

Delta R.T. 0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48


Tgt Ion:202 Resp: 728380

Ion Ratio Lower Upper

202 100

200 19.8 16.8 25.2



203 17.5 14.4 21.6




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E





#78

Terphenyl-d14

Concen: 94.79 ng

RT: 19.61 min Scan# 2879

Delta R.T. -0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48

Instrument :

BNA_G

ClientSampleId :

MW-19MS


Tgt Ion:244 Resp: 1153853

Ion Ratio Lower Upper

244 100

212 9.6 6.3 9.5#




122 6.3 4.6 6.8



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#79

Butylbenzylphthalate

Concen: 51.43 ng

RT: 20.30 min Scan# 2998

Delta R.T. -0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48


Tgt Ion:149 Resp: 266180

Ion Ratio Lower Upper

149 100

91 87.0 66.2 99.4



206 30.7 24.2 36.4




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BG

Ion 206.00 (205.70 to 206.70): E





#80

Benzo(a)anthracene

Concen: 48.78 ng

RT: 21.15 min Scan# 3142

Delta R.T. 0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48

Instrument :

BNA_G

ClientSampleId :

MW-19MS


Tgt Ion:228 Resp: 764210

Ion Ratio Lower Upper

228 100

226 27.8 22.1 33.1




229 20.2 16.2 24.4



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#81

3,3'-Dichlorobenzidine

Concen: 15.56 ng

RT: 21.09 min Scan# 3131

Delta R.T. -0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48


Tgt Ion:252 Resp: 87238

Ion Ratio Lower Upper

252 100

254 64.7 54.3 81.5



126 5.6 6.4 9.6#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 254.00 (253.70 to 254.70): E

Ion 126.00 (125.70 to 126.70): E





#82

Chrysene

Concen: 50.23 ng

RT: 21.20 min Scan# 3151

Delta R.T. 0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48

Instrument :

BNA_G

ClientSampleId :

MW-19MS


Tgt Ion:228 Resp: 709735

Ion Ratio Lower Upper

228 100

226 29.4 24.1 36.1




229 21.1 16.1 24.1



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#83

Bis(2-ethylhexyl)phthalate

Concen: 51.08 ng

RT: 21.09 min Scan# 3131

Delta R.T. 0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48


Tgt Ion:149 Resp: 388427

Ion Ratio Lower Upper

149 100

167 34.2 24.6 37.0



279 13.6 10.0 15.0




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E





#84

Di-n-octyl phthalate

Concen: 51.02 ng

RT: 21.94 min Scan# 3277

Delta R.T. -0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48

Instrument :

BNA_G

ClientSampleId :

MW-19MS


Tgt Ion:149 Resp: 646336

Ion Ratio Lower Upper

149 100

167 1.7 1.5 2.3




43 8.4 6.6 10.0



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 43.00 (42.70 to 43.70): BGC



#85

Indeno(1,2,3-cd)pyrene

Concen: 47.88 ng

RT: 25.60 min Scan# 3899

Delta R.T. 0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48


Tgt Ion:276 Resp: 879880

Ion Ratio Lower Upper

276 100

138 11.1 9.0 13.4



277 26.3 21.8 32.8




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#86
Perylene-d12
Concen: 20.00 ng
RT: 23.36 min Scan# 3518
Delta R.T. 0.00 min
Lab File: BG017663.D
Acq: 13 Jun 2015 7:48




Instrument :
BNA_G
ClientSampleId :
MW-19MS


Tgt Ion	Ratio	Lower	Upper
264	100		
260	20.3	18.3	27.5
265	21.4	16.2	24.2



#87
Benzo(b)fluoranthene
Concen: 49.49 ng
RT: 22.71 min Scan# 3407
Delta R.T. 0.00 min
Lab File: BG017663.D
Acq: 13 Jun 2015 7:48

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.9	18.0	27.0
125	3.9	3.4	5.0





#88

Benzo(k)fluoranthene

Concen: 48.98 ng

RT: 22.75 min Scan# 3414

Delta R.T. 0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48





Instrument :

BNA_G

ClientSampleId :

MW-19MS

Tgt Ion:	252	Resp:	739834
Ion Ratio	Lower	Upper	
252	100		
253	23.3	18.1	27.1
125	4.1	3.2	4.8



#89

Benzo(a)pyrene

Concen: 50.39 ng




RT: 23.27 min Scan# 3503


Delta R.T. 0.00 min

Lab File: BG017663.D

Acq: 13 Jun 2015 7:48

Tgt Ion:	252	Resp:	730689
Ion Ratio	Lower	Upper	
252	100		
253	22.2	17.6	26.4
125	4.4	4.2	6.2









#90
Dibenzo(a,h)anthracene
Concen: 49.87 ng
RT: 25.60 min Scan# 3900
Delta R.T. 0.00 min
Lab File: BG017663.D
Acq: 13 Jun 2015 7:48

Instrument :
BNA_G
ClientSampleId :
MW-19MS

Tgt Ion	Ratio	Lower	Upper
278	100		
139	6.1	6.2	9.4#
279	23.9	19.8	29.6



#91
Benzo(g,h,i)perylene
Concen: 49.14 ng
RT: 26.28 min Scan# 4015
Delta R.T. 0.00 min
Lab File: BG017663.D
Acq: 13 Jun 2015 7:48

Tgt Ion	Ratio	Lower	Upper
276	100		
277	24.2	18.6	28.0
138	8.1	6.6	9.8

