


Data Path : Z:\HPCHEM1\BNA_G\Data\BG061217\
 Data File : BG027376.D
 Acq On : 12 Jun 2017 15:31
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG060517.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jun 05 16:00:35 2017



AutoFind: Scans 2481, 2482, 2483; Background Corrected with Scan 2474

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.4	13683	PASS
68	69	0.00	2	1.1	135	PASS
69	198	0.00	100	34.9	12770	PASS
70	69	0.00	2	0.4	50	PASS
127	198	10	80	44.0	16106	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	36573	PASS
199	198	5	9	6.9	2537	PASS
275	198	10	60	28.3	10357	PASS
365	198	1	100	3.9	1421	PASS
441	442	0.01	24	15.7	6005	PASS
442	198	50	100	104.4	38184	FAIL *
443	442	15	24	19.4	7424	PASS