

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG061220\
 Data File : BG045752.D
 Acq On : 12 Jun 2020 22:45
 Operator : CG/JU
 Sample : L2986-08
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 16-20-COMP

Quant Time: Jun 13 03:04:45 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG051520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 01 15:07:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.94	152	53907	20.00	ng	0.00
21) Naphthalene-d8	10.75	136	205221	20.00	ng	0.00
39) Acenaphthene-d10	14.59	164	128523	20.00	ng	0.00
64) Phenanthrene-d10	17.34	188	263944	20.00	ng	0.00
76) Chrysene-d12	21.61	240	257717	20.00	ng	0.00
87) Perylene-d12	24.75	264	282969	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.56	112	24016	8.71	ng	0.03
7) Phenol-d6	7.18	99	22635	5.77	ng	0.03
23) Nitrobenzene-d5	9.11	82	52535	13.96	ng	0.00
42) 2,4,6-Tribromophenol	16.09	330	25687	15.63	ng	0.00
45) 2-Fluorobiphenyl	13.21	172	117680	15.53	ng	0.00
79) Terphenyl-d14	19.96	244	178422	16.15	ng	0.00
Target Compounds						
10) Phenol	7.20	94	20435	5.050	ng	95
20) 3+4-Methylphenols	8.77	107	37138	9.005	ng	91
50) Dimethylphthalate	14.04	163	142085	16.116	ng	98
84) Bis(2-ethylhexyl)phthalate	21.50	149	597241	68.517	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG061220\
Data File : BG045752.D
Acq On : 12 Jun 2020 22:45
Operator : CG/JU
Sample : L2986-08
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
16-20-COMP

Quant Time: Jun 13 03:04:45 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG051520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 01 15:07:00 2020
Response via : Initial Calibration

