

Data Path : Z:\HPCHEM1\BNA G\DATA\BG061316\
 Data File : BG022323.D
 Acq On : 13 Jun 2016 13:39
 Operator : UM/SJ
 Sample : H3448-03
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 2-VE-PILE-WALL(0-2)

Quant Time: Jun 13 18:15:53 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG060616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 13 17:40:23 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.09	152	38949	20.00	ng	0.00
21) Naphthalene-d8	10.91	136	190192	20.00	ng	0.00
38) Acenaphthene-d10	14.72	164	103346	20.00	ng	0.00
63) Phenanthrene-d10	17.47	188	201606	20.00	ng	0.00
75) Chrysene-d12	21.73	240	165433	20.00	ng	0.00
86) Perylene-d12	25.01	264	157798	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.63	112	358126	177.78	ng	0.00
7) Phenol-d6	7.25	99	521230	164.57	ng	0.00
23) Nitrobenzene-d5	9.26	82	290185	106.37	ng	0.00
41) 2,4,6-Tribromophenol	16.21	330	149724	128.21	ng	0.00
44) 2-Fluorobiphenyl	13.34	172	665839	98.19	ng	0.00
78) Terphenyl-d14	20.06	244	715679	102.71	ng	0.00
Target Compounds						
49) Dimethylphthalate	14.17	163	68655	8.01	ng	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG061316\
Data File : BG022323.D
Acq On : 13 Jun 2016 13:39
Operator : UM/SJ
Sample : H3448-03
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
2-VE-PILE-WALL(0-2)

Quant Time: Jun 13 18:15:53 2016
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG060616.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 13 17:40:23 2016
Response via : Initial Calibration

