

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG061322\
 Data File : BG053895.D
 Acq On : 13 Jun 2022 15:31
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG061322

Quant Time: Jun 13 16:54:39 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 13 15:49:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
2	1,4-Dioxane	0.457	0.462	-1.1	97	0.00
3	Pyridine	1.220	1.299	-6.5	100	0.00
4	n-Nitrosodimethylamine	0.537	0.569	-6.0	98	0.00
5 S	2-Fluorophenol	1.046	1.072	-2.5	98	0.00
6	Aniline	1.682	1.722	-2.4	97	0.00
7 S	Phenol-d6	1.419	1.446	-1.9	98	0.00
8	2-Chlorophenol	1.117	1.155	-3.4	99	0.00
9	Benzaldehyde	0.838	0.786	6.2	98	0.00
10 C	Phenol	1.438	1.473	-2.4	96	0.00
11	bis(2-Chloroethyl)ether	1.104	1.125	-1.9	100	0.00
12	1,3-Dichlorobenzene	1.400	1.410	-0.7	100	0.00
13 C	1,4-Dichlorobenzene	1.406	1.432	-1.8	102	0.00
14	1,2-Dichlorobenzene	1.350	1.339	0.8	97	0.00
15	Benzyl Alcohol	1.061	1.120	-5.6	99	0.00
16	2,2'-oxybis(1-Chloropropane	1.650	1.640	0.6	96	0.00
17	2-Methylphenol	0.999	1.009	-1.0	97	0.00
18	Hexachloroethane	0.466	0.478	-2.6	99	0.00
19 P	n-Nitroso-di-n-propylamine	0.953	0.979	-2.7	98	0.00
20	3+4-Methylphenols	1.401	1.443	-3.0	99	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	97	0.00
22	Acetophenone	0.486	0.486	0.0	96	0.00
23 S	Nitrobenzene-d5	0.356	0.370	-3.9	99	0.00
24	Nitrobenzene	0.350	0.363	-3.7	99	0.00
25	Isophorone	0.665	0.675	-1.5	98	0.00
26 C	2-Nitrophenol	0.155	0.173	-11.6	104	0.00
27	2,4-Dimethylphenol	0.251	0.257	-2.4	98	0.00
28	bis(2-Chloroethoxy)methane	0.358	0.361	-0.8	97	0.00
29 C	2,4-Dichlorophenol	0.336	0.350	-4.2	99	0.00
30	1,2,4-Trichlorobenzene	0.416	0.426	-2.4	98	0.00
31	Naphthalene	1.010	1.017	-0.7	97	0.00
32	Benzoic acid	0.087	0.100	-14.9	105	0.00
33	4-Chloroaniline	0.419	0.429	-2.4	97	0.00
34 C	Hexachlorobutadiene	0.320	0.324	-1.3	96	0.00
35	Caprolactam	0.092	0.097	-5.4	96	0.00
36 C	4-Chloro-3-methylphenol	0.327	0.335	-2.4	96	0.00
37	2-Methylnaphthalene	0.747	0.752	-0.7	96	0.00
38	1-Methylnaphthalene	0.727	0.731	-0.6	97	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	96	0.00
40	1,2,4,5-Tetrachlorobenzene	0.771	0.783	-1.6	97	0.00
41 P	Hexachlorocyclopentadiene	0.359	0.386	-7.5	100	0.00
42 S	2,4,6-Tribromophenol	0.321	0.342	-6.5	100	0.00
43 C	2,4,6-Trichlorophenol	0.433	0.455	-5.1	97	0.00
44	2,4,5-Trichlorophenol	0.483	0.506	-4.8	97	0.00
45 S	2-Fluorobiphenyl	1.375	1.393	-1.3	97	0.00
46	1,1'-Biphenyl	1.393	1.394	-0.1	95	0.00
47	2-Chloronaphthalene	1.127	1.147	-1.8	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG061322\
 Data File : BG053895.D
 Acq On : 13 Jun 2022 15:31
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG061322

Quant Time: Jun 13 16:54:39 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 13 15:49:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.300	0.329	-9.7	100	0.00
49	Acenaphthylene	1.753	1.768	-0.9	96	0.00
50	Dimethylphthalate	1.523	1.540	-1.1	97	0.00
51	2,6-Dinitrotoluene	0.308	0.325	-5.5	98	0.00
52 C	Acenaphthene	1.118	1.150	-2.9	98	0.00
53	3-Nitroaniline	0.298	0.312	-4.7	96	0.00
54 P	2,4-Dinitrophenol	0.140	0.161	-15.0	108	0.00
55	Dibenzofuran	1.801	1.848	-2.6	98	0.00
56 P	4-Nitrophenol	0.212	0.233	-9.9	101	0.00
57	2,4-Dinitrotoluene	0.435	0.464	-6.7	98	0.00
58	Fluorene	1.478	1.499	-1.4	97	0.00
59	2,3,4,6-Tetrachlorophenol	0.473	0.500	-5.7	98	0.00
60	Diethylphthalate	1.475	1.502	-1.8	98	0.00
61	4-Chlorophenyl-phenylether	0.885	0.884	0.1	96	0.00
62	4-Nitroaniline	0.311	0.328	-5.5	99	0.00
63	Azobenzene	1.187	1.204	-1.4	98	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
65	4,6-Dinitro-2-methylphenol	0.100	0.111	-11.0	106	0.00
66 c	n-Nitrosodiphenylamine	0.521	0.524	-0.6	97	0.00
67	4-Bromophenyl-phenylether	0.250	0.255	-2.0	99	0.00
68	Hexachlorobenzene	0.286	0.284	0.7	98	0.00
69	Atrazine	0.216	0.225	-4.2	99	0.00
70 C	Pentachlorophenol	0.133	0.138	-3.8	98	0.00
71	Phenanthrene	1.028	1.031	-0.3	98	0.00
72	Anthracene	1.003	1.015	-1.2	98	0.00
73	Carbazole	0.861	0.878	-2.0	100	0.00
74	Di-n-butylphthalate	0.996	1.030	-3.4	100	0.00
75 C	Fluoranthene	1.349	1.360	-0.8	99	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
77	Benzydine	0.414	0.391	5.6	93	0.00
78	Pyrene	1.208	1.202	0.5	99	0.00
79 S	Terphenyl-d14	0.993	0.987	0.6	99	0.00
80	Butylbenzylphthalate	0.387	0.405	-4.7	103	0.00
81	Benzo(a)anthracene	1.295	1.296	-0.1	101	0.00
82	3,3'-Dichlorobenzidine	0.387	0.394	-1.8	100	0.00
83	Chrysene	1.234	1.233	0.1	101	0.00
84	Bis(2-ethylhexyl)phthalate	0.554	0.572	-3.2	101	0.00
85 c	Di-n-octyl phthalate	0.952	0.999	-4.9	103	0.00
86 I	Perylene-d12	1.000	1.000	0.0	101	0.00
87	Indeno(1,2,3-cd)pyrene	1.509	1.530	-1.4	101	0.00
88	Benzo(b)fluoranthene	1.216	1.237	-1.7	101	0.00
89	Benzo(k)fluoranthene	1.204	1.201	0.2	102	0.00
90 C	Benzo(a)pyrene	1.026	1.053	-2.6	102	0.00
91	Dibenzo(a,h)anthracene	1.248	1.254	-0.5	101	0.00
92	Benzo(g,h,i)perylene	1.232	1.245	-1.1	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG061322\
Data File : BG053895.D
Acq On : 13 Jun 2022 15:31
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG061322

Quant Time: Jun 13 16:54:39 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 13 15:49:31 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0