


Data Path : Z:\HPCHEM1\BNA G\DATA\BG061416\
 Data File : BG022333.D
 Acq On : 14 Jun 2016 12:13
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG060616.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jun 13 17:40:23 2016



AutoFind: Scans 2557, 2558, 2559; Background Corrected with Scan 2549

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	28.8	12780	PASS
68	69	0.00	2	0.6	98	PASS
69	198	0.00	100	35.4	15691	PASS
70	69	0.00	2	0.8	133	PASS
127	198	10	80	54.6	24221	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	44373	PASS
199	198	5	9	6.8	3016	PASS
275	198	10	60	26.2	11620	PASS
365	198	1	100	3.3	1470	PASS
441	442	0.01	24	16.5	5418	PASS
442	198	50	100	74.2	32936	PASS
443	442	15	24	19.3	6368	PASS