

Data Path : Z:\HPCHEM1\BNA G\DATA\BG061417\
 Data File : BG027426.D
 Acq On : 14 Jun 2017 20:47
 Operator : SJ/MA
 Sample : I3635-03
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 CL-02-061217-A

Quant Time: Jun 15 05:24:58 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG061417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 14 19:09:08 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.53	152	40414	20.00	ng	0.00
21) Naphthalene-d8	11.34	136	182998	20.00	ng	0.00
38) Acenaphthene-d10	15.11	164	123572	20.00	ng	0.00
63) Phenanthrene-d10	17.86	188	361764	20.00	ng	0.00
75) Chrysene-d12	22.24	240	415890	20.00	ng	0.00
86) Perylene-d12	25.89	264	408682	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	6.11	112	108371	37.90	ng	0.00
7) Phenol-d6	7.66	99	140597	31.99	ng	0.00
23) Nitrobenzene-d5	9.69	82	94442	26.31	ng	0.00
41) 2,4,6-Tribromophenol	16.59	330	81412	43.90	ng	0.00
44) 2-Fluorobiphenyl	13.73	172	218904	25.59	ng	0.00
78) Terphenyl-d14	20.43	244	487247	30.03	ng	0.00
Target Compounds						
76) Benzidine	20.06	184	166501	18.26	ng	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG061417\
Data File : BG027426.D
Acq On : 14 Jun 2017 20:47
Operator : SJ/MA
Sample : I3635-03
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
CL-02-061217-A

Quant Time: Jun 15 05:24:58 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG061417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 14 19:09:08 2017
Response via : Initial Calibration

