

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG061422\
 Data File : BG053899.D
 Acq On : 14 Jun 2022 12:21
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 15 02:58:18 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 13 15:49:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	103	0.00
2	1,4-Dioxane	40.000	41.372	-3.4	105	0.00
3	Pyridine	40.000	42.852	-7.1	106	0.00
4	n-Nitrosodimethylamine	40.000	41.798	-4.5	102	0.00
5 S	2-Fluorophenol	80.000	82.983	-3.7	105	0.00
6	Aniline	40.000	40.969	-2.4	103	0.00
7 S	Phenol-d6	80.000	83.360	-4.2	106	0.00
8	2-Chlorophenol	40.000	41.498	-3.7	105	0.00
9	Benzaldehyde	40.000	37.570	6.1	104	0.00
10 C	Phenol	40.000	41.234	-3.1	102	0.00
11	bis(2-Chloroethyl)ether	40.000	40.727	-1.8	105	0.00
12	1,3-Dichlorobenzene	40.000	39.745	0.6	104	0.00
13 C	1,4-Dichlorobenzene	40.000	40.125	-0.3	106	0.00
14	1,2-Dichlorobenzene	40.000	40.360	-0.9	104	0.00
15	Benzyl Alcohol	40.000	41.490	-3.7	103	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.751	-1.9	104	0.00
17	2-Methylphenol	40.000	40.233	-0.6	102	0.00
18	Hexachloroethane	40.000	41.475	-3.7	106	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.288	-0.7	102	0.00
20	3+4-Methylphenols	40.000	40.456	-1.1	102	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	102	0.00
22	Acetophenone	40.000	40.369	-0.9	102	0.00
23 S	Nitrobenzene-d5	80.000	82.561	-3.2	104	0.00
24	Nitrobenzene	40.000	41.372	-3.4	104	0.00
25	Isophorone	40.000	40.413	-1.0	103	0.00
26 C	2-Nitrophenol	40.000	41.075	-2.7	101	0.00
27	2,4-Dimethylphenol	40.000	40.643	-1.6	102	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.331	-0.8	103	0.00
29 C	2,4-Dichlorophenol	40.000	42.222	-5.6	106	0.00
30	1,2,4-Trichlorobenzene	40.000	40.308	-0.8	102	0.00
31	Naphthalene	40.000	39.841	0.4	101	0.00
32	Benzoic acid	40.000	33.858	15.4	90	0.00
33	4-Chloroaniline	40.000	40.765	-1.9	102	0.00
34 C	Hexachlorobutadiene	40.000	40.881	-2.2	102	0.00
35	Caprolactam	40.000	42.217	-5.5	101	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.597	-1.5	100	0.00
37	2-Methylnaphthalene	40.000	40.249	-0.6	102	0.00
38	1-Methylnaphthalene	40.000	39.885	0.3	102	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	103	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.997	0.0	102	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.098	-2.7	102	0.00
42 S	2,4,6-Tribromophenol	80.000	80.117	-0.1	100	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.628	-1.6	100	0.00
44	2,4,5-Trichlorophenol	40.000	40.783	-2.0	101	0.00
45 S	2-Fluorobiphenyl	80.000	79.533	0.6	102	0.00
46	1,1'-Biphenyl	40.000	39.093	2.3	99	0.00
47	2-Chloronaphthalene	40.000	40.030	-0.1	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG061422\
 Data File : BG053899.D
 Acq On : 14 Jun 2022 12:21
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 15 02:58:18 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 13 15:49:31 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.382	-3.5	101	0.00
49	Acenaphthylene	40.000	39.986	0.0	101	0.00
50	Dimethylphthalate	40.000	39.813	0.5	102	0.00
51	2,6-Dinitrotoluene	40.000	42.170	-5.4	104	0.00
52 C	Acenaphthene	40.000	40.011	-0.0	101	0.00
53	3-Nitroaniline	40.000	41.168	-2.9	100	0.00
54 P	2,4-Dinitrophenol	40.000	36.986	7.5	99	0.00
55	Dibenzofuran	40.000	39.902	0.2	102	0.00
56 P	4-Nitrophenol	40.000	44.074	-10.2	108	0.00
57	2,4-Dinitrotoluene	40.000	41.881	-4.7	103	0.00
58	Fluorene	40.000	40.282	-0.7	103	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.107	-0.3	99	0.00
60	Diethylphthalate	40.000	40.316	-0.8	104	0.00
61	4-Chlorophenyl-phenylether	40.000	39.280	1.8	101	0.00
62	4-Nitroaniline	40.000	41.184	-3.0	103	0.00
63	Azobenzene	40.000	40.814	-2.0	105	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.241	-3.1	103	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.290	-0.7	102	0.00
67	4-Bromophenyl-phenylether	40.000	40.193	-0.5	102	0.00
68	Hexachlorobenzene	40.000	39.423	1.4	101	0.00
69	Atrazine	40.000	41.217	-3.0	102	0.00
70 C	Pentachlorophenol	40.000	39.662	0.8	98	0.00
71	Phenanthrene	40.000	40.132	-0.3	102	0.00
72	Anthracene	40.000	40.459	-1.1	103	0.00
73	Carbazole	40.000	40.827	-2.1	104	0.00
74	Di-n-butylphthalate	40.000	41.345	-3.4	104	0.00
75 C	Fluoranthene	40.000	40.426	-1.1	103	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	107	0.00
77	Benzydine	40.000	44.286	-10.7	114	0.00
78	Pyrene	40.000	39.748	0.6	103	0.00
79 S	Terphenyl-d14	80.000	78.827	1.5	102	0.00
80	Butylbenzylphthalate	40.000	41.497	-3.7	106	0.00
81	Benzo(a)anthracene	40.000	39.971	0.1	105	0.00
82	3,3'-Dichlorobenzidine	40.000	40.955	-2.4	105	0.00
83	Chrysene	40.000	39.138	2.2	104	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.246	-3.1	106	0.00
85 c	Di-n-octyl phthalate	40.000	41.710	-4.3	107	0.00
86 I	Perylene-d12	20.000	20.000	0.0	105	0.01
87	Indeno(1,2,3-cd)pyrene	40.000	40.677	-1.7	105	0.00
88	Benzo(b)fluoranthene	40.000	39.435	1.4	101	0.00
89	Benzo(k)fluoranthene	40.000	40.869	-2.2	109	0.00
90 C	Benzo(a)pyrene	40.000	40.580	-1.4	105	0.00
91	Dibenzo(a,h)anthracene	40.000	40.283	-0.7	105	0.00
92	Benzo(g,h,i)perylene	40.000	40.516	-1.3	105	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG061422\
Data File : BG053899.D
Acq On : 14 Jun 2022 12:21
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jun 15 02:58:18 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 13 15:49:31 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0