

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG061515\
 Data File : BG017691.D
 Acq On : 15 Jun 2015 15:46
 Operator : TP/IZ
 Sample : G2522-23DL2 400X
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 MW-11CDL2

Quant Time: Jun 16 12:05:57 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG061215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jun 13 02:22:44 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.52	152	14780	20.00	ng	0.00
21) Naphthalene-d8	10.32	136	58084	20.00	ng	0.00
38) Acenaphthene-d10	14.21	164	45111	20.00	ng	0.00
63) Phenanthrene-d10	16.97	188	130503	20.00	ng	0.00
75) Chrysene-d12	21.16	240	180208	20.00	ng	0.00
86) Perylene-d12	23.36	264	190051	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	0.00	112	0d	0.00	ng	
7) Phenol-d6	0.00	99	0d	0.00	ng	
23) Nitrobenzene-d5	0.00	82	0d	0.00	ng	
41) 2,4,6-Tribromophenol	0.00	330	0d	0.00	ng	
44) 2-Fluorobiphenyl	0.00	172	0d	0.00	ng	
78) Terphenyl-d14	0.00	244	0d	0.00	ng	
Target Compounds						
31) Naphthalene	10.37	128	119571	38.14	ng	# 87
37) 2-Methylnaphthalene	12.01	142	5523	2.20	ng	# 68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG061515\
Data File : BG017691.D
Acq On : 15 Jun 2015 15:46
Operator : TP/IZ
Sample : G2522-23DL2 400X
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
MW-11CDL2

Quant Time: Jun 16 12:05:57 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG061215.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jun 13 02:22:44 2015
Response via : Initial Calibration

