

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG061522\
 Data File : BG053920.D
 Acq On : 15 Jun 2022 15:56
 Operator : CG/JU
 Sample : N3339-02
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 CB-226-(26-28)

Quant Time: Jun 15 17:17:12 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 13 15:49:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.083	152	19374	20.000	ng	0.00
21) Naphthalene-d8	10.892	136	71726	20.000	ng	# 0.00
39) Acenaphthene-d10	14.705	164	54838	20.000	ng	0.00
64) Phenanthrene-d10	17.449	188	151322	20.000	ng	# 0.00
76) Chrysene-d12	21.726	240	192904	20.000	ng	# 0.00
86) Perylene-d12	24.993	264	199617	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.645	112	119292	117.723	ng	0.00
7) Phenol-d6	7.231	99	155823	113.354	ng	0.00
23) Nitrobenzene-d5	9.241	82	101657	79.647	ng	0.00
42) 2,4,6-Tribromophenol	16.191	330	115908	131.554	ng	0.00
45) 2-Fluorobiphenyl	13.336	172	292117	77.460	ng	0.00
79) Terphenyl-d14	20.057	244	725202	75.717	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG061522\
Data File : BG053920.D
Acq On : 15 Jun 2022 15:56
Operator : CG/JU
Sample : N3339-02
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
CB-226-(26-28)

Quant Time: Jun 15 17:17:12 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 13 15:49:31 2022
Response via : Initial Calibration

