

Method Path : Z:\HPCHEM1\BNA_G\METHODS\
 Method File : 625-TERPINEOL-BG061615.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Jun 17 07:04:19 2015
 Response Via : Initial Calibration

Calibration Files

2.5 =BG017706.D 10 =BG017707.D 25 =BG017708.D 40 =BG017709.D 50 =BG017710.D 60 =BG017711.D 80 =BG017712.D

	Compound	2.5	10	25	40	50	60	80	Avg	%RSD
1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2) S	2-Fluorophenol	0.827	0.898	0.978	0.928	1.012	0.933	0.932	0.930	6.29
3) S	Phenol-d6	0.999	1.050	1.226	1.159	1.304	1.198	1.227	1.166	9.16
4) I	Naphthalene-d8	-----ISTD-----								
5) S	Nitrobenzene-d5	0.263	0.395	0.403	0.415	0.413	0.426	0.426	0.392	14.75
6)	Terpineol			0.019	0.014	0.020	0.015	0.020	0.018	17.74
7) I	Acenaphthene-d10	-----ISTD-----								
8) S	2,4,6-Tribromo...	0.200	0.216	0.246	0.240	0.265	0.290	0.276	0.247	13.09
9) S	2-Fluorobiphenyl	1.195	1.153	1.276	1.175	1.290	1.279	1.245	1.230	4.52
10) I	Phenanthrene-d10	-----ISTD-----								
11) I	Chrysene-d12	-----ISTD-----								
12) S	Terphenyl-d14	0.830	0.842	0.877	0.867	0.897	0.888	0.892	0.871	2.95
13) I	Perylene-d12	-----ISTD-----								

(#) = Out of Range