

Data Path : Z:\HPCHEM1\BNA\_G\DATA\BG061615\  
Data File : BG017715.D  
Acq On : 17 Jun 2015 7:54  
Operator : TP/IZ  
Sample : SSTDSCC040  
Misc :  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
BNA\_G  
LabSampleId :  
SSTDSCC040

Quant Time: Jun 17 08:25:11 2015  
Quant Method : Z:\HPCHEM1\BNA\_G\METHODS\625-TERPINEOL-BG061615.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Jun 17 07:04:19 2015  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound               | AvgRF | CCRF  | %Dev | Area% | Dev(min) |
|------|------------------------|-------|-------|------|-------|----------|
| 1 I  | 1,4-Dichlorobenzene-d4 | 1.000 | 1.000 | 0.0  | 94    | 0.00     |
| 2 S  | 2-Fluorophenol         | 0.930 | 0.944 | -1.5 | 96    | 0.00     |
| 3 S  | Phenol-d6              | 1.166 | 1.177 | -0.9 | 96    | 0.00     |
| 4 I  | Naphthalene-d8         | 1.000 | 1.000 | 0.0  | 102   | 0.00     |
| 5 S  | Nitrobenzene-d5        | 0.392 | 0.409 | -4.3 | 101   | 0.00     |
| 6    | Terpineol              | 0.018 | 0.018 | 0.0  | 132   | 0.01     |
| 7 I  | Acenaphthene-d10       | 1.000 | 1.000 | 0.0  | 97    | 0.00     |
| 8 S  | 2,4,6-Tribromophenol   | 0.247 | 0.247 | 0.0  | 100   | 0.00     |
| 9 S  | 2-Fluorobiphenyl       | 1.230 | 1.246 | -1.3 | 102   | 0.00     |
| 10 I | Phenanthrene-d10       | 1.000 | 1.000 | 0.0  | 108   | 0.00     |
| 11 I | Chrysene-d12           | 1.000 | 1.000 | 0.0  | 99    | 0.00     |
| 12 S | Terphenyl-d14          | 0.871 | 0.880 | -1.0 | 101   | 0.00     |
| 13 I | Perylene-d12           | 1.000 | 1.000 | 0.0  | 97    | 0.00     |

(#) = Out of Range

SPCC's out = 0 CCC's out = 0