

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG061615\
 Data File : BG017715.D
 Acq On : 17 Jun 2015 7:54
 Operator : TP/IZ
 Sample : SSTDSCC040
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDSCC040

Quant Time: Jun 17 08:25:11 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\625-TERPINEOL-BG061615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 17 07:04:19 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	94	0.00
2 S	2-Fluorophenol	80.000	81.201	-1.5	96	0.00
3 S	Phenol-d6	80.000	80.766	-1.0	96	0.00
4 I	Naphthalene-d8	20.000	20.000	0.0	102	0.00
5 S	Nitrobenzene-d5	80.000	83.532	-4.4	101	0.00
6	Terpineol	40.000	40.415	-1.0	132	0.01
7 I	Acenaphthene-d10	20.000	20.000	0.0	97	0.00
8 S	2,4,6-Tribromophenol	80.000	79.919	0.1	100	0.00
9 S	2-Fluorobiphenyl	80.000	81.044	-1.3	102	0.00
10 I	Phenanthrene-d10	20.000	20.000	0.0	108	0.00
11 I	Chrysene-d12	20.000	20.000	0.0	99	0.00
12 S	Terphenyl-d14	80.000	80.915	-1.1	101	0.00
13 I	Perylene-d12	20.000	20.000	0.0	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0