


Data Path : Z:\HPCHEM1\BNA\_G\DATA\BG061616\  
 Data File : BG022364.D  
 Acq On : 16 Jun 2016 13:31  
 Operator : UM/SJ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_G\METHODS\8270-BG060616.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Mon Jun 13 17:40:23 2016



AutoFind: Scans 2554, 2555, 2556; Background Corrected with Scan 2548

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 32.3      | 6684    | PASS             |
| 68          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 69          | 198          | 0.00         | 100          | 40.0      | 8279    | PASS             |
| 70          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 127         | 198          | 10           | 80           | 59.7      | 12352   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.3       | 68      | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 20682   | PASS             |
| 199         | 198          | 5            | 9            | 6.6       | 1367    | PASS             |
| 275         | 198          | 10           | 60           | 23.8      | 4914    | PASS             |
| 365         | 198          | 1            | 100          | 3.3       | 673     | PASS             |
| 441         | 442          | 0.01         | 24           | 16.1      | 2029    | PASS             |
| 442         | 198          | 50           | 100          | 61.0      | 12610   | PASS             |
| 443         | 442          | 15           | 24           | 18.5      | 2330    | PASS             |