

Data File : BG027475.D
Acq On : 16 Jun 2017 17:27
Operator : SJ/MA
Sample : PB99751BL
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK51

Quant Time: Jun 16 22:56:05 2017
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG061617.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jun 16 22:56:07 2017
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.52	152	41407	20.00	ng/ul	0.00
18) Naphthalene-d8	11.34	136	201683	20.00	ng/ul	0.00
35) Acenaphthene-d10	15.10	164	132152	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.86	188	345992	20.00	ng/ul	0.00
75) Chrysene-d12	22.23	240	372915	20.00	ng/ul	0.00
83) Perylene-d12	25.89	264	356052	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	4.08	96	7375	7.12	ng/uL	0.00
5) Phenol-d5	7.65	99	143854	32.04	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.83	67	110764	36.19	ng/ul	0.00
9) 2-Chlorophenol-d4	8.05	132	103667	34.94	ng/ul	0.00
13) 4-Methylphenol-d8	9.20	113	114361	32.57	ng/ul	0.00
19) Nitrobenzene-d5	9.68	128	53495	36.51	ng/ul	0.00
22) 2-Nitrophenol-d4	10.41	143	58775	37.76	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.94	165	96764	32.91	ng/ul	0.00
29) 4-Chloroaniline-d4	11.46	131	159170	42.76	ng/ul	0.00
43) Dimethylphthalate-d6	14.49	166	434272	39.25	ng/ul	0.00
46) Acenaphthylene-d8	14.80	160	459901	36.09	ng/ul	0.00
51) 4-Nitrophenol-d4	15.27	143	69725	32.21	ng/ul	0.00
57) Fluorene-d10	16.09	176	360414	34.72	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	16.19	200	68535	31.90	ng/ul	0.00
70) Anthracene-d10	17.96	188	615156	38.13	ng/ul	0.00
76) Pyrene-d10	20.22	212	749007	41.13	ng/ul	0.00
87) Benzo(a)pyrene-d12	25.64	264	620134	38.68	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BG027475.D

Acq On : 16 Jun 2017 17:27

Operator : SJ/MA

Sample : PB99751BL

Misc :

ALS Vial : 9 Sample Multiplier: 1

Instrument :

BNA_G

ClientSampleId :

SBLK51

Quant Time: Jun 16 22:56:05 2017

Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG061617.M

Quant Title : SVOA CALIBRATION

QLast Update : Fri Jun 16 22:56:07 2017

Response via : Initial Calibration

