


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG061915\
Data File : BG017725.D
Acq On : 19 Jun 2015 11:12
Operator : TP/IZ
Sample : DFTPP05
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP05

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG061915.M
Title : SVOA CALIBRATION
Last Update : Fri Jun 19 17:30:37 2015



AutoFind: Scans 2517, 2518, 2519; Background Corrected with Scan 2512

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	54.8	32816	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	55.0	32896	PASS
70	69	0.00	2	0.9	298	PASS
127	198	10	80	59.6	35640	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	59832	PASS
199	198	5	9	7.2	4322	PASS
275	198	10	60	25.7	15364	PASS
365	198	1	100	3.8	2276	PASS
441	442	0.01	24	13.9	5316	PASS
442	198	50	100	63.7	38133	PASS
443	442	15	24	17.3	6615	PASS

m/z	Abundance
35.90	173.0
36.90	238.0
38.10	464.0
39.10	2274.0
40.00	1247.0
41.00	342.0
43.00	250.0
43.90	736.0
44.90	194.0
49.00	569.0
50.00	7854.0
51.00	28704.0
52.00	1602.0
55.10	457.0
56.00	848.0
57.00	2301.0
57.90	203.0
61.00	523.0
62.00	424.0
63.00	1078.0
64.00	232.0
65.00	373.0
68.90	25544.0
69.90	306.0
71.00	195.0
73.00	715.0
74.00	3427.0
75.00	4643.0
76.10	1708.0
77.00	30800.0
78.00	2234.0
78.90	1927.0
79.90	1641.0
81.00	2499.0
81.90	555.0
83.00	505.0
84.10	287.0
85.00	619.0
86.00	665.0
86.90	157.0
90.90	326.0
92.00	806.0
92.90	3974.0
93.90	257.0
96.00	408.0
98.00	2965.0
99.00	1789.0
100.90	1537.0
103.00	500.0
104.00	1097.0
105.00	957.0
107.00	9000.0
108.00	1479.0
110.00	14212.0
111.00	2645.0
111.90	209.0
115.00	352.0
115.90	621.0
117.00	8679.0
117.90	428.0
121.90	639.0
122.90	778.0
123.80	455.0
125.00	399.0
127.00	27776.0
127.90	2662.0
129.00	11280.0
129.90	1133.0
131.00	327.0
131.90	238.0
132.80	427.0
133.90	632.0
134.90	1211.0
135.90	393.0
137.00	336.0
137.70	232.0
138.20	163.0
140.10	237.0
140.90	1789.0
141.90	306.0
142.90	227.0
145.90	608.0
147.00	1051.0
147.90	1398.0
148.90	336.0
152.00	289.0
153.00	729.0
154.00	311.0
154.90	991.0
156.00	774.0
156.90	250.0

Instrument :
BNA_G
ClientSampleId :
DFTPP05

158.00	380.0
158.90	268.0
160.00	544.0
160.90	338.0
161.50	223.0
164.90	693.0
165.80	382.0
167.00	2384.0
168.00	1795.0
168.70	152.0
169.00	183.0
171.90	177.0
172.90	378.0
173.90	765.0
175.00	1711.0
176.00	679.0
177.00	565.0
177.70	167.0
178.00	168.0
178.20	168.0
178.90	2019.0
180.00	1586.0
181.00	586.0
182.00	210.0
184.00	248.0
184.90	1086.0
186.00	6569.0
187.00	1658.0
187.90	199.0
189.00	462.0
190.00	179.0
190.90	368.0
191.90	620.0
193.00	708.0
193.90	181.0
194.70	201.0
195.90	1531.0
197.90	41328.0
199.00	2978.0
201.60	194.0
202.70	253.0
203.90	1646.0
205.00	2842.0
206.00	12144.0
207.00	3348.0
208.00	1146.0
209.00	501.0
209.90	172.0
210.70	632.0
212.60	153.0
215.20	155.0
216.00	286.0
216.90	3828.0
217.90	401.0
220.90	1931.0
222.90	1004.0
224.00	6987.0
225.00	1939.0
225.90	206.0
227.00	3040.0
228.10	589.0
228.90	571.0
231.00	237.0
232.80	169.0
233.90	176.0
234.80	249.0
235.90	167.0
237.00	310.0
241.00	252.0
242.00	204.0
242.70	260.0
244.00	4454.0
244.90	637.0
246.00	980.0
247.00	370.0
248.70	249.0
252.70	240.0
253.10	284.0
253.80	330.0
255.00	24328.0
255.90	4183.0
256.90	361.0
258.00	1978.0
258.90	363.0
264.90	816.0
273.00	1008.0
273.90	2149.0
275.00	9548.0
276.10	1268.0
277.00	1331.0
277.90	266.0
280.90	229.0
292.90	335.0
295.90	4206.0
296.90	603.0

Instrument :
BNA_G
ClientSampleId :
DFTPP05

302.90	485.0
307.90	161.0
313.80	193.0
314.20	171.0
314.90	284.0
321.90	156.0
323.10	1198.0
326.80	204.0
334.00	653.0
334.90	193.0
341.00	166.0
343.70	157.0
352.00	497.0
352.80	219.0
354.00	326.0
355.00	179.0
364.90	1420.0
366.00	325.0
372.00	344.0
373.10	208.0
402.10	183.0
403.00	175.0
421.00	237.0
423.00	1380.0
423.90	230.0
436.40	171.0
438.20	171.0
441.00	2889.0
442.00	20616.0
443.00	3273.0
444.10	538.0

Instrument :

BNA_G

ClientSampleId :

DFTPP05

m/z	Abundance
36.90	245.0
38.00	630.0
39.00	3202.0
40.00	775.0
41.10	513.0
42.10	200.0
43.00	627.0
44.00	865.0
46.80	164.0
49.00	768.0
50.00	11384.0
51.00	39424.0
52.00	1543.0
55.10	248.0
56.00	1329.0
57.00	2898.0
57.90	199.0
60.90	518.0
62.00	575.0
62.90	1462.0
64.00	360.0
65.00	982.0
67.00	245.0
69.00	40168.0
70.00	386.0
73.00	413.0
74.00	4303.0
75.00	6687.0
76.00	2348.0
77.00	45736.0
78.00	2824.0
79.00	3356.0
80.00	2265.0
81.00	3926.0
82.00	921.0
83.00	754.0
84.00	506.0
84.90	767.0
86.00	1215.0
87.00	335.0
87.90	280.0
89.10	205.0
91.00	873.0
92.00	1022.0
93.00	5983.0
93.90	533.0
96.00	360.0
96.90	207.0
97.10	209.0
98.00	4479.0
98.90	2911.0
100.00	264.0
101.00	1862.0
103.00	571.0
104.00	1185.0
104.90	887.0
107.00	13398.0
108.00	1717.0
109.00	540.0
110.00	21944.0
111.00	3476.0
112.00	514.0
112.90	152.0
114.80	305.0
115.90	811.0
117.00	13701.0
118.00	919.0
118.90	412.0
119.90	323.0
120.90	223.0
121.90	1253.0
122.90	1728.0
124.00	643.0
125.00	658.0
127.00	43248.0
128.00	2835.0
129.00	19632.0
129.90	1444.0
131.00	472.0
132.00	164.0
132.30	178.0
133.90	356.0
135.00	1572.0
136.00	535.0
137.00	1000.0
137.80	415.0
139.70	168.0
140.90	2156.0
142.00	781.0
142.90	603.0
143.90	188.0

Instrument :
BNA_G
ClientSampleId :
DFTPP05

146.00	534.0
146.90	1270.0
147.90	3172.0
149.00	704.0
151.00	279.0
151.70	358.0
153.00	1061.0
154.00	559.0
154.90	1451.0
156.00	2341.0
157.10	237.0
157.90	414.0
158.80	531.0
160.00	700.0
160.90	1281.0
161.60	170.0
161.90	187.0
163.90	152.0
165.00	938.0
165.90	593.0
167.00	4996.0
167.90	2054.0
169.00	471.0
169.90	192.0
171.10	323.0
171.90	914.0
172.90	743.0
173.90	503.0
175.00	1584.0
176.00	427.0
177.00	979.0
177.90	315.0
178.90	3140.0
179.90	1852.0
180.90	1138.0
182.00	352.0
183.80	209.0
184.90	1898.0
186.00	12262.0
187.00	3479.0
187.90	350.0
189.00	972.0
189.80	228.0
190.90	431.0
192.00	1157.0
193.00	1032.0
194.00	236.0
196.00	2355.0
197.90	71512.0
198.90	4974.0
200.00	355.0
201.50	335.0
202.00	309.0
202.90	960.0
204.00	2956.0
205.00	4566.0
206.00	19416.0
207.00	4982.0
208.00	1540.0
208.80	339.0
210.00	388.0
211.00	1036.0
212.00	327.0
214.90	232.0
215.90	554.0
216.90	5983.0
217.80	659.0
219.00	229.0
219.90	218.0
220.90	4456.0
223.00	1433.0
224.00	12535.0
224.90	2563.0
225.90	434.0
227.00	5571.0
227.90	834.0
229.00	1073.0
231.00	788.0
232.90	157.0
234.10	192.0
234.90	429.0
236.00	415.0
236.90	283.0
238.90	356.0
240.10	296.0
240.90	302.0
241.80	638.0
243.00	912.0
244.00	8005.0
244.90	956.0
245.90	2507.0
246.90	415.0
247.90	224.0
248.90	509.0
251.90	258.0

Instrument :
BNA_G
ClientSampleId :
DFTPP05

252.70	237.0
255.00	43872.0
255.90	6267.0
256.90	804.0
258.00	3281.0
258.90	457.0
264.90	1573.0
265.90	343.0
272.20	184.0
272.90	1472.0
273.90	4144.0
275.00	18976.0
276.00	2307.0
277.00	1830.0
277.90	251.0
281.00	411.0
281.90	200.0
282.80	388.0
285.00	438.0
289.80	158.0
293.00	422.0
296.00	6764.0
296.90	1210.0
302.00	176.0
302.90	535.0
304.10	207.0
314.00	263.0
315.00	797.0
316.10	321.0
323.00	1576.0
323.90	259.0
326.90	423.0
328.00	310.0
331.90	250.0
332.90	223.0
333.90	1503.0
335.00	262.0
341.00	163.0
345.90	280.0
352.00	884.0
352.90	495.0
354.20	212.0
355.00	294.0
364.90	2454.0
365.90	452.0
371.00	232.0
372.00	738.0
373.10	259.0
382.90	246.0
402.00	383.0
403.00	383.0
404.00	497.0
421.00	175.0
422.00	476.0
423.00	1821.0
424.00	545.0
440.00	321.0
441.00	6536.0
442.00	42608.0
443.00	8121.0
444.10	581.0

Instrument :
BNA_G
ClientSampleId :
DFTPP05

m/z	Abundance
36.10	195.0
37.00	237.0
37.70	279.0
38.10	504.0
39.00	2862.0
40.00	1300.0
41.10	385.0
41.70	199.0
43.00	379.0
44.00	1064.0
47.90	160.0
49.00	760.0
50.00	9392.0
51.00	30320.0
52.00	1534.0
53.10	227.0
55.10	495.0
55.90	1126.0
57.00	2187.0
57.90	181.0
61.00	431.0
62.10	327.0
63.00	1214.0
63.90	192.0
65.00	595.0
69.00	32976.0
70.00	203.0
73.00	426.0
74.00	3783.0
75.00	5342.0
76.10	2031.0
77.00	34784.0
78.00	3028.0
79.00	2020.0
80.00	2692.0
81.00	3035.0
82.00	768.0
83.00	616.0
83.90	393.0
85.00	539.0
86.00	808.0
87.00	315.0
87.90	196.0
90.40	187.0
91.00	570.0
91.90	609.0
92.90	4520.0
93.90	345.0
96.10	418.0
96.70	219.0
98.00	3635.0
99.00	2314.0
99.90	224.0
100.90	1668.0
103.00	354.0
104.00	807.0
105.00	1086.0
107.00	11645.0
108.00	1218.0
109.00	584.0
110.00	17584.0
111.00	2902.0
112.00	578.0
112.90	198.0
114.10	198.0
115.90	831.0
116.90	11028.0
117.90	750.0
118.90	217.0
122.00	804.0
123.00	1051.0
123.90	357.0
125.00	557.0
127.00	36456.0
128.00	2584.0
128.90	15659.0
129.90	1156.0
130.70	269.0
130.90	265.0
132.90	358.0
134.00	537.0
134.90	1339.0
136.00	304.0
136.80	646.0
138.90	160.0
140.90	2335.0
142.00	694.0
143.00	496.0
144.00	260.0
145.90	688.0
147.00	1328.0

Instrument :
BNA_G
ClientSampleId :
DFTPP05

147.90	2650.0
148.90	610.0
149.90	205.0
151.00	169.0
153.00	520.0
153.90	385.0
155.00	1078.0
155.90	1524.0
157.10	259.0
157.90	377.0
159.90	786.0
161.00	955.0
161.80	315.0
163.80	213.0
164.90	664.0
166.00	519.0
167.00	4102.0
168.00	2302.0
168.90	287.0
170.20	156.0
170.90	233.0
171.80	326.0
173.10	347.0
173.90	710.0
175.00	1388.0
175.90	294.0
176.90	642.0
177.60	154.0
177.90	190.0
178.90	2946.0
180.00	2054.0
181.00	554.0
184.00	336.0
185.00	1273.0
186.00	10260.0
187.00	2521.0
188.00	229.0
188.90	652.0
190.90	452.0
191.90	1117.0
193.00	1521.0
194.00	280.0
195.10	248.0
196.00	2292.0
197.90	66656.0
199.00	5014.0
199.90	293.0
201.40	422.0
203.00	783.0
204.00	2577.0
205.00	5304.0
206.00	19504.0
207.00	4314.0
208.00	908.0
208.90	432.0
209.90	388.0
211.00	569.0
211.80	190.0
214.80	184.0
215.90	536.0
216.90	5183.0
218.00	693.0
218.90	205.0
221.00	3394.0
221.70	1382.0
223.00	1391.0
224.00	10346.0
225.00	2721.0
226.00	230.0
226.90	5104.0
227.90	616.0
228.90	683.0
229.90	150.0
231.00	405.0
233.90	522.0
235.00	528.0
235.90	310.0
236.90	539.0
238.80	220.0
239.80	213.0
241.00	254.0
241.90	619.0
243.00	725.0
244.00	7892.0
245.00	1186.0
246.00	1747.0
247.10	495.0
248.90	261.0
249.80	165.0
251.80	160.0
252.60	172.0
253.00	198.0
254.90	41232.0
255.90	5153.0
257.00	336.0

Instrument :
BNA_G
ClientSampleId :
DFTPP05

257.90	2979.0
259.00	459.0
264.90	1274.0
266.00	343.0
267.00	213.0
272.90	1667.0
274.00	4210.0
275.00	17568.0
276.00	2081.0
276.90	2336.0
278.00	282.0
279.80	184.0
281.00	166.0
282.00	222.0
282.90	270.0
284.00	177.0
285.00	262.0
288.90	223.0
293.00	476.0
295.90	6553.0
297.00	756.0
303.00	609.0
304.10	298.0
314.10	248.0
314.90	455.0
315.70	295.0
320.90	344.0
323.00	1404.0
324.00	275.0
326.20	176.0
326.80	311.0
328.00	219.0
331.70	162.0
334.00	1405.0
335.10	182.0
341.00	296.0
345.90	438.0
347.00	276.0
352.00	614.0
352.80	362.0
354.00	734.0
365.00	2956.0
366.00	537.0
370.80	254.0
372.00	803.0
372.90	320.0
382.90	397.0
402.00	349.0
402.90	445.0
403.90	171.0
420.90	490.0
421.90	315.0
423.00	3081.0
424.00	291.0
431.50	190.0
434.70	151.0
437.40	221.0
438.30	187.0
441.00	6525.0
442.00	51176.0
443.00	8453.0
443.90	668.0

Instrument :
BNA_G
ClientSampleId :
DFTPP05

m/z	Abundance
37.90	154.0
39.10	208.0
39.90	1176.0
41.00	381.0
42.10	259.0
43.10	303.0
44.00	883.0
48.90	310.0
55.00	221.0
56.80	231.0
60.10	162.0
71.00	217.0
73.00	470.0
79.80	171.0
83.00	179.0
83.90	321.0
85.70	189.0
96.00	243.0
98.10	233.0
127.10	186.0
132.90	199.0
134.90	158.0
147.20	209.0
176.90	174.0
190.90	210.0
207.00	1593.0
207.80	338.0
209.10	203.0
213.30	155.0
221.00	182.0
252.20	162.0

Instrument :
BNA_G
ClientSampleId :
DFTPP05