

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062022\
 Data File : BG053945.D
 Acq On : 20 Jun 2022 12:47
 Operator : CG/JU
 Sample : SST02030
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SST020430

Quant Time: Jun 20 14:09:29 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG062022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 20 14:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.077	152	17337	20.000	ng/u1	0.00
20) Naphthalene-d8	10.885	136	71951	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.705	164	55187	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.443	188	151860	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.720	240	177684	20.000	ng/u1	# 0.00
88) Perylene-d12	24.981	264	184268	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.506	96	2936	6.405	ng/uL	0.00
4) Pyridine-d5	3.917	84	19929	19.720	ng/u1	0.00
7) Phenol-d5	7.231	99	25057	19.005	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.401	67	16164	18.967	ng/u1	0.00
11) 2-Chlorophenol-d4	7.607	132	20751	19.795	ng/u1	0.00
15) 4-Methylphenol-d8	8.776	113	22085	20.222	ng/u1	0.00
21) Nitrobenzene-d5	9.234	128	10634	24.724	ng/u1	0.00
24) 2-Nitrophenol-d4	9.963	143	10841	25.395	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.504	165	25450	21.951	ng/u1	0.00
31) 4-Chloroaniline-d4	11.015	131	31533	19.825	ng/u1	0.00
46) Dimethylphthalate-d6	14.099	166	89579	21.123	ng/u1	0.00
49) Acenaphthylene-d8	14.399	160	95419	20.136	ng/u1	0.00
54) 4-Nitrophenol-d4	14.881	143	11802	20.240	ng/u1	0.00
60) Fluorene-d10	15.692	176	78835	21.226	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.803	200	15730	25.907	ng/u1	0.00
73) Anthracene-d10	17.542	188	140834	20.532	ng/u1	0.00
81) Pyrene-d10	19.828	212	193218	20.431	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.757	264	189105	20.520	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.541	88	3403	7.058	ng/uL#	84
5) Pyridine	3.935	79	20715	19.385	ng/u1	94
6) Benzaldehyde	7.213	77	15858	22.836	ng/u1	96
8) Phenol	7.254	94	26584	18.837	ng/u1	92
10) Bis(2-Chloroethyl)ether	7.495	93	20578	19.040	ng/u1	88
12) 2-Chlorophenol	7.642	128	21344	19.895	ng/u1	97
13) 2-Methylphenol	8.512	108	20635	19.034	ng/u1	94
14) 2,2'-oxybis(1-Chloropr...	8.606	45	30823	16.969	ng/u1	97
16) Acetophenone	8.900	105	34896	20.458	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.882	70	17827	18.724	ng/u1#	94
18) 4-Methylphenol	8.841	108	22798	19.894	ng/u1	90
19) Hexachloroethane	9.170	117	9114	20.040	ng/u1	94
22) Nitrobenzene	9.276	77	25459	22.697	ng/u1	92
23) Isophorone	9.804	82	49678	18.780	ng/u1	98
25) 2-Nitrophenol	9.998	139	11881	26.155	ng/u1#	83
26) 2,4-Dimethylphenol	10.051	107	26128	19.053	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.286	93	28151	18.507	ng/u1#	98
29) 2,4-Dichlorophenol	10.533	162	25811	22.448	ng/u1	95
30) Naphthalene	10.938	128	72017	19.776	ng/u1	96
32) 4-Chloroaniline	11.038	127	31335	19.995	ng/u1	97
33) Hexachlorobutadiene	11.226	225	24261	25.424	ng/u1	96
34) Caprolactam	11.790	113	4635	14.403	ng/u1#	69
35) 4-Chloro-3-methylphenol	12.155	107	24221	19.963	ng/u1	98
36) 2-Methylnaphthalene	12.537	142	53182	20.676	ng/u1	94


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062022\
 Data File : BG053945.D
 Acq On : 20 Jun 2022 12:47
 Operator : CG/JU
 Sample : SST02030
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SST020430

Quant Time: Jun 20 14:09:29 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG062022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 20 14:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.754	142	52222	20.407	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.907	216	45397	23.517	ng/ul	99
40) Hexachlorocyclopentadiene	12.889	237	23061	17.572	ng/ul#	86
41) 2,4,6-Trichlorophenol	13.136	196	23439	21.309	ng/ul	87
42) 2,4,5-Trichlorophenol	13.206	196	27021	22.838	ng/ul	97
43) 1,1'-Biphenyl	13.541	154	78741	19.724	ng/ul	98
44) 2-Chloronaphthalene	13.582	162	64422	20.717	ng/ul	98
45) 2-Nitroaniline	13.776	65	16788	25.393	ng/ul	97
47) Dimethylphthalate	14.146	163	86047	21.191	ng/ul	98
48) 2,6-Dinitrotoluene	14.264	165	17656	29.798	ng/ul	93
50) Acenaphthylene	14.423	152	99793	19.644	ng/ul	99
51) 3-Nitroaniline	14.593	138	14474	25.511	ng/ul#	88
52) Acenaphthene	14.769	153	65948	19.559	ng/ul	97
53) 2,4-Dinitrophenol	14.799	184	6197	19.824	ng/ul#	75
55) 4-Nitrophenol	14.893	109	11775	21.398	ng/ul	92
56) Dibenzofuran	15.098	168	102001	20.915	ng/ul	100
57) 2,4-Dinitrotoluene	15.051	165	25157	30.394	ng/ul#	97
58) 2,3,4,6-Tetrachlorophenol	15.321	232	23028	21.967	ng/ul	93
59) Diethylphthalate	15.509	149	84962	20.222	ng/ul	98
61) Fluorene	15.750	166	82714	20.697	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.739	204	52219	23.290	ng/ul	98
63) 4-Nitroaniline	15.750	138	14581	26.622	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.821	198	14900	26.385	ng/ul#	93
67) N-Nitrosodiphenylamine	15.950	169	75710	18.258	ng/ul	98
68) 4-Bromophenyl-phenylether	16.632	248	38041	22.590	ng/ul	93
69) Hexachlorobenzene	16.755	284	44305	24.367	ng/ul#	93
70) Atrazine	16.890	200	37103	20.284	ng/ul	98
71) Pentachlorophenol	17.096	266	17055	15.783	ng/ul	94
72) Phenanthrene	17.490	178	155736	19.878	ng/ul	97
74) Anthracene	17.578	178	154856	19.370	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.506	216	49273	20.516	ng/ul	93
76) Pentachlorobenzene	15.022	250	47770	22.145	ng/ul	95
77) Carbazole	17.842	167	131566	18.879	ng/ul	98
78) Di-n-butylphthalate	18.406	149	147415	16.918	ng/ul	99
80) Fluoranthene	19.493	202	212887	19.365	ng/ul#	93
82) Pyrene	19.857	202	213139	19.560	ng/ul#	93
83) Butylbenzylphthalate	20.739	149	62619	15.791	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.614	252	77827	19.686	ng/ul#	96
85) Benzo(a)anthracene	21.702	228	224963	20.024	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.608	149	90851	15.602	ng/ul#	95
87) Chrysene	21.767	228	212838	19.820	ng/ul	99
89) Di-n-octyl phthalate	22.836	149	157853	15.745	ng/ul	100
90) Benzo(b)fluoranthene	23.941	252	230421	19.932	ng/ul#	97
91) Benzo(k)fluoranthene	24.017	252	224090	20.352	ng/ul#	98
93) Benzo(a)pyrene	24.828	252	223753	20.040	ng/ul#	97
94) Indeno(1,2,3-cd)pyrene	28.700	276	266677	20.822	ng/ul#	97
95) Dibenzo(a,h)anthracene	28.765	278	227260	20.665	ng/ul#	97
96) Benzo(g,h,i)perylene	29.857	276	225728	20.896	ng/ul#	94





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 7.058 ng/uL
 RT: 3.541 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BG053945.D
 Acq: 20 Jun 2022 12:47




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020430


Tgt Ion: 88 Resp: 3403
 Ion Ratio Lower Upper
 88 100
 43 37.6 48.1 72.1#
 58 78.5 59.8 89.6



#5
 Pyridine
 Concen: 19.385 ng/uL
 RT: 3.935 min Scan# 93
 Delta R.T. 0.000 min
 Lab File: BG053945.D
 Acq: 20 Jun 2022 12:47

Tgt Ion: 79 Resp: 20715
 Ion Ratio Lower Upper
 79 100
 52 90.0 75.4 113.2
 51 37.6 35.2 52.8









#6
Benzaldehyde
Concen: 22.836 ng/ul
RT: 7.213 min Scan# 651
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47




Instrument :
BNA_G
ClientSampleId :
SSTD020430


Tgt Ion: 77 Resp: 15858
Ion Ratio Lower Upper
77 100
105 95.3 79.7 119.5
106 92.6 71.0 106.4



#8
Phenol
Concen: 18.837 ng/ul
RT: 7.254 min Scan# 658
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47

Tgt Ion: 94 Resp: 26584
Ion Ratio Lower Upper
94 100
65 33.1 24.1 36.1
66 46.6 32.5 48.7









#10
Bis(2-Chloroethyl)ether
Concen: 19.040 ng/ul
RT: 7.495 min Scan# 699
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47




Instrument :
BNA_G
ClientSampleId :
SSTD020430


Tgt Ion: 93	Resp:	20578
Ion	Ratio	Lower Upper
93	100	
63	76.5	72.3 108.5
95	33.8	29.4 44.0



#12
2-Chlorophenol
Concen: 19.895 ng/ul
RT: 7.642 min Scan# 724
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47

Tgt Ion: 128	Resp:	21344
Ion	Ratio	Lower Upper
128	100	
64	54.9	41.4 62.2
130	33.8	27.0 40.6









#13
2-Methylphenol
Concen: 19.034 ng/ul
RT: 8.512 min Scan# 872
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47




Instrument :
BNA_G
ClientSampleId :
SSTD020430


Tgt Ion:108 Resp: 20635
Ion Ratio Lower Upper
108 100
107 88.9 75.8 113.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 16.969 ng/ul
RT: 8.606 min Scan# 888
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47

Tgt Ion: 45 Resp: 30823
Ion Ratio Lower Upper
45 100
77 14.0 10.2 15.4
79 11.0 8.1 12.1









#16
Acetophenone
Concen: 20.458 ng/ul
RT: 8.900 min Scan# 91
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47




Instrument :
BNA_G
ClientSampleId :
SSTD020430


Tgt Ion:105 Resp: 34896
Ion Ratio Lower Upper
105 100
77 77.4 61.0 91.6
51 29.6 22.8 34.2



#17
N-Nitroso-di-n-propylamine
Concen: 18.724 ng/ul
RT: 8.882 min Scan# 935
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47

Tgt Ion: 70 Resp: 17827
Ion Ratio Lower Upper
70 100
42 68.3 50.2 75.4
101 10.8 7.1 10.7#
130 21.2 15.5 23.3









#18
4-Methylphenol
Concen: 19.894 ng/ul
RT: 8.841 min Scan# 91
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47




Instrument :
BNA_G
ClientSampleId :
SSTD020430


Tgt Ion:108 Resp: 22798
Ion Ratio Lower Upper
108 100
107 102.2 90.7 136.1



#19
Hexachloroethane
Concen: 20.040 ng/ul
RT: 9.170 min Scan# 984
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47

Tgt Ion:117 Resp: 9114
Ion Ratio Lower Upper
117 100
201 148.7 117.4 176.0
199 92.9 64.0 96.0









#22
Nitrobenzene
Concen: 22.697 ng/ul
RT: 9.276 min Scan# 1002
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47




Instrument :
BNA_G
ClientSampleId :
SSTD020430


Tgt Ion: 77 Resp: 25459
Ion Ratio Lower Upper
77 100
123 44.3 31.2 46.8
65 15.0 10.5 15.7



#23
Isophorone
Concen: 18.780 ng/ul
RT: 9.804 min Scan# 1092
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47

Tgt Ion: 82 Resp: 49678
Ion Ratio Lower Upper
82 100
95 7.9 6.1 9.1
138 17.8 13.6 20.4









#25
2-Nitrophenol
Concen: 26.155 ng/ul
RT: 9.998 min Scan# 1125
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47




Instrument :
BNA_G
ClientSampleId :
SSTD020430


Tgt Ion:139 Resp: 11881
Ion Ratio Lower Upper
139 100
65 44.3 47.7 71.5#
109 40.6 27.1 40.7



#26
2,4-Dimethylphenol
Concen: 19.053 ng/ul
RT: 10.051 min Scan# 1134
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47

Tgt Ion:107 Resp: 26128
Ion Ratio Lower Upper
107 100
121 52.6 39.9 59.9
122 84.8 66.7 100.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 18.507 ng/ul
 RT: 10.286 min Scan# 1174
 Delta R.T. 0.000 min
 Lab File: BG053945.D
 Acq: 20 Jun 2022 12:47




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020430


Tgt Ion: 93 Resp: 28151
 Ion Ratio Lower Upper
 93 100
 95 33.0 26.5 39.7
 123 13.7 8.8 13.2#



#29
 2,4-Dichlorophenol
 Concen: 22.448 ng/ul
 RT: 10.533 min Scan# 1216
 Delta R.T. 0.000 min
 Lab File: BG053945.D
 Acq: 20 Jun 2022 12:47

Tgt Ion:162 Resp: 25811
 Ion Ratio Lower Upper
 162 100
 164 61.8 50.9 76.3
 98 29.3 28.5 42.7









#30
Naphthalene
Concen: 19.776 ng/ul
RT: 10.938 min Scan# 1285
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47




Instrument :
BNA_G
ClientSampleId :
SSTD020430


Tgt Ion:128 Resp: 72017
Ion Ratio Lower Upper
128 100
129 10.4 9.0 13.6
127 12.0 11.1 16.7



#32
4-Chloroaniline
Concen: 19.995 ng/ul
RT: 11.038 min Scan# 1302
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47

Tgt Ion:127 Resp: 31335
Ion Ratio Lower Upper
127 100
129 33.5 25.4 38.2









#33
Hexachlorobutadiene
Concen: 25.424 ng/ul
RT: 11.226 min Scan# 11
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47




Instrument :
BNA_G
ClientSampleId :
SSTD020430


Tgt Ion:225 Resp: 24261
Ion Ratio Lower Upper
225 100
223 61.3 50.2 75.2
227 59.4 51.8 77.6



#34
Caprolactam
Concen: 14.403 ng/ul
RT: 11.790 min Scan# 1430
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47

Tgt Ion:113 Resp: 4635
Ion Ratio Lower Upper
113 100
55 149.3 167.8 251.6#
56 132.2 127.6 191.4









#35
4-Chloro-3-methylphenol
Concen: 19.963 ng/ul
RT: 12.155 min Scan# 1492
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47




Instrument :
BNA_G
ClientSampleId :
SSTD020430


Tgt Ion:	107	Resp:	24221
Ion	Ratio	Lower	Upper
107	100		
144	26.0	21.3	31.9
142	82.7	65.0	97.4



#36
2-Methylnaphthalene
Concen: 20.676 ng/ul
RT: 12.537 min Scan# 1557
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47

Tgt Ion:	142	Resp:	53182
Ion	Ratio	Lower	Upper
142	100		
141	89.1	76.3	114.5





#37

1-Methylnaphthalene

Concen: 20.407 ng/ul

RT: 12.754 min Scan# 11

Delta R.T. 0.000 min

Lab File: BG053945.D

Acq: 20 Jun 2022 12:47

Instrument :

BNA_G

ClientSampleId :

SSTD020430





Tgt Ion:142 Resp: 52222

Ion Ratio Lower Upper

142 100

141 93.0 73.4 110.2

116 4.0 3.2 4.8



#39

1,2,4,5-Tetrachlorobenzene

Concen: 23.517 ng/ul

RT: 12.907 min Scan# 1620

Delta R.T. 0.000 min

Lab File: BG053945.D

Acq: 20 Jun 2022 12:47

Tgt Ion:216 Resp: 45397




Ion Ratio Lower Upper


216 100

214 72.6 58.1 87.1

179 16.7 14.8 22.2

108 10.7 10.4 15.6









#40
Hexachlorocyclopentadiene
Concen: 17.572 ng/ul
RT: 12.889 min Scan# 1617
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47




Instrument :
BNA_G
ClientSampleId :
SSTD020430


Tgt Ion:237 Resp: 23061
Ion Ratio Lower Upper
237 100
235 70.6 47.0 70.6#
272 12.7 8.6 12.8



#41
2,4,6-Trichlorophenol
Concen: 21.309 ng/ul
RT: 13.136 min Scan# 1659
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47

Tgt Ion:196 Resp: 23439
Ion Ratio Lower Upper
196 100
198 85.8 81.9 122.9
200 33.0 25.7 38.5









#42
2,4,5-Trichlorophenol
Concen: 22.838 ng/ul
RT: 13.206 min Scan# 1671
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47




Instrument :
BNA_G
ClientSampleId :
SSTD020430


Tgt Ion:196 Resp: 27021
Ion Ratio Lower Upper
196 100
198 93.7 76.3 114.5
200 27.8 25.5 38.3



#43
1,1'-Biphenyl
Concen: 19.724 ng/ul
RT: 13.541 min Scan# 1728
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47

Tgt Ion:154 Resp: 78741
Ion Ratio Lower Upper
154 100
153 39.8 32.7 49.1
76 10.9 9.0 13.6









#44
2-Chloronaphthalene
Concen: 20.717 ng/ul
RT: 13.582 min Scan# 1735
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47




Instrument :
BNA_G
ClientSampleId :
SSTD020430


Tgt Ion:	162	Resp:	64422
Ion	Ratio	Lower	Upper
162	100		
164	32.1	25.7	38.5
127	33.6	28.3	42.5



#45
2-Nitroaniline
Concen: 25.393 ng/ul
RT: 13.776 min Scan# 1768
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47

Tgt Ion:	65	Resp:	16788
Ion	Ratio	Lower	Upper
65	100		
92	61.2	52.2	78.4
138	99.8	79.0	118.4









#47
Dimethylphthalate
Concen: 21.191 ng/ul
RT: 14.146 min Scan# 1831
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47




Instrument :
BNA_G
ClientSampleId :
SSTD020430


Tgt Ion:163 Resp: 86047
Ion Ratio Lower Upper
163 100
194 4.2 3.7 5.5
164 11.0 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 29.798 ng/ul
RT: 14.264 min Scan# 1851
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47

Tgt Ion:165 Resp: 17656
Ion Ratio Lower Upper
165 100
89 49.8 44.7 67.1
121 19.9 16.9 25.3









#50
Acenaphthylene
Concen: 19.644 ng/ul
RT: 14.423 min Scan# 1878
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47




Instrument :
BNA_G
ClientSampleId :
SSTD020430


Tgt Ion:152 Resp: 99793
Ion Ratio Lower Upper
152 100
151 19.8 16.2 24.2
153 14.1 10.5 15.7



#51
3-Nitroaniline
Concen: 25.511 ng/ul
RT: 14.593 min Scan# 1907
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47

Tgt Ion:138 Resp: 14474
Ion Ratio Lower Upper
138 100
108 9.2 17.8 26.6#
92 107.8 94.2 141.4









#52
Acenaphthene
Concen: 19.559 ng/ul
RT: 14.769 min Scan# 1937
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47




Instrument :
BNA_G
ClientSampleId :
SSTD020430


Tgt Ion:153 Resp: 65948
Ion Ratio Lower Upper
153 100
152 44.5 37.5 56.3
154 83.8 68.8 103.2



#53
2,4-Dinitrophenol
Concen: 19.824 ng/ul
RT: 14.799 min Scan# 1942
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47

Tgt Ion:184 Resp: 6197
Ion Ratio Lower Upper
184 100
63 50.6 34.7 52.1
154 76.6 87.3 130.9#









#55
4-Nitrophenol
Concen: 21.398 ng/ul
RT: 14.893 min Scan# 1958
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47




Instrument :
BNA_G
ClientSampleId :
SSTD020430


Tgt Ion:109 Resp: 11775
Ion Ratio Lower Upper
109 100
139 107.2 92.9 139.3
65 114.3 97.8 146.6



#56
Dibenzofuran
Concen: 20.915 ng/ul
RT: 15.098 min Scan# 1993
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47

Tgt Ion:168 Resp: 102001
Ion Ratio Lower Upper
168 100
139 35.3 28.2 42.2









#57
2,4-Dinitrotoluene
Concen: 30.394 ng/ul
RT: 15.051 min Scan# 1985
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47




Instrument :
BNA_G
ClientSampleId :
SSTD020430


Tgt Ion:165 Resp: 25157
Ion Ratio Lower Upper
165 100
63 43.0 33.0 49.4
182 2.6 2.8 4.2#



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.967 ng/ul
RT: 15.321 min Scan# 2031
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47

Tgt Ion:232 Resp: 23028
Ion Ratio Lower Upper
232 100
131 29.0 27.5 41.3
130 2.1 1.8 2.8
166 23.3 20.7 31.1









#59
Diethylphthalate
Concen: 20.222 ng/ul
RT: 15.509 min Scan# 2063
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47




Instrument :
BNA_G
ClientSampleId :
SSTD020430


Tgt Ion:149 Resp: 84962
Ion Ratio Lower Upper
149 100
177 24.6 18.9 28.3
150 12.6 10.6 15.8



#61
Fluorene
Concen: 20.697 ng/ul
RT: 15.750 min Scan# 2104
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47

Tgt Ion:166 Resp: 82714
Ion Ratio Lower Upper
166 100
165 101.3 81.7 122.5
167 13.2 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 23.290 ng/ul
RT: 15.739 min Scan# 2102
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47




Instrument :
BNA_G
ClientSampleId :
SSTD020430


Tgt Ion:204 Resp: 52219
Ion Ratio Lower Upper
204 100
206 33.1 27.1 40.7
141 48.7 40.7 61.1



#63
4-Nitroaniline
Concen: 26.622 ng/ul
RT: 15.750 min Scan# 2104
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47

Tgt Ion:138 Resp: 14581
Ion Ratio Lower Upper
138 100
92 51.5 42.0 63.0
108 97.0 79.0 118.6









#66
 4,6-Dinitro-2-methylphenol
 Concen: 26.385 ng/ul
 RT: 15.821 min Scan# 2116
 Delta R.T. 0.000 min
 Lab File: BG053945.D
 Acq: 20 Jun 2022 12:47




Instrument :
 BNA_G
 ClientSampleId :
 SST020430


Tgt Ion:198 Resp: 14900
 Ion Ratio Lower Upper
 198 100
 182 3.0 3.4 5.0#
 77 15.2 14.7 22.1



#67
 N-Nitrosodiphenylamine
 Concen: 18.258 ng/ul
 RT: 15.950 min Scan# 2138
 Delta R.T. 0.000 min
 Lab File: BG053945.D
 Acq: 20 Jun 2022 12:47

Tgt Ion:169 Resp: 75710
 Ion Ratio Lower Upper
 169 100
 168 64.8 51.0 76.4
 167 37.5 28.7 43.1









#68
4-Bromophenyl-phenylether
Concen: 22.590 ng/ul
RT: 16.632 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47




Instrument :
BNA_G
ClientSampleId :
SSTD020430


Tgt Ion:248 Resp: 38041
Ion Ratio Lower Upper
248 100
250 100.2 78.2 117.2
141 48.3 46.7 70.1



#69
Hexachlorobenzene
Concen: 24.367 ng/ul
RT: 16.755 min Scan# 2275
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47

Tgt Ion:284 Resp: 44305
Ion Ratio Lower Upper
284 100
142 20.3 20.8 31.2#
249 31.7 27.4 41.0









#70
Atrazine
Concen: 20.284 ng/ul
RT: 16.890 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47




Instrument :
BNA_G
ClientSampleId :
SSTD020430


Tgt Ion:200 Resp: 37103
Ion Ratio Lower Upper
200 100
173 22.0 19.8 29.6
215 48.4 38.8 58.2



#71
Pentachlorophenol
Concen: 15.783 ng/ul
RT: 17.096 min Scan# 2333
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47

Tgt Ion:266 Resp: 17055
Ion Ratio Lower Upper
266 100
264 64.2 56.8 85.2
268 62.5 48.0 72.0









#72
Phenanthrene
Concen: 19.878 ng/ul
RT: 17.490 min Scan# 2400
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47




Instrument :
BNA_G
ClientSampleId :
SSTD020430


Tgt Ion:178 Resp: 155736
Ion Ratio Lower Upper
178 100
179 14.5 12.6 19.0
176 18.5 15.7 23.5



#74
Anthracene
Concen: 19.370 ng/ul
RT: 17.578 min Scan# 2415
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47

Tgt Ion:178 Resp: 154856
Ion Ratio Lower Upper
178 100
179 16.4 12.4 18.6
176 18.5 14.6 22.0









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.516 ng/uL
RT: 13.506 min Scan# 1722
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47




Instrument :
BNA_G
ClientSampleId :
SSTD020430


Tgt Ion:216 Resp: 49273
Ion Ratio Lower Upper
216 100
214 74.0 64.2 96.4
218 43.8 39.0 58.4



#76
Pentachlorobenzene
Concen: 22.145 ng/uL
RT: 15.022 min Scan# 1980
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47

Tgt Ion:250 Resp: 47770
Ion Ratio Lower Upper
250 100
248 65.9 47.7 71.5
252 63.4 50.0 75.0









#77
 Carbazole
 Concen: 18.879 ng/ul
 RT: 17.842 min Scan# 2460
 Delta R.T. 0.000 min
 Lab File: BG053945.D
 Acq: 20 Jun 2022 12:47




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020430


Tgt Ion:167	Resp:	131566
Ion	Ratio	Lower Upper
167	100	
166	21.6	18.2 27.2
139	11.6	9.0 13.6



#78
 Di-n-butylphthalate
 Concen: 16.918 ng/ul
 RT: 18.406 min Scan# 2556
 Delta R.T. 0.000 min
 Lab File: BG053945.D
 Acq: 20 Jun 2022 12:47

Tgt Ion:149	Resp:	147415
Ion	Ratio	Lower Upper
149	100	
150	9.4	7.1 10.7
104	5.0	4.2 6.2









#80
Fluoranthene
Concen: 19.365 ng/u1
RT: 19.493 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47




Instrument :
BNA_G
ClientSampleId :
SSTD020430


Tgt Ion:202 Resp: 212887
Ion Ratio Lower Upper
202 100
101 6.0 6.6 9.8#
100 4.6 5.8 8.8#



#82
Pyrene
Concen: 19.560 ng/u1
RT: 19.857 min Scan# 2803
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47

Tgt Ion:202 Resp: 213139
Ion Ratio Lower Upper
202 100
101 6.5 7.3 10.9#
100 5.7 6.2 9.4#





#83

Butylbenzylphthalate

Concen: 15.791 ng/ul

RT: 20.739 min Scan# 2953

Delta R.T. 0.000 min

Lab File: BG053945.D

Acq: 20 Jun 2022 12:47

Instrument :

BNA_G

ClientSampleId :

SSTD020430





Tgt Ion:149 Resp: 62619

Ion Ratio Lower Upper

149 100

91 64.6 54.0 81.0

206 28.1 19.5 29.3



#84

3,3'-Dichlorobenzidine

Concen: 19.686 ng/ul

RT: 21.614 min Scan# 3102

Delta R.T. 0.000 min

Lab File: BG053945.D

Acq: 20 Jun 2022 12:47




Tgt Ion:252 Resp: 77827


Ion Ratio Lower Upper

252 100

254 65.3 49.6 74.4

126 5.6 6.0 9.0#





#85

Benzo(a)anthracene

Concen: 20.024 ng/ul

RT: 21.702 min Scan# 3117

Delta R.T. 0.000 min

Lab File: BG053945.D

Acq: 20 Jun 2022 12:47

Instrument :

BNA_G

ClientSampleId :

SSTD020430





Tgt Ion:228 Resp: 224963

Ion Ratio Lower Upper

228 100

229 20.5 16.5 24.7

226 27.2 22.1 33.1



#86

Bis(2-ethylhexyl)phthalate

Concen: 15.602 ng/ul

RT: 21.608 min Scan# 3101

Delta R.T. 0.000 min

Lab File: BG053945.D

Acq: 20 Jun 2022 12:47




Tgt Ion:149 Resp: 90851


Ion Ratio Lower Upper

149 100

167 30.0 21.7 32.5

279 6.0 3.9 5.9#









#87
Chrysene
Concen: 19.820 ng/ul
RT: 21.767 min Scan# 3128
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47




Instrument :
BNA_G
ClientSampleId :
SSTD020430


Tgt Ion:228 Resp: 212838
Ion Ratio Lower Upper
228 100
226 30.6 24.2 36.4
229 19.5 16.3 24.5



#89
Di-n-octyl phthalate
Concen: 15.745 ng/ul
RT: 22.836 min Scan# 3310
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47

Tgt Ion:149 Resp: 157853









#90
Benzo(b)fluoranthene
Concen: 19.932 ng/ul
RT: 23.941 min Scan# 3498
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47




Instrument :
BNA_G
ClientSampleId :
SSTD020430


Tgt Ion:252 Resp: 230421
Ion Ratio Lower Upper
252 100
253 21.9 18.2 27.4
125 5.2 6.1 9.1#



#91
Benzo(k)fluoranthene
Concen: 20.352 ng/ul
RT: 24.017 min Scan# 3511
Delta R.T. 0.000 min
Lab File: BG053945.D
Acq: 20 Jun 2022 12:47

Tgt Ion:252 Resp: 224090
Ion Ratio Lower Upper
252 100
253 21.2 17.4 26.2
125 4.5 5.4 8.0#





#93

Benzo(a)pyrene

Concen: 20.040 ng/ul

RT: 24.828 min Scan# 3649

Delta R.T. 0.000 min

Lab File: BG053945.D

Acq: 20 Jun 2022 12:47

Instrument :

BNA_G

ClientSampleId :

SSTD020430





Tgt Ion:252 Resp: 223753

Ion Ratio Lower Upper

252 100

253 22.4 17.6 26.4

125 5.5 6.8 10.2#



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.822 ng/ul

RT: 28.700 min Scan# 4308

Delta R.T. 0.000 min

Lab File: BG053945.D

Acq: 20 Jun 2022 12:47




Tgt Ion:276 Resp: 266677


Ion Ratio Lower Upper

276 100

138 9.7 10.1 15.1#

277 24.3 19.9 29.9





#95

Dibenzo(a,h)anthracene

Concen: 20.665 ng/ul

RT: 28.765 min Scan# 4319

Delta R.T. 0.000 min

Lab File: BG053945.D

Acq: 20 Jun 2022 12:47

Instrument :

BNA_G

ClientSampleId :

SSTD020430





Tgt Ion:278 Resp: 227260

Ion Ratio Lower Upper

278 100

139 7.3 8.6 13.0#

279 23.7 19.4 29.0



#96

Benzo(g,h,i)perylene

Concen: 20.896 ng/ul

RT: 29.857 min Scan# 4505

Delta R.T. 0.000 min

Lab File: BG053945.D

Acq: 20 Jun 2022 12:47

Tgt Ion:276 Resp: 225728

Ion Ratio Lower Upper

276 100

138 9.9 11.3 16.9#

277 24.5 18.6 27.8

