

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062022\
 Data File : BG053946.D
 Acq On : 20 Jun 2022 13:26
 Operator : CG/JU
 Sample : SST04031
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SST040431

Quant Time: Jun 20 14:10:03 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG062022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 20 14:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.077	152	18803	20.000	ng/u1	0.00
20) Naphthalene-d8	10.891	136	72951	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.704	164	55920	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.448	188	150878	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.725	240	170652	20.000	ng/u1	# 0.00
88) Perylene-d12	24.980	264	175081	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.500	96	6656	13.389	ng/uL	0.00
4) Pyridine-d5	3.917	84	45758	41.749	ng/u1	0.00
7) Phenol-d5	7.230	99	56166	39.279	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.401	67	33372	36.106	ng/u1	0.00
11) 2-Chlorophenol-d4	7.612	132	45449	39.975	ng/u1	0.00
15) 4-Methylphenol-d8	8.776	113	46917	39.611	ng/u1	0.00
21) Nitrobenzene-d5	9.240	128	23291	53.410	ng/u1	0.00
24) 2-Nitrophenol-d4	9.963	143	25686	59.345	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.503	165	55614	47.311	ng/u1	0.00
31) 4-Chloroaniline-d4	11.014	131	67600	41.919	ng/u1	0.00
46) Dimethylphthalate-d6	14.105	166	187032	43.525	ng/u1	0.00
49) Acenaphthylene-d8	14.399	160	202197	42.110	ng/u1	0.00
54) 4-Nitrophenol-d4	14.880	143	26330	44.563	ng/u1	0.00
60) Fluorene-d10	15.691	176	166320	44.195	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.809	200	34666	57.467	ng/u1	0.00
73) Anthracene-d10	17.548	188	284700	41.777	ng/u1	0.00
81) Pyrene-d10	19.827	212	379080	41.736	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.757	264	387483	44.253	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.541	88	7149	13.671	ng/uL#	78
5) Pyridine	3.934	79	45490	39.250	ng/u1	94
6) Benzaldehyde	7.213	77	33069	43.908	ng/u1	95
8) Phenol	7.260	94	58924	38.497	ng/u1	96
10) Bis(2-Chloroethyl)ether	7.495	93	44442	37.914	ng/u1	88
12) 2-Chlorophenol	7.642	128	46088	39.609	ng/u1	99
13) 2-Methylphenol	8.511	108	45724	38.888	ng/u1	94
14) 2,2'-oxybis(1-Chloropr...	8.605	45	64563	32.773	ng/u1	97
16) Acetophenone	8.899	105	72896	39.404	ng/u1	96
17) N-Nitroso-di-n-propyla...	8.887	70	38486	37.270	ng/u1	100
18) 4-Methylphenol	8.840	108	47916	38.553	ng/u1	97
19) Hexachloroethane	9.169	117	19129	38.782	ng/u1	88
22) Nitrobenzene	9.281	77	55109	48.457	ng/u1	96
23) Isophorone	9.804	82	106225	39.606	ng/u1	99
25) 2-Nitrophenol	9.992	139	26570	57.691	ng/u1#	90
26) 2,4-Dimethylphenol	10.051	107	56242	40.451	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.286	93	60105	38.972	ng/u1	96
29) 2,4-Dichlorophenol	10.527	162	54463	46.718	ng/u1#	93
30) Naphthalene	10.938	128	150398	40.733	ng/u1	99
32) 4-Chloroaniline	11.038	127	66530	41.871	ng/u1	99
33) Hexachlorobutadiene	11.232	225	53024	54.803	ng/u1	97
34) Caprolactam	11.790	113	9161	28.076	ng/u1	80
35) 4-Chloro-3-methylphenol	12.154	107	51727	42.050	ng/u1	97
36) 2-Methylnaphthalene	12.536	142	112153	43.006	ng/u1	91

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062022\
 Data File : BG053946.D
 Acq On : 20 Jun 2022 13:26
 Operator : CG/JU
 Sample : SST04031
 Misc :
 ALS Vial : 5 Sample Multiplier: 1


Instrument :
 BNA_G
 ClientSampleId :
 SST040431


Quant Time: Jun 20 14:10:03 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG062022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 20 14:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.759	142	112111	43.209	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.906	216	95935	49.046	ng/ul	96
40) Hexachlorocyclopentadiene	12.889	237	54791	41.202	ng/ul#	95
41) 2,4,6-Trichlorophenol	13.135	196	52164	46.802	ng/ul	95
42) 2,4,5-Trichlorophenol	13.212	196	57501	47.963	ng/ul	95
43) 1,1'-Biphenyl	13.541	154	162314	40.126	ng/ul	98
44) 2-Chloronaphthalene	13.582	162	133859	42.482	ng/ul	98
45) 2-Nitroaniline	13.776	65	36084	53.864	ng/ul	98
47) Dimethylphthalate	14.152	163	177869	43.230	ng/ul	100
48) 2,6-Dinitrotoluene	14.269	165	37433	62.347	ng/ul	90
50) Acenaphthylene	14.428	152	207318	40.275	ng/ul	99
51) 3-Nitroaniline	14.598	138	31141	54.167	ng/ul#	91
52) Acenaphthene	14.769	153	136668	40.002	ng/ul	100
53) 2,4-Dinitrophenol	14.804	184	16775	52.960	ng/ul#	62
55) 4-Nitrophenol	14.898	109	26094	46.797	ng/ul	89
56) Dibenzofuran	15.098	168	210285	42.552	ng/ul	98
57) 2,4-Dinitrotoluene	15.057	165	55117	65.719	ng/ul#	93
58) 2,3,4,6-Tetrachlorophenol	15.327	232	53042	49.936	ng/ul#	94
59) Diethylphthalate	15.515	149	177041	41.586	ng/ul	97
61) Fluorene	15.750	166	171853	42.437	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.738	204	108105	47.583	ng/ul	94
63) 4-Nitroaniline	15.756	138	30975	55.813	ng/ul	95
66) 4,6-Dinitro-2-methylph...	15.820	198	34245	61.035	ng/ul	100
67) N-Nitrosodiphenylamine	15.950	169	155726	37.799	ng/ul	97
68) 4-Bromophenyl-phenylether	16.631	248	80204	47.938	ng/ul	96
69) Hexachlorobenzene	16.755	284	91259	50.518	ng/ul	93
70) Atrazine	16.896	200	78303	43.086	ng/ul	96
71) Pentachlorophenol	17.095	266	39924	37.186	ng/ul	89
72) Phenanthrene	17.489	178	315948	40.590	ng/ul	99
74) Anthracene	17.577	178	315901	39.772	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.505	216	103294	43.290	ng/ul	93
76) Pentachlorobenzene	15.021	250	100483	46.885	ng/ul	96
77) Carbazole	17.842	167	266206	38.449	ng/ul	99
78) Di-n-butylphthalate	18.406	149	308205	35.601	ng/ul	99
80) Fluoranthene	19.498	202	426946	40.438	ng/ul#	92
82) Pyrene	19.857	202	415393	39.692	ng/ul#	94
83) Butylbenzylphthalate	20.738	149	131322	34.480	ng/ul	95
84) 3,3'-Dichlorobenzidine	21.614	252	156898	41.323	ng/ul	96
85) Benzo(a)anthracene	21.702	228	439379	40.720	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.608	149	187705	33.562	ng/ul	98
87) Chrysene	21.772	228	422902	41.004	ng/ul	99
89) Di-n-octyl phthalate	22.836	149	326982	34.326	ng/ul	100
90) Benzo(b)fluoranthene	23.946	252	472607	43.026	ng/ul#	97
91) Benzo(k)fluoranthene	24.017	252	439332	41.994	ng/ul#	98
93) Benzo(a)pyrene	24.833	252	453167	42.717	ng/ul#	97
94) Indeno(1,2,3-cd)pyrene	28.711	276	551041	45.282	ng/ul#	97
95) Dibenzo(a,h)anthracene	28.776	278	462489	44.261	ng/ul#	96
96) Benzo(g,h,i)perylene	29.875	276	460481	44.865	ng/ul#	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 20 14:10:03 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG062022.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jun 20 14:06:57 2022
Response via : Initial Calibration





#2

1,4-Dioxane

Concen: 13.671 ng/uL

RT: 3.541 min Scan# 2

Delta R.T. -0.000 min

Lab File: BG053946.D

Acq: 20 Jun 2022 13:26

Instrument :

BNA_G

ClientSampleId :

SSTD040431





Tgt Ion: 88 Resp: 7149

Ion Ratio Lower Upper

88 100

43 31.3 48.1 72.1#

58 67.7 59.8 89.6



#5

Pyridine

Concen: 39.250 ng/uL

RT: 3.934 min Scan# 93

Delta R.T. -0.000 min

Lab File: BG053946.D

Acq: 20 Jun 2022 13:26




Tgt Ion: 79 Resp: 45490


Ion Ratio Lower Upper

79 100

52 87.5 75.4 113.2

51 41.1 35.2 52.8









#6
Benzaldehyde
Concen: 43.908 ng/ul
RT: 7.213 min Scan# 61
Delta R.T. -0.000 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26




Instrument :
BNA_G
ClientSampleId :
SSTD040431


Tgt Ion: 77 Resp: 33069
Ion Ratio Lower Upper
77 100
105 98.0 79.7 119.5
106 96.3 71.0 106.4



#8
Phenol
Concen: 38.497 ng/ul
RT: 7.260 min Scan# 659
Delta R.T. 0.005 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26

Tgt Ion: 94 Resp: 58924
Ion Ratio Lower Upper
94 100
65 31.7 24.1 36.1
66 43.4 32.5 48.7









#10
Bis(2-Chloroethyl)ether
Concen: 37.914 ng/ul
RT: 7.495 min Scan# 699
Delta R.T. -0.000 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26




Instrument :
BNA_G
ClientSampleId :
SSTD040431


Tgt Ion: 93 Resp: 44442
Ion Ratio Lower Upper
93 100
63 76.1 72.3 108.5
95 33.7 29.4 44.0



#12
2-Chlorophenol
Concen: 39.609 ng/ul
RT: 7.642 min Scan# 724
Delta R.T. -0.000 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26

Tgt Ion:128 Resp: 46088
Ion Ratio Lower Upper
128 100
64 52.6 41.4 62.2
130 34.8 27.0 40.6





#13

2-Methylphenol

Concen: 38.888 ng/ul

RT: 8.511 min Scan# 81

Delta R.T. -0.000 min

Lab File: BG053946.D

Acq: 20 Jun 2022 13:26

Instrument :

BNA_G

ClientSampleId :





SSTD040431

Tgt Ion:108 Resp: 45724

Ion Ratio Lower Upper

108 100

107 89.3 75.8 113.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 32.773 ng/ul

RT: 8.605 min Scan# 888

Delta R.T. -0.000 min

Lab File: BG053946.D

Acq: 20 Jun 2022 13:26




Tgt Ion: 45 Resp: 64563


Ion Ratio Lower Upper

45 100

77 14.9 10.2 15.4

79 10.1 8.1 12.1









#16
Acetophenone
Concen: 39.404 ng/ul
RT: 8.899 min Scan# 91
Delta R.T. -0.000 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26




Instrument :
BNA_G
ClientSampleId :
SSTD040431


Tgt Ion:105 Resp: 72896
Ion Ratio Lower Upper
105 100
77 79.3 61.0 91.6
51 31.1 22.8 34.2



#17
N-Nitroso-di-n-propylamine
Concen: 37.270 ng/ul
RT: 8.887 min Scan# 936
Delta R.T. 0.005 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26

Tgt Ion: 70 Resp: 38486
Ion Ratio Lower Upper
70 100
42 62.8 50.2 75.4
101 9.6 7.1 10.7
130 19.3 15.5 23.3









#18
4-Methylphenol
Concen: 38.553 ng/ul
RT: 8.840 min Scan# 911
Delta R.T. -0.000 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26




Instrument :
BNA_G
ClientSampleId :
SSTD040431


Tgt Ion:108 Resp: 47916
Ion Ratio Lower Upper
108 100
107 116.9 90.7 136.1



#19
Hexachloroethane
Concen: 38.782 ng/ul
RT: 9.169 min Scan# 984
Delta R.T. -0.000 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26

Tgt Ion:117 Resp: 19129
Ion Ratio Lower Upper
117 100
201 157.3 117.4 176.0
199 95.8 64.0 96.0









#22
Nitrobenzene
Concen: 48.457 ng/ul
RT: 9.281 min Scan# 1003
Delta R.T. 0.005 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26




Instrument :
BNA_G
ClientSampleId :
SSTD040431


Tgt Ion: 77 Resp: 55109
Ion Ratio Lower Upper
77 100
123 41.7 31.2 46.8
65 13.8 10.5 15.7



#23
Isophorone
Concen: 39.606 ng/ul
RT: 9.804 min Scan# 1092
Delta R.T. -0.000 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26

Tgt Ion: 82 Resp: 106225
Ion Ratio Lower Upper
82 100
95 8.1 6.1 9.1
138 17.6 13.6 20.4









#25
2-Nitrophenol
Concen: 57.691 ng/ul
RT: 9.992 min Scan# 1125
Delta R.T. -0.006 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26




Instrument :
BNA_G
ClientSampleId :
SSTD040431


Tgt Ion:139 Resp: 26570
Ion Ratio Lower Upper
139 100
65 53.4 47.7 71.5
109 40.8 27.1 40.7#



#26
2,4-Dimethylphenol
Concen: 40.451 ng/ul
RT: 10.051 min Scan# 1134
Delta R.T. -0.000 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26

Tgt Ion:107 Resp: 56242
Ion Ratio Lower Upper
107 100
121 50.3 39.9 59.9
122 81.3 66.7 100.1





#27

Bis(2-Chloroethoxy)methane

Concen: 38.972 ng/ul

RT: 10.286 min Scan# 1174

Delta R.T. -0.000 min

Lab File: BG053946.D

Acq: 20 Jun 2022 13:26

Instrument :

BNA_G

ClientSampleId :

SSTD040431





Tgt Ion: 93 Resp: 60105

Ion Ratio Lower Upper

93 100

95 30.4 26.5 39.7

123 11.7 8.8 13.2



#29

2,4-Dichlorophenol

Concen: 46.718 ng/ul

RT: 10.527 min Scan# 1215

Delta R.T. -0.006 min

Lab File: BG053946.D

Acq: 20 Jun 2022 13:26




Tgt Ion: 162 Resp: 54463


Ion Ratio Lower Upper

162 100

164 66.8 50.9 76.3

98 27.9 28.5 42.7#









#30
Naphthalene
Concen: 40.733 ng/ul
RT: 10.938 min Scan# 1285
Delta R.T. -0.000 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26




Instrument :
BNA_G
ClientSampleId :
SSTD040431


Tgt Ion:128 Resp: 150398
Ion Ratio Lower Upper
128 100
129 11.5 9.0 13.6
127 14.8 11.1 16.7



#32
4-Chloroaniline
Concen: 41.871 ng/ul
RT: 11.038 min Scan# 1302
Delta R.T. -0.000 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26

Tgt Ion:127 Resp: 66530
Ion Ratio Lower Upper
127 100
129 32.2 25.4 38.2









#33
Hexachlorobutadiene
Concen: 54.803 ng/ul
RT: 11.232 min Scan# 11
Delta R.T. 0.005 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26




Instrument :
BNA_G
ClientSampleId :
SSTD040431


Tgt Ion:225 Resp: 53024
Ion Ratio Lower Upper
225 100
223 59.9 50.2 75.2
227 63.0 51.8 77.6



#34
Caprolactam
Concen: 28.076 ng/ul
RT: 11.790 min Scan# 1430
Delta R.T. -0.000 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26

Tgt Ion:113 Resp: 9161
Ion Ratio Lower Upper
113 100
55 174.3 167.8 251.6
56 137.1 127.6 191.4









#35
4-Chloro-3-methylphenol
Concen: 42.050 ng/ul
RT: 12.154 min Scan# 1492
Delta R.T. -0.000 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26




Instrument :
BNA_G
ClientSampleId :
SSTD040431


Tgt Ion:107	Resp:	51727
Ion Ratio	Lower	Upper
107	100	
144	28.6	21.3 31.9
142	83.3	65.0 97.4



#36
2-Methylnaphthalene
Concen: 43.006 ng/ul
RT: 12.536 min Scan# 1557
Delta R.T. -0.000 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26

Tgt Ion:142	Resp:	112153
Ion Ratio	Lower	Upper
142	100	
141	86.2	76.3 114.5





#37

1-Methylnaphthalene

Concen: 43.209 ng/ul

RT: 12.759 min Scan# 1

Delta R.T. 0.005 min

Lab File: BG053946.D

Acq: 20 Jun 2022 13:26

Instrument :

BNA_G

ClientSampleId :

SSTD040431





Tgt Ion:142 Resp: 112111

Ion Ratio Lower Upper

142 100

141 89.8 73.4 110.2

116 4.4 3.2 4.8



#39

1,2,4,5-Tetrachlorobenzene

Concen: 49.046 ng/ul

RT: 12.906 min Scan# 1620

Delta R.T. -0.000 min

Lab File: BG053946.D

Acq: 20 Jun 2022 13:26

Tgt Ion:216 Resp: 95935




Ion Ratio Lower Upper


216 100

214 75.8 58.1 87.1

179 16.5 14.8 22.2

108 10.8 10.4 15.6









#40
Hexachlorocyclopentadiene
Concen: 41.202 ng/ul
RT: 12.889 min Scan# 1617
Delta R.T. -0.000 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26




Instrument :
BNA_G
ClientSampleId :
SSTD040431


Tgt Ion:237 Resp: 54791
Ion Ratio Lower Upper
237 100
235 61.9 47.0 70.6
272 13.5 8.6 12.8#



#41
2,4,6-Trichlorophenol
Concen: 46.802 ng/ul
RT: 13.135 min Scan# 1659
Delta R.T. -0.000 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26

Tgt Ion:196 Resp: 52164
Ion Ratio Lower Upper
196 100
198 95.3 81.9 122.9
200 32.2 25.7 38.5









#42
2,4,5-Trichlorophenol
Concen: 47.963 ng/ul
RT: 13.212 min Scan# 1
Delta R.T. 0.005 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26




Instrument :
BNA_G
ClientSampleId :
SSTD040431


Tgt Ion:196 Resp: 57501
Ion Ratio Lower Upper
196 100
198 89.8 76.3 114.5
200 30.2 25.5 38.3



#43
1,1'-Biphenyl
Concen: 40.126 ng/ul
RT: 13.541 min Scan# 1728
Delta R.T. -0.000 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26

Tgt Ion:154 Resp: 162314
Ion Ratio Lower Upper
154 100
153 40.2 32.7 49.1
76 9.5 9.0 13.6









#44
2-Chloronaphthalene
Concen: 42.482 ng/ul
RT: 13.582 min Scan# 1735
Delta R.T. -0.000 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26




Instrument :
BNA_G
ClientSampleId :
SSTD040431


Tgt Ion:162 Resp: 133859
Ion Ratio Lower Upper
162 100
164 31.1 25.7 38.5
127 33.8 28.3 42.5



#45
2-Nitroaniline
Concen: 53.864 ng/ul
RT: 13.776 min Scan# 1768
Delta R.T. -0.000 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26

Tgt Ion: 65 Resp: 36084
Ion Ratio Lower Upper
65 100
92 66.5 52.2 78.4
138 101.0 79.0 118.4





#47

Dimethylphthalate

Concen: 43.230 ng/ul

RT: 14.152 min Scan# 1832

Delta R.T. 0.005 min

Lab File: BG053946.D

Acq: 20 Jun 2022 13:26

Instrument :

BNA_G

ClientSampleId :

SSTD040431





Tgt Ion:163 Resp: 177869

Ion Ratio Lower Upper

163 100

194 4.4 3.7 5.5

164 9.8 7.9 11.9



#48

2,6-Dinitrotoluene

Concen: 62.347 ng/ul

RT: 14.269 min Scan# 1852

Delta R.T. 0.005 min

Lab File: BG053946.D

Acq: 20 Jun 2022 13:26




Tgt Ion:165 Resp: 37433


Ion Ratio Lower Upper

165 100

89 48.5 44.7 67.1

121 17.1 16.9 25.3









#50
Acenaphthylene
Concen: 40.275 ng/ul
RT: 14.428 min Scan# 11
Delta R.T. 0.005 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26




Instrument :
BNA_G
ClientSampleId :
SSTD040431


Tgt Ion:152 Resp: 207318
Ion Ratio Lower Upper
152 100
151 19.5 16.2 24.2
153 13.1 10.5 15.7



#51
3-Nitroaniline
Concen: 54.167 ng/ul
RT: 14.598 min Scan# 1908
Delta R.T. 0.005 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26

Tgt Ion:138 Resp: 31141
Ion Ratio Lower Upper
138 100
108 9.5 17.8 26.6#
92 111.6 94.2 141.4









#52
Acenaphthene
Concen: 40.002 ng/ul
RT: 14.769 min Scan# 1937
Delta R.T. -0.000 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26




Instrument :
BNA_G
ClientSampleId :
SSTD040431


Tgt Ion:153 Resp: 136668
Ion Ratio Lower Upper
153 100
152 46.9 37.5 56.3
154 86.2 68.8 103.2



#53
2,4-Dinitrophenol
Concen: 52.960 ng/ul
RT: 14.804 min Scan# 1943
Delta R.T. 0.005 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26

Tgt Ion:184 Resp: 16775
Ion Ratio Lower Upper
184 100
63 52.9 34.7 52.1#
154 60.1 87.3 130.9#









#55
4-Nitrophenol
Concen: 46.797 ng/ul
RT: 14.898 min Scan# 1959
Delta R.T. 0.005 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26




Instrument :
BNA_G
ClientSampleId :
SSTD040431


Tgt Ion:	109	Resp:	26094
Ion	Ratio	Lower	Upper
109	100		
139	103.1	92.9	139.3
65	110.1	97.8	146.6



#56
Dibenzofuran
Concen: 42.552 ng/ul
RT: 15.098 min Scan# 1993
Delta R.T. -0.000 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26

Tgt Ion:	168	Resp:	210285
Ion	Ratio	Lower	Upper
168	100		
139	33.9	28.2	42.2





#57

2,4-Dinitrotoluene

Concen: 65.719 ng/ul

RT: 15.057 min Scan# 1985

Delta R.T. 0.005 min

Lab File: BG053946.D

Acq: 20 Jun 2022 13:26

Instrument :

BNA_G

ClientSampleId :

SSTD040431





Tgt Ion:165 Resp: 55117

Ion Ratio Lower Upper

165 100

63 36.6 33.0 49.4

182 2.7 2.8 4.2#



#58

2,3,4,6-Tetrachlorophenol

Concen: 49.936 ng/ul

RT: 15.327 min Scan# 2032

Delta R.T. 0.005 min

Lab File: BG053946.D

Acq: 20 Jun 2022 13:26

Tgt Ion:232 Resp: 53042




Ion Ratio Lower Upper


232 100

131 29.8 27.5 41.3

130 1.7 1.8 2.8#

166 24.5 20.7 31.1









#59
Diethylphthalate
Concen: 41.586 ng/ul
RT: 15.515 min Scan# 2064
Delta R.T. 0.005 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26




Instrument :
BNA_G
ClientSampleId :
SSTD040431


Tgt Ion:149 Resp: 177041
Ion Ratio Lower Upper
149 100
177 24.9 18.9 28.3
150 12.2 10.6 15.8



#61
Fluorene
Concen: 42.437 ng/ul
RT: 15.750 min Scan# 2104
Delta R.T. -0.000 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26

Tgt Ion:166 Resp: 171853
Ion Ratio Lower Upper
166 100
165 99.7 81.7 122.5
167 13.8 11.0 16.4





#62

4-Chlorophenyl-phenylether

Concen: 47.583 ng/ul

RT: 15.738 min Scan# 2102

Delta R.T. -0.000 min

Lab File: BG053946.D

Acq: 20 Jun 2022 13:26

Instrument :

BNA_G

ClientSampleId :

SSTD040431





Tgt Ion:204 Resp: 108105

Ion Ratio Lower Upper

204 100

206 33.4 27.1 40.7

141 44.1 40.7 61.1



#63

4-Nitroaniline

Concen: 55.813 ng/ul

RT: 15.756 min Scan# 2105

Delta R.T. 0.005 min

Lab File: BG053946.D

Acq: 20 Jun 2022 13:26




Tgt Ion:138 Resp: 30975


Ion Ratio Lower Upper

138 100

92 50.0 42.0 63.0

108 92.4 79.0 118.6









#66
 4,6-Dinitro-2-methylphenol
 Concen: 61.035 ng/ul
 RT: 15.820 min Scan# 2116
 Delta R.T. -0.000 min
 Lab File: BG053946.D
 Acq: 20 Jun 2022 13:26




Instrument :
 BNA_G
 ClientSampleId :
 SSTD040431


Tgt Ion:198 Resp: 34245
 Ion Ratio Lower Upper
 198 100
 182 3.8 3.4 5.0
 77 18.5 14.7 22.1



#67
 N-Nitrosodiphenylamine
 Concen: 37.799 ng/ul
 RT: 15.950 min Scan# 2138
 Delta R.T. -0.000 min
 Lab File: BG053946.D
 Acq: 20 Jun 2022 13:26

Tgt Ion:169 Resp: 155726
 Ion Ratio Lower Upper
 169 100
 168 66.9 51.0 76.4
 167 36.9 28.7 43.1









#68
4-Bromophenyl-phenylether
Concen: 47.938 ng/ul
RT: 16.631 min Scan# 21
Delta R.T. -0.000 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26




Instrument :
BNA_G
ClientSampleId :
SSTD040431


Tgt Ion:248 Resp: 80204
Ion Ratio Lower Upper
248 100
250 99.4 78.2 117.2
141 51.8 46.7 70.1



#69
Hexachlorobenzene
Concen: 50.518 ng/ul
RT: 16.755 min Scan# 2275
Delta R.T. -0.000 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26

Tgt Ion:284 Resp: 91259
Ion Ratio Lower Upper
284 100
142 21.3 20.8 31.2
249 31.6 27.4 41.0









#70
 Atrazine
 Concen: 43.086 ng/ul
 RT: 16.896 min Scan# 21
 Delta R.T. 0.005 min
 Lab File: BG053946.D
 Acq: 20 Jun 2022 13:26




Instrument :
 BNA_G
 ClientSampleId :
 SST040431


Tgt Ion:200 Resp: 78303
 Ion Ratio Lower Upper
 200 100
 173 20.9 19.8 29.6
 215 49.6 38.8 58.2



#71
 Pentachlorophenol
 Concen: 37.186 ng/ul
 RT: 17.095 min Scan# 2333
 Delta R.T. -0.000 min
 Lab File: BG053946.D
 Acq: 20 Jun 2022 13:26

Tgt Ion:266 Resp: 39924
 Ion Ratio Lower Upper
 266 100
 264 64.8 56.8 85.2
 268 71.5 48.0 72.0









#72
Phenanthrene
Concen: 40.590 ng/u1
RT: 17.489 min Scan# 2400
Delta R.T. -0.000 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26




Instrument :
BNA_G
ClientSampleId :
SSTD040431


Tgt Ion:178 Resp: 315948
Ion Ratio Lower Upper
178 100
179 15.7 12.6 19.0
176 18.6 15.7 23.5



#74
Anthracene
Concen: 39.772 ng/u1
RT: 17.577 min Scan# 2415
Delta R.T. -0.000 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26

Tgt Ion:178 Resp: 315901
Ion Ratio Lower Upper
178 100
179 16.0 12.4 18.6
176 18.1 14.6 22.0









#75
1,2,3,4-Tetrachlorobenzene
Concen: 43.290 ng/uL
RT: 13.505 min Scan# 1722
Delta R.T. -0.000 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26




Instrument :
BNA_G
ClientSampleId :
SSTD040431


Tgt Ion:216 Resp: 103294
Ion Ratio Lower Upper
216 100
214 72.8 64.2 96.4
218 46.5 39.0 58.4



#76
Pentachlorobenzene
Concen: 46.885 ng/uL
RT: 15.021 min Scan# 1980
Delta R.T. -0.000 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26

Tgt Ion:250 Resp: 100483
Ion Ratio Lower Upper
250 100
248 62.0 47.7 71.5
252 66.2 50.0 75.0





#77

Carbazole

Concen: 38.449 ng/ul

RT: 17.842 min Scan# 2460

Delta R.T. -0.000 min

Lab File: BG053946.D

Acq: 20 Jun 2022 13:26





Instrument :

BNA_G

ClientSampleId :

SSTD040431

Tgt Ion:167	Resp:	266206	
Ion	Ratio	Lower	Upper
167	100		
166	22.2	18.2	27.2
139	11.5	9.0	13.6



#78

Di-n-butylphthalate

Concen: 35.601 ng/ul




RT: 18.406 min Scan# 2556


Delta R.T. -0.000 min

Lab File: BG053946.D

Acq: 20 Jun 2022 13:26

Tgt Ion:149	Resp:	308205	
Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.1	10.7
104	4.8	4.2	6.2









#80
Fluoranthene
Concen: 40.438 ng/u1
RT: 19.498 min Scan# 21
Delta R.T. 0.005 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26




Instrument :
BNA_G
ClientSampleId :
SSTD040431


Tgt Ion:202 Resp: 426946
Ion Ratio Lower Upper
202 100
101 5.6 6.6 9.8#
100 4.6 5.8 8.8#



#82
Pyrene
Concen: 39.692 ng/u1
RT: 19.857 min Scan# 2803
Delta R.T. -0.000 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26

Tgt Ion:202 Resp: 415393
Ion Ratio Lower Upper
202 100
101 7.0 7.3 10.9#
100 6.0 6.2 9.4#





#83

Butylbenzylphthalate

Concen: 34.480 ng/ul

RT: 20.738 min Scan# 2953

Delta R.T. -0.000 min

Lab File: BG053946.D

Acq: 20 Jun 2022 13:26

Instrument :

BNA_G

ClientSampleId :

SSTD040431





Tgt Ion:149 Resp: 131322

Ion Ratio Lower Upper

149 100

91 63.8 54.0 81.0

206 28.0 19.5 29.3



#84

3,3'-Dichlorobenzidine

Concen: 41.323 ng/ul

RT: 21.614 min Scan# 3102

Delta R.T. -0.000 min

Lab File: BG053946.D

Acq: 20 Jun 2022 13:26




Tgt Ion:252 Resp: 156898


Ion Ratio Lower Upper

252 100

254 65.0 49.6 74.4

126 6.4 6.0 9.0





#85

Benzo(a)anthracene

Concen: 40.720 ng/ul

RT: 21.702 min Scan# 3117

Delta R.T. -0.000 min

Lab File: BG053946.D

Acq: 20 Jun 2022 13:26

Instrument :

BNA_G

ClientSampleId :

SSTD040431





Tgt Ion:228 Resp: 439379

Ion Ratio Lower Upper

228 100

229 20.8 16.5 24.7

226 28.0 22.1 33.1



#86

Bis(2-ethylhexyl)phthalate

Concen: 33.562 ng/ul

RT: 21.608 min Scan# 3101

Delta R.T. -0.000 min

Lab File: BG053946.D

Acq: 20 Jun 2022 13:26




Tgt Ion:149 Resp: 187705


Ion Ratio Lower Upper

149 100

167 28.0 21.7 32.5

279 5.5 3.9 5.9





#87

Chrysene

Concen: 41.004 ng/ul

RT: 21.772 min Scan# 3129

Delta R.T. 0.005 min

Lab File: BG053946.D

Acq: 20 Jun 2022 13:26

Instrument :

BNA_G

ClientSampleId :

SSTD040431





Tgt Ion:228 Resp: 422902

Ion Ratio Lower Upper

228 100

226 31.1 24.2 36.4

229 20.0 16.3 24.5



#89

Di-n-octyl phthalate

Concen: 34.326 ng/ul




RT: 22.836 min Scan# 3310


Delta R.T. -0.000 min

Lab File: BG053946.D

Acq: 20 Jun 2022 13:26

Tgt Ion:149 Resp: 326982









#90
Benzo(b)fluoranthene
Concen: 43.026 ng/ul
RT: 23.946 min Scan# 3499
Delta R.T. 0.005 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26




Instrument :
BNA_G
ClientSampleId :
SSTD040431


Tgt Ion:252 Resp: 472607
Ion Ratio Lower Upper
252 100
253 21.9 18.2 27.4
125 4.9 6.1 9.1#



#91
Benzo(k)fluoranthene
Concen: 41.994 ng/ul
RT: 24.017 min Scan# 3511
Delta R.T. -0.000 min
Lab File: BG053946.D
Acq: 20 Jun 2022 13:26

Tgt Ion:252 Resp: 439332
Ion Ratio Lower Upper
252 100
253 21.7 17.4 26.2
125 4.7 5.4 8.0#





#93

Benzo(a)pyrene

Concen: 42.717 ng/ul

RT: 24.833 min Scan# 3

Delta R.T. 0.005 min

Lab File: BG053946.D

Acq: 20 Jun 2022 13:26

Instrument :

BNA_G

ClientSampleId :

SSTD040431





Tgt Ion:252 Resp: 453167

Ion Ratio Lower Upper

252 100

253 21.6 17.6 26.4

125 5.5 6.8 10.2#



#94

Indeno(1,2,3-cd)pyrene

Concen: 45.282 ng/ul

RT: 28.711 min Scan# 4310

Delta R.T. 0.011 min

Lab File: BG053946.D

Acq: 20 Jun 2022 13:26




Tgt Ion:276 Resp: 551041


Ion Ratio Lower Upper

276 100

138 9.3 10.1 15.1#

277 24.4 19.9 29.9









#95
 Dibenzo(a,h)anthracene
 Concen: 44.261 ng/ul
 RT: 28.776 min Scan# 41
 Delta R.T. 0.011 min
 Lab File: BG053946.D
 Acq: 20 Jun 2022 13:26

Instrument :
 BNA_G
 ClientSampleId :
 SST040431

Tgt Ion:278 Resp: 462489
 Ion Ratio Lower Upper
 278 100
 139 6.9 8.6 13.0#
 279 24.7 19.4 29.0



#96
 Benzo(g,h,i)perylene
 Concen: 44.865 ng/ul
 RT: 29.875 min Scan# 4508
 Delta R.T. 0.017 min
 Lab File: BG053946.D
 Acq: 20 Jun 2022 13:26

Tgt Ion:276 Resp: 460481
 Ion Ratio Lower Upper
 276 100
 138 10.1 11.3 16.9#
 277 25.0 18.6 27.8

