

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062022\
 Data File : BG053947.D
 Acq On : 20 Jun 2022 14:05
 Operator : CG/JU
 Sample : SSTD08032
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD080432

Quant Time: Jun 20 15:09:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG062022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 20 14:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.083	152	19619	20.000	ng/u1	0.00
20) Naphthalene-d8	10.892	136	76159	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.705	164	58190	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.449	188	156912	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.726	240	171970	20.000	ng/u1	# 0.00
88) Perylene-d12	24.987	264	183723	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.506	96	14423	27.807	ng/uL	0.00
4) Pyridine-d5	3.912	84	98175	85.847	ng/u1	0.00
7) Phenol-d5	7.231	99	118066	79.133	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.402	67	69630	72.200	ng/u1	0.00
11) 2-Chlorophenol-d4	7.613	132	96276	81.157	ng/u1	0.00
15) 4-Methylphenol-d8	8.782	113	98967	80.080	ng/u1	0.00
21) Nitrobenzene-d5	9.241	128	48425	106.368	ng/u1	0.00
24) 2-Nitrophenol-d4	9.969	143	56051	124.045	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.510	165	115928	94.466	ng/u1	0.00
31) 4-Chloroaniline-d4	11.021	131	139515	82.869	ng/u1	0.00
46) Dimethylphthalate-d6	14.105	166	376949	84.300	ng/u1	0.00
49) Acenaphthylene-d8	14.399	160	409921	82.041	ng/u1	0.00
54) 4-Nitrophenol-d4	14.893	143	55493	90.258	ng/u1	0.01
60) Fluorene-d10	15.698	176	335178	85.590	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.809	200	77682	123.823	ng/u1	0.00
73) Anthracene-d10	17.549	188	554045	78.174	ng/u1	0.00
81) Pyrene-d10	19.834	212	724492	79.153	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.775	264	778098	84.683	ng/u1	0.02
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.535	88	14960	27.418	ng/uL#	81
5) Pyridine	3.935	79	97990	81.032	ng/u1	93
6) Benzaldehyde	7.214	77	51330	65.320	ng/u1	96
8) Phenol	7.261	94	121643	76.167	ng/u1	92
10) Bis(2-Chloroethyl)ether	7.496	93	90790	74.232	ng/u1	88
12) 2-Chlorophenol	7.642	128	97425	80.247	ng/u1	96
13) 2-Methylphenol	8.518	108	93970	76.596	ng/u1	93
14) 2,2'-oxybis(1-Chloropr...	8.606	45	133769	65.078	ng/u1	99
16) Acetophenone	8.906	105	151787	78.637	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.894	70	76603	71.097	ng/u1	97
18) 4-Methylphenol	8.847	108	99672	76.861	ng/u1	99
19) Hexachloroethane	9.170	117	39875	77.479	ng/u1#	91
22) Nitrobenzene	9.282	77	113618	95.695	ng/u1	94
23) Isophorone	9.816	82	220899	78.893	ng/u1	98
25) 2-Nitrophenol	9.999	139	57049	118.651	ng/u1	93
26) 2,4-Dimethylphenol	10.051	107	116051	79.952	ng/u1	97
27) Bis(2-Chloroethoxy)met...	10.286	93	122171	75.879	ng/u1	97
29) 2,4-Dichlorophenol	10.533	162	112752	92.644	ng/u1	95
30) Naphthalene	10.944	128	311073	80.700	ng/u1	99
32) 4-Chloroaniline	11.044	127	137181	82.700	ng/u1	99
33) Hexachlorobutadiene	11.232	225	107207	106.137	ng/u1	98
34) Caprolactam	11.861	113	1932	5.672	ng/u1	88
35) 4-Chloro-3-methylphenol	12.161	107	109597	85.341	ng/u1	99
36) 2-Methylnaphthalene	12.543	142	227286	83.483	ng/u1	94


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062022\
 Data File : BG053947.D
 Acq On : 20 Jun 2022 14:05
 Operator : CG/JU
 Sample : SST08032
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SST080432

Quant Time: Jun 20 15:09:58 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG062022.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jun 20 14:06:57 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.760	142	227932	84.147	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.907	216	195080	95.843	ng/ul	94
40) Hexachlorocyclopentadiene	12.889	237	122312	88.389	ng/ul#	97
41) 2,4,6-Trichlorophenol	13.142	196	110311	95.111	ng/ul	94
42) 2,4,5-Trichlorophenol	13.212	196	120473	96.570	ng/ul	99
43) 1,1'-Biphenyl	13.541	154	325640	77.362	ng/ul	98
44) 2-Chloronaphthalene	13.583	162	269354	82.148	ng/ul	98
45) 2-Nitroaniline	13.782	65	77017	110.482	ng/ul	97
47) Dimethylphthalate	14.152	163	356302	83.218	ng/ul	99
48) 2,6-Dinitrotoluene	14.276	165	79406	127.097	ng/ul#	89
50) Acenaphthylene	14.429	152	413399	77.176	ng/ul	99
51) 3-Nitroaniline	14.599	138	57629	96.330	ng/ul#	91
52) Acenaphthene	14.769	153	277562	78.071	ng/ul	99
53) 2,4-Dinitrophenol	14.805	184	42184	127.984	ng/ul#	64
55) 4-Nitrophenol	14.910	109	53593	92.365	ng/ul	94
56) Dibenzofuran	15.104	168	417808	81.248	ng/ul	100
57) 2,4-Dinitrotoluene	15.063	165	113374	129.908	ng/ul	92
58) 2,3,4,6-Tetrachlorophenol	15.328	232	110802	100.244	ng/ul#	93
59) Diethylphthalate	15.521	149	351814	79.415	ng/ul	98
61) Fluorene	15.751	166	336355	79.819	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.739	204	214585	90.767	ng/ul	95
63) 4-Nitroaniline	15.768	138	54461	94.304	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.827	198	72907	124.946	ng/ul#	96
67) N-Nitrosodiphenylamine	15.950	169	308158	71.923	ng/ul	98
68) 4-Bromophenyl-phenylether	16.632	248	161685	92.923	ng/ul	96
69) Hexachlorobenzene	16.761	284	185892	98.946	ng/ul#	90
70) Atrazine	16.902	200	153571	81.253	ng/ul	96
71) Pentachlorophenol	17.096	266	88530	79.288	ng/ul	91
72) Phenanthrene	17.490	178	612702	75.688	ng/ul	99
74) Anthracene	17.584	178	607974	73.600	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.512	216	212556	85.655	ng/uL	96
76) Pentachlorobenzene	15.028	250	202571	90.883	ng/uL	98
77) Carbazole	17.848	167	521634	72.444	ng/ul	99
78) Di-n-butylphthalate	18.406	149	598973	66.528	ng/ul	99
80) Fluoranthene	19.499	202	809599	76.093	ng/ul#	93
82) Pyrene	19.863	202	778723	73.839	ng/ul#	94
83) Butylbenzylphthalate	20.739	149	264081	68.807	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.620	252	295178	77.146	ng/ul#	96
85) Benzo(a)anthracene	21.708	228	855340	78.663	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.614	149	377495	66.980	ng/ul#	97
87) Chrysene	21.773	228	816892	78.599	ng/ul	98
89) Di-n-octyl phthalate	22.842	149	663586	66.385	ng/ul	100
90) Benzo(b)fluoranthene	23.959	252	929922	80.678	ng/ul#	98
91) Benzo(k)fluoranthene	24.029	252	879897	80.149	ng/ul#	97
93) Benzo(a)pyrene	24.846	252	894005	80.308	ng/ul#	97
94) Indeno(1,2,3-cd)pyrene	28.730	276	1115767	87.377	ng/ul#	97
95) Dibenzo(a,h)anthracene	28.794	278	942837	85.988	ng/ul#	96
96) Benzo(g,h,i)perylene	29.899	276	934728	86.788	ng/ul#	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed






#2
1,4-Dioxane
Concen: 27.418 ng/uL
RT: 3.535 min Scan# 21
Delta R.T. -0.006 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05

Instrument :
BNA_G
ClientSampleId :
SSTD080432







Tgt Ion: 88 Resp: 14960
Ion Ratio Lower Upper
88 100
43 32.7 48.1 72.1#
58 71.0 59.8 89.6



#5
Pyridine
Concen: 81.032 ng/uL
RT: 3.935 min Scan# 93
Delta R.T. 0.000 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05

Tgt Ion: 79 Resp: 97990
Ion Ratio Lower Upper
79 100
52 84.8 75.4 113.2
51 43.2 35.2 52.8









#6
Benzaldehyde
Concen: 65.320 ng/ul
RT: 7.214 min Scan# 61
Delta R.T. 0.000 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05




Instrument :
BNA_G
ClientSampleId :
SSTD080432


Tgt Ion: 77 Resp: 51330
Ion Ratio Lower Upper
77 100
105 101.6 79.7 119.5
106 94.8 71.0 106.4



#8
Phenol
Concen: 76.167 ng/ul
RT: 7.261 min Scan# 659
Delta R.T. 0.006 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05

Tgt Ion: 94 Resp: 121643
Ion Ratio Lower Upper
94 100
65 34.5 24.1 36.1
66 45.3 32.5 48.7









#10
 Bis(2-Chloroethyl)ether
 Concen: 74.232 ng/ul
 RT: 7.496 min Scan# 699
 Delta R.T. 0.000 min
 Lab File: BG053947.D
 Acq: 20 Jun 2022 14:05




Instrument :
 BNA_G
 ClientSampleId :
 SST080432


Tgt Ion: 93	Resp: 90790
Ion Ratio	Lower Upper
93 100	
63 76.8	72.3 108.5
95 32.8	29.4 44.0



#12
 2-Chlorophenol
 Concen: 80.247 ng/ul
 RT: 7.642 min Scan# 724
 Delta R.T. 0.000 min
 Lab File: BG053947.D
 Acq: 20 Jun 2022 14:05

Tgt Ion:128	Resp:	97425	
Ion Ratio	Lower	Upper	
128 100			
64 48.2	41.4	62.2	
130 31.9	27.0	40.6	





#13

2-Methylphenol

Concen: 76.596 ng/ul

RT: 8.518 min Scan# 81

Delta R.T. 0.006 min

Lab File: BG053947.D

Acq: 20 Jun 2022 14:05

Instrument :

BNA_G

ClientSampleId :





SSTD080432

Tgt Ion:108 Resp: 93970

Ion Ratio Lower Upper

108 100

107 88.2 75.8 113.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 65.078 ng/ul

RT: 8.606 min Scan# 888

Delta R.T. 0.000 min

Lab File: BG053947.D

Acq: 20 Jun 2022 14:05




Tgt Ion: 45 Resp: 133769


Ion Ratio Lower Upper

45 100

77 13.1 10.2 15.4

79 10.2 8.1 12.1









#16
 Acetophenone
 Concen: 78.637 ng/ul
 RT: 8.906 min Scan# 91
 Delta R.T. 0.006 min
 Lab File: BG053947.D
 Acq: 20 Jun 2022 14:05




Instrument :
 BNA_G
 ClientSampleId :
 SSTD080432


Tgt Ion:105 Resp: 151787
 Ion Ratio Lower Upper
 105 100
 77 76.9 61.0 91.6
 51 28.4 22.8 34.2



#17
 N-Nitroso-di-n-propylamine
 Concen: 71.097 ng/ul
 RT: 8.894 min Scan# 937
 Delta R.T. 0.012 min
 Lab File: BG053947.D
 Acq: 20 Jun 2022 14:05

Tgt Ion: 70 Resp: 76603
 Ion Ratio Lower Upper
 70 100
 42 64.2 50.2 75.4
 101 10.1 7.1 10.7
 130 21.6 15.5 23.3









#18
4-Methylphenol
Concen: 76.861 ng/ul
RT: 8.847 min Scan# 911
Delta R.T. 0.006 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05




Instrument :
BNA_G
ClientSampleId :
SSTD080432


Tgt Ion:108 Resp: 99672
Ion Ratio Lower Upper
108 100
107 112.5 90.7 136.1



#19
Hexachloroethane
Concen: 77.479 ng/ul
RT: 9.170 min Scan# 984
Delta R.T. 0.000 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05

Tgt Ion:117 Resp: 39875
Ion Ratio Lower Upper
117 100
201 152.0 117.4 176.0
199 96.8 64.0 96.0#









#22
Nitrobenzene
Concen: 95.695 ng/ul
RT: 9.282 min Scan# 1003
Delta R.T. 0.006 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05




Instrument :
BNA_G
ClientSampleId :
SSTD080432


Tgt Ion: 77 Resp: 113618
Ion Ratio Lower Upper
77 100
123 43.7 31.2 46.8
65 14.1 10.5 15.7



#23
Isophorone
Concen: 78.893 ng/ul
RT: 9.816 min Scan# 1094
Delta R.T. 0.012 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05

Tgt Ion: 82 Resp: 220899
Ion Ratio Lower Upper
82 100
95 8.7 6.1 9.1
138 17.6 13.6 20.4









#25
2-Nitrophenol
Concen: 118.651 ng/ul
RT: 9.999 min Scan# 1125
Delta R.T. 0.000 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05




Instrument :
BNA_G
ClientSampleId :
SSTD080432


Tgt Ion:139 Resp: 57049
Ion Ratio Lower Upper
139 100
65 53.5 47.7 71.5
109 36.9 27.1 40.7



#26
2,4-Dimethylphenol
Concen: 79.952 ng/ul
RT: 10.051 min Scan# 1134
Delta R.T. 0.000 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05

Tgt Ion:107 Resp: 116051
Ion Ratio Lower Upper
107 100
121 48.0 39.9 59.9
122 80.1 66.7 100.1





#27

Bis(2-Chloroethoxy)methane

Concen: 75.879 ng/ul

RT: 10.286 min Scan# 1174

Delta R.T. 0.000 min

Lab File: BG053947.D

Acq: 20 Jun 2022 14:05

Instrument :

BNA_G

ClientSampleId :

SSTD080432





Tgt Ion: 93 Resp: 122171

Ion Ratio Lower Upper

93 100

95 30.8 26.5 39.7

123 11.0 8.8 13.2



#29

2,4-Dichlorophenol

Concen: 92.644 ng/ul

RT: 10.533 min Scan# 1216

Delta R.T. 0.000 min

Lab File: BG053947.D

Acq: 20 Jun 2022 14:05




Tgt Ion:162 Resp: 112752


Ion Ratio Lower Upper

162 100

164 65.5 50.9 76.3

98 30.1 28.5 42.7









#30
Naphthalene
Concen: 80.700 ng/ul
RT: 10.944 min Scan# 11073
Delta R.T. 0.006 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05




Instrument :
BNA_G
ClientSampleId :
SSTD080432


Tgt Ion:128 Resp: 311073
Ion Ratio Lower Upper
128 100
129 10.8 9.0 13.6
127 13.6 11.1 16.7



#32
4-Chloroaniline
Concen: 82.700 ng/ul
RT: 11.044 min Scan# 1303
Delta R.T. 0.006 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05

Tgt Ion:127 Resp: 137181
Ion Ratio Lower Upper
127 100
129 32.2 25.4 38.2









#33
Hexachlorobutadiene
Concen: 106.137 ng/ul
RT: 11.232 min Scan# 11
Delta R.T. 0.006 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05




Instrument :
BNA_G
ClientSampleId :
SSTD080432


Tgt Ion:225 Resp: 107207
Ion Ratio Lower Upper
225 100
223 64.2 50.2 75.2
227 62.5 51.8 77.6



#34
Caprolactam
Concen: 5.672 ng/ul
RT: 11.861 min Scan# 1442
Delta R.T. 0.071 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05

Tgt Ion:113 Resp: 1932
Ion Ratio Lower Upper
113 100
55 192.8 167.8 251.6
56 142.3 127.6 191.4









#35
 4-Chloro-3-methylphenol
 Concen: 85.341 ng/ul
 RT: 12.161 min Scan# 1493
 Delta R.T. 0.006 min
 Lab File: BG053947.D
 Acq: 20 Jun 2022 14:05




Instrument :
 BNA_G
 ClientSampleId :
 SSTD080432


Tgt Ion:107 Resp: 109597
 Ion Ratio Lower Upper
 107 100
 144 27.3 21.3 31.9
 142 80.3 65.0 97.4



#36
 2-Methylnaphthalene
 Concen: 83.483 ng/ul
 RT: 12.543 min Scan# 1558
 Delta R.T. 0.006 min
 Lab File: BG053947.D
 Acq: 20 Jun 2022 14:05

Tgt Ion:142 Resp: 227286
 Ion Ratio Lower Upper
 142 100
 141 89.2 76.3 114.5





#37

1-Methylnaphthalene

Concen: 84.147 ng/ul

RT: 12.760 min Scan# 11

Delta R.T. 0.006 min

Lab File: BG053947.D

Acq: 20 Jun 2022 14:05

Instrument :

BNA_G

ClientSampleId :

SSTD080432





Tgt Ion:142 Resp: 227932

Ion Ratio Lower Upper

142 100

141 92.0 73.4 110.2

116 4.0 3.2 4.8



#39

1,2,4,5-Tetrachlorobenzene

Concen: 95.843 ng/ul

RT: 12.907 min Scan# 1620

Delta R.T. 0.000 min

Lab File: BG053947.D

Acq: 20 Jun 2022 14:05

Tgt Ion:216 Resp: 195080




Ion Ratio Lower Upper


216 100

214 77.8 58.1 87.1

179 16.9 14.8 22.2

108 10.6 10.4 15.6









#40
Hexachlorocyclopentadiene
Concen: 88.389 ng/ul
RT: 12.889 min Scan# 1617
Delta R.T. 0.000 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05




Instrument :
BNA_G
ClientSampleId :
SSTD080432


Tgt Ion:237 Resp: 122312
Ion Ratio Lower Upper
237 100
235 60.2 47.0 70.6
272 13.4 8.6 12.8#



#41
2,4,6-Trichlorophenol
Concen: 95.111 ng/ul
RT: 13.142 min Scan# 1660
Delta R.T. 0.006 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05

Tgt Ion:196 Resp: 110311
Ion Ratio Lower Upper
196 100
198 94.7 81.9 122.9
200 32.3 25.7 38.5





#42

2,4,5-Trichlorophenol

Concen: 96.570 ng/ul

RT: 13.212 min Scan# 1

Delta R.T. 0.006 min

Lab File: BG053947.D

Acq: 20 Jun 2022 14:05

Instrument :

BNA_G

ClientSampleId :

SSTD080432





Tgt Ion:196 Resp: 120473

Ion Ratio Lower Upper

196 100

198 94.1 76.3 114.5

200 32.0 25.5 38.3



#43

1,1'-Biphenyl

Concen: 77.362 ng/ul

RT: 13.541 min Scan# 1728

Delta R.T. 0.000 min

Lab File: BG053947.D

Acq: 20 Jun 2022 14:05




Tgt Ion:154 Resp: 325640


Ion Ratio Lower Upper

154 100

153 41.9 32.7 49.1

76 10.7 9.0 13.6





#44

2-Chloronaphthalene

Concen: 82.148 ng/ul

RT: 13.583 min Scan# 1

Delta R.T. 0.000 min

Lab File: BG053947.D

Acq: 20 Jun 2022 14:05

Instrument :

BNA_G

ClientSampleId :

SSTD080432





Tgt Ion:162 Resp: 269354

Ion Ratio Lower Upper

162 100

164 33.1 25.7 38.5

127 34.7 28.3 42.5



#45

2-Nitroaniline

Concen: 110.482 ng/ul

RT: 13.782 min Scan# 1769

Delta R.T. 0.006 min

Lab File: BG053947.D

Acq: 20 Jun 2022 14:05




Tgt Ion: 65 Resp: 77017


Ion Ratio Lower Upper

65 100

92 65.2 52.2 78.4

138 104.2 79.0 118.4









#47
Dimethylphthalate
Concen: 83.218 ng/ul
RT: 14.152 min Scan# 1832
Delta R.T. 0.006 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05




Instrument :
BNA_G
ClientSampleId :
SSTD080432


Tgt Ion:163 Resp: 356302
Ion Ratio Lower Upper
163 100
194 4.3 3.7 5.5
164 10.3 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 127.097 ng/ul
RT: 14.276 min Scan# 1853
Delta R.T. 0.012 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05

Tgt Ion:165 Resp: 79406
Ion Ratio Lower Upper
165 100
89 47.1 44.7 67.1
121 16.7 16.9 25.3#









#50
Acenaphthylene
Concen: 77.176 ng/ul
RT: 14.429 min Scan# 11
Delta R.T. 0.006 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05




Instrument :
BNA_G
ClientSampleId :
SSTD080432


Tgt Ion:152 Resp: 413399
Ion Ratio Lower Upper
152 100
151 20.5 16.2 24.2
153 14.0 10.5 15.7



#51
3-Nitroaniline
Concen: 96.330 ng/ul
RT: 14.599 min Scan# 1908
Delta R.T. 0.006 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05

Tgt Ion:138 Resp: 57629
Ion Ratio Lower Upper
138 100
108 10.5 17.8 26.6#
92 111.0 94.2 141.4









#52
Acenaphthene
Concen: 78.071 ng/ul
RT: 14.769 min Scan# 1937
Delta R.T. 0.000 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05




Instrument :
BNA_G
ClientSampleId :
SSTD080432


Tgt Ion:153 Resp: 277562
Ion Ratio Lower Upper
153 100
152 46.0 37.5 56.3
154 85.8 68.8 103.2



#53
2,4-Dinitrophenol
Concen: 127.984 ng/ul
RT: 14.805 min Scan# 1943
Delta R.T. 0.006 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05

Tgt Ion:184 Resp: 42184
Ion Ratio Lower Upper
184 100
63 51.3 34.7 52.1
154 61.1 87.3 130.9#









#55
4-Nitrophenol
Concen: 92.365 ng/ul
RT: 14.910 min Scan# 1961
Delta R.T. 0.018 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05




Instrument :
BNA_G
ClientSampleId :
SSTD080432


Tgt Ion:109 Resp: 53593
Ion Ratio Lower Upper
109 100
139 109.7 92.9 139.3
65 114.5 97.8 146.6



#56
Dibenzofuran
Concen: 81.248 ng/ul
RT: 15.104 min Scan# 1994
Delta R.T. 0.006 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05

Tgt Ion:168 Resp: 417808
Ion Ratio Lower Upper
168 100
139 35.0 28.2 42.2





#57

2,4-Dinitrotoluene

Concen: 129.908 ng/ul

RT: 15.063 min Scan# 1985

Delta R.T. 0.012 min

Lab File: BG053947.D

Acq: 20 Jun 2022 14:05

Instrument :

BNA_G

ClientSampleId :

SSTD080432





Tgt Ion:165 Resp: 113374

Ion Ratio Lower Upper

165 100

63 36.2 33.0 49.4

182 2.9 2.8 4.2



#58

2,3,4,6-Tetrachlorophenol

Concen: 100.244 ng/ul

RT: 15.328 min Scan# 2032

Delta R.T. 0.006 min

Lab File: BG053947.D

Acq: 20 Jun 2022 14:05

Tgt Ion:232 Resp: 110802




Ion Ratio Lower Upper


232 100

131 29.8 27.5 41.3

130 1.7 1.8 2.8#

166 22.3 20.7 31.1









#59
Diethylphthalate
Concen: 79.415 ng/ul
RT: 15.521 min Scan# 2065
Delta R.T. 0.012 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05




Instrument :
BNA_G
ClientSampleId :
SSTD080432


Tgt Ion:149 Resp: 351814
Ion Ratio Lower Upper
149 100
177 25.2 18.9 28.3
150 12.9 10.6 15.8



#61
Fluorene
Concen: 79.819 ng/ul
RT: 15.751 min Scan# 2104
Delta R.T. 0.000 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05

Tgt Ion:166 Resp: 336355
Ion Ratio Lower Upper
166 100
165 100.2 81.7 122.5
167 13.7 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 90.767 ng/ul
RT: 15.739 min Scan# 2102
Delta R.T. 0.000 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05




Instrument :
BNA_G
ClientSampleId :
SSTD080432


Tgt Ion:204 Resp: 214585
Ion Ratio Lower Upper
204 100
206 33.7 27.1 40.7
141 45.3 40.7 61.1



#63
4-Nitroaniline
Concen: 94.304 ng/ul
RT: 15.768 min Scan# 2107
Delta R.T. 0.018 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05

Tgt Ion:138 Resp: 54461
Ion Ratio Lower Upper
138 100
92 50.2 42.0 63.0
108 94.9 79.0 118.6









#66
 4,6-Dinitro-2-methylphenol
 Concen: 124.946 ng/ul
 RT: 15.827 min Scan# 2116
 Delta R.T. 0.006 min
 Lab File: BG053947.D
 Acq: 20 Jun 2022 14:05




Instrument :
 BNA_G
 ClientSampleId :
 SST080432


Tgt Ion:198 Resp: 72907
 Ion Ratio Lower Upper
 198 100
 182 3.3 3.4 5.0#
 77 20.3 14.7 22.1



#67
 N-Nitrosodiphenylamine
 Concen: 71.923 ng/ul
 RT: 15.950 min Scan# 2138
 Delta R.T. 0.000 min
 Lab File: BG053947.D
 Acq: 20 Jun 2022 14:05

Tgt Ion:169 Resp: 308158
 Ion Ratio Lower Upper
 169 100
 168 65.8 51.0 76.4
 167 36.4 28.7 43.1









#68
 4-Bromophenyl-phenylether
 Concen: 92.923 ng/ul
 RT: 16.632 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BG053947.D
 Acq: 20 Jun 2022 14:05




Instrument :
 BNA_G
 ClientSampleId :
 SSTD080432


Tgt Ion:248 Resp: 161685
 Ion Ratio Lower Upper
 248 100
 250 97.2 78.2 117.2
 141 51.5 46.7 70.1



#69
 Hexachlorobenzene
 Concen: 98.946 ng/ul
 RT: 16.761 min Scan# 2276
 Delta R.T. 0.006 min
 Lab File: BG053947.D
 Acq: 20 Jun 2022 14:05

Tgt Ion:284 Resp: 185892
 Ion Ratio Lower Upper
 284 100
 142 19.3 20.8 31.2#
 249 30.3 27.4 41.0





#70

Atrazine

Concen: 81.253 ng/ul

RT: 16.902 min Scan# 21

Delta R.T. 0.012 min

Lab File: BG053947.D

Acq: 20 Jun 2022 14:05

Instrument :

BNA_G

ClientSampleId :

SSTD080432





Tgt Ion:200 Resp: 153571

Ion Ratio Lower Upper

200 100

173 21.9 19.8 29.6

215 51.0 38.8 58.2



#71

Pentachlorophenol

Concen: 79.288 ng/ul

RT: 17.096 min Scan# 2333

Delta R.T. 0.000 min

Lab File: BG053947.D

Acq: 20 Jun 2022 14:05




Tgt Ion:266 Resp: 88530


Ion Ratio Lower Upper

266 100

264 65.3 56.8 85.2

268 69.1 48.0 72.0









#72
Phenanthrene
Concen: 75.688 ng/u1
RT: 17.490 min Scan# 2400
Delta R.T. 0.000 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05




Instrument :
BNA_G
ClientSampleId :
SSTD080432


Tgt Ion:178 Resp: 612702
Ion Ratio Lower Upper
178 100
179 16.2 12.6 19.0
176 19.3 15.7 23.5



#74
Anthracene
Concen: 73.600 ng/u1
RT: 17.584 min Scan# 2416
Delta R.T. 0.006 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05

Tgt Ion:178 Resp: 607974
Ion Ratio Lower Upper
178 100
179 15.9 12.4 18.6
176 18.3 14.6 22.0









#75
1,2,3,4-Tetrachlorobenzene
Concen: 85.655 ng/uL
RT: 13.512 min Scan# 1723
Delta R.T. 0.006 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05




Instrument :
BNA_G
ClientSampleId :
SSTD080432


Tgt Ion:216 Resp: 212556
Ion Ratio Lower Upper
216 100
214 75.9 64.2 96.4
218 50.3 39.0 58.4



#76
Pentachlorobenzene
Concen: 90.883 ng/uL
RT: 15.028 min Scan# 1981
Delta R.T. 0.006 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05

Tgt Ion:250 Resp: 202571
Ion Ratio Lower Upper
250 100
248 58.2 47.7 71.5
252 60.2 50.0 75.0









#77
 Carbazole
 Concen: 72.444 ng/ul
 RT: 17.848 min Scan# 2461
 Delta R.T. 0.006 min
 Lab File: BG053947.D
 Acq: 20 Jun 2022 14:05




Instrument :
 BNA_G
 ClientSampleId :
 SSTD080432


Tgt Ion:167 Resp: 521634
 Ion Ratio Lower Upper
 167 100
 166 22.8 18.2 27.2
 139 11.8 9.0 13.6



#78
 Di-n-butylphthalate
 Concen: 66.528 ng/ul
 RT: 18.406 min Scan# 2556
 Delta R.T. 0.000 min
 Lab File: BG053947.D
 Acq: 20 Jun 2022 14:05

Tgt Ion:149 Resp: 598973
 Ion Ratio Lower Upper
 149 100
 150 9.7 7.1 10.7
 104 5.2 4.2 6.2





#80

Fluoranthene

Concen: 76.093 ng/u1

RT: 19.499 min Scan# 21

Delta R.T. 0.006 min

Lab File: BG053947.D

Acq: 20 Jun 2022 14:05

Instrument :

BNA_G

ClientSampleId :

SSTD080432





Tgt Ion:202 Resp: 809599

Ion Ratio Lower Upper

202 100

101 5.8 6.6 9.8#

100 4.8 5.8 8.8#



#82

Pyrene

Concen: 73.839 ng/u1

RT: 19.863 min Scan# 2804

Delta R.T. 0.006 min

Lab File: BG053947.D

Acq: 20 Jun 2022 14:05




Tgt Ion:202 Resp: 778723


Ion Ratio Lower Upper

202 100

101 6.7 7.3 10.9#

100 5.8 6.2 9.4#









#83
Butylbenzylphthalate
Concen: 68.807 ng/ul
RT: 20.739 min Scan# 2953
Delta R.T. 0.000 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05




Instrument :
BNA_G
ClientSampleId :
SSTD080432


Tgt Ion:149 Resp: 264081
Ion Ratio Lower Upper
149 100
91 66.2 54.0 81.0
206 26.9 19.5 29.3



#84
3,3'-Dichlorobenzidine
Concen: 77.146 ng/ul
RT: 21.620 min Scan# 3103
Delta R.T. 0.006 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05

Tgt Ion:252 Resp: 295178
Ion Ratio Lower Upper
252 100
254 64.8 49.6 74.4
126 5.9 6.0 9.0#





#85

Benzo(a)anthracene

Concen: 78.663 ng/ul

RT: 21.708 min Scan# 3117

Delta R.T. 0.006 min

Lab File: BG053947.D

Acq: 20 Jun 2022 14:05

Instrument :

BNA_G

ClientSampleId :

SSTD080432





Tgt Ion:228 Resp: 855340

Ion Ratio Lower Upper

228 100

229 20.9 16.5 24.7

226 28.6 22.1 33.1



#86

Bis(2-ethylhexyl)phthalate

Concen: 66.980 ng/ul

RT: 21.614 min Scan# 3102

Delta R.T. 0.006 min

Lab File: BG053947.D

Acq: 20 Jun 2022 14:05




Tgt Ion:149 Resp: 377495


Ion Ratio Lower Upper

149 100

167 28.6 21.7 32.5

279 6.3 3.9 5.9#





#87

Chrysene

Concen: 78.599 ng/ul

RT: 21.773 min Scan# 31

Delta R.T. 0.006 min

Lab File: BG053947.D

Acq: 20 Jun 2022 14:05

Instrument :

BNA_G

ClientSampleId :

SSTD080432





Tgt Ion:228 Resp: 816892

Ion Ratio Lower Upper

228 100

226 31.7 24.2 36.4

229 21.2 16.3 24.5



#89

Di-n-octyl phthalate

Concen: 66.385 ng/ul




RT: 22.842 min Scan# 3311


Delta R.T. 0.006 min

Lab File: BG053947.D

Acq: 20 Jun 2022 14:05

Tgt Ion:149 Resp: 663586









#90
Benzo(b)fluoranthene
Concen: 80.678 ng/ul
RT: 23.959 min Scan# 311
Delta R.T. 0.018 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05




Instrument :
BNA_G
ClientSampleId :
SSTD080432


Tgt Ion:252 Resp: 929922
Ion Ratio Lower Upper
252 100
253 23.1 18.2 27.4
125 5.0 6.1 9.1#



#91
Benzo(k)fluoranthene
Concen: 80.149 ng/ul
RT: 24.029 min Scan# 3513
Delta R.T. 0.012 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05

Tgt Ion:252 Resp: 879897
Ion Ratio Lower Upper
252 100
253 22.7 17.4 26.2
125 4.8 5.4 8.0#





#93

Benzo(a)pyrene

Concen: 80.308 ng/ul

RT: 24.846 min Scan# 3

Delta R.T. 0.018 min

Lab File: BG053947.D

Acq: 20 Jun 2022 14:05

Instrument :

BNA_G

ClientSampleId :

SSTD080432





Tgt Ion:252 Resp: 894005

Ion Ratio Lower Upper

252 100

253 21.9 17.6 26.4

125 5.2 6.8 10.2#



#94

Indeno(1,2,3-cd)pyrene

Concen: 87.377 ng/ul

RT: 28.730 min Scan# 4313

Delta R.T. 0.030 min

Lab File: BG053947.D

Acq: 20 Jun 2022 14:05




Tgt Ion:276 Resp: 1115767


Ion Ratio Lower Upper

276 100

138 8.9 10.1 15.1#

277 24.9 19.9 29.9









#95
Dibenzo(a,h)anthracene
Concen: 85.988 ng/ul
RT: 28.794 min Scan# 41
Delta R.T. 0.030 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05

Instrument :
BNA_G
ClientSampleId :
SSTD080432

Tgt Ion:278 Resp: 942837
Ion Ratio Lower Upper
278 100
139 7.1 8.6 13.0#
279 23.8 19.4 29.0



#96
Benzo(g,h,i)perylene
Concen: 86.788 ng/ul
RT: 29.899 min Scan# 4512
Delta R.T. 0.041 min
Lab File: BG053947.D
Acq: 20 Jun 2022 14:05

Tgt Ion:276 Resp: 934728
Ion Ratio Lower Upper
276 100
138 9.8 11.3 16.9#
277 24.4 18.6 27.8

