


Data Path : Z:\HPCHEM1\BNA G\DATA\BG062116\
 Data File : BG022467.D
 Acq On : 21 Jun 2016 15:12
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG060616.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Jun 22 01:27:56 2016



AutoFind: Scans 2544, 2545, 2546; Background Corrected with Scan 2537

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.0	5952	PASS
68	69	0.00	2	0.9	66	PASS
69	198	0.00	100	38.6	7653	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	58.6	11627	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	19840	PASS
199	198	5	9	6.3	1256	PASS
275	198	10	60	22.6	4488	PASS
365	198	1	100	2.9	572	PASS
441	442	0.01	24	19.2	2027	PASS
442	198	50	100	53.3	10565	PASS
443	442	15	24	20.1	2120	PASS