


Data Path : Z:\HPCHEM1\BNA_G\Data\BG062117\
 Data File : BG027577.D
 Acq On : 22 Jun 2017 2:01
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG062117.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Jun 21 18:52:15 2017



AutoFind: Scans 2440, 2441, 2442; Background Corrected with Scan 2434

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	45.4	12491	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	45.4	12491	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	50.5	13890	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	27498	PASS
199	198	5	9	6.7	1830	PASS
275	198	10	60	28.7	7885	PASS
365	198	1	100	3.9	1068	PASS
441	442	0.01	24	14.9	3747	PASS
442	198	50	100	91.2	25080	PASS
443	442	15	24	19.8	4973	PASS