

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062121\
 Data File : BG049028.D
 Acq On : 21 Jun 2021 11:34
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 21 13:28:09 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG061121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 16 12:04:50 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	71	0.00
2	1,4-Dioxane	40.000	37.343	6.6	76	0.00
3	Pyridine	40.000	38.644	3.4	74	0.00
4	n-Nitrosodimethylamine	40.000	44.710	-11.8	83	0.00
5 S	2-Fluorophenol	80.000	78.838	1.5	79	0.00
6	Aniline	40.000	38.483	3.8	76	0.00
7 S	Phenol-d6	80.000	77.542	3.1	76	0.00
8	2-Chlorophenol	40.000	39.608	1.0	80	0.00
9	Benzaldehyde	40.000	38.814	3.0	81	0.00
10 C	Phenol	40.000	38.954	2.6	78	0.00
11	bis(2-Chloroethyl)ether	40.000	37.537	6.2	74	0.00
12	1,3-Dichlorobenzene	40.000	40.097	-0.2	79	0.00
13 C	1,4-Dichlorobenzene	40.000	39.120	2.2	79	0.00
14	1,2-Dichlorobenzene	40.000	38.830	2.9	78	0.00
15	Benzyl Alcohol	40.000	37.884	5.3	77	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	35.362	11.6	71	0.00
17	2-Methylphenol	40.000	39.652	0.9	78	0.00
18	Hexachloroethane	40.000	39.034	2.4	77	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	35.791	10.5	72	0.00
20	3+4-Methylphenols	40.000	39.800	0.5	78	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	74	0.00
22	Acetophenone	40.000	35.959	10.1	77	0.00
23 S	Nitrobenzene-d5	80.000	76.250	4.7	80	0.00
24	Nitrobenzene	40.000	36.434	8.9	77	0.00
25	Isophorone	40.000	34.535	13.7	74	0.00
26 C	2-Nitrophenol	40.000	40.768	-1.9	85	0.00
27	2,4-Dimethylphenol	40.000	35.599	11.0	76	0.00
28	bis(2-Chloroethoxy)methane	40.000	34.876	12.8	74	0.00
29 C	2,4-Dichlorophenol	40.000	36.813	8.0	78	0.00
30	1,2,4-Trichlorobenzene	40.000	37.710	5.7	81	0.00
31	Naphthalene	40.000	36.043	9.9	77	0.00
32	Benzoic acid	40.000	41.154	-2.9	85	0.00
33	4-Chloroaniline	40.000	36.621	8.4	79	0.00
34 C	Hexachlorobutadiene	40.000	37.756	5.6	81	0.00
35	Caprolactam	40.000	42.723	-6.8	91	0.00
36 C	4-Chloro-3-methylphenol	40.000	37.227	6.9	79	0.00
37	2-Methylnaphthalene	40.000	36.343	9.1	78	0.00
38	1-Methylnaphthalene	40.000	36.671	8.3	78	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	75	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	37.097	7.3	83	0.00
41 P	Hexachlorocyclopentadiene	40.000	34.994	12.5	77	0.00
42 S	2,4,6-Tribromophenol	80.000	81.250	-1.6	89	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.315	-3.3	89	0.00
44	2,4,5-Trichlorophenol	40.000	42.460	-6.2	92	0.00
45 S	2-Fluorobiphenyl	80.000	71.633	10.5	78	0.00
46	1,1'-Biphenyl	40.000	35.610	11.0	78	0.00
47	2-Chloronaphthalene	40.000	37.062	7.3	83	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062121\
 Data File : BG049028.D
 Acq On : 21 Jun 2021 11:34
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 21 13:28:09 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG061121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 16 12:04:50 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.293	-3.2	90	0.00
49	Acenaphthylene	40.000	36.294	9.3	80	0.00
50	Dimethylphthalate	40.000	37.858	5.4	83	0.00
51	2,6-Dinitrotoluene	40.000	41.301	-3.3	89	0.00
52 C	Acenaphthene	40.000	37.107	7.2	81	0.00
53	3-Nitroaniline	40.000	41.729	-4.3	90	0.00
54 P	2,4-Dinitrophenol	40.000	46.970	-17.4	100	0.00
55	Dibenzofuran	40.000	36.795	8.0	81	0.00
56 P	4-Nitrophenol	40.000	43.703	-9.3	91	0.00
57	2,4-Dinitrotoluene	40.000	49.382	-23.5	95	0.00
58	Fluorene	40.000	36.882	7.8	82	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.168	2.1	84	0.00
60	Diethylphthalate	40.000	37.857	5.4	84	0.00
61	4-Chlorophenyl-phenylether	40.000	38.831	2.9	87	0.00
62	4-Nitroaniline	40.000	44.198	-10.5	94	0.00
63	Azobenzene	40.000	34.304	14.2	77	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	79	0.00
65	4,6-Dinitro-2-methylphenol	40.000	47.314	-18.3	100	0.00
66 c	n-Nitrosodiphenylamine	40.000	34.890	12.8	81	0.00
67	4-Bromophenyl-phenylether	40.000	35.525	11.2	82	0.00
68	Hexachlorobenzene	40.000	36.802	8.0	87	0.00
69	Atrazine	40.000	36.416	9.0	79	0.00
70 C	Pentachlorophenol	40.000	38.794	3.0	88	0.00
71	Phenanthrene	40.000	35.548	11.1	84	0.00
72	Anthracene	40.000	35.526	11.2	85	0.00
73	Carbazole	40.000	36.274	9.3	85	0.00
74	Di-n-butylphthalate	40.000	37.873	5.3	90	0.00
75 C	Fluoranthene	40.000	37.594	6.0	90	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	83	0.00
77	Benzydine	40.000	42.663	-6.7	90	0.00
78	Pyrene	40.000	36.864	7.8	90	0.00
79 S	Terphenyl-d14	80.000	80.965	-1.2	98	0.00
80	Butylbenzylphthalate	40.000	38.731	3.2	94	0.00
81	Benzo(a)anthracene	40.000	37.504	6.2	91	0.00
82	3,3'-Dichlorobenzidine	40.000	41.963	-4.9	100	0.00
83	Chrysene	40.000	37.541	6.1	91	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	38.137	4.7	93	0.00
85 c	Di-n-octyl phthalate	40.000	38.354	4.1	93	0.00
86 I	Perylene-d12	20.000	20.000	0.0	86	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.252	-0.6	96	0.00
88	Benzo(b)fluoranthene	40.000	38.644	3.4	94	0.00
89	Benzo(k)fluoranthene	40.000	38.379	4.1	92	0.00
90 C	Benzo(a)pyrene	40.000	39.113	2.2	95	0.00
91	Dibenzo(a,h)anthracene	40.000	40.771	-1.9	96	0.00
92	Benzo(g,h,i)perylene	40.000	39.508	1.2	95	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062121\
Data File : BG049028.D
Acq On : 21 Jun 2021 11:34
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jun 21 13:28:09 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG061121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 16 12:04:50 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0