

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA G\Data\BG062217\
 Data File : BG027596.D
 Acq On : 22 Jun 2017 14:57
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 23 02:07:36 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG062117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 21 18:52:15 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	90	0.00
2	1,4-Dioxane	40.000	39.406	1.5	90	0.00
3	Pyridine	40.000	38.739	3.2	83	0.00
4	n-Nitrosodimethylamine	40.000	43.687	-9.2	93	0.00
5 S	2-Fluorophenol	80.000	82.000	-2.5	88	0.00
6	Aniline	40.000	39.928	0.2	85	0.00
7 S	Phenol-d6	80.000	80.083	-0.1	85	0.00
8	2-Chlorophenol	40.000	41.323	-3.3	90	0.00
9	Benzaldehyde	40.000	40.808	-2.0	86	0.00
10 C	Phenol	40.000	40.519	-1.3	87	0.00
11	bis(2-Chloroethyl)ether	40.000	38.174	4.6	80	0.00
12	1,3-Dichlorobenzene	40.000	40.318	-0.8	89	0.00
13 C	1,4-Dichlorobenzene	40.000	40.425	-1.1	89	0.00
14	1,2-Dichlorobenzene	40.000	41.556	-3.9	92	0.00
15	Benzyl Alcohol	40.000	41.472	-3.7	88	0.00
16	2,2'-oxybis(1-Chloropropane)	40.000	35.342	11.6	76	0.00
17	2-Methylphenol	40.000	38.927	2.7	85	0.00
18	Hexachloroethane	40.000	41.415	-3.5	91	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.085	4.8	80	0.00
20	3+4-Methylphenols	40.000	40.923	-2.3	87	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	87	0.00
22	Acetophenone	40.000	39.380	1.5	85	0.00
23 S	Nitrobenzene-d5	80.000	81.083	-1.4	87	0.00
24	Nitrobenzene	40.000	39.616	1.0	86	0.00
25	Isophorone	40.000	38.538	3.7	82	0.00
26 C	2-Nitrophenol	40.000	40.615	-1.5	89	0.00
27	2,4-Dimethylphenol	40.000	39.641	0.9	86	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.890	2.8	83	0.00
29 C	2,4-Dichlorophenol	40.000	42.714	-6.8	91	0.00
30	1,2,4-Trichlorobenzene	40.000	41.259	-3.1	93	0.00
31	Naphthalene	40.000	40.096	-0.2	88	0.00
32	Benzoic acid	40.000	38.506	3.7	87	0.00
33	4-Chloroaniline	40.000	39.707	0.7	85	0.00
34 C	Hexachlorobutadiene	40.000	43.083	-7.7	96	0.00
35	Caprolactam	40.000	37.714	5.7	80	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.323	-0.8	85	0.00
37	2-Methylnaphthalene	40.000	41.016	-2.5	88	0.00
38 I	Acenaphthene-d10	20.000	20.000	0.0	86	0.00
39	1,2,4,5-Tetrachlorobenzene	40.000	42.740	-6.9	93	0.00
40 P	Hexachlorocyclopentadiene	40.000	41.757	-4.4	90	0.00
41 S	2,4,6-Tribromophenol	80.000	88.009	-10.0	93	0.00
42 C	2,4,6-Trichlorophenol	40.000	42.975	-7.4	89	0.00
43	2,4,5-Trichlorophenol	40.000	41.697	-4.2	89	0.00
44 S	2-Fluorobiphenyl	80.000	82.642	-3.3	89	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA G\Data\BG062217\
 Data File : BG027596.D
 Acq On : 22 Jun 2017 14:57
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 23 02:07:36 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG062117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 21 18:52:15 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	40.000	40.921	-2.3	88	0.00
46	2-Chloronaphthalene	40.000	41.145	-2.9	88	0.00
47	2-Nitroaniline	40.000	39.785	0.5	82	0.00
48	Acenaphthylene	40.000	40.716	-1.8	86	0.00
49	Dimethylphthalate	40.000	39.725	0.7	84	0.00
50	2,6-Dinitrotoluene	40.000	42.412	-6.0	89	0.00
51 C	Acenaphthene	40.000	40.713	-1.8	87	0.00
52	3-Nitroaniline	40.000	39.117	2.2	81	0.00
53 P	2,4-Dinitrophenol	40.000	40.396	-1.0	86	0.00
54	Dibenzofuran	40.000	40.170	-0.4	86	0.00
55 P	4-Nitrophenol	40.000	38.823	2.9	82	0.00
56	2,4-Dinitrotoluene	40.000	42.039	-5.1	85	0.00
57	Fluorene	40.000	40.165	-0.4	86	0.00
58	2,3,4,6-Tetrachlorophenol	40.000	43.306	-8.3	91	0.00
59	Diethylphthalate	40.000	39.828	0.4	84	0.00
60	4-Chlorophenyl-phenylether	40.000	40.878	-2.2	87	0.00
61	4-Nitroaniline	40.000	40.767	-1.9	83	0.00
62	Azobenzene	40.000	38.388	4.0	81	0.00
63 I	Phenanthrene-d10	20.000	20.000	0.0	85	0.00
64	4,6-Dinitro-2-methylphenol	40.000	40.884	-2.2	86	0.00
65 c	n-Nitrosodiphenylamine	40.000	40.358	-0.9	84	0.00
66	4-Bromophenyl-phenylether	40.000	42.359	-5.9	88	0.00
67	Hexachlorobenzene	40.000	42.055	-5.1	89	0.00
68	Atrazine	40.000	41.210	-3.0	85	0.00
69 C	Pentachlorophenol	40.000	41.325	-3.3	87	0.00
70	Phenanthrene	40.000	40.055	-0.1	85	0.00
71	Anthracene	40.000	40.688	-1.7	85	0.00
72	Carbazole	40.000	39.731	0.7	84	0.00
73	Di-n-butylphthalate	40.000	39.786	0.5	83	0.00
74 C	Fluoranthene	40.000	40.017	-0.0	86	0.00
75 I	Chrysene-d12	20.000	20.000	0.0	87	0.00
76	Benzidine	40.000	39.533	1.2	80	0.00
77	Pyrene	40.000	41.080	-2.7	87	0.00
78 S	Terphenyl-d14	80.000	84.793	-6.0	89	0.00
79	Butylbenzylphthalate	40.000	39.829	0.4	83	0.00
80	Benzo(a)anthracene	40.000	41.154	-2.9	88	0.00
81	3,3'-Dichlorobenzidine	40.000	42.275	-5.7	87	0.00
82	Chrysene	40.000	40.933	-2.3	88	0.00
83	Bis(2-ethylhexyl)phthalate	40.000	39.215	2.0	81	0.00
84 c	Di-n-octyl phthalate	40.000	39.911	0.2	82	0.00
85	Indeno(1,2,3-cd)pyrene	40.000	42.737	-6.8	90	0.00
86 I	Perylene-d12	20.000	20.000	0.0	89	0.00
87	Benzo(b)fluoranthene	40.000	41.004	-2.5	91	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA G\Data\BG062217\
Data File : BG027596.D
Acq On : 22 Jun 2017 14:57
Operator : SJ/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jun 23 02:07:36 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG062117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 21 18:52:15 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	40.000	40.626	-1.6	90	0.00
89 C	Benzo(a)pyrene	40.000	40.142	-0.4	87	0.00
90	Dibenzo(a,h)anthracene	40.000	41.230	-3.1	91	-0.01
91	Benzo(a,h,i)perylene	40.000	40.865	-2.2	89	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0