

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062221\
 Data File : BG049048.D
 Acq On : 22 Jun 2021 16:33
 Operator : CG/JU
 Sample : M2777-02RE
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 X-2


Quant Time: Jun 22 17:12:51 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG061121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 16 12:04:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.116	152	39530	20.000	ng	0.00
21) Naphthalene-d8	10.936	136	144040	20.000	ng	# 0.00
39) Acenaphthene-d10	14.767	164	106010	20.000	ng	0.00
64) Phenanthrene-d10	17.517	188	204520	20.000	ng	# 0.00
76) Chrysene-d12	21.794	240	221467	20.000	ng	0.00
86) Perylene-d12	25.120	264	243674	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.666	112	240145	94.568	ng	0.00
7) Phenol-d6	7.264	99	316823	83.242	ng	0.00
23) Nitrobenzene-d5	9.279	82	240980	72.140	ng	0.00
42) 2,4,6-Tribromophenol	16.259	330	159179	101.349	ng	0.01
45) 2-Fluorobiphenyl	13.386	172	448836	59.811	ng	0.00
79) Terphenyl-d14	20.114	244	726692	60.094	ng	0.00
Target Compounds						
						Qvalue
9) Benzaldehyde	7.258	77	5324	2.703	ng	# 1
18) Hexachloroethane	9.226	117	37375	29.887	ng	# 1
19) n-Nitroso-di-n-propyla...	8.886	70	15101	5.076	ng	# 37
22) Acetophenone	8.927	105	168691	38.494	ng	# 54
24) Nitrobenzene	9.320	77	11694	3.108	ng	# 37
25) Isophorone	9.855	82	66696	9.165	ng	# 1
26) 2-Nitrophenol	10.049	139	4960	3.886	ng	# 1
29) 2,4-Dichlorophenol	10.554	162	11054	4.258	ng	# 49
31) Naphthalene	11.001	128	23031	2.684	ng	# 1
33) 4-Chloroaniline	11.054	127	8452	2.320	ng	# 1
35) Caprolactam	11.853	113	12383	16.500	ng	# 1
37) 2-Methylnaphthalene	12.593	142	85887	13.251	ng	# 79
38) 1-Methylnaphthalene	12.816	142	97798	16.072	ng	# 92
46) 1,1'-Biphenyl	13.598	154	59532	7.617	ng	# 89
48) 2-Nitroaniline	13.856	65	6643	3.866	ng	# 48
52) Acenaphthene	14.832	154	17887	2.668	ng	# 44
53) 3-Nitroaniline	14.679	138	4883	2.868	ng	# 19
55) Dibenzofuran	15.161	168	22309	2.141	ng	# 1
56) 4-Nitrophenol	14.973	139	28913	20.830	ng	# 82
58) Fluorene	15.813	166	18144	2.129	ng	# 52
65) 4,6-Dinitro-2-methylph...	15.848	198	2276	2.347	ng	# 1
66) n-Nitrosodiphenylamine	16.024	169	155448	24.389	ng	# 55
69) Atrazine	17.288	200	27548	13.466	ng	# 79
71) Phenanthrene	17.558	178	37645	3.271	ng	# 76
75) Fluoranthene	19.926	202	62600	4.511	ng	# 84
78) Pyrene	19.926	202	62600	3.894	ng	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jun 22 17:12:51 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG061121.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 16 12:04:50 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.116 min Scan# 864
Delta R.T. 0.002 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33




Instrument :
BNA_G
ClientSampled :
X-2


Tgt Ion:152 Resp: 39530
Ion Ratio Lower Upper
152 100
150 139.8 119.4 179.0
115 62.0 60.2 90.4



#5
2-Fluorophenol
Concen: 94.568 ng
RT: 5.666 min Scan# 447
Delta R.T. 0.002 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33

Tgt Ion:112 Resp: 240145
Ion Ratio Lower Upper
112 100
64 71.7 46.7 70.1#
63 41.9 35.3 52.9









#7
Phenol-d6
Concen: 83.242 ng
RT: 7.264 min Scan# 719
Delta R.T. -0.003 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33




Instrument :
BNA_G
ClientSampleId :
X-2


Tgt Ion: 99 Resp: 316823
Ion Ratio Lower Upper
99 100
42 27.2 20.5 30.7
71 58.5 41.5 62.3



#9
Benzaldehyde
Concen: 2.703 ng
RT: 7.258 min Scan# 718
Delta R.T. 0.014 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33

Tgt Ion: 77 Resp: 5324
Ion Ratio Lower Upper
77 100
105 321.6 68.8 108.8#
106 31.9 47.9 87.9#









#18
Hexachloroethane
Concen: 29.887 ng
RT: 9.226 min Scan# 1053
Delta R.T. 0.020 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33




Instrument :
BNA_G
ClientSampled :
X-2


Tgt Ion:117 Resp: 37375
Ion Ratio Lower Upper
117 100
119 527.4 74.4 111.6#
201 0.0 88.6 133.0#



#19
n-Nitroso-di-n-propylamine
Concen: 5.076 ng
RT: 8.886 min Scan# 995
Delta R.T. -0.039 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33

Tgt Ion: 70 Resp: 15101
Ion Ratio Lower Upper
70 100
42 128.0 54.5 81.7#
101 0.0 9.0 13.6#
130 0.0 17.0 25.6#









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.936 min Scan# 1344
Delta R.T. -0.003 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33




Instrument :
BNA_G
ClientSampled :
X-2


Tgt Ion:	136	Resp:	144040
Ion Ratio	Lower	Upper	
136	100		
137	13.4	8.2	12.2#
54	17.5	5.5	8.3#
68	14.4	2.3	3.5#



#22
Acetophenone
Concen: 38.494 ng
RT: 8.927 min Scan# 1002
Delta R.T. -0.009 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33

Tgt Ion:	105	Resp:	168691
Ion Ratio	Lower	Upper	
105	100		
71	4.3	0.6	0.8#
51	5.8	26.6	40.0#
120	0.2	16.2	24.4#









#23
Nitrobenzene-d5
Concen: 72.140 ng
RT: 9.279 min Scan# 1062
Delta R.T. -0.003 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33




Instrument :
BNA_G
ClientSampled :
X-2


Tgt Ion: 82 Resp: 240980
Ion Ratio Lower Upper
82 100
128 37.2 29.8 44.6
54 60.8 45.8 68.8



#24
Nitrobenzene
Concen: 3.108 ng
RT: 9.320 min Scan# 1069
Delta R.T. -0.003 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33

Tgt Ion: 77 Resp: 11694
Ion Ratio Lower Upper
77 100
123 17.4 27.5 41.3#
65 73.9 12.6 18.8#









#25
Isophorone
Concen: 9.165 ng
RT: 9.855 min Scan# 1160
Delta R.T. 0.008 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33




Instrument :
BNA_G
ClientSampled :
X-2


Tgt Ion: 82 Resp: 66696
Ion Ratio Lower Upper
82 100
95 166.8 8.5 12.7#
138 8.7 11.8 17.6#



#26
2-Nitrophenol
Concen: 3.886 ng
RT: 10.049 min Scan# 1193
Delta R.T. 0.008 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33

Tgt Ion:139 Resp: 4960
Ion Ratio Lower Upper
139 100
109 278.1 37.8 56.6#
65 201.9 51.8 77.8#









#29
2,4-Dichlorophenol
Concen: 4.258 ng
RT: 10.554 min Scan# 1279
Delta R.T. -0.027 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33




Instrument :
BNA_G
ClientSampled :
X-2


Tgt Ion:162 Resp: 11054
Ion Ratio Lower Upper
162 100
164 3.4 39.4 79.4#
98 30.2 18.7 58.7



#31
Naphthalene
Concen: 2.684 ng
RT: 11.001 min Scan# 1355
Delta R.T. 0.014 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33

Tgt Ion:128 Resp: 23031
Ion Ratio Lower Upper
128 100
129 72.2 8.5 12.7#
127 39.4 10.2 15.4#









#33
4-Chloroaniline
Concen: 2.320 ng
RT: 11.054 min Scan# 1364
Delta R.T. -0.039 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33




Instrument :
BNA_G
ClientSampled :
X-2


Tgt Ion:127	Resp:	8452
Ion Ratio	Lower	Upper
127	100	
129	223.9	26.7 40.1#
65	138.3	32.8 49.2#
92	44.5	16.2 24.4#



#35
Caprolactam
Concen: 16.500 ng
RT: 11.853 min Scan# 1500
Delta R.T. -0.009 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33

Tgt Ion:113	Resp:	12383
Ion Ratio	Lower	Upper
113	100	
55	1173.0	113.8 153.8#
56	587.5	105.6 145.6#









#37
2-Methylnaphthalene
Concen: 13.251 ng
RT: 12.593 min Scan# 1626
Delta R.T. 0.002 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33




Instrument :
BNA_G
ClientSampled :
X-2


Tgt Ion:142 Resp: 85887
Ion Ratio Lower Upper
142 100
141 93.9 68.8 103.2
115 73.9 36.4 54.6#



#38
1-Methylnaphthalene
Concen: 16.072 ng
RT: 12.816 min Scan# 1664
Delta R.T. 0.008 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33


Tgt Ion:142 Resp: 97798
Ion Ratio Lower Upper
142 100
141 93.8 70.6 106.0
116 17.7 4.0 6.0#








#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.767 min Scan# 1996
Delta R.T. 0.008 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33


Instrument :
BNA_G
ClientSampled :
X-2





Tgt Ion:164 Resp: 106010
Ion Ratio Lower Upper
164 100
162 98.8 81.4 122.0
160 51.6 34.9 52.3




#42
2,4,6-Tribromophenol
Concen: 101.349 ng
RT: 16.259 min Scan# 2250
Delta R.T. 0.014 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33



Tgt Ion:330 Resp: 159179
Ion Ratio Lower Upper
330 100
332 94.6 77.3 115.9
141 48.8 35.6 53.4









#45
2-Fluorobiphenyl
Concen: 59.811 ng
RT: 13.386 min Scan# 1761
Delta R.T. 0.002 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33




Instrument :
BNA_G
ClientSampled :
X-2


Tgt Ion:172 Resp: 448836
Ion Ratio Lower Upper
172 100
171 37.1 28.6 43.0
170 24.4 18.6 27.8



#46
1,1'-Biphenyl
Concen: 7.617 ng
RT: 13.598 min Scan# 1797
Delta R.T. 0.008 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33

Tgt Ion:154 Resp: 59532
Ion Ratio Lower Upper
154 100
153 33.4 20.5 60.5
76 13.2 0.0 29.3









#48
2-Nitroaniline
Concen: 3.866 ng
RT: 13.856 min Scan# 1841
Delta R.T. 0.026 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33




Instrument :
BNA_G
ClientSampled :
X-2


Tgt Ion: 65 Resp: 6643
Ion Ratio Lower Upper
65 100
92 72.9 47.2 70.8#
138 16.7 68.0 102.0#



#52
Acenaphthene
Concen: 2.668 ng
RT: 14.832 min Scan# 2007
Delta R.T. 0.008 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33

Tgt Ion:154 Resp: 17887
Ion Ratio Lower Upper
154 100
153 146.4 87.0 130.4#
152 118.9 40.1 60.1#









#53
3-Nitroaniline
Concen: 2.868 ng
RT: 14.679 min Scan# 1981
Delta R.T. 0.026 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33




Instrument :
BNA_G
ClientSampled :
X-2


Tgt Ion:138 Resp: 4883
Ion Ratio Lower Upper
138 100
108 108.7 8.3 12.5#
92 193.7 96.4 144.6#



#55
Dibenzofuran
Concen: 2.141 ng
RT: 15.161 min Scan# 2063
Delta R.T. 0.008 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33

Tgt Ion:168 Resp: 22309
Ion Ratio Lower Upper
168 100
139 103.1 33.4 50.2#
169 193.4 12.3 18.5#









#56
4-Nitrophenol
Concen: 20.830 ng
RT: 14.973 min Scan# 2031
Delta R.T. 0.002 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33




Instrument :
BNA_G
ClientSampled :
X-2


Tgt Ion:139 Resp: 28913
Ion Ratio Lower Upper
139 100
109 133.9 116.3 156.3
65 84.1 105.0 145.0#



#58
Fluorene
Concen: 2.129 ng
RT: 15.813 min Scan# 2174
Delta R.T. 0.008 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33

Tgt Ion:166 Resp: 18144
Ion Ratio Lower Upper
166 100
165 136.5 77.7 116.5#
167 55.8 10.8 16.2#









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.517 min Scan# 2464
Delta R.T. 0.008 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33




Instrument :
BNA_G
ClientSampled :
X-2


Tgt Ion:188 Resp: 204520
Ion Ratio Lower Upper
188 100
94 9.6 5.4 8.0#
80 11.1 3.7 5.5#



#65
4,6-Dinitro-2-methylphenol
Concen: 2.347 ng
RT: 15.848 min Scan# 2180
Delta R.T. -0.027 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33

Tgt Ion:198 Resp: 2276
Ion Ratio Lower Upper
198 100
51 938.5 29.4 69.4#
105 7232.5 15.1 55.1#









#66
n-Nitrosodiphenylamine
Concen: 24.389 ng
RT: 16.024 min Scan# 2210
Delta R.T. 0.014 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33




Instrument :
BNA_G
ClientSampled :
X-2


Tgt Ion:169 Resp: 155448
Ion Ratio Lower Upper
169 100
168 30.7 56.3 84.5#
167 13.6 28.4 42.6#



#69
Atrazine
Concen: 13.466 ng
RT: 17.288 min Scan# 2425
Delta R.T. 0.331 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33

Tgt Ion:200 Resp: 27548
Ion Ratio Lower Upper
200 100
173 43.6 3.7 43.7
215 39.8 26.9 66.9









#71
Phenanthrene
Concen: 3.271 ng
RT: 17.558 min Scan# 2471
Delta R.T. 0.008 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33




Instrument :
BNA_G
ClientSampled :
X-2


Tgt Ion:178 Resp: 37645
Ion Ratio Lower Upper
178 100
176 20.0 16.0 24.0
179 39.9 14.2 21.4#



#75
Fluoranthene
Concen: 4.511 ng
RT: 19.926 min Scan# 2874
Delta R.T. 0.373 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33

Tgt Ion:202 Resp: 62600
Ion Ratio Lower Upper
202 100
101 9.6 0.0 24.7
203 25.5 0.0 38.5









#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.794 min Scan# 3192
Delta R.T. -0.003 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33




Instrument :
BNA_G
ClientSampled :
X-2


Tgt Ion:240 Resp: 221467
Ion Ratio Lower Upper
240 100
120 5.6 4.2 6.2
236 26.1 21.8 32.6



#78
Pyrene
Concen: 3.894 ng
RT: 19.926 min Scan# 2874
Delta R.T. 0.008 min
Lab File: BG049048.D
Acq: 22 Jun 2021 16:33

Tgt Ion:202 Resp: 62600
Ion Ratio Lower Upper
202 100
200 19.4 15.8 23.8
203 25.5 14.6 21.8#









#79
 Terphenyl-d14
 Concen: 60.094 ng
 RT: 20.114 min Scan# 2906
 Delta R.T. 0.002 min
 Lab File: BG049048.D
 Acq: 22 Jun 2021 16:33

Instrument :
 BNA_G
 ClientSampled :
 X-2

Tgt Ion:244 Resp: 726692
 Ion Ratio Lower Upper
 244 100
 212 8.6 6.9 10.3
 122 6.8 5.0 7.6



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 25.120 min Scan# 3758
 Delta R.T. -0.009 min
 Lab File: BG049048.D
 Acq: 22 Jun 2021 16:33

Tgt Ion:264 Resp: 243674
 Ion Ratio Lower Upper
 264 100
 260 22.4 19.4 29.2
 265 20.6 16.9 25.3

