

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062224\
 Data File : BG061655.D
 Acq On : 22 Jun 2024 19:12
 Operator : MA/JU
 Sample : P2776-11
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 A4CA6

Quant Time: Jun 23 00:17:11 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG062024.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 21 05:03:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.076	152	67815	20.000	ng/ul	0.00
20) Naphthalene-d8	10.890	136	330529	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.703	164	234905	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.447	188	512140	20.000	ng/ul	0.00
79) Chrysene-d12	21.707	240	429365	20.000	ng/ul	0.00
88) Perylene-d12	24.939	264	515494	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.475	96	7647	4.073	ng/uL	0.01
4) Pyridine-d5	3.887	84	90950	16.024	ng/ul	0.00
7) Phenol-d5	7.218	99	169891	23.077	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.400	67	95454	20.799	ng/ul	0.00
11) 2-Chlorophenol-d4	7.600	132	119239	23.602	ng/ul	0.00
15) 4-Methylphenol-d8	8.775	113	130277	22.891	ng/ul	0.01
21) Nitrobenzene-d5	9.239	128	58852	26.223	ng/ul	0.00
24) 2-Nitrophenol-d4	9.968	143	64741	28.482	ng/ul	0.01
28) 2,4-Dichlorophenol-d3	10.497	165	134871	25.059	ng/ul	0.00
31) 4-Chloroaniline-d4	11.020	131	114990	14.111	ng/ul	0.00
46) Dimethylphthalate-d6	14.110	166	399604	22.106	ng/ul	0.00
49) Acenaphthylene-d8	14.398	160	479916	24.225	ng/ul	0.00
54) 4-Nitrophenol-d4	14.880	143	67087	21.609	ng/ul	0.01
60) Fluorene-d10	15.691	176	374996	23.696	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.796	200	39972	18.252	ng/ul	0.00
73) Anthracene-d10	17.541	188	552872	23.650	ng/ul	0.00
81) Pyrene-d10	19.821	212	604352	25.199	ng/ul	0.00
92) Benzo(a)pyrene-d12	24.715	264	614208	24.947	ng/ul	0.00
Target Compounds						
					Qvalue	
80) Fluoranthene	19.492	202	51561	1.853	ng/ul#	96
82) Pyrene	19.850	202	49745	1.681	ng/ul	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062224\
Data File : BG061655.D
Acq On : 22 Jun 2024 19:12
Operator : MA/JU
Sample : P2776-11
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
A4CA6

Quant Time: Jun 23 00:17:11 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG062024.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jun 21 05:03:55 2024
Response via : Initial Calibration





#80

Fluoranthene

Concen: 1.853 ng/ul

RT: 19.492 min Scan# 21

Delta R.T. -0.000 min

Lab File: BG061655.D

Acq: 22 Jun 2024 19:12

Instrument :

BNA_G

ClientSampleId :

A4CA6





Tgt Ion:202 Resp: 51561

Ion Ratio Lower Upper

202 100

101 12.1 7.8 11.6#

100 10.3 7.7 11.5



#82

Pyrene

Concen: 1.681 ng/ul

RT: 19.850 min Scan# 2819

Delta R.T. -0.000 min

Lab File: BG061655.D

Acq: 22 Jun 2024 19:12

Tgt Ion:202 Resp: 49745

Ion Ratio Lower Upper

202 100

101 10.4 10.1 15.1

100 10.7 9.9 14.9

