

Data Path : Z:\HPCHEM1\BNA G\DATA\BG062316\
 Data File : BG022508.D
 Acq On : 23 Jun 2016 7:09
 Operator : UM/SJ
 Sample : H3623-03
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SC-STA-1447(0-4)

Quant Time: Jun 23 07:52:46 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG060616.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 22 01:27:56 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.97	152	27331	20.00	ng	-0.02
21) Naphthalene-d8	10.79	136	161615	20.00	ng	-0.02
38) Acenaphthene-d10	14.62	164	110895	20.00	ng	-0.01
63) Phenanthrene-d10	17.36	188	237156	20.00	ng	-0.02
75) Chrysene-d12	21.62	240	145339	20.00	ng	-0.02
86) Perylene-d12	24.80	264	123321	20.00	ng	-0.04
System Monitoring Compounds						
5) 2-Fluorophenol	5.51	112	271835	192.30	ng	-0.03
7) Phenol-d6	7.15	99	388467	174.79	ng	-0.03
23) Nitrobenzene-d5	9.14	82	237714	102.55	ng	-0.02
41) 2,4,6-Tribromophenol	16.11	330	163882	130.79	ng	-0.02
44) 2-Fluorobiphenyl	13.24	172	695908	95.63	ng	-0.01
78) Terphenyl-d14	19.96	244	805461	131.57	ng	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG062316\
Data File : BG022508.D
Acq On : 23 Jun 2016 7:09
Operator : UM/SJ
Sample : H3623-03
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SC-STA-1447(0-4)

Quant Time: Jun 23 07:52:46 2016
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG060616.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 22 01:27:56 2016
Response via : Initial Calibration

