


Data Path : Z:\HPCHEM1\BNA G\Data\BG062317\
 Data File : BG027633.D
 Acq On : 23 Jun 2017 15:52
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG062117.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Jun 21 18:52:15 2017



AutoFind: Scans 2439, 2440, 2441; Background Corrected with Scan 2433

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	42.7	9740	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	41.5	9475	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	50.1	11426	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	22815	PASS
199	198	5	9	6.5	1480	PASS
275	198	10	60	30.4	6939	PASS
365	198	1	100	4.4	994	PASS
441	442	0.01	24	15.3	3482	PASS
442	198	50	100	99.5	22706	PASS
443	442	15	24	19.8	4487	PASS