

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG062320\
 Data File : BG045831.D
 Acq On : 23 Jun 2020 8:47
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 23 09:23:20 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG061520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 15 18:34:19 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.542	0.557	-2.8	99	0.00
3	Pyridine	1.607	1.701	-5.8	100	0.00
4	n-Nitrosodimethylamine	0.539	0.613	-13.7	109	0.00
5 S	2-Fluorophenol	1.184	1.304	-10.1	104	-0.01
6	Aniline	2.121	2.311	-9.0	104	0.00
7 S	Phenol-d6	1.801	1.928	-7.1	101	-0.02
8	2-Chlorophenol	1.267	1.373	-8.4	105	0.00
9	Benzaldehyde	1.082	1.129	-4.3	101	0.00
10 C	Phenol	1.745	1.855	-6.3	101	-0.02
11	bis(2-Chloroethyl)ether	1.325	1.420	-7.2	104	0.00
12	1,3-Dichlorobenzene	1.510	1.630	-7.9	103	0.00
13 C	1,4-Dichlorobenzene	1.509	1.631	-8.1	103	0.00
14	1,2-Dichlorobenzene	1.444	1.559	-8.0	104	0.00
15	Benzyl Alcohol	1.394	1.563	-12.1	106	0.00
16	2,2'-oxybis(1-Chloropropane	1.249	1.323	-5.9	104	0.00
17	2-Methylphenol	1.299	1.397	-7.5	102	-0.01
18	Hexachloroethane	0.573	0.623	-8.7	104	0.00
19 P	n-Nitroso-di-n-propylamine	1.184	1.271	-7.3	102	0.00
20	3+4-Methylphenols	1.708	1.864	-9.1	103	-0.02
21 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
22	Acetophenone	0.549	0.590	-7.5	102	0.00
23 S	Nitrobenzene-d5	0.438	0.479	-9.4	105	0.00
24	Nitrobenzene	0.467	0.508	-8.8	103	0.00
25	Isophorone	0.848	0.910	-7.3	101	0.00
26 C	2-Nitrophenol	0.194	0.210	-8.2	100	0.00
27	2,4-Dimethylphenol	0.297	0.320	-7.7	104	-0.01
28	bis(2-Chloroethoxy)methane	0.443	0.473	-6.8	103	0.00
29 C	2,4-Dichlorophenol	0.346	0.380	-9.8	105	0.00
30	1,2,4-Trichlorobenzene	0.410	0.449	-9.5	106	0.00
31	Naphthalene	1.093	1.174	-7.4	102	0.00
32	Benzoic acid	0.170	0.213	-25.3#	124	-0.02
33	4-Chloroaniline	0.458	0.498	-8.7	103	0.00
34 C	Hexachlorobutadiene	0.292	0.332	-13.7	108	0.00
35	Caprolactam	0.112	0.127	-13.4	110	0.00
36 C	4-Chloro-3-methylphenol	0.400	0.429	-7.2	103	-0.02
37	2-Methylnaphthalene	0.814	0.879	-8.0	103	0.00
38	1-Methylnaphthalene	0.771	0.829	-7.5	103	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
40	1,2,4,5-Tetrachlorobenzene	0.659	0.723	-9.7	104	0.00
41 P	Hexachlorocyclopentadiene	0.299	0.329	-10.0	103	0.00
42 S	2,4,6-Tribromophenol	0.302	0.322	-6.6	105	0.00
43 C	2,4,6-Trichlorophenol	0.446	0.477	-7.0	103	0.00
44	2,4,5-Trichlorophenol	0.509	0.533	-4.7	100	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG062320\
 Data File : BG045831.D
 Acq On : 23 Jun 2020 8:47
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 23 09:23:20 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG061520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 15 18:34:19 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.361	1.487	-9.3	105	0.00
46	1,1'-Biphenyl	1.432	1.525	-6.5	103	0.00
47	2-Chloronaphthalene	1.157	1.239	-7.1	104	0.00
48	2-Nitroaniline	0.385	0.414	-7.5	103	0.00
49	Acenaphthylene	1.853	1.974	-6.5	103	0.00
50	Dimethylphthalate	1.567	1.687	-7.7	104	0.00
51	2,6-Dinitrotoluene	0.352	0.376	-6.8	103	0.00
52 C	Acenaphthene	1.123	1.178	-4.9	102	0.00
53	3-Nitroaniline	0.351	0.375	-6.8	104	-0.01
54 P	2,4-Dinitrophenol	0.152	0.198	-30.3#	130	0.00
55	Dibenzofuran	1.824	1.964	-7.7	104	0.00
56 P	4-Nitrophenol	0.226	0.229	-1.3	97	-0.03
57	2,4-Dinitrotoluene	0.504	0.544	-7.9	104	0.00
58	Fluorene	1.517	1.620	-6.8	103	0.00
59	2,3,4,6-Tetrachlorophenol	0.442	0.449	-1.6	98	0.00
60	Diethylphthalate	1.602	1.726	-7.7	104	0.00
61	4-Chlorophenyl-phenylether	0.874	0.950	-8.7	105	0.00
62	4-Nitroaniline	0.361	0.390	-8.0	103	0.00
63	Azobenzene	1.538	1.656	-7.7	105	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	104	0.00
65	4,6-Dinitro-2-methylphenol	0.124	0.144	-16.1	112	0.00
66 c	n-Nitrosodiphenylamine	0.566	0.597	-5.5	102	0.00
67	4-Bromophenyl-phenylether	0.260	0.279	-7.3	104	0.00
68	Hexachlorobenzene	0.269	0.293	-8.9	108	0.00
69	Atrazine	0.220	0.229	-4.1	102	0.00
70 C	Pentachlorophenol	0.111	0.110	0.9	93	0.00
71	Phenanthrene	1.031	1.099	-6.6	103	0.00
72	Anthracene	1.027	1.100	-7.1	103	0.00
73	Carbazole	0.978	1.052	-7.6	103	0.00
74	Di-n-butylphthalate	1.158	1.247	-7.7	103	0.00
75 C	Fluoranthene	1.268	1.363	-7.5	104	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
77	Benzidine	0.520	0.471	9.4	82	0.00
78	Pyrene	1.339	1.466	-9.5	104	0.00
79 S	Terphenyl-d14	0.987	1.091	-10.5	103	0.00
80	Butylbenzylphthalate	0.561	0.611	-8.9	104	0.00
81	Benzo(a)anthracene	1.297	1.403	-8.2	104	0.00
82	3,3'-Dichlorobenzidine	0.490	0.525	-7.1	101	0.00
83	Chrysene	1.254	1.348	-7.5	102	0.00
84	Bis(2-ethylhexyl)phthalate	0.780	0.852	-9.2	104	0.00
85 c	Di-n-octyl phthalate	1.345	1.456	-8.3	104	0.00
86 I	Perylene-d12	1.000	1.000	0.0	104	0.00
87	Indeno(1,2,3-cd)pyrene	1.450	1.570	-8.3	105	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG062320\
Data File : BG045831.D
Acq On : 23 Jun 2020 8:47
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jun 23 09:23:20 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG061520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 15 18:34:19 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.254	1.339	-6.8	103	0.00
89	Benzo(k)fluoranthene	1.200	1.297	-8.1	103	0.00
90 C	Benzo(a)pyrene	1.171	1.265	-8.0	104	0.00
91	Dibenzo(a,h)anthracene	1.163	1.282	-10.2	107	0.00
92	Benzo(q,h,i)perylene	1.164	1.251	-7.5	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0