


Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG062320\
 Data File : BG045830.D
 Acq On : 23 Jun 2020 8:07
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG061520.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jun 15 18:34:19 2020



AutoFind: Scans 2497, 2498, 2499; Background Corrected with Scan 2491

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	26.2	11977	PASS
68	69	0.00	2	1.0	168	PASS
69	198	0.00	100	35.4	16190	PASS
70	69	0.00	2	1.4	232	PASS
127	198	10	80	42.4	19426	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	45768	PASS
199	198	5	9	6.6	3007	PASS
275	198	10	60	30.0	13737	PASS
365	198	1	100	4.9	2238	PASS
441	198	0.01	100	14.2	6505	PASS
442	198	50	100	90.4	41397	PASS
443	442	15	24	18.5	7643	PASS