


Data Path : Z:\HPCHEM1\BNA G\DATA\BG062416\
 Data File : BG022514.D
 Acq On : 23 Jun 2016 21:29
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG062416.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jun 24 04:19:57 2016



AutoFind: Scans 2579, 2580, 2581; Background Corrected with Scan 2573

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.6	51134	PASS
68	69	0.00	2	0.5	294	PASS
69	198	0.00	100	45.0	64623	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	43.8	62913	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	143730	PASS
199	198	5	9	6.7	9671	PASS
275	198	10	60	31.6	45469	PASS
365	198	1	100	4.6	6656	PASS
441	442	0.01	24	16.1	16477	PASS
442	198	50	100	71.1	102162	PASS
443	442	15	24	20.1	20552	PASS