

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062424\
 Data File : BG061688.D
 Acq On : 24 Jun 2024 23:39
 Operator : MA/JU
 Sample : P2856-20
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 COSJ6

Quant Time: Jun 25 01:19:41 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG062024.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jun 21 05:03:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.073	152	62101	20.000	ng/u1	0.00
20) Naphthalene-d8	10.887	136	332043	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.701	164	254426	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.444	188	587859	20.000	ng/u1	0.00
79) Chrysene-d12	21.704	240	483707	20.000	ng/u1	0.00
88) Perylene-d12	24.936	264	555837	20.000	ng/u1	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.473	96	8009	4.659	ng/uL	0.00
4) Pyridine-d5	3.890	84	109406	21.049	ng/u1	0.00
7) Phenol-d5	7.215	99	177665	26.353	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.397	67	109850	26.139	ng/u1	0.00
11) 2-Chlorophenol-d4	7.597	132	120791	26.110	ng/u1	0.00
15) 4-Methylphenol-d8	8.766	113	147029	28.211	ng/u1	0.00
21) Nitrobenzene-d5	9.236	128	68695	30.469	ng/u1	0.00
24) 2-Nitrophenol-d4	9.959	143	81078	35.507	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.500	165	147449	27.271	ng/u1	0.00
31) 4-Chloroaniline-d4	11.017	131	216342	26.428	ng/u1	0.00
46) Dimethylphthalate-d6	14.107	166	555493	28.372	ng/u1	0.00
49) Acenaphthylene-d8	14.395	160	624419	29.101	ng/u1	0.00
54) 4-Nitrophenol-d4	14.877	143	85768	25.507	ng/u1	0.00
60) Fluorene-d10	15.694	176	501131	29.237	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.799	200	62840	24.999	ng/u1	0.00
73) Anthracene-d10	17.544	188	794188	29.597	ng/u1	0.00
81) Pyrene-d10	19.824	212	849356	31.436	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.718	264	818240	30.822	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062424\
Data File : BG061688.D
Acq On : 24 Jun 2024 23:39
Operator : MA/JU
Sample : P2856-20
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
C0SJ6

Quant Time: Jun 25 01:19:41 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG062024.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Jun 21 05:03:55 2024
Response via : Initial Calibration

