


Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG062518\
 Data File : BG035370.D
 Acq On : 25 Jun 2018 9:50
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG060718.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jun 08 17:16:28 2018



AutoFind: Scans 2477, 2478, 2479; Background Corrected with Scan 2471

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	25.7	12714	PASS
68	69	0.00	2	0.7	102	PASS
69	198	0.00	100	29.6	14682	PASS
70	69	0.00	2	0.5	67	PASS
127	198	10	80	36.5	18090	PASS
197	198	0.00	2	0.4	192	PASS
198	198	100	100	100.0	49520	PASS
199	198	5	9	8.0	3965	PASS
275	198	10	60	28.9	14311	PASS
365	198	1	100	3.5	1752	PASS
441	443	0.01	100	82.7	5729	PASS
442	198	50	100	70.9	35117	PASS
443	442	15	24	19.7	6931	PASS