


Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG062518\
 Data File : BG035389.D
 Acq On : 25 Jun 2018 23:22
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG060718.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jun 08 17:16:28 2018



AutoFind: Scans 2477, 2478, 2479; Background Corrected with Scan 2471

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	26.0	15376	PASS
68	69	0.00	2	0.6	110	PASS
69	198	0.00	100	30.9	18247	PASS
70	69	0.00	2	0.3	51	PASS
127	198	10	80	37.8	22346	PASS
197	198	0.00	2	0.4	243	PASS
198	198	100	100	100.0	59122	PASS
199	198	5	9	6.7	3969	PASS
275	198	10	60	30.7	18125	PASS
365	198	1	100	3.7	2203	PASS
441	443	0.01	100	78.0	6158	PASS
442	198	50	100	68.8	40685	PASS
443	442	15	24	19.4	7894	PASS