



Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG062719\
Data File : BG041491.D
Acq On : 27 Jun 2019 16:24
Operator : HP/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 28 08:15:47 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG062619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 27 13:25:55 2019
Response via : Initial Calibration



TIC: BG041491.D

(77) Benzidine



19.667min (+0.003) 585109.03ng

response 701047

Ion	Exp%	Act%
184.00	100	100
185.00	15.90	15.19
183.00	11.80	12.27
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG062719\
Data File : BG041491.D
Acq On : 27 Jun 2019 16:24
Operator : HP/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 28 08:15:47 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG062619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 27 13:25:55 2019
Response via : Initial Calibration



TIC: BG041491.D

(70) Pentachlorophenol (C)

17.076min (+0.003) 32615.96ng

response 122289

Ion	Exp%	Act%
265.70	100	100
268.00	58.70	65.24
264.00	59.20	61.18
0.00	0.00	0.00