



Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG062719\
Data File : BG041506.D
Acq On : 28 Jun 2019 5:06
Operator : HP/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 28 08:18:45 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG062619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 27 13:25:55 2019
Response via : Initial Calibration



TIC: BG041506.D

(70) Pentachlorophenol (C)



17.080min (+0.006) 54866.10ng

response 19766

Ion	Exp%	Act%
265.70	100	100
268.00	58.70	72.12#
264.00	59.20	68.34
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA G\Data\BG062719\
Data File : BG041506.D
Acq On : 28 Jun 2019 5:06
Operator : HP/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jun 28 08:18:45 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG062619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 27 13:25:55 2019
Response via : Initial Calibration



TIC: BG041506.D

(77) Benzidine

19.659min (-0.005) 0.00ng

response 164070

Ion	Exp%	Act%
184.00	100	100
185.00	15.90	14.68
183.00	11.80	12.91
0.00	0.00	0.00