

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG062719\
 Data File : BG041480.D
 Acq On : 27 Jun 2019 8:02
 Operator : HP/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 27 10:07:40 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG062619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 27 06:49:14 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	76	0.00
2	1,4-Dioxane	0.508	0.563	-10.8	80	0.00
3	Pyridine	1.288	1.480	-14.9	86	0.00
4	n-Nitrosodimethylamine	0.717	0.791	-10.3	82	0.00
5 S	2-Fluorophenol	1.122	1.212	-8.0	81	0.00
6	Aniline	2.036	2.187	-7.4	81	0.00
7 S	Phenol-d6	1.579	1.700	-7.7	83	0.00
8	2-Chlorophenol	1.258	1.337	-6.3	79	0.00
9	Benzaldehyde	0.954	0.995	-4.3	81	0.00
10 C	Phenol	1.592	1.727	-8.5	82	0.00
11	bis(2-Chloroethyl)ether	1.262	1.304	-3.3	80	0.00
12	1,3-Dichlorobenzene	1.627	1.711	-5.2	79	0.00
13 C	1,4-Dichlorobenzene	1.641	1.757	-7.1	82	0.00
14	1,2-Dichlorobenzene	1.571	1.665	-6.0	80	0.00
15	Benzyl Alcohol	1.258	1.388	-10.3	83	0.00
16	2,2'-oxybis(1-Chloropropane	1.772	1.851	-4.5	79	0.00
17	2-Methylphenol	1.159	1.185	-2.2	78	0.00
18	Hexachloroethane	0.647	0.705	-9.0	83	0.00
19 P	n-Nitroso-di-n-propylamine	1.178	1.241	-5.3	82	0.00
20	3+4-Methylphenols	1.544	1.749	-13.3	86	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	80	0.00
22	Acetophenone	0.600	0.616	-2.7	81	0.00
23 S	Nitrobenzene-d5	0.480	0.509	-6.0	83	0.00
24	Nitrobenzene	0.481	0.496	-3.1	82	0.00
25	Isophorone	0.861	0.894	-3.8	83	0.00
26 C	2-Nitrophenol	0.201	0.197	2.0	79	0.00
27	2,4-Dimethylphenol	0.280	0.278	0.7	76	0.00
28	bis(2-Chloroethoxy)methane	0.455	0.466	-2.4	82	0.00
29 C	2,4-Dichlorophenol	0.378	0.398	-5.3	82	0.00
30	1,2,4-Trichlorobenzene	0.484	0.492	-1.7	81	0.00
31	Naphthalene	1.106	1.161	-5.0	84	0.00
32	Benzoic acid	0.154	0.159	-3.2	89	-0.01
33	4-Chloroaniline	0.466	0.484	-3.9	81	0.00
34 C	Hexachlorobutadiene	0.384	0.405	-5.5	84	0.00
35	Caprolactam	0.118	0.127	-7.6	86	0.00
36 C	4-Chloro-3-methylphenol	0.409	0.426	-4.2	84	0.00
37	2-Methylnaphthalene	0.817	0.857	-4.9	83	0.00
38	1-Methylnaphthalene	0.780	0.808	-3.6	83	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	86	0.00
40	1,2,4,5-Tetrachlorobenzene	0.881	0.875	0.7	82	0.00
41 P	Hexachlorocyclopentadiene	0.436	0.449	-3.0	81	0.00
42 S	2,4,6-Tribromophenol	0.341	0.354	-3.8	86	0.00
43 C	2,4,6-Trichlorophenol	0.523	0.537	-2.7	84	0.00
44	2,4,5-Trichlorophenol	0.521	0.532	-2.1	87	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG062719\
 Data File : BG041480.D
 Acq On : 27 Jun 2019 8:02
 Operator : HP/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 27 10:07:40 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG062619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 27 06:49:14 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.559	1.605	-3.0	85	0.00
46	1,1'-Biphenyl	1.579	1.597	-1.1	84	0.00
47	2-Chloronaphthalene	1.355	1.379	-1.8	84	0.00
48	2-Nitroaniline	0.459	0.481	-4.8	87	0.00
49	Acenaphthylene	2.026	2.090	-3.2	86	0.00
50	Dimethylphthalate	1.820	1.915	-5.2	88	0.00
51	2,6-Dinitrotoluene	0.386	0.403	-4.4	88	0.00
52 C	Acenaphthene	1.201	1.222	-1.7	85	0.00
53	3-Nitroaniline	0.369	0.394	-6.8	87	0.00
54 P	2,4-Dinitrophenol	0.208	0.217	-4.3	90	0.00
55	Dibenzofuran	2.092	2.187	-4.5	86	0.00
56 P	4-Nitrophenol	0.238	0.240	-0.8	81	0.00
57	2,4-Dinitrotoluene	0.560	0.588	-5.0	87	0.00
58	Fluorene	1.664	1.762	-5.9	88	0.00
59	2,3,4,6-Tetrachlorophenol	0.549	0.584	-6.4	87	0.00
60	Diethylphthalate	1.863	1.998	-7.2	89	0.00
61	4-Chlorophenyl-phenylether	1.073	1.144	-6.6	89	0.00
62	4-Nitroaniline	0.359	0.398	-10.9	91	0.00
63	Azobenzene	1.571	1.685	-7.3	90	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	91	0.00
65	4,6-Dinitro-2-methylphenol	0.124	0.131	-5.6	93	0.00
66 c	n-Nitrosodiphenylamine	0.550	0.551	-0.2	89	0.00
67	4-Bromophenyl-phenylether	0.273	0.271	0.7	90	0.00
68	Hexachlorobenzene	0.294	0.298	-1.4	91	0.00
69	Atrazine	0.233	0.232	0.4	92	0.00
70 C	Pentachlorophenol	0.133	0.141	-6.0	92	0.00
71	Phenanthrene	1.102	1.144	-3.8	92	0.00
72	Anthracene	1.103	1.146	-3.9	92	0.00
73	Carbazole	0.937	1.011	-7.9	95	0.00
74	Di-n-butylphthalate	1.178	1.280	-8.7	95	0.00
75 C	Fluoranthene	1.469	1.608	-9.5	95	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	97	0.00
77	Benzidine	0.444	0.524	-18.0	97	0.00
78	Pyrene	1.365	1.414	-3.6	96	0.00
79 S	Terphenyl-d14	0.941	0.977	-3.8	96	0.00
80	Butylbenzylphthalate	0.510	0.527	-3.3	97	0.00
81	Benzo(a)anthracene	1.390	1.446	-4.0	97	0.00
82	3,3'-Dichlorobenzidine	0.481	0.503	-4.6	96	0.00
83	Chrysene	1.346	1.398	-3.9	98	0.00
84	Bis(2-ethylhexyl)phthalate	0.723	0.730	-1.0	95	0.00
85 c	Di-n-octyl phthalate	1.223	1.222	0.1	94	0.00
86	Indeno(1,2,3-cd)pyrene	1.630	1.633	-0.2	94	0.00
87 I	Perylene-d12	1.000	1.000	0.0	93	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG062719\
Data File : BG041480.D
Acq On : 27 Jun 2019 8:02
Operator : HP/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jun 27 10:07:40 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG062619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 27 06:49:14 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.276	1.304	-2.2	92	0.00
89	Benzo(k)fluoranthene	1.252	1.314	-5.0	96	0.00
90 C	Benzo(a)pyrene	1.212	1.251	-3.2	95	0.00
91	Dibenzo(a,h)anthracene	1.214	1.263	-4.0	95	0.00
92	Benzo(q,h,i)perylene	1.207	1.243	-3.0	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0