

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_G\DATA\BG062817\
 Data File : BG027723.D
 Acq On : 29 Jun 2017 2:43
 Operator : SJ/JU
 Sample : SSTDC0040
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDC0040

Quant Time: Jun 29 07:16:24 2017
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG062817.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 28 19:26:37 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	107	0.00
2	1,4-Dioxane	40.000	42.830	-7.1	113	0.00
3	Pyridine	40.000	41.532	-3.8	106	0.00
4	n-Nitrosodimethylamine	40.000	40.109	-0.3	109	0.00
5 S	2-Fluorophenol	80.000	85.693	-7.1	110	0.00
6	Aniline	40.000	40.029	-0.1	100	0.00
7 S	Phenol-d6	80.000	78.902	1.4	101	0.00
8	2-Chlorophenol	40.000	40.938	-2.3	106	0.00
9	Benzaldehyde	40.000	39.454	1.4	101	0.00
10 C	Phenol	40.000	39.804	0.5	101	0.00
11	bis(2-Chloroethyl)ether	40.000	39.945	0.1	104	0.00
12	1,3-Dichlorobenzene	40.000	41.765	-4.4	108	0.00
13 C	1,4-Dichlorobenzene	40.000	41.176	-2.9	107	0.00
14	1,2-Dichlorobenzene	40.000	40.890	-2.2	106	0.00
15	Benzyl Alcohol	40.000	37.667	5.8	95	0.00
16	2,2'-oxybis(1-Chloropropane)	40.000	39.922	0.2	103	0.00
17	2-Methylphenol	40.000	38.685	3.3	96	0.00
18	Hexachloroethane	40.000	42.372	-5.9	113	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	36.235	9.4	92	0.00
20	3+4-Methylphenols	40.000	37.313	6.7	94	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	96	0.00
22	Acetophenone	40.000	39.457	1.4	93	0.00
23 S	Nitrobenzene-d5	80.000	81.403	-1.8	96	0.00
24	Nitrobenzene	40.000	40.425	-1.1	95	0.00
25	Isophorone	40.000	37.577	6.1	89	0.00
26 C	2-Nitrophenol	40.000	40.954	-2.4	93	0.00
27	2,4-Dimethylphenol	40.000	39.468	1.3	91	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.815	3.0	92	0.00
29 C	2,4-Dichlorophenol	40.000	39.450	1.4	92	0.00
30	1,2,4-Trichlorobenzene	40.000	41.360	-3.4	99	0.00
31	Naphthalene	40.000	39.770	0.6	94	0.00
32	Benzoic acid	40.000	33.517	16.2	81	0.00
33	4-Chloroaniline	40.000	38.016	5.0	88	0.00
34 C	Hexachlorobutadiene	40.000	43.698	-9.2	107	0.00
35	Caprolactam	40.000	33.903	15.2	80	0.00
36 C	4-Chloro-3-methylphenol	40.000	35.888	10.3	84	0.00
37	2-Methylnaphthalene	40.000	38.354	4.1	91	0.00
38 I	Acenaphthene-d10	20.000	20.000	0.0	81	0.00
39	1,2,4,5-Tetrachlorobenzene	40.000	44.524	-11.3	90	0.00
40 P	Hexachlorocyclopentadiene	40.000	47.664	-19.2	96	0.00
41 S	2,4,6-Tribromophenol	80.000	81.468	-1.8	81	0.00
42 C	2,4,6-Trichlorophenol	40.000	41.734	-4.3	84	0.00
43	2,4,5-Trichlorophenol	40.000	42.150	-5.4	82	0.00
44 S	2-Fluorobiphenyl	80.000	86.681	-8.4	87	0.00

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_G\DATA\BG062817\
 Data File : BG027723.D
 Acq On : 29 Jun 2017 2:43
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 29 07:16:24 2017
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG062817.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 28 19:26:37 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	40.000	42.944	-7.4	87	0.00
46	2-Chloronaphthalene	40.000	42.224	-5.6	85	0.00
47	2-Nitroaniline	40.000	41.281	-3.2	81	0.00
48	Acenaphthylene	40.000	41.652	-4.1	83	0.00
49	Dimethylphthalate	40.000	39.498	1.3	80	0.00
50	2,6-Dinitrotoluene	40.000	41.102	-2.8	81	0.00
51 C	Acenaphthene	40.000	41.506	-3.8	83	0.00
52	3-Nitroaniline	40.000	40.834	-2.1	81	0.00
53 P	2,4-Dinitrophenol	40.000	38.521	3.7	83	0.00
54	Dibenzofuran	40.000	40.194	-0.5	81	0.00
55 P	4-Nitrophenol	40.000	44.120	-10.3	86	0.00
56	2,4-Dinitrotoluene	40.000	42.046	-5.1	82	0.00
57	Fluorene	40.000	40.465	-1.2	81	0.00
58	2,3,4,6-Tetrachlorophenol	40.000	39.175	2.1	79	0.00
59	Diethylphthalate	40.000	39.704	0.7	81	0.00
60	4-Chlorophenyl-phenylether	40.000	40.774	-1.9	81	0.00
61	4-Nitroaniline	40.000	44.057	-10.1	88	0.00
62	Azobenzene	40.000	40.252	-0.6	81	0.00
63 I	Phenanthrene-d10	20.000	20.000	0.0	82	0.00
64	4,6-Dinitro-2-methylphenol	40.000	39.976	0.1	86	0.00
65 c	n-Nitrosodiphenylamine	40.000	40.285	-0.7	81	0.00
66	4-Bromophenyl-phenylether	40.000	40.402	-1.0	81	0.00
67	Hexachlorobenzene	40.000	40.689	-1.7	83	0.00
68	Atrazine	40.000	44.474	-11.2	88	0.00
69 C	Pentachlorophenol	40.000	39.272	1.8	84	0.00
70	Phenanthrene	40.000	41.081	-2.7	83	0.00
71	Anthracene	40.000	41.676	-4.2	83	0.00
72	Carbazole	40.000	44.188	-10.5	89	0.00
73	Di-n-butylphthalate	40.000	45.663	-14.2	91	0.00
74 C	Fluoranthene	40.000	45.032	-12.6	92	0.00
75 I	Chrysene-d12	20.000	20.000	0.0	95	0.00
76	Benzidine	40.000	45.355	-13.4	104	0.00
77	Pyrene	40.000	40.016	-0.0	94	0.00
78 S	Terphenyl-d14	80.000	83.460	-4.3	95	0.00
79	Butylbenzylphthalate	40.000	40.924	-2.3	95	0.00
80	Benzo(a)anthracene	40.000	40.735	-1.8	95	0.00
81	3,3'-Dichlorobenzidine	40.000	41.683	-4.2	96	0.00
82	Chrysene	40.000	41.167	-2.9	96	0.00
83	Bis(2-ethylhexyl)phthalate	40.000	40.791	-2.0	94	0.00
84 c	Di-n-octyl phthalate	40.000	41.350	-3.4	94	0.00
85	Indeno(1,2,3-cd)pyrene	40.000	41.866	-4.7	96	0.00
86 I	Perylene-d12	20.000	20.000	0.0	96	0.00
87	Benzo(b)fluoranthene	40.000	40.783	-2.0	96	0.00

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_G\DATA\BG062817\
Data File : BG027723.D
Acq On : 29 Jun 2017 2:43
Operator : SJ/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jun 29 07:16:24 2017
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG062817.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 28 19:26:37 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	40.000	40.903	-2.3	96	0.00
89 C	Benzo(a)pyrene	40.000	40.931	-2.3	96	0.00
90	Dibenzo(a,h)anthracene	40.000	41.125	-2.8	97	0.00
91	Benzo(g,h,i)perylene	40.000	40.812	-2.0	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0