

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062822\
 Data File : BG054115.D
 Acq On : 28 Jun 2022 11:49
 Operator : CG/JU
 Sample : PB145785TB
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB145785TB

Quant Time: Jun 28 14:38:48 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 27 00:27:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.071	152	19264	20.000	ng	0.00
21) Naphthalene-d8	10.880	136	85464	20.000	ng	0.00
39) Acenaphthene-d10	14.699	164	73191	20.000	ng	0.00
64) Phenanthrene-d10	17.442	188	195724	20.000	ng	0.00
76) Chrysene-d12	21.720	240	192682	20.000	ng	0.00
86) Perylene-d12	24.981	264	194051	20.000	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.639	112	155782	154.610	ng	0.00
7) Phenol-d6	7.231	99	218981	160.208	ng	0.00
23) Nitrobenzene-d5	9.229	82	138048	90.772	ng	0.00
42) 2,4,6-Tribromophenol	16.185	330	160507	136.492	ng	0.00
45) 2-Fluorobiphenyl	13.324	172	430801	85.590	ng	0.00
79) Terphenyl-d14	20.051	244	922623	96.441	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG062822\
Data File : BG054115.D
Acq On : 28 Jun 2022 11:49
Operator : CG/JU
Sample : PB145785TB
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB145785TB

Quant Time: Jun 28 14:38:48 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG061322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 27 00:27:54 2022
Response via : Initial Calibration

