

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG062916\
 Data File : BG022713.D
 Acq On : 30 Jun 2016 3:58
 Operator : UM/SJ
 Sample : H3828-09
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 S8-0579

Quant Time: Jun 30 05:24:05 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG062516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 24 20:14:59 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.16	152	118435	20.00	ng	-0.02
21) Naphthalene-d8	11.00	136	560538	20.00	ng	0.00
38) Acenaphthene-d10	14.80	164	554649	20.00	ng	0.00
63) Phenanthrene-d10	17.56	188	958122	20.00	ng	0.00
75) Chrysene-d12	21.85	240	1045648	20.00	ng	0.00
86) Perylene-d12	25.22	264	1080568	20.00	ng	0.01

System Monitoring Compounds

5) 2-Fluorophenol	5.71	112	458318	62.07	ng	0.00
7) Phenol-d6	7.33	99	541477	52.03	ng	0.03
23) Nitrobenzene-d5	9.34	82	1091351	82.40	ng	0.00
41) 2,4,6-Tribromophenol	16.30	330	880784	100.96	ng	0.01
44) 2-Fluorobiphenyl	13.43	172	2213754	63.30	ng	0.00
78) Terphenyl-d14	20.16	244	2658501	67.36	ng	0.00

Target Compounds


						Qvalue
9) Benzaldehyde	7.30	77	1316620	359.46	ng	# 88
15) Benzyl Alcohol	8.41	79	186404	19.36	ng	94
16) 2,2'-oxybis(1-Chloropropan	8.58	45	113953	11.38	ng	75
17) 2-Methylphenol	8.63	107	31444	4.34	ng	# 86
18) Hexachloroethane	9.27	117	15709	4.07	ng	# 1
20) 3+4-Methylphenols	8.95	107	60487	6.06	ng	97
22) Acetophenone	9.02	105	5802908	358.10	ng	# 67
24) Nitrobenzene	9.32	77	56601	3.87	ng	# 1
27) 2,4-Dimethylphenol	10.17	122	310489	30.82	ng	91
28) bis(2-Chloroethoxy)methane	10.42	93	200092	14.25	ng	# 61
31) Naphthalene	11.04	128	1260080	44.72	ng	98
32) Benzoic acid	10.17	122	310489	47.24	ng	# 46
33) 4-Chloroaniline	11.05	127	197154	16.24	ng	# 37
37) 2-Methylnaphthalene	12.64	142	282340	12.92	ng	98
50) 2,6-Dinitrotoluene	14.32	165	48737	4.87	ng	# 31


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG062916\
Data File : BG022713.D
Acq On : 30 Jun 2016 3:58
Operator : UM/SJ
Sample : H3828-09
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
S8-0579

Quant Time: Jun 30 05:24:05 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG062516.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 24 20:14:59 2016
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 8.16 min Scan# 905

Delta R.T. -0.02 min

Lab File: BG022713.D

Acq: 30 Jun 2016 3:58

Instrument :

BNA_G

ClientSampleId :

S8-0579


Tgt Ion:152 Resp: 118435

Ion Ratio Lower Upper

152 100

150 151.3 144.7 217.1




115 68.0 64.0 96.0



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#5

2-Fluorophenol

Concen: 62.07 ng

RT: 5.71 min Scan# 489

Delta R.T. 0.01 min

Lab File: BG022713.D

Acq: 30 Jun 2016 3:58


Tgt Ion:112 Resp: 458318

Ion Ratio Lower Upper

112 100

64 72.3 59.0 88.4



63 42.1 37.8 56.8




Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BGC

Ion 63.00 (62.70 to 63.70): BGC





#7

Phenol-d6

Concen: 52.03 ng

RT: 7.33 min Scan# 764

Delta R.T. 0.03 min

Lab File: BG022713.D

Acq: 30 Jun 2016 3:58

Instrument :

BNA_G

ClientSampleId :

S8-0579


Tgt Ion: 99 Resp: 541477

Ion Ratio Lower Upper




99 100

42 18.8 17.8 26.6

71 47.8 41.5 62.3



Abundance Ion 99.00 (98.70 to 99.70): BGC
Ion 42.00 (41.70 to 42.70): BGC
Ion 71.00 (70.70 to 71.70): BGC



#9

Benzaldehyde

Concen: 359.46 ng

RT: 7.30 min Scan# 759

Delta R.T. 0.00 min

Lab File: BG022713.D

Acq: 30 Jun 2016 3:58


Tgt Ion: 77 Resp: 1316620

Ion Ratio Lower Upper



77 100


105 101.2 58.7 98.7#

106 87.2 67.5 107.5



Abundance Ion 77.00 (76.70 to 77.70): BGC
Ion 105.00 (104.70 to 105.70): BGC
Ion 106.00 (105.70 to 106.70): BGC









#15
Benzyl Alcohol
Concen: 19.36 ng
RT: 8.41 min Scan# 948
Delta R.T. 0.01 min
Lab File: BG022713.D
Acq: 30 Jun 2016 3:58




Instrument :
BNA_G
ClientSampleId :
S8-0579


Tgt Ion: 79 Resp: 186404
Ion Ratio Lower Upper
79 100
108 62.7 46.5 69.7
77 60.9 52.3 78.5



#16
2,2'-oxybis(1-Chloropropane)
Concen: 11.38 ng
RT: 8.58 min Scan# 977
Delta R.T. -0.12 min
Lab File: BG022713.D
Acq: 30 Jun 2016 3:58

Tgt Ion: 45 Resp: 113953
Ion Ratio Lower Upper
45 100
77 7.9 0.6 40.6
79 7.2 0.0 36.2









#17
2-Methylphenol
Concen: 4.34 ng
RT: 8.63 min Scan# 985
Delta R.T. 0.03 min
Lab File: BG022713.D
Acq: 30 Jun 2016 3:58




Instrument :
BNA_G
ClientSampleId :
S8-0579


Tgt Ion	Ratio	Lower	Upper
107	100		
108	107.7	92.0	138.0
77	51.0	55.8	83.6#
79	53.4	55.0	82.6#



#18
Hexachloroethane
Concen: 4.07 ng
RT: 9.27 min Scan# 1095
Delta R.T. 0.01 min
Lab File: BG022713.D
Acq: 30 Jun 2016 3:58

Tgt Ion	Ratio	Lower	Upper
117	100		
119	312.7	73.7	110.5#
201	0.0	88.7	133.1#









#20
3+4-Methylphenols
Concen: 6.06 ng
RT: 8.95 min Scan# 1040
Delta R.T. 0.03 min
Lab File: BG022713.D
Acq: 30 Jun 2016 3:58




Instrument :
BNA_G
ClientSampleId :
S8-0579


Tgt Ion:	107	Resp:	60487
Ion	Ratio	Lower	Upper
107	100		
108	88.9	72.5	112.5
77	50.2	30.1	70.1
79	41.1	24.2	64.2



#21
Naphthalene-d8
Concen: 20.00 ng
RT: 11.00 min Scan# 1388
Delta R.T. 0.00 min
Lab File: BG022713.D
Acq: 30 Jun 2016 3:58

Tgt Ion:	136	Resp:	560538
Ion	Ratio	Lower	Upper
136	100		
137	11.4	9.6	14.4
54	9.5	7.3	10.9
68	7.7	5.3	7.9





#22

Acetophenone

Concen: 358.10 ng

RT: 9.02 min Scan# 1051

Delta R.T. 0.03 min

Lab File: BG022713.D

Acq: 30 Jun 2016 3:58

Instrument :

BNA_G

ClientSampleId :

S8-0579

Tgt Ion:105 Resp: 5802908


Ion Ratio Lower Upper

105 100

71 0.1 2.6 3.8#

51 51.1 27.1 40.7#

120 38.3 15.1 22.7#






Abundance Ion 105.00 (104.70 to 105.70): E

Ion 71.00 (70.70 to 71.70): BGC

Ion 51.00 (50.70 to 51.70): BGC

Ion 120.00 (119.70 to 120.70): E



#23

Nitrobenzene-d5

Concen: 82.40 ng

RT: 9.34 min Scan# 1107

Delta R.T. 0.01 min

Lab File: BG022713.D

Acq: 30 Jun 2016 3:58


Tgt Ion: 82 Resp: 1091351

Ion Ratio Lower Upper

82 100

128 35.2 24.4 36.6



54 49.7 41.4 62.0




Abundance Ion 82.00 (81.70 to 82.70): BGC

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BGC









#24
 Nitrobenzene
 Concen: 3.87 ng
 RT: 9.32 min Scan# 1103
 Delta R.T. -0.06 min
 Lab File: BG022713.D
 Acq: 30 Jun 2016 3:58




Instrument :
 BNA_G
 ClientSampleId :
 S8-0579


Tgt Ion: 77 Resp: 56601
 Ion Ratio Lower Upper
 77 100
 123 2.5 25.0 37.6#
 65 1558.4 13.4 20.0#



#27
 2,4-Dimethylphenol
 Concen: 30.82 ng
 RT: 10.17 min Scan# 1247
 Delta R.T. 0.02 min
 Lab File: BG022713.D
 Acq: 30 Jun 2016 3:58

Tgt Ion: 122 Resp: 310489
 Ion Ratio Lower Upper
 122 100
 107 157.8 117.0 175.4
 121 56.3 50.1 75.1









#28
bis(2-Chloroethoxy)methane
Concen: 14.25 ng
RT: 10.42 min Scan# 1290
Delta R.T. 0.04 min
Lab File: BG022713.D
Acq: 30 Jun 2016 3:58




Instrument :
BNA_G
ClientSampleId :
S8-0579


Tgt Ion: 93 Resp: 200092
Ion Ratio Lower Upper
93 100
95 6.8 25.4 38.2#
123 3.0 8.2 12.2#



#31
Naphthalene
Concen: 44.72 ng
RT: 11.04 min Scan# 1396
Delta R.T. 0.00 min
Lab File: BG022713.D
Acq: 30 Jun 2016 3:58

Tgt Ion: 128 Resp: 1260080
Ion Ratio Lower Upper
128 100
129 11.5 9.4 14.0
127 15.2 11.3 16.9









#32
Benzoic acid
Concen: 47.24 ng
RT: 10.17 min Scan# 1247
Delta R.T. -0.03 min
Lab File: BG022713.D
Acq: 30 Jun 2016 3:58




Instrument :
BNA_G
ClientSampleId :
S8-0579


Tgt Ion	Ratio	Lower	Upper
122	100		
105	28.2	117.2	157.2#
77	111.1	109.2	149.2



#33
4-Chloroaniline
Concen: 16.24 ng
RT: 11.05 min Scan# 1397
Delta R.T. -0.10 min
Lab File: BG022713.D
Acq: 30 Jun 2016 3:58

Tgt Ion	Ratio	Lower	Upper
127	100		
129	75.4	27.9	41.9#
65	3.9	36.8	55.2#
92	0.3	21.0	31.6#









#37
2-Methylnaphthalene
Concen: 12.92 ng
RT: 12.64 min Scan# 1668
Delta R.T. 0.00 min
Lab File: BG022713.D
Acq: 30 Jun 2016 3:58




Instrument :
BNA_G
ClientSampleId :
S8-0579


Tgt Ion:142 Resp: 282340
Ion Ratio Lower Upper
142 100
141 88.3 70.2 105.2
115 48.7 41.7 62.5



#38
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.80 min Scan# 2036
Delta R.T. 0.00 min
Lab File: BG022713.D
Acq: 30 Jun 2016 3:58

Tgt Ion:164 Resp: 554649
Ion Ratio Lower Upper
164 100
162 80.3 87.7 131.5#
160 36.4 37.2 55.8#









#41
 2,4,6-Tribromophenol
 Concen: 100.96 ng
 RT: 16.30 min Scan# 2291
 Delta R.T. 0.01 min
 Lab File: BG022713.D
 Acq: 30 Jun 2016 3:58

Instrument :
 BNA_G
 ClientSampleId :
 S8-0579

Tgt Ion:330 Resp: 880784
 Ion Ratio Lower Upper
 330 100
 332 96.0 77.4 116.2
 141 35.9 31.7 47.5




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#44
 2-Fluorobiphenyl
 Concen: 63.30 ng
 RT: 13.43 min Scan# 1802
 Delta R.T. 0.00 min
 Lab File: BG022713.D
 Acq: 30 Jun 2016 3:58

Tgt Ion:172 Resp: 2213754
 Ion Ratio Lower Upper
 172 100
 171 43.4 31.6 47.4
 170 28.9 21.3 31.9



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E





#50

2,6-Dinitrotoluene

Concen: 4.87 ng

RT: 14.32 min Scan# 1954

Delta R.T. -0.04 min

Lab File: BG022713.D

Acq: 30 Jun 2016 3:58

Instrument :

BNA_G

ClientSampleId :

S8-0579


Tgt Ion:165 Resp: 48737

Ion Ratio Lower Upper

165 100

63 23.6 64.3 96.5#




89 15.9 64.3 96.5#



Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BGC

Ion 89.00 (88.70 to 89.70): BGC



#63

Phenanthrene-d10

Concen: 20.00 ng

RT: 17.56 min Scan# 2505

Delta R.T. 0.01 min

Lab File: BG022713.D

Acq: 30 Jun 2016 3:58


Tgt Ion:188 Resp: 958122

Ion Ratio Lower Upper

188 100

94 7.9 5.2 7.8#



80 9.2 7.2 10.8




Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BGC

Ion 80.00 (79.70 to 80.70): BGC





#75

Chrysene-d12

Concen: 20.00 ng

RT: 21.85 min Scan# 3236

Delta R.T. 0.01 min

Lab File: BG022713.D

Acq: 30 Jun 2016 3:58

Instrument :

BNA_G

ClientSampleId :

S8-0579


Tgt Ion:240 Resp: 1045648

Ion Ratio Lower Upper

240 100

120 5.3 4.1 6.1




236 27.5 21.1 31.7



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#78

Terphenyl-d14

Concen: 67.36 ng

RT: 20.16 min Scan# 2948

Delta R.T. 0.01 min

Lab File: BG022713.D

Acq: 30 Jun 2016 3:58


Tgt Ion:244 Resp: 2658501

Ion Ratio Lower Upper

244 100

212 13.2 10.8 16.2



122 8.8 8.0 12.0




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E





#86
Perylene-d12
Concen: 20.00 ng
RT: 25.22 min Scan# 3809
Delta R.T. 0.01 min
Lab File: BG022713.D
Acq: 30 Jun 2016 3:58

Instrument :
BNA_G
ClientSampleId :
S8-0579

Tgt Ion: 264 Resp: 1080568

Ion	Ratio	Lower	Upper
264	100		
260	25.1	18.4	27.6
265	22.7	18.9	28.3

