

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG063016\
 Data File : BG022745.D
 Acq On : 1 Jul 2016 3:00
 Operator : UM/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleID :
 SSTD02008

Quant Time: Jul 01 07:59:23 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG063016.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jul 01 05:25:08 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.16	152	100185	20.00	ng/ul	0.00
18) Naphthalene-d8	10.60	136	9947	20.00	ng/ul	-0.38
35) Acenaphthene-d10	14.28	164	68	20.00	ng/ul	-0.51
61) Phenanthrene-d10	16.94	188	520	20.00	ng/ul	-0.62
75) Chrysene-d12	21.08	240	121	20.00	ng/ul	-0.77
83) Perylene-d12	24.33	264	174	20.00	ng/ul	-0.89

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.54	96	15128	8.44	ng/uL	0.00
5) Phenol-d5	7.31	99	199067	18.89	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.47	67	108264	18.22	ng/ul	0.00
9) 2-Chlorophenol-d4	7.69	132	124940	17.75	ng/ul	0.00
13) 4-Methylphenol-d8	8.87	113	144225	17.62	ng/ul	0.00
19) Nitrobenzene-d5	8.95	128	163	2.06	ng/ul	-0.38
22) 2-Nitrophenol-d4	9.47	143	140	1.46	ng/ul	-0.58
26) 2,4-Dichlorophenol-d3	10.22	165	223	1.11	ng/ul	-0.38
29) 4-Chloroaniline-d4	10.61	131	2451	14.47	ng/ul	-0.51
43) Dimethylphthalate-d6	13.69	166	789	142.82	ng/ul	-0.51
46) Acenaphthylene-d8	14.03	160	136	21.48	ng/ul	-0.47
51) 4-Nitrophenol-d4	14.32	143	125	144.11	ng/ul	-0.68
57) Fluorene-d10	15.15	176	395	75.47	ng/ul	-0.64
62) 4,6-Dinitro-2-methylphenol	15.42	200	1437	429.67	ng/ul	-0.49
70) Anthracene-d10	17.00	188	2145	88.75	ng/ul	-0.66
76) Pyrene-d10	19.03	212	57	9.86	ng/ul	-0.91
87) Benzo(a)pyrene-d12	24.12	264	106	12.82	ng/ul	-0.87

Target Compounds

					Qvalue
2) 1,4-Dioxane	3.58	88	16523	7.86 ng/uL#	78
4) Benzaldehyde	7.29	77	129024	19.26 ng/ul#	75
6) Phenol	7.34	94	192393	17.82 ng/ul#	66
8) Bis(2-Chloroethyl)ether	7.57	93	137414	18.62 ng/ul	89
10) 2-Chlorophenol	7.72	128	125083	17.47 ng/ul#	85
11) 2-Methylphenol	8.60	108	148463	18.31 ng/ul	96
12) 2,2'-oxybis(1-Chloropropan	8.69	45	151175	18.22 ng/ul#	89
14) Acetophenone	8.98	105	240771	18.21 ng/ul#	77
15) N-Nitroso-di-n-propylamine	8.96	70	139337	17.54 ng/ul#	87
16) 4-Methylphenol	8.93	108	158791	17.61 ng/ul	99
17) Hexachloroethane	9.25	117	56696	18.60 ng/ul#	77
20) Nitrobenzene	9.05	77	486	1.99 ng/ul#	36
21) Isophorone	9.33	82	201721	463.64 ng/ul#	64
23) 2-Nitrophenol	9.89	139	5580	54.96 ng/ul#	28
24) 2,4-Dimethylphenol	10.14	107	202667	833.90 ng/ul	85
25) Bis(2-Chloroethoxy)methane	10.38	93	207130	806.82 ng/ul	97
27) 2,4-Dichlorophenol	10.63	162	162146	806.31 ng/ul	100
28) Naphthalene	11.03	128	423530	832.66 ng/ul	99
30) 4-Chloroaniline	11.14	127	143650	830.25 ng/ul	96
32) Caprolactam	11.48	113	128	1.91 ng/ul#	1
33) 4-Chloro-3-methylphenol	11.67	107	286	1.12 ng/ul#	63
34) 2-Methylnaphthalene	12.63	142	328146	779.76 ng/ul	100
36) 1,2,4,5-Tetrachlorobenzene	12.99	216	217732	85664.57 ng/ul#	94

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG063016\
 Data File : BG022745.D
 Acq On : 1 Jul 2016 3:00
 Operator : UM/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_G
 ClientSampleId :
 SSTD02008

Quant Time: Jul 01 07:59:23 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG063016.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jul 01 05:25:08 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)

37) Hexachlorocyclopentadiene	12.97	237	124496	81114.70	ng/ul	98
39) 2,4,5-Trichlorophenol	13.23	196	174856	94776.56	ng/ul	95
40) 1,1'-Biphenyl	13.00	154	793	153.37	ng/ul#	60
41) 2-Chloronaphthalene	13.23	162	14389	3436.37	ng/ul#	53
42) 2-Nitroaniline	13.33	65	136	83.98	ng/ul#	1
44) Dimethylphthalate	13.67	163	49449	8766.51	ng/ul#	1
45) 2,6-Dinitrotoluene	13.97	165	139	113.83	ng/ul#	34
47) Acenaphthylene	13.76	152	184	28.87	ng/ul#	59
48) 3-Nitroaniline	14.20	138	33804	35425.35	ng/ul#	1
49) Acenaphthene	14.37	153	232	53.57	ng/ul#	15
50) 2,4-Dinitrophenol	14.34	184	58	85.47	ng/ul#	1
52) 4-Nitrophenol	14.48	109	555	503.16	ng/ul#	27
53) Dibenzofuran	14.31	168	109	16.29	ng/ul	93
54) 2,4-Dinitrotoluene	14.62	165	186	105.33	ng/ul#	46
55) 2,3,4,6-Tetrachlorophenol	15.20	232	117	62.30	ng/ul#	1
56) Diethylphthalate	15.05	149	194	33.85	ng/ul#	59
58) Fluorene	15.38	166	143	26.60	ng/ul#	8
59) 4-Chlorophenyl-phenylether	15.41	204	361	112.11	ng/ul#	44
60) 4-Nitroaniline	15.22	138	232	202.40	ng/ul#	24
63) 4,6-Dinitro-2-methylphenol	15.42	198	11227	3167.53	ng/ul#	86
64) N-Nitrosodiphenylamine	15.42	169	18178	1187.14	ng/ul#	1
65) 4-Bromophenyl-phenylether	16.45	248	1334	191.78	ng/ul#	63
66) Hexachlorobenzene	16.53	284	102	13.85	ng/ul#	42
67) Atrazine	16.39	200	340	53.25	ng/ul#	36
68) Pentachlorophenol	16.79	266	121	28.58	ng/ul#	36
69) Phenanthrene	16.98	178	191	7.18	ng/ul#	18
71) Anthracene	17.09	178	112	4.10	ng/ul#	59
72) Carbazole	17.34	167	130	6.43	ng/ul#	33
73) Di-n-butylphthalate	17.86	149	79	3.04	ng/ul#	1
74) Fluoranthene	18.87	202	97	3.50	ng/ul#	69
77) Pyrene	19.60	202	1071092	158131.09	ng/ul#	82
78) Butylbenzylphthalate	20.14	149	548	208.61	ng/ul#	26
79) 3,3'-Dichlorobenzidine	20.99	252	211	92.08	ng/ul#	26
80) Benzo(a)anthracene	21.13	228	277	39.03	ng/ul#	51
81) Bis(2-ethylhexyl)phthalate	20.93	149	677	199.71	ng/ul#	52
82) Chrysene	21.13	228	277	42.88	ng/ul#	50
84) Di-n-octyl phthalate	22.19	149	719	93.76	ng/ul	100
85) Benzo(b)fluoranthene	23.29	252	230	21.15	ng/ul#	73
86) Benzo(k)fluoranthene	23.34	252	729	71.37	ng/ul#	1
88) Benzo(a)pyrene	24.14	252	1184085	114618.83	ng/ul#	88
89) Indeno(1,2,3-cd)pyrene	28.02	276	82	7.51	ng/ul#	53
90) Dibenzo(a,h)anthracene	28.11	278	73	8.08	ng/ul#	1
91) Benzo(g,h,i)perylene	29.28	276	2470	288.46	ng/ul#	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2

1,4-Dioxane

Concen: 7.86 ng/uL

RT: 3.58 min Scan# 126

Delta R.T. 0.00 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00

Instrument :

BNA_G

ClientSampleId :

SSTD02008





Tgt Ion: 88 Resp: 16523

Ion Ratio Lower Upper

88 100

43 38.7 16.6 24.8#

58 71.2 47.2 70.8#



#4

Benzaldehyde

Concen: 19.26 ng/uL

RT: 7.29 min Scan# 758

Delta R.T. 0.00 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00




Tgt Ion: 77 Resp: 129024


Ion Ratio Lower Upper

77 100

105 77.6 83.8 125.8#

106 74.4 77.9 116.9#





#6

Phenol

Concen: 17.82 ng/ul

RT: 7.34 min Scan# 766

Delta R.T. 0.00 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00

Instrument :

BNA_G

ClientSampleId :

SSTD02008


Tgt Ion: 94 Resp: 192393

Ion Ratio Lower Upper




94 100

65 35.5 16.8 25.2#

66 47.1 22.6 33.8#



Abundance Ion 94.00 (93.70 to 94.70): BGC
Ion 65.00 (64.70 to 65.70): BGC
Ion 66.00 (65.70 to 66.70): BGC



#8

Bis(2-Chloroethyl)ether

Concen: 18.62 ng/ul

RT: 7.57 min Scan# 805

Delta R.T. -0.00 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00


Tgt Ion: 93 Resp: 137414

Ion Ratio Lower Upper



93 100


63 81.7 56.0 84.0

95 31.4 27.7 41.5



Abundance Ion 93.00 (92.70 to 93.70): BGC
Ion 63.00 (62.70 to 63.70): BGC
Ion 95.00 (94.70 to 95.70): BGC









#10
2-Chlorophenol
Concen: 17.47 ng/ul
RT: 7.72 min Scan# 831
Delta R.T. -0.00 min
Lab File: BG022745.D
Acq: 1 Jul 2016 3:00




Instrument :
BNA_G
ClientSampleId :
SSTD02008


Tgt Ion:128 Resp: 125083
Ion Ratio Lower Upper
128 100
64 58.0 34.2 51.4#
130 32.1 27.0 40.4



#11
2-Methylphenol
Concen: 18.31 ng/ul
RT: 8.60 min Scan# 980
Delta R.T. -0.00 min
Lab File: BG022745.D
Acq: 1 Jul 2016 3:00

Tgt Ion:108 Resp: 148463
Ion Ratio Lower Upper
108 100
107 93.9 78.6 117.8









#12
2,2'-oxybis(1-Chloropropane)
Concen: 18.22 ng/ul
RT: 8.69 min Scan# 996
Delta R.T. -0.00 min
Lab File: BG022745.D
Acq: 1 Jul 2016 3:00




Instrument :
BNA_G
ClientSampleId :
SSTD02008


Tgt Ion: 45 Resp: 151175
Ion Ratio Lower Upper
45 100
77 24.2 13.4 20.2#
79 14.0 10.2 15.4



#14
Acetophenone
Concen: 18.21 ng/ul
RT: 8.98 min Scan# 1046
Delta R.T. -0.00 min
Lab File: BG022745.D
Acq: 1 Jul 2016 3:00

Tgt Ion: 105 Resp: 240771
Ion Ratio Lower Upper
105 100
77 89.8 56.1 84.1#
51 27.5 14.6 22.0#









#15
N-Nitroso-di-n-propylamine
Concen: 17.54 ng/ul
RT: 8.96 min Scan# 1042
Delta R.T. -0.00 min
Lab File: BG022745.D
Acq: 1 Jul 2016 3:00

Instrument :
BNA_G
ClientSampleId :
SSTD02008

Tgt Ion	Ratio	Lower	Upper
70	100		
42	48.7	31.4	47.2#
101	8.7	8.6	13.0
130	17.6	19.0	28.4#




Abundance Ion 70.00 (69.70 to 70.70): BGC
Ion 42.00 (41.70 to 42.70): BGC
Ion 101.00 (100.70 to 101.70): BGC
Ion 130.00 (129.70 to 130.70): BGC






#16
4-Methylphenol
Concen: 17.61 ng/ul
RT: 8.93 min Scan# 1036
Delta R.T. -0.01 min
Lab File: BG022745.D
Acq: 1 Jul 2016 3:00

Tgt Ion	Ratio	Lower	Upper
108	100		
107	109.4	88.3	132.5



Abundance Ion 108.00 (107.70 to 108.70): BGC
Ion 107.00 (106.70 to 107.70): BGC









#17
Hexachloroethane
Concen: 18.60 ng/ul
RT: 9.25 min Scan# 1091
Delta R.T. -0.00 min
Lab File: BG022745.D
Acq: 1 Jul 2016 3:00




Instrument :
BNA_G
ClientSampleId :
SSTD02008


Tgt Ion: 117 Resp: 56696
Ion Ratio Lower Upper
117 100
201 111.8 76.8 115.2
199 78.6 43.5 65.3#



#20
Nitrobenzene
Concen: 1.99 ng/ul
RT: 9.05 min Scan# 1058
Delta R.T. -0.32 min
Lab File: BG022745.D
Acq: 1 Jul 2016 3:00

Tgt Ion: 77 Resp: 486
Ion Ratio Lower Upper
77 100
123 0.0 39.5 59.3#
65 0.0 9.0 13.4#





#21

Isophorone

Concen: 463.64 ng/ul

RT: 9.33 min Scan# 1105

Delta R.T. -0.56 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00

Instrument :

BNA_G

ClientSampled :

SSTD02008





Tgt Ion: 82 Resp: 201721

Ion Ratio Lower Upper

82 100

95 0.3 5.1 7.7#

138 0.0 15.0 22.6#



#23

2-Nitrophenol

Concen: 54.96 ng/ul

RT: 9.89 min Scan# 1200

Delta R.T. -0.20 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00




Tgt Ion: 139 Resp: 5580


Ion Ratio Lower Upper

139 100

65 107.4 31.8 47.8#

109 22.5 17.1 25.7









#24
 2,4-Dimethylphenol
 Concen: 833.90 ng/ul
 RT: 10.14 min Scan# 1243
 Delta R.T. -0.00 min
 Lab File: BG022745.D
 Acq: 1 Jul 2016 3:00




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02008


Tgt Ion	Ratio	Lower	Upper
107	100		
121	44.9	43.7	65.5
122	72.6	70.3	105.5



#25
 Bis(2-Chloroethoxy)methane
 Concen: 806.82 ng/ul
 RT: 10.38 min Scan# 1283
 Delta R.T. -0.00 min
 Lab File: BG022745.D
 Acq: 1 Jul 2016 3:00

Tgt Ion	Ratio	Lower	Upper
93	100		
95	30.9	26.0	39.0
123	10.7	10.1	15.1









#27
 2,4-Dichlorophenol
 Concen: 806.31 ng/ul
 RT: 10.63 min Scan# 1326
 Delta R.T. 0.00 min
 Lab File: BG022745.D
 Acq: 1 Jul 2016 3:00




Instrument :
 BNA_G
 ClientSampleId :
 SST02008


Tgt Ion	Ratio	Lower	Upper
162	100		
164	64.9	51.6	77.4
98	35.7	28.3	42.5



#28
 Naphthalene
 Concen: 832.66 ng/ul
 RT: 11.03 min Scan# 1395
 Delta R.T. -0.00 min
 Lab File: BG022745.D
 Acq: 1 Jul 2016 3:00

Tgt Ion	Ratio	Lower	Upper
128	100		
129	11.0	9.4	14.2
127	14.4	11.2	16.8









#30
4-Chloroaniline
Concen: 830.25 ng/ul
RT: 11.14 min Scan# 1413
Delta R.T. -0.00 min
Lab File: BG022745.D
Acq: 1 Jul 2016 3:00




Instrument :
BNA_G
ClientSampleId :
SSTD02008


Tgt Ion:127 Resp: 143650
Ion Ratio Lower Upper
127 100
129 31.1 26.9 40.3



#32
Caprolactam
Concen: 1.91 ng/ul
RT: 11.48 min Scan# 1470
Delta R.T. -0.42 min
Lab File: BG022745.D
Acq: 1 Jul 2016 3:00

Tgt Ion:113 Resp: 128
Ion Ratio Lower Upper
113 100
55 0.0 99.6 149.4#
56 0.0 66.4 99.6#









#33
 4-Chloro-3-methylphenol
 Concen: 1.12 ng/ul
 RT: 11.67 min Scan# 1503
 Delta R.T. -0.58 min
 Lab File: BG022745.D
 Acq: 1 Jul 2016 3:00




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02008


Tgt Ion:107 Resp: 286
 Ion Ratio Lower Upper
 107 100
 144 0.0 22.7 34.1#
 142 59.9 71.6 107.4#



#34
 2-Methylnaphthalene
 Concen: 779.76 ng/ul
 RT: 12.63 min Scan# 1667
 Delta R.T. 0.00 min
 Lab File: BG022745.D
 Acq: 1 Jul 2016 3:00

Tgt Ion:142 Resp: 328146
 Ion Ratio Lower Upper
 142 100
 141 94.4 75.2 112.8









#36
 1,2,4,5-Tetrachlorobenzene
 Concen: 85664.57 ng/ul
 RT: 12.99 min Scan# 1728
 Delta R.T. -0.00 min
 Lab File: BG022745.D
 Acq: 1 Jul 2016 3:00




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02008


Tgt Ion	Ratio	Lower	Upper
216	100		
214	81.3	61.7	92.5
179	22.4	17.0	25.4
108	11.4	14.2	21.2



#37
 Hexachlorocyclopentadiene
 Concen: 81114.70 ng/ul
 RT: 12.97 min Scan# 1724
 Delta R.T. -0.00 min
 Lab File: BG022745.D
 Acq: 1 Jul 2016 3:00

Tgt Ion	Ratio	Lower	Upper
237	100		
235	64.0	49.6	74.4
272	11.7	9.3	13.9









#39
 2,4,5-Trichlorophenol
 Concen: 94776.56 ng/ul
 RT: 13.23 min Scan# 1769
 Delta R.T. -0.08 min
 Lab File: BG022745.D
 Acq: 1 Jul 2016 3:00




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02008


Tgt Ion	Ratio	Lower	Upper
196	100		
198	93.4	79.8	119.8
200	30.4	24.0	36.0



#40
 1,1'-Biphenyl
 Concen: 153.37 ng/ul
 RT: 13.00 min Scan# 1729
 Delta R.T. -0.64 min
 Lab File: BG022745.D
 Acq: 1 Jul 2016 3:00

Tgt Ion	Ratio	Lower	Upper
154	100		
153	17.1	34.6	52.0#
76	0.0	11.9	17.9#









#41
 2-Chloronaphthalene
 Concen: 3436.37 ng/ul
 RT: 13.23 min Scan# 1768
 Delta R.T. -0.45 min
 Lab File: BG022745.D
 Acq: 1 Jul 2016 3:00

Instrument :
 BNA_G
 ClientSampleId :
 SSTD02008

Tgt Ion	Ratio	Lower	Upper
162	100		
164	9.9	25.4	38.2#
127	6.2	30.2	45.4#




Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 164.00 (163.70 to 164.70): E
 Ion 127.00 (126.70 to 127.70): E






#42
 2-Nitroaniline
 Concen: 83.98 ng/ul
 RT: 13.33 min Scan# 1785
 Delta R.T. -0.56 min
 Lab File: BG022745.D
 Acq: 1 Jul 2016 3:00

Tgt Ion	Ratio	Lower	Upper
65	100		
92	158.0	67.1	100.7#
138	0.0	109.6	164.4#



Abundance Ion 65.00 (64.70 to 65.70): BGC
 Ion 92.00 (91.70 to 92.70): BGC
 Ion 138.00 (137.70 to 138.70): E





#44

Dimethylphthalate

Concen: 8766.51 ng/ul

RT: 13.67 min Scan# 1844

Delta R.T. -0.57 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00

Instrument :

BNA_G

ClientSampled :

SSTD02008


Tgt Ion:163 Resp: 49449

Ion Ratio Lower Upper

163 100

194 0.0 3.4 5.2#




164 230.7 8.8 13.2#



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#45

2,6-Dinitrotoluene

Concen: 113.83 ng/ul

RT: 13.97 min Scan# 1894

Delta R.T. -0.40 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00


Tgt Ion:165 Resp: 139

Ion Ratio Lower Upper

165 100

89 0.0 41.5 62.3#



121 0.0 17.0 25.4#




Abundance Ion 165.00 (164.70 to 165.70): E

Ion 89.00 (88.70 to 89.70): BG

Ion 121.00 (120.70 to 121.70): E





#47

Acenaphthylene

Concen: 28.87 ng/ul

RT: 13.76 min Scan# 1859

Delta R.T. -0.77 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00

Instrument :

BNA_G

ClientSampleId :

SSTD02008


Tgt Ion:152 Resp: 184

Ion Ratio Lower Upper

152 100

151 0.0 17.5 26.3#




153 0.0 11.1 16.7#



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



#48

3-Nitroaniline

Concen: 35425.35 ng/ul

RT: 14.20 min Scan# 1933

Delta R.T. -0.51 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00


Tgt Ion:138 Resp: 33804

Ion Ratio Lower Upper

138 100

108 0.0 6.2 9.2#



92 207.0 79.3 118.9#




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BGC





#49

Acenaphthene

Concen: 53.57 ng/ul

RT: 14.37 min Scan# 1962

Delta R.T. -0.50 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00

Instrument :

BNA_G

ClientSampleId :

SSTD02008





Tgt Ion:153 Resp: 232

Ion Ratio Lower Upper

153 100

152 0.0 37.3 55.9#

154 0.0 68.2 102.2#



#50

2,4-Dinitrophenol

Concen: 85.47 ng/ul

RT: 14.34 min Scan# 1958

Delta R.T. -0.57 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00




Tgt Ion:184 Resp: 58


Ion Ratio Lower Upper

184 100

63 2222.6 40.5 60.7#

154 0.0 48.5 72.7#





#52

4-Nitrophenol

Concen: 503.16 ng/ul

RT: 14.48 min Scan# 1982

Delta R.T. -0.53 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00

Instrument :





BNA_G

ClientSampleId :

SSTD02008

Tgt Ion:109 Resp: 555

Ion	Ratio	Lower	Upper
109	100		
139	0.0	136.1	204.1#
65	124.5	106.0	159.0



#53

Dibenzofuran

Concen: 16.29 ng/ul

RT: 14.31 min Scan# 1953




Delta R.T. -0.88 min


Lab File: BG022745.D

Acq: 1 Jul 2016 3:00

Tgt Ion:168 Resp: 109

Ion	Ratio	Lower	Upper
168	100		
139	38.0	27.4	41.0





#54

2,4-Dinitrotoluene

Concen: 105.33 ng/ul

RT: 14.62 min Scan# 2006

Delta R.T. -0.53 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00

Instrument :

BNA_G

ClientSampleId :

SSTD02008


Tgt Ion:165 Resp: 186

Ion Ratio Lower Upper

165 100

63 0.0 25.7 38.5#




182 0.0 2.1 3.1#



Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BG

Ion 182.00 (181.70 to 182.70): E



#55

2,3,4,6-Tetrachlorophenol

Concen: 62.30 ng/ul

RT: 15.20 min Scan# 2104

Delta R.T. -0.23 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00

Tgt Ion:232 Resp: 117


Ion Ratio Lower Upper

232 100

131 0.0 34.8 52.2#

130 0.0 2.2 3.2#

166 891.0 23.8 35.6#






Abundance Ion 232.00 (231.70 to 232.70): E

Ion 131.00 (130.70 to 131.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 166.00 (165.70 to 166.70): E





#56

Diethylphthalate

Concen: 33.85 ng/ul

RT: 15.05 min Scan# 2078

Delta R.T. -0.56 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00

Instrument :

BNA_G

ClientSampleId :

SSTD02008


Tgt Ion:149 Resp: 194

Ion Ratio Lower Upper

149 100

177 0.0 17.4 26.2#




150 0.0 9.6 14.4#



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 177.00 (176.70 to 177.70): E

Ion 150.00 (149.70 to 150.70): E



#58

Fluorene

Concen: 26.60 ng/ul

RT: 15.38 min Scan# 2134

Delta R.T. -0.47 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00


Tgt Ion:166 Resp: 143

Ion Ratio Lower Upper

166 100

165 0.0 80.3 120.5#



167 0.0 10.3 15.5#




Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E









#59
 4-Chlorophenyl-phenylether
 Concen: 112.11 ng/ul
 RT: 15.41 min Scan# 2139
 Delta R.T. -0.43 min
 Lab File: BG022745.D
 Acq: 1 Jul 2016 3:00




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02008


Tgt Ion:204 Resp: 361
 Ion Ratio Lower Upper
 204 100
 206 0.0 27.4 41.2#
 141 99.4 47.1 70.7#



#60
 4-Nitroaniline
 Concen: 202.40 ng/ul
 RT: 15.22 min Scan# 2108
 Delta R.T. -0.64 min
 Lab File: BG022745.D
 Acq: 1 Jul 2016 3:00

Tgt Ion:138 Resp: 232
 Ion Ratio Lower Upper
 138 100
 92 0.0 39.0 58.4#
 108 0.0 49.2 73.8#





#63

4,6-Dinitro-2-methylphenol

Concen: 3167.53 ng/ul

RT: 15.42 min Scan# 2142

Delta R.T. -0.50 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00

Instrument :

BNA_G

ClientSampleId :

SSTD02008


Tgt Ion:198 Resp: 11227

Ion Ratio Lower Upper

198 100

182 0.0 4.2 6.2#




77 14.8 16.7 25.1#



Abundance Ion 198.00 (197.70 to 198.70): E

Ion 182.00 (181.70 to 182.70): E

Ion 77.00 (76.70 to 77.70): E



#64

N-Nitrosodiphenylamine

Concen: 1187.14 ng/ul

RT: 15.42 min Scan# 2142

Delta R.T. -0.63 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00


Tgt Ion:169 Resp: 18178

Ion Ratio Lower Upper

169 100

168 271.7 59.0 88.6#



167 91.3 30.6 46.0#




Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E





#65

4-Bromophenyl-phenylether

Concen: 191.78 ng/ul

RT: 16.45 min Scan# 2316

Delta R.T. -0.29 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00





Instrument :

BNA_G

ClientSampleId :

SSTD02008

Tgt Ion:	248	Resp:	1334
Ion Ratio		Lower	Upper
248	100		
250	66.7	76.6	114.8#
141	36.9	61.8	92.6#



#66

Hexachlorobenzene

Concen: 13.85 ng/ul




RT: 16.53 min Scan# 2330


Delta R.T. -0.33 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00

Tgt Ion:	284	Resp:	102
Ion Ratio		Lower	Upper
284	100		
142	0.0	27.8	41.8#
249	0.0	24.4	36.6#









#67
 Atrazine
 Concen: 53.25 ng/ul
 RT: 16.39 min Scan# 2307
 Delta R.T. -0.61 min
 Lab File: BG022745.D
 Acq: 1 Jul 2016 3:00

Instrument :
 BNA_G
 ClientSampleId :
 SSTD02008

Tgt Ion:200 Resp: 340
 Ion Ratio Lower Upper
 200 100
 173 0.0 21.0 31.6#
 215 0.0 37.9 56.9#




Abundance Ion 200.00 (199.70 to 200.70): E
 Ion 173.00 (172.70 to 173.70): E
 Ion 215.00 (214.70 to 215.70): E






#68
 Pentachlorophenol
 Concen: 28.58 ng/ul
 RT: 16.79 min Scan# 2374
 Delta R.T. -0.41 min
 Lab File: BG022745.D
 Acq: 1 Jul 2016 3:00

Tgt Ion:266 Resp: 121
 Ion Ratio Lower Upper
 266 100
 264 0.0 54.7 82.1#
 268 110.0 57.8 86.8#



Abundance Ion 265.70 (265.40 to 266.40): E
 Ion 264.00 (263.70 to 264.70): E
 Ion 268.00 (267.70 to 268.70): E





#69

Phenanthrene

Concen: 7.18 ng/ul

RT: 16.98 min Scan# 2407

Delta R.T. -0.62 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00

Instrument :

BNA_G

ClientSampleId :

SSTD02008


Tgt Ion:178 Resp: 191

Ion Ratio Lower Upper

178 100

179 0.0 12.9 19.3#




176 73.7 15.8 23.8#



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E



#71

Anthracene

Concen: 4.10 ng/ul

RT: 17.09 min Scan# 2426

Delta R.T. -0.60 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00


Tgt Ion:178 Resp: 112

Ion Ratio Lower Upper

178 100

179 0.0 12.3 18.5#



176 0.0 16.3 24.5#




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E





#72

Carbazole

Concen: 6.43 ng/ul

RT: 17.34 min Scan# 2468

Delta R.T. -0.63 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00

Instrument :

BNA_G

ClientSampleId :

SSTD02008


Tgt Ion:167 Resp: 130

Ion Ratio Lower Upper

167 100

166 0.0 19.2 28.8#

139 52.4 10.2 15.4#




Abundance Ion 167.00 (166.70 to 167.70): E


Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E

Time-->



Time-->



#73

Di-n-butylphthalate

Concen: 3.04 ng/ul

RT: 17.86 min Scan# 2556

Delta R.T. -0.65 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00


Tgt Ion:149 Resp: 79

Ion Ratio Lower Upper

149 100

150 146.2 7.1 10.7#

104 0.0 3.3 4.9#




Abundance Ion 149.00 (148.70 to 149.70): E


Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E

Time-->



Time-->



#74

Fluoranthene

Concen: 3.50 ng/ul

RT: 18.87 min Scan# 2729

Delta R.T. -0.73 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00

Instrument :

BNA_G

ClientSampleId :

SSTD02008


Tgt Ion:202 Resp: 97

Ion Ratio Lower Upper

202 100

101 0.0 10.6 16.0#




100 0.0 8.7 13.1#



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): B



#77

Pyrene

Concen: 158131.09 ng/ul

RT: 19.60 min Scan# 2853

Delta R.T. -0.37 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00


Tgt Ion:202 Resp: 1071092

Ion Ratio Lower Upper

202 100

101 8.2 12.5 18.7#



100 5.4 10.2 15.4#




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): B





#78

Butylbenzylphthalate

Concen: 208.61 ng/ul

RT: 20.14 min Scan# 2944

Delta R.T. -0.70 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00


Instrument :

BNA_G

ClientSampleId :

SSTD02008




Tgt Ion:	149	Resp:	548
Ion Ratio	Lower	Upper	
149	100		
91	0.0	54.2	81.4#
206	33.8	14.1	21.1#



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 91.00 (90.70 to 91.70): BG

Ion 206.00 (205.70 to 206.70): E



#79

3,3'-Dichlorobenzidine

Concen: 92.08 ng/ul


RT: 20.99 min Scan# 3089

Delta R.T. -0.76 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00



Tgt Ion:	252	Resp:	211
Ion Ratio	Lower	Upper	
252	100		
254	0.0	51.4	77.2#
126	0.0	10.7	16.1#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 254.00 (253.70 to 254.70): E

Ion 126.00 (125.70 to 126.70): E





#80

Benzo(a)anthracene

Concen: 39.03 ng/ul

RT: 21.13 min Scan# 3113

Delta R.T. -0.70 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00





Instrument :

BNA_G

ClientSampleId :

SSTD02008

Tgt Ion:	228	Resp:	277
Ion Ratio	Lower	Upper	
228	100		
229	0.0	15.4	23.2#
226	0.0	22.8	34.2#



#81

Bis(2-ethylhexyl)phthalate

Concen: 199.71 ng/ul




RT: 20.93 min Scan# 3079


Delta R.T. -0.79 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00

Tgt Ion:	149	Resp:	677
Ion Ratio	Lower	Upper	
149	100		
167	0.0	21.9	32.9#
279	0.0	2.9	4.3#





#82

Chrysene

Concen: 42.88 ng/ul

RT: 21.13 min Scan# 3113

Delta R.T. -0.77 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00

Instrument :

BNA_G

ClientSampleId :

SSTD02008





Tgt Ion:228 Resp: 277

Ion Ratio Lower Upper

228 100

226 0.0 24.2 36.4#

229 0.0 15.0 22.6#



#84

Di-n-octyl phthalate

Concen: 93.76 ng/ul




RT: 22.19 min Scan# 3293


Delta R.T. -0.80 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00

Tgt Ion:149 Resp: 719





#85

Benzo(b)fluoranthene

Concen: 21.15 ng/ul

RT: 23.29 min Scan# 3481

Delta R.T. -0.85 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00





Instrument :

BNA_G

ClientSampleId :

SSTD02008

Tgt Ion:	252	Resp:	230
Ion Ratio	Lower	Upper	
252	100		
253	34.1	18.2	27.2#
125	0.0	10.8	16.2#



#86

Benzo(k)fluoranthene

Concen: 71.37 ng/ul




RT: 23.34 min Scan# 3490


Delta R.T. -0.87 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00

Tgt Ion:	252	Resp:	729
Ion Ratio	Lower	Upper	
252	100		
253	106.8	15.8	23.6#
125	147.4	10.6	16.0#





#88

Benzo(a)pyrene

Concen: 114618.83 ng/u1

RT: 24.14 min Scan# 3626

Delta R.T. -0.92 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00

Instrument :

BNA_G

ClientSampleId :

SSTD02008


Tgt Ion:252 Resp: 1184085

Ion Ratio Lower Upper




252 100

253 21.3 19.0 28.4

125 4.9 12.0 18.0#



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E



#89

Indeno(1,2,3-cd)pyrene

Concen: 7.51 ng/u1

RT: 28.02 min Scan# 4286

Delta R.T. -1.05 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00


Tgt Ion:276 Resp: 82

Ion Ratio Lower Upper



276 100


138 53.6 20.8 31.2#

277 40.9 17.6 26.4#



Abundance Ion 276.00 (275.70 to 276.70): E
Ion 138.00 (137.70 to 138.70): E
Ion 277.00 (276.70 to 277.70): E





#90

Dibenzo(a,h)anthracene

Concen: 8.08 ng/ul

RT: 28.11 min Scan# 4302

Delta R.T. -1.03 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00

Instrument :

BNA_G

ClientSampleId :

SSTD02008


Tgt Ion:278 Resp: 73

Ion Ratio Lower Upper

278 100

139 0.0 16.3 24.5#




279 173.9 19.1 28.7#



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E



#91

Benzo(g,h,i)perylene

Concen: 288.46 ng/ul

RT: 29.28 min Scan# 4501

Delta R.T. -1.00 min

Lab File: BG022745.D

Acq: 1 Jul 2016 3:00


Tgt Ion:276 Resp: 2470

Ion Ratio Lower Upper

276 100

138 17.0 22.1 33.1#

277 13.7 17.0 25.4#



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E

