

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG070116\
 Data File : BG022756.D
 Acq On : 1 Jul 2016 10:30
 Operator : UM/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD02010

Manual Integrations

sohil
 7/4/2016 1:30:04 PM

Quant Time: Jul 02 00:06:13 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG063016.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jul 01 08:12:10 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.16	152	103169	20.00	ng/ul	0.00
18) Naphthalene-d8	10.99	136	462112	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.80	164	354548	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.55	188	934123	20.00	ng/ul	0.00
75) Chrysene-d12	21.85	240	1054024	20.00	ng/ul	0.00
83) Perylene-d12	25.22	264	1088844	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.54	96	16461	8.92	ng/uL	0.00
5) Phenol-d5	7.31	99	195896	18.05	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.47	67	118960	19.44	ng/ul	0.00
9) 2-Chlorophenol-d4	7.69	132	132063	18.22	ng/ul	0.00
13) 4-Methylphenol-d8	8.87	113	142628	16.92	ng/ul	0.00
19) Nitrobenzene-d5	9.33	128	66367	18.01	ng/ul	0.00
22) 2-Nitrophenol-d4	10.06	143	77255	17.38	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.60	165	156758	16.87	ng/ul	0.00
29) 4-Chloroaniline-d4	11.12	131	149044	18.94	ng/ul	0.00
43) Dimethylphthalate-d6	14.20	166	545331	18.93	ng/ul	0.00
46) Acenaphthylene-d8	14.49	160	631763	19.14	ng/ul	0.00
51) 4-Nitrophenol-d4	14.99	143	84211	18.62	ng/ul	0.00
57) Fluorene-d10	15.79	176	527249	19.32	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.90	200	93112	15.50	ng/ul	0.00
70) Anthracene-d10	17.65	188	794319	18.29	ng/ul	0.00
76) Pyrene-d10	19.93	212	951265	18.89	ng/ul	0.00
87) Benzo(a)pyrene-d12	24.98	264	926202	17.90	ng/ul	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.58	88	16473	7.61	ng/uL#	64
4) Benzaldehyde	7.29	77	136410	19.77	ng/ul#	83
6) Phenol	7.34	94	198494	17.86	ng/ul#	66
8) Bis(2-Chloroethyl)ether	7.57	93	146468	19.27	ng/ul	93
10) 2-Chlorophenol	7.72	128	126395	17.14	ng/ul#	82
11) 2-Methylphenol	8.60	108	149295	17.88	ng/ul	89
12) 2,2'-oxybis(1-Chloropropan	8.69	45	160656	18.80	ng/ul#	91
14) Acetophenone	8.98	105	245587	18.04	ng/ul#	68
15) N-Nitroso-di-n-propylamine	8.96	70	141893	17.34	ng/ul#	85
16) 4-Methylphenol	8.93	108	164291	17.70	ng/ul	94
17) Hexachloroethane	9.25	117	60736	19.34	ng/ul#	68
20) Nitrobenzene	9.38	77	210060	18.54	ng/ul#	81
21) Isophorone	9.89	82	363974	18.01	ng/ul#	89
23) 2-Nitrophenol	10.09	139	83888	17.78	ng/ul#	42
24) 2,4-Dimethylphenol	10.14	107	203204	18.00	ng/ul	83
25) Bis(2-Chloroethoxy)methane	10.38	93	212727	17.84	ng/ul	94
27) 2,4-Dichlorophenol	10.63	162	165185	17.68	ng/ul#	93
28) Naphthalene	11.03	128	432254	18.29	ng/ul	99
30) 4-Chloroaniline	11.14	127	146635	18.24	ng/ul	99
31) Hexachlorobutadiene	11.32	225	132939	18.47	ng/ul	100
32) Caprolactam	11.90	113	55134m	17.69	ng/ul	
33) 4-Chloro-3-methylphenol	12.26	107	200943	17.00	ng/ul	86
34) 2-Methylnaphthalene	12.63	142	350275	17.92	ng/ul	91

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG070116\
 Data File : BG022756.D
 Acq On : 1 Jul 2016 10:30
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD02010


Manual Integrations

sohil
 7/4/2016 1:30:04 PM

Quant Time: Jul 02 00:06:13 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG063016.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Jul 01 08:12:10 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
36) 1,2,4,5-Tetrachlorobenzene	13.00	216	252353	19.04	ng/ul#	95
37) Hexachlorocyclopentadiene	12.97	237	141733	17.71	ng/ul	97
38) 2,4,6-Trichlorophenol	13.23	196	162151	17.74	ng/ul	96
39) 2,4,5-Trichlorophenol	13.31	196	173540	18.04	ng/ul	95
40) 1,1'-Biphenyl	13.63	154	496148	18.40	ng/ul	96
41) 2-Chloronaphthalene	13.68	162	401265	18.38	ng/ul	94
42) 2-Nitroaniline	13.88	65	154974	18.35	ng/ul#	61
44) Dimethylphthalate	14.24	163	548531	18.65	ng/ul	100
45) 2,6-Dinitrotoluene	14.37	165	119741	18.81	ng/ul#	86
47) Acenaphthylene	14.52	152	630818	18.98	ng/ul	97
48) 3-Nitroaniline	14.71	138	98683	19.83	ng/ul#	61
49) Acenaphthene	14.86	153	414402	18.35	ng/ul	92
50) 2,4-Dinitrophenol	14.90	184	49022	13.85	ng/ul#	69
52) 4-Nitrophenol	15.01	109	106059	18.44	ng/ul#	59
53) Dibenzofuran	15.19	168	668633	19.17	ng/ul	91
54) 2,4-Dinitrotoluene	15.15	165	173354	18.83	ng/ul#	72
55) 2,3,4,6-Tetrachlorophenol	15.42	232	178656	18.24	ng/ul#	93
56) Diethylphthalate	15.60	149	572316	19.15	ng/ul	97
58) Fluorene	15.85	166	532397	18.99	ng/ul	95
59) 4-Chlorophenyl-phenylether	15.83	204	316578	18.86	ng/ul	94
60) 4-Nitroaniline	15.86	138	117577	19.67	ng/ul#	35
63) 4,6-Dinitro-2-methylphenol	15.92	198	101782	15.99	ng/ul#	92
64) N-Nitrosodiphenylamine	16.05	169	488270	17.75	ng/ul	95
65) 4-Bromophenyl-phenylether	16.73	248	224431	17.96	ng/ul	97
66) Hexachlorobenzene	16.85	284	243190	18.38	ng/ul#	89
67) Atrazine	16.99	200	216352	18.86	ng/ul	91
68) Pentachlorophenol	17.20	266	124574	16.38	ng/ul	87
69) Phenanthrene	17.59	178	895143	18.73	ng/ul	99
71) Anthracene	17.69	178	925999	18.86	ng/ul	98
72) Carbazole	17.96	167	758072	20.86	ng/ul	98
73) Di-n-butylphthalate	18.50	149	921104	19.75	ng/ul#	96
74) Fluoranthene	19.59	202	1100906	22.12	ng/ul#	86
77) Pyrene	19.96	202	1112880	18.86	ng/ul#	84
78) Butylbenzylphthalate	20.83	149	424316	18.54	ng/ul#	83
79) 3,3'-Dichlorobenzidine	21.74	252	403231	20.20	ng/ul#	96
80) Benzo(a)anthracene	21.83	228	1179351	19.08	ng/ul	97
81) Bis(2-ethylhexyl)phthalate	21.72	149	594946	20.15	ng/ul#	99
82) Chrysene	21.90	228	1082074	19.23	ng/ul	98
84) Di-n-octyl phthalate	22.97	149	1007745	21.00	ng/ul	100
85) Benzo(b)fluoranthene	24.14	252	1240388m	18.23	ng/ul	
86) Benzo(k)fluoranthene	24.21	252	1184610m	18.53	ng/ul	
88) Benzo(a)pyrene	25.05	252	1167251	18.06	ng/ul#	90
89) Indeno(1,2,3-cd)pyrene	29.06	276	1308677	19.15	ng/ul#	81
90) Dibenzo(a,h)anthracene	29.12	278	1058810	18.74	ng/ul#	87
91) Benzo(g,h,i)perylene	30.27	276	1079189	20.14	ng/ul#	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2

1,4-Dioxane

Concen: 7.61 ng/uL

RT: 3.58 min Scan# 126

Delta R.T. -0.00 min

Lab File: BG022756.D

Acq: 1 Jul 2016 10:30

Instrument :

BNA_G

ClientSampleId :





SSTD02010

Tgt Ion:	88	Resp:	16473
Ion Ratio	Lower	Upper	
88	100		
43	40.7	16.6	24.8#
58	84.4	47.2	70.8#

Manual Integrations

sohil

7/4/2016 1:30:04 PM



#4

Benzaldehyde

Concen: 19.77 ng/uL




RT: 7.29 min Scan# 758


Delta R.T. -0.00 min

Lab File: BG022756.D

Acq: 1 Jul 2016 10:30

Tgt Ion:	77	Resp:	136410
Ion Ratio	Lower	Upper	
77	100		
105	83.8	83.8	125.8#
106	83.4	77.9	116.9





#6

Phenol

Concen: 17.86 ng/ul

RT: 7.34 min Scan# 766

Delta R.T. -0.00 min

Lab File: BG022756.D

Acq: 1 Jul 2016 10:30

Instrument :

BNA_G

ClientSampleId :

SSTD02010

Tgt Ion: 94 Resp: 198494

Ion Ratio Lower Upper

94 100


65 31.6 16.8 25.2#

66 50.9 22.6 33.8#

Manual Integrations

sohil




7/4/2016 1:30:04 PM



Abundance Ion 94.00 (93.70 to 94.70): BGC

Ion 65.00 (64.70 to 65.70): BGC

Ion 66.00 (65.70 to 66.70): BGC



#8

Bis(2-Chloroethyl)ether

Concen: 19.27 ng/ul

RT: 7.57 min Scan# 806

Delta R.T. 0.01 min

Lab File: BG022756.D

Acq: 1 Jul 2016 10:30


Tgt Ion: 93 Resp: 146468

Ion Ratio Lower Upper

93 100

63 74.1 56.0 84.0



95 28.2 27.7 41.5




Abundance Ion 93.00 (92.70 to 93.70): BGC

Ion 63.00 (62.70 to 63.70): BGC

Ion 95.00 (94.70 to 95.70): BGC









#10
2-Chlorophenol
Concen: 17.14 ng/ul
RT: 7.72 min Scan# 831
Delta R.T. -0.00 min
Lab File: BG022756.D
Acq: 1 Jul 2016 10:30

Instrument :
BNA_G
ClientSampleId :
SSTD02010

Tgt Ion	Ratio	Resp	Lower	Upper
128	100	126395		
64	59.1	34.2	51.4#	
130	39.0	27.0	40.4	




Manual Integrations


sohil
7/4/2016 1:30:04 PM



#11
2-Methylphenol
Concen: 17.88 ng/ul
RT: 8.60 min Scan# 981
Delta R.T. 0.01 min
Lab File: BG022756.D
Acq: 1 Jul 2016 10:30

Tgt Ion	Ratio	Resp	Lower	Upper
108	100	149295		
107	87.0	78.6	117.8	









#12
2,2'-oxybis(1-chloropropane)
Concen: 18.80 ng/ul
RT: 8.69 min Scan# 996
Delta R.T. -0.00 min
Lab File: BG022756.D
Acq: 1 Jul 2016 10:30

Instrument :
BNA_G
ClientSampleId :
SSTD02010

Tgt Ion	Ratio	Lower	Upper
45	100		
77	21.4	13.4	20.2#
79	15.5	10.2	15.4#




Manual Integrations


sohil
7/4/2016 1:30:04 PM



#14
Acetophenone
Concen: 18.04 ng/ul
RT: 8.98 min Scan# 1046
Delta R.T. -0.00 min
Lab File: BG022756.D
Acq: 1 Jul 2016 10:30

Tgt Ion	Ratio	Lower	Upper
105	100		
77	98.4	56.1	84.1#
51	29.2	14.6	22.0#









#15
N-Nitroso-di-n-propylamine
Concen: 17.34 ng/ul
RT: 8.96 min Scan# 1042
Delta R.T. -0.00 min
Lab File: BG022756.D
Acq: 1 Jul 2016 10:30

Instrument :
BNA_G
ClientSampleId :
SSTD02010

Tgt Ion	Ratio	Lower	Upper
70	100		
42	48.3	31.4	47.2#
101	7.3	8.6	13.0#
130	15.1	19.0	28.4#




Manual Integrations


sohil
7/4/2016 1:30:04 PM



#16
4-Methylphenol
Concen: 17.70 ng/ul
RT: 8.93 min Scan# 1037
Delta R.T. 0.01 min
Lab File: BG022756.D
Acq: 1 Jul 2016 10:30

Tgt Ion	Ratio	Lower	Upper
108	100		
107	103.8	88.3	132.5





#17

Hexachloroethane

Concen: 19.34 ng/ul

RT: 9.25 min Scan# 1092

Delta R.T. 0.01 min

Lab File: BG022756.D


Acq: 1 Jul 2016 10:30

Instrument :

BNA_G

ClientSampleId :

SSTD02010






Tgt Ion	Ratio	Lower	Upper
117	100		
201	126.0	76.8	115.2#
199	79.3	43.5	65.3#

Manual Integrations

sohil

7/4/2016 1:30:04 PM



#20

Nitrobenzene

Concen: 18.54 ng/ul




RT: 9.38 min Scan# 1113


Delta R.T. 0.01 min

Lab File: BG022756.D

Acq: 1 Jul 2016 10:30

Tgt Ion	Ratio	Lower	Upper
77	100		
123	34.6	39.5	59.3#
65	14.2	9.0	13.4#





#21

Isophorone

Concen: 18.01 ng/ul

RT: 9.89 min Scan# 1201

Delta R.T. -0.00 min

Lab File: BG022756.D


Acq: 1 Jul 2016 10:30

Instrument :

BNA_G

ClientSampleId :

SSTD02010



Tgt Ion: 82 Resp: 363974

Ion Ratio Lower Upper

82 100




95 8.2 5.1 7.7#

138 13.0 15.0 22.6#

Manual Integrations

sohil

7/4/2016 1:30:04 PM



#23

2-Nitrophenol

Concen: 17.78 ng/ul

RT: 10.09 min Scan# 1234

Delta R.T. -0.00 min

Lab File: BG022756.D

Acq: 1 Jul 2016 10:30




Tgt Ion: 139 Resp: 83888


Ion Ratio Lower Upper

139 100

65 75.3 31.8 47.8#

109 48.5 17.1 25.7#









#24
 2,4-Dimethylphenol
 Concen: 18.00 ng/ul
 RT: 10.14 min Scan# 1243
 Delta R.T. -0.00 min
 Lab File: BG022756.D
 Acq: 1 Jul 2016 10:30

Instrument :
 BNA_G
 ClientSampleId :
 SST02010

Tgt Ion	Ratio	Lower	Upper
107	100		
121	44.8	43.7	65.5
122	70.8	70.3	105.5




Manual Integrations


sohil
 7/4/2016 1:30:04 PM



#25
 Bis(2-Chloroethoxy)methane
 Concen: 17.84 ng/ul
 RT: 10.38 min Scan# 1283
 Delta R.T. -0.00 min
 Lab File: BG022756.D
 Acq: 1 Jul 2016 10:30

Tgt Ion	Ratio	Lower	Upper
93	100		
95	29.1	26.0	39.0
123	10.1	10.1	15.1









#27
 2,4-Dichlorophenol
 Concen: 17.68 ng/ul
 RT: 10.63 min Scan# 1326
 Delta R.T. -0.00 min
 Lab File: BG022756.D
 Acq: 1 Jul 2016 10:30

Instrument :
 BNA_G
 ClientSampleId :
 SST02010

Tgt Ion	Ratio	Resp	Lower	Upper
162	100	165185		
164	66.7	51.6	77.4	
98	43.4	28.3	42.5	




Manual Integrations


sohil
 7/4/2016 1:30:04 PM



#28
 Naphthalene
 Concen: 18.29 ng/ul
 RT: 11.03 min Scan# 1395
 Delta R.T. -0.00 min
 Lab File: BG022756.D
 Acq: 1 Jul 2016 10:30

Tgt Ion	Ratio	Resp	Lower	Upper
128	100	432254		
129	11.3	9.4	14.2	
127	14.4	11.2	16.8	









#30
4-Chloroaniline
Concen: 18.24 ng/ul
RT: 11.14 min Scan# 1413
Delta R.T. -0.00 min
Lab File: BG022756.D
Acq: 1 Jul 2016 10:30

Instrument :
BNA_G
ClientSampleId :
SSTD02010

Tgt Ion:127 Resp: 146635
Ion Ratio Lower Upper
127 100
129 33.9 26.9 40.3




Manual Integrations


sohil
7/4/2016 1:30:04 PM



#31
Hexachlorobutadiene
Concen: 18.47 ng/ul
RT: 11.32 min Scan# 1443
Delta R.T. 0.01 min
Lab File: BG022756.D
Acq: 1 Jul 2016 10:30

Tgt Ion:225 Resp: 132939
Ion Ratio Lower Upper
225 100
223 68.9 54.7 82.1
227 65.1 52.2 78.4





#32

Caprolactam

Concen: 17.69 ng/ul m

RT: 11.90 min Scan# 1542

Delta R.T. 0.01 min

Lab File: BG022756.D


Acq: 1 Jul 2016 10:30

Instrument :

BNA_G

ClientSampleId :

SSTD02010



Tgt Ion:113 Resp: 55134

Ion Ratio Lower Upper

113 100




55 164.3 99.6 149.4#

56 130.5 66.4 99.6#

Manual Integrations

sohil

7/4/2016 1:30:04 PM



#33

4-Chloro-3-methylphenol

Concen: 17.00 ng/ul

RT: 12.26 min Scan# 1603

Delta R.T. -0.00 min

Lab File: BG022756.D

Acq: 1 Jul 2016 10:30




Tgt Ion:107 Resp: 200943


Ion Ratio Lower Upper

107 100

144 23.5 22.7 34.1

142 74.4 71.6 107.4









#34
 2-Methylnaphthalene
 Concen: 17.92 ng/ul
 RT: 12.63 min Scan# 1667
 Delta R.T. -0.00 min
 Lab File: BG022756.D
 Acq: 1 Jul 2016 10:30

Instrument :
 BNA_G
 ClientSampleId :
 SST02010

Tgt Ion	Ratio	Lower	Upper
142	100		
141	84.9	75.2	112.8




Manual Integrations


sohil
 7/4/2016 1:30:04 PM



#36
 1,2,4,5-Tetrachlorobenzene
 Concen: 19.04 ng/ul
 RT: 13.00 min Scan# 1729
 Delta R.T. 0.01 min
 Lab File: BG022756.D
 Acq: 1 Jul 2016 10:30

Tgt Ion	Ratio	Lower	Upper
216	100		
214	79.7	61.7	92.5
179	17.7	17.0	25.4
108	12.9	14.2	21.2#





#37

Hexachlorocyclopentadiene

Concen: 17.71 ng/ul

RT: 12.97 min Scan# 1725

Delta R.T. 0.01 min

Lab File: BG022756.D


Acq: 1 Jul 2016 10:30

Instrument :

BNA_G

ClientSampleId :

SSTD02010






Tgt Ion	Ratio	Lower	Upper
237	100		
235	64.4	49.6	74.4
272	10.7	9.3	13.9

Manual Integrations

sohil

7/4/2016 1:30:04 PM



#38

2,4,6-Trichlorophenol

Concen: 17.74 ng/ul




RT: 13.23 min Scan# 1769


Delta R.T. -0.00 min

Lab File: BG022756.D

Acq: 1 Jul 2016 10:30

Tgt Ion	Ratio	Lower	Upper
196	100		
198	101.7	77.4	116.0
200	33.7	26.6	39.8









#39
 2,4,5-Trichlorophenol
 Concen: 18.04 ng/ul
 RT: 13.31 min Scan# 1782
 Delta R.T. -0.00 min
 Lab File: BG022756.D
 Acq: 1 Jul 2016 10:30

Instrument :
 BNA_G
 ClientSampleId :
 SSTD02010

Tgt Ion	Ratio	Resp	Lower	Upper
196	100	173540		
198	94.8	79.8	119.8	
200	32.3	24.0	36.0	




Manual Integrations


sohil
 7/4/2016 1:30:04 PM



#40
 1,1'-Biphenyl
 Concen: 18.40 ng/ul
 RT: 13.63 min Scan# 1837
 Delta R.T. -0.00 min
 Lab File: BG022756.D
 Acq: 1 Jul 2016 10:30

Tgt Ion	Ratio	Resp	Lower	Upper
154	100	496148		
153	41.3	34.6	52.0	
76	12.9	11.9	17.9	









#41
 2-Chloronaphthalene
 Concen: 18.38 ng/ul
 RT: 13.68 min Scan# 1845
 Delta R.T. -0.00 min
 Lab File: BG022756.D
 Acq: 1 Jul 2016 10:30

Instrument :
 BNA_G
 ClientSampleId :
 SST02010

Tgt Ion	Ratio	Lower	Upper
162	100		
164	34.7	25.4	38.2
127	41.4	30.2	45.4




Manual Integrations


sohil
 7/4/2016 1:30:04 PM



#42
 2-Nitroaniline
 Concen: 18.35 ng/ul
 RT: 13.88 min Scan# 1880
 Delta R.T. 0.01 min
 Lab File: BG022756.D
 Acq: 1 Jul 2016 10:30

Tgt Ion	Ratio	Lower	Upper
65	100		
92	66.8	67.1	100.7#
138	75.4	109.6	164.4#









#44
Dimethylphthalate
Concen: 18.65 ng/ul
RT: 14.24 min Scan# 1941
Delta R.T. -0.00 min
Lab File: BG022756.D
Acq: 1 Jul 2016 10:30

Instrument :
BNA_G
ClientSampleId :
SSTD02010

Tgt Ion	Ratio	Lower	Upper
163	100		
194	4.1	3.4	5.2
164	10.8	8.8	13.2




Manual Integrations


sohil
7/4/2016 1:30:04 PM



#45
2,6-Dinitrotoluene
Concen: 18.81 ng/ul
RT: 14.37 min Scan# 1963
Delta R.T. 0.01 min
Lab File: BG022756.D
Acq: 1 Jul 2016 10:30

Tgt Ion	Ratio	Lower	Upper
165	100		
89	65.1	41.5	62.3#
121	19.5	17.0	25.4









#47
Acenaphthylene
Concen: 18.98 ng/ul
RT: 14.52 min Scan# 1989
Delta R.T. -0.00 min
Lab File: BG022756.D
Acq: 1 Jul 2016 10:30

Instrument :
BNA_G
ClientSampleId :
SSTD02010

Tgt Ion	Ratio	Lower	Upper
152	100		
151	20.1	17.5	26.3
153	14.3	11.1	16.7




Manual Integrations


sohil
7/4/2016 1:30:04 PM



#48
3-Nitroaniline
Concen: 19.83 ng/ul
RT: 14.71 min Scan# 2020
Delta R.T. 0.01 min
Lab File: BG022756.D
Acq: 1 Jul 2016 10:30

Tgt Ion	Ratio	Lower	Upper
138	100		
108	13.3	6.2	9.2#
92	140.1	79.3	118.9#





#49

Acenaphthene

Concen: 18.35 ng/ul

RT: 14.86 min Scan# 2047

Delta R.T. -0.00 min

Lab File: BG022756.D

Acq: 1 Jul 2016 10:30

Instrument :

BNA_G

ClientSampleId :





SSTD02010

Tgt Ion:	153	Resp:	414402
Ion Ratio	Lower	Upper	
153	100		
152	47.7	37.3	55.9
154	95.8	68.2	102.2

Manual Integrations

sohil

7/4/2016 1:30:04 PM



#50

2,4-Dinitrophenol

Concen: 13.85 ng/ul




RT: 14.90 min Scan# 2053


Delta R.T. -0.00 min

Lab File: BG022756.D

Acq: 1 Jul 2016 10:30

Tgt Ion:	184	Resp:	49022
Ion Ratio	Lower	Upper	
184	100		
63	72.7	40.5	60.7#
154	83.7	48.5	72.7#





#52

4-Nitrophenol

Concen: 18.44 ng/ul

RT: 15.01 min Scan# 2071

Delta R.T. -0.00 min

Lab File: BG022756.D


Acq: 1 Jul 2016 10:30

Instrument :

BNA_G

ClientSampleId :

SSTD02010






Tgt Ion	Ratio	Lower	Upper
109	100		
139	87.5	136.1	204.1#
65	112.5	106.0	159.0

Manual Integrations

sohil

7/4/2016 1:30:04 PM



#53

Dibenzofuran

Concen: 19.17 ng/ul




RT: 15.19 min Scan# 2103


Delta R.T. -0.00 min

Lab File: BG022756.D

Acq: 1 Jul 2016 10:30

Tgt Ion	Ratio	Lower	Upper
168	100		
139	39.4	27.4	41.0





#54

2,4-Dinitrotoluene

Concen: 18.83 ng/ul

RT: 15.15 min Scan# 2096

Delta R.T. -0.00 min

Lab File: BG022756.D


Acq: 1 Jul 2016 10:30

Instrument :

BNA_G

ClientSampleId :

SSTD02010






Tgt Ion:	165	Resp:	173354
Ion Ratio	Lower	Upper	
165	100		
63	49.1	25.7	38.5#
182	2.5	2.1	3.1

Manual Integrations

sohil

7/4/2016 1:30:04 PM



#55

2,3,4,6-Tetrachlorophenol

Concen: 18.24 ng/ul




RT: 15.42 min Scan# 2142


Delta R.T. -0.00 min

Lab File: BG022756.D

Acq: 1 Jul 2016 10:30

Tgt Ion:	232	Resp:	178656
Ion Ratio	Lower	Upper	
232	100		
131	36.8	34.8	52.2
130	3.2	2.2	3.2#
166	28.9	23.8	35.6





#56

Diethylphthalate

Concen: 19.15 ng/ul

RT: 15.60 min Scan# 2172

Delta R.T. -0.01 min

Lab File: BG022756.D


Acq: 1 Jul 2016 10:30

Instrument :

BNA_G

ClientSampleId :




SSTD02010



Tgt Ion:	149	Resp:	572316
Ion Ratio	Lower	Upper	
149	100		
177	23.2	17.4	26.2
150	13.5	9.6	14.4

Manual Integrations

sohil
7/4/2016 1:30:04 PM



#58

Fluorene

Concen: 18.99 ng/ul




RT: 15.85 min Scan# 2214


Delta R.T. -0.00 min

Lab File: BG022756.D

Acq: 1 Jul 2016 10:30

Tgt Ion:	166	Resp:	532397
Ion Ratio	Lower	Upper	
166	100		
165	94.9	80.3	120.5
167	12.5	10.3	15.5









#59
4-Chlorophenyl-phenylether
Concen: 18.86 ng/ul
RT: 15.83 min Scan# 2212
Delta R.T. -0.00 min
Lab File: BG022756.D
Acq: 1 Jul 2016 10:30

Instrument :
BNA_G
ClientSampleId :
SSTD02010

Tgt Ion	Ratio	Resp	Lower	Upper
204	100	316578		
206	33.2	27.4	41.2	
141	52.4	47.1	70.7	




Manual Integrations


sohil
7/4/2016 1:30:04 PM



#60
4-Nitroaniline
Concen: 19.67 ng/ul
RT: 15.86 min Scan# 2217
Delta R.T. -0.00 min
Lab File: BG022756.D
Acq: 1 Jul 2016 10:30

Tgt Ion	Ratio	Resp	Lower	Upper
138	100	117577		
92	68.1	39.0	58.4#	
108	133.0	49.2	73.8#	









#63
 4,6-Dinitro-2-methylphenol
 Concen: 15.99 ng/ul
 RT: 15.92 min Scan# 2226
 Delta R.T. -0.00 min
 Lab File: BG022756.D
 Acq: 1 Jul 2016 10:30

Instrument :
 BNA_G
 ClientSampleId :
 SST02010

Tgt Ion	Ratio	Lower	Upper
198	100		
182	3.2	4.2	6.2#
77	25.1	16.7	25.1#




Manual Integrations


sohil
 7/4/2016 1:30:04 PM



#64
 N-Nitrosodiphenylamine
 Concen: 17.75 ng/ul
 RT: 16.05 min Scan# 2248
 Delta R.T. -0.01 min
 Lab File: BG022756.D
 Acq: 1 Jul 2016 10:30

Tgt Ion	Ratio	Lower	Upper
169	100		
168	67.9	59.0	88.6
167	36.8	30.6	46.0









#65
4-Bromophenyl-phenylether
Concen: 17.96 ng/ul
RT: 16.73 min Scan# 2365
Delta R.T. -0.00 min
Lab File: BG022756.D
Acq: 1 Jul 2016 10:30

Instrument :
BNA_G
ClientSampleId :
SSTD02010

Tgt Ion	Ratio	Lower	Upper
248	100		
250	96.7	76.6	114.8
141	73.1	61.8	92.6




Manual Integrations


sohil
7/4/2016 1:30:04 PM



#66
Hexachlorobenzene
Concen: 18.38 ng/ul
RT: 16.85 min Scan# 2385
Delta R.T. -0.00 min
Lab File: BG022756.D
Acq: 1 Jul 2016 10:30

Tgt Ion	Ratio	Lower	Upper
284	100		
142	27.1	27.8	41.8#
249	34.9	24.4	36.6









#67
Atrazine
Concen: 18.86 ng/ul
RT: 16.99 min Scan# 2409
Delta R.T. -0.01 min
Lab File: BG022756.D
Acq: 1 Jul 2016 10:30

Instrument :
BNA_G
ClientSampleId :
SSTD02010

Tgt Ion	Ratio	Resp	Lower	Upper
200	100	216352		
173	21.1	21.0	31.6	
215	52.8	37.9	56.9	




Manual Integrations


sohil
7/4/2016 1:30:04 PM



#68
Pentachlorophenol
Concen: 16.38 ng/ul
RT: 17.20 min Scan# 2444
Delta R.T. -0.00 min
Lab File: BG022756.D
Acq: 1 Jul 2016 10:30

Tgt Ion	Ratio	Resp	Lower	Upper
266	100	124574		
264	59.3	54.7	82.1	
268	60.1	57.8	86.8	





#69

Phenanthrene

Concen: 18.73 ng/ul

RT: 17.59 min Scan# 2511

Delta R.T. -0.01 min

Lab File: BG022756.D


Acq: 1 Jul 2016 10:30

Instrument :

BNA_G

ClientSampleId :

SSTD02010






Tgt Ion:	178	Resp:	895143
Ion Ratio	Lower	Upper	
178	100		
179	15.9	12.9	19.3
176	19.3	15.8	23.8

Manual Integrations

sohil

7/4/2016 1:30:04 PM



#71

Anthracene

Concen: 18.86 ng/ul




RT: 17.69 min Scan# 2527


Delta R.T. -0.00 min

Lab File: BG022756.D

Acq: 1 Jul 2016 10:30

Tgt Ion:	178	Resp:	925999
Ion Ratio	Lower	Upper	
178	100		
179	17.1	12.3	18.5
176	20.7	16.3	24.5





#72

Carbazole

Concen: 20.86 ng/ul

RT: 17.96 min Scan# 2573

Delta R.T. -0.01 min

Lab File: BG022756.D


Acq: 1 Jul 2016 10:30

Instrument :

BNA_G

ClientSampleId :

SSTD02010






Tgt Ion:	167	Resp:	758072
Ion Ratio	Lower	Upper	
167	100		
166	23.9	19.2	28.8
139	14.8	10.2	15.4

Manual Integrations

sohil

7/4/2016 1:30:04 PM



#73

Di-n-butylphthalate

Concen: 19.75 ng/ul




RT: 18.50 min Scan# 2665


Delta R.T. -0.01 min

Lab File: BG022756.D

Acq: 1 Jul 2016 10:30

Tgt Ion:	149	Resp:	921104
Ion Ratio	Lower	Upper	
149	100		
150	9.4	7.1	10.7
104	7.0	3.3	4.9





#74

Fluoranthene

Concen: 22.12 ng/ul

RT: 19.59 min Scan# 2852

Delta R.T. -0.01 min

Lab File: BG022756.D


Acq: 1 Jul 2016 10:30

Instrument :

BNA_G

ClientSampleId :

SSTD02010



Tgt Ion:202 Resp: 1100906

Ion Ratio Lower Upper

202 100


101 6.5 10.6 16.0#

100 6.9 8.7 13.1#

Manual Integrations

sohil



7/4/2016 1:30:04 PM



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): B



#77

Pyrene

Concen: 18.86 ng/ul

RT: 19.96 min Scan# 2914

Delta R.T. -0.01 min

Lab File: BG022756.D

Acq: 1 Jul 2016 10:30


Tgt Ion:202 Resp: 1112880

Ion Ratio Lower Upper

202 100

101 8.0 12.5 18.7#



100 7.1 10.2 15.4#




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): B





#78

Butylbenzylphthalate

Concen: 18.54 ng/ul

RT: 20.83 min Scan# 3062

Delta R.T. -0.01 min

Lab File: BG022756.D


Acq: 1 Jul 2016 10:30

Instrument :

BNA_G

ClientSampleId :




SSTD02010



Tgt Ion	Ratio	Lower	Upper
149	100		
91	82.8	54.2	81.4#
206	22.5	14.1	21.1#

Manual Integrations

sohil
7/4/2016 1:30:04 PM



#79

3,3'-Dichlorobenzidine

Concen: 20.20 ng/ul




RT: 21.74 min Scan# 3217


Delta R.T. -0.01 min

Lab File: BG022756.D

Acq: 1 Jul 2016 10:30

Tgt Ion	Ratio	Lower	Upper
252	100		
254	65.5	51.4	77.2
126	6.3	10.7	16.1#





#80

Benzo(a)anthracene

Concen: 19.08 ng/ul

RT: 21.83 min Scan# 3232

Delta R.T. -0.01 min

Lab File: BG022756.D

Acq: 1 Jul 2016 10:30

Instrument :

BNA_G

ClientSampleId :

SSTD02010





Tgt Ion:228 Resp: 1179351

Ion	Ratio	Lower	Upper
228	100		
229	22.2	15.4	23.2
226	28.6	22.8	34.2

Manual Integrations

sohil

7/4/2016 1:30:04 PM



#81

Bis(2-ethylhexyl)phthalate

Concen: 20.15 ng/ul

RT: 21.72 min Scan# 3213




Delta R.T. -0.00 min


Lab File: BG022756.D

Acq: 1 Jul 2016 10:30

Tgt Ion:149 Resp: 594946

Ion	Ratio	Lower	Upper
149	100		
167	27.4	21.9	32.9
279	5.7	2.9	4.3#





#82

Chrysene

Concen: 19.23 ng/ul

RT: 21.90 min Scan# 3244

Delta R.T. -0.01 min

Lab File: BG022756.D

Acq: 1 Jul 2016 10:30

Instrument :

BNA_G

ClientSampleId :

SSTD02010

Tgt Ion:228 Resp: 1082074

Ion Ratio Lower Upper

228 100


226 30.2 24.2 36.4

229 21.6 15.0 22.6

Manual Integrations

sohil

7/4/2016 1:30:04 PM






Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E

21.90



#84

Di-n-octyl phthalate

Concen: 21.00 ng/ul


RT: 22.97 min Scan# 3427

Delta R.T. -0.01 min

Lab File: BG022756.D



Acq: 1 Jul 2016 10:30


Tgt Ion:149 Resp: 1007745



Abundance Ion 149.00 (148.70 to 149.70): E

22.97









#85
Benzo(b)fluoranthene
Concen: 18.23 ng/ul m
RT: 24.14 min Scan# 3626
Delta R.T. -0.00 min
Lab File: BG022756.D
Acq: 1 Jul 2016 10:30

Instrument :
BNA_G
ClientSampleId :
SSTD02010

Tgt Ion:252 Resp: 1240388
Ion Ratio Lower Upper
252 100
253 23.5 18.2 27.2
125 5.4 10.8 16.2#




Manual Integrations


sohil
7/4/2016 1:30:04 PM



#86
Benzo(k)fluoranthene
Concen: 18.53 ng/ul m
RT: 24.21 min Scan# 3638
Delta R.T. -0.01 min
Lab File: BG022756.D
Acq: 1 Jul 2016 10:30

Tgt Ion:252 Resp: 1184610
Ion Ratio Lower Upper
252 100
253 21.8 15.8 23.6
125 5.1 10.6 16.0#









#88
Benzo(a)pyrene
Concen: 18.06 ng/ul
RT: 25.05 min Scan# 3781
Delta R.T. -0.01 min
Lab File: BG022756.D
Acq: 1 Jul 2016 10:30

Instrument :
BNA_G
ClientSampleId :
SSTD02010

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.7	19.0	28.4
125	6.0	12.0	18.0




Manual Integrations


sohil
7/4/2016 1:30:04 PM



#89
Indeno(1,2,3-cd)pyrene
Concen: 19.15 ng/ul
RT: 29.06 min Scan# 4463
Delta R.T. -0.00 min
Lab File: BG022756.D
Acq: 1 Jul 2016 10:30

Tgt Ion	Ratio	Lower	Upper
276	100		
138	10.0	20.8	31.2
277	24.4	17.6	26.4









#90
Dibenzo(a,h)anthracene
Concen: 18.74 ng/ul
RT: 29.12 min Scan# 4474
Delta R.T. -0.01 min
Lab File: BG022756.D
Acq: 1 Jul 2016 10:30

Instrument :
BNA_G
ClientSampleId :
SSTD02010

Tgt Ion:278 Resp: 1058810
Ion Ratio Lower Upper
278 100
139 8.5 16.3 24.5#
279 23.1 19.1 28.7

Manual Integrations

sohil
7/4/2016 1:30:04 PM



#91
Benzo(g,h,i)perylene
Concen: 20.14 ng/ul
RT: 30.27 min Scan# 4669
Delta R.T. -0.00 min
Lab File: BG022756.D
Acq: 1 Jul 2016 10:30

Tgt Ion:276 Resp: 1079189
Ion Ratio Lower Upper
276 100
138 9.9 22.1 33.1#
277 25.5 17.0 25.4#

