

Data Path : Z:\HPCHEM1\BNA G\DATA\BG070117\
 Data File : BG027791.D
 Acq On : 1 Jul 2017 19:52
 Operator : SJ/JU
 Sample : I3982-17
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :

Quant Time: Jul 03 06:55:46 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG062817.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 28 19:26:37 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.48	152	30333	20.00	ng	0.00
21) Naphthalene-d8	11.30	136	144624	20.00	ng	0.00
38) Acenaphthene-d10	15.07	164	97877	20.00	ng	0.00
63) Phenanthrene-d10	17.81	188	248840	20.00	ng	0.00
75) Chrysene-d12	22.18	240	245576	20.00	ng	0.00
86) Perylene-d12	25.79	264	244604	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	6.06	112	281325	142.82	ng	0.00
7) Phenol-d6	7.62	99	354956	117.96	ng	0.00
23) Nitrobenzene-d5	9.64	82	247245	95.29	ng	0.00
41) 2,4,6-Tribromophenol	16.56	330	247389	184.28	ng	0.00
44) 2-Fluorobiphenyl	13.69	172	663029	96.75	ng	0.00
78) Terphenyl-d14	20.39	244	1195725	114.15	ng	0.00


Target Compounds Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG070117\
Data File : BG027791.D
Acq On : 1 Jul 2017 19:52
Operator : SJ/JU
Sample : I3982-17
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :

Quant Time: Jul 03 06:55:46 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG062817.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 28 19:26:37 2017
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 8.48 min Scan# 781
Delta R.T. -0.00 min
Lab File: BG027791.D
Acq: 1 Jul 2017 19:52

Instrument :
BNA_G
ClientSampleId :

Tot Ion:152 Resp: 30333
Ion Ratio Lower Upper
152 100
150 151.6 114.0 171.0
115 60.9 48.1 72.1




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 142.82 ng
RT: 6.06 min Scan# 369
Delta R.T. -0.00 min
Lab File: BG027791.D
Acq: 1 Jul 2017 19:52

Tgt Ion:112 Resp: 281325
Ion Ratio Lower Upper
112 100
64 71.5 61.8 92.6
63 39.2 37.2 55.8



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BGC
Ion 63.00 (62.70 to 63.70): BGC





#7

Phenol-d6

Concen: 117.96 ng

RT: 7.62 min Scan# 635

Delta R.T. 0.00 min

Lab File: BG027791.D

Acq: 1 Jul 2017 19:52

Instrument :

BNA_G

ClientSampled :


Tot Ion: 99 Resp: 354956

Ion Ratio Lower Upper




99 100

42 18.8 20.5 30.7#

71 33.9 37.6 56.4#



Abundance Ion 99.00 (98.70 to 99.70): BGC
Ion 42.00 (41.70 to 42.70): BGC
Ion 71.00 (70.70 to 71.70): BGC



#21

Naphthalene-d8

Concen: 20.00 ng

RT: 11.30 min Scan# 1261

Delta R.T. 0.00 min

Lab File: BG027791.D

Acq: 1 Jul 2017 19:52

Tgt Ion: 136 Resp: 144624


Ion Ratio Lower Upper

136 100



137 11.2 8.9 13.3


54 10.3 9.3 13.9

68 4.8 5.1 7.7#



Abundance Ion 136.00 (135.70 to 136.70): BGC
Ion 137.00 (136.70 to 137.70): BGC
Ion 54.00 (53.70 to 54.70): BGC
Ion 68.00 (67.70 to 68.70): BGC









#23
 Nitrobenzene-d5
 Concen: 95.29 ng
 RT: 9.64 min Scan# 979
 Delta R.T. -0.00 min
 Lab File: BG027791.D
 Acq: 1 Jul 2017 19:52




Instrument :
 BNA_G
 ClientSampled :


Tot Ion: 82 Resp: 247245
 Ion Ratio Lower Upper
 82 100
 128 43.9 26.4 39.6#
 54 60.8 49.4 74.2



#38
 Acenaphthene-d10
 Concen: 20.00 ng
 RT: 15.07 min Scan# 1903
 Delta R.T. 0.00 min
 Lab File: BG027791.D
 Acq: 1 Jul 2017 19:52

Tgt Ion:164 Resp: 97877
 Ion Ratio Lower Upper
 164 100
 162 101.3 85.1 127.7
 160 46.3 34.7 52.1









#41
 2,4,6-Tribromophenol
 Concen: 184.28 ng
 RT: 16.56 min Scan# 2156
 Delta R.T. 0.00 min
 Lab File: BG027791.D
 Acq: 1 Jul 2017 19:52




Instrument :
 BNA_G
 ClientSampled :


Tot Ion	Ratio	Lower	Upper
330	100		
332	95.5	77.3	115.9
141	34.2	32.1	48.1



#44
 2-Fluorobiphenyl
 Concen: 96.75 ng
 RT: 13.69 min Scan# 1668
 Delta R.T. -0.00 min
 Lab File: BG027791.D
 Acq: 1 Jul 2017 19:52

Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.5	32.2	48.2
170	24.2	21.8	32.6









#63
Phenanthrene-d10
Concen: 20.00 ng
RT: 17.81 min Scan# 2370
Delta R.T. -0.00 min
Lab File: BG027791.D
Acq: 1 Jul 2017 19:52

Instrument :
BNA_G
ClientSampleId :

Tot Ion:188 Resp: 248840
Ion Ratio Lower Upper
188 100
94 9.2 5.8 8.8#
80 11.0 7.5 11.3




Abundance Ion 188.00 (187.70 to 188.70): E
Ion 94.00 (93.70 to 94.70): BGC
Ion 80.00 (79.70 to 80.70): BGC






#75
Chrysene-d12
Concen: 20.00 ng
RT: 22.18 min Scan# 3113
Delta R.T. 0.00 min
Lab File: BG027791.D
Acq: 1 Jul 2017 19:52

Tgt Ion:240 Resp: 245576
Ion Ratio Lower Upper
240 100
120 11.5 5.7 8.5#
236 26.1 22.2 33.4



Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E









#78
 Terphenyl-d14
 Concen: 114.15 ng
 RT: 20.39 min Scan# 2809
 Delta R.T. 0.00 min
 Lab File: BG027791.D
 Acq: 1 Jul 2017 19:52

Instrument :
 BNA_G
 ClientSampled :

Tot Ion:244 Resp: 1195725
 Ion Ratio Lower Upper
 244 100
 212 7.6 10.0 15.0#
 122 11.0 8.6 12.8



#86
 Perylene-d12
 Concen: 20.00 ng
 RT: 25.79 min Scan# 3728
 Delta R.T. 0.00 min
 Lab File: BG027791.D
 Acq: 1 Jul 2017 19:52

Tgt Ion:264 Resp: 244604
 Ion Ratio Lower Upper
 264 100
 260 22.7 21.8 32.8
 265 20.9 18.2 27.4

