

Data Path : Z:\HPCHEM1\BNA_G\Data\BG070217\
 Data File : BG027812.D
 Acq On : 2 Jul 2017 10:37
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 03 08:18:07 2017
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG062817.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 28 19:26:37 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
2	1,4-Dioxane	0.484	0.399	17.6	65	0.00
3	Pyridine	1.489	1.203	19.2	61	0.00
4	n-Nitrosodimethylamine	0.730	0.660	9.6	73	0.00
5 S	2-Fluorophenol	1.299	1.264	2.7	75	0.00
6	Aniline	2.300	1.614	29.8#	52	0.00
7 S	Phenol-d6	1.984	1.891	4.7	72	0.00
8	2-Chlorophenol	1.451	1.458	-0.5	78	0.00
9	Benzaldehyde	1.181	1.022	13.5	66	0.00
10 C	Phenol	1.963	1.868	4.8	72	0.00
11	bis(2-Chloroethyl)ether	1.491	1.549	-3.9	80	0.00
12	1,3-Dichlorobenzene	1.544	1.613	-4.5	81	0.00
13 C	1,4-Dichlorobenzene	1.600	1.664	-4.0	81	0.00
14	1,2-Dichlorobenzene	1.551	1.669	-7.6	83	0.00
15	Benzyl Alcohol	1.373	0.215	84.3#	12#	0.00
16	2,2'-oxybis(1-Chloropropane	3.180	2.980	6.3	72	0.00
17	2-Methylphenol	1.389	1.379	0.7	74	0.00
18	Hexachloroethane	0.563	0.563	0.0	79	0.00
19 P	n-Nitroso-di-n-propylamine	1.443	1.321	8.5	69	0.00
20	3+4-Methylphenols	2.003	1.904	4.9	71	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	77	0.00
22	Acetophenone	0.483	0.482	0.2	76	0.00
23 S	Nitrobenzene-d5	0.359	0.361	-0.6	76	0.00
24	Nitrobenzene	0.375	0.374	0.3	76	0.00
25	Isophorone	0.782	0.727	7.0	71	0.00
26 C	2-Nitrophenol	0.169	0.180	-6.5	78	0.00
27	2,4-Dimethylphenol	0.317	0.320	-0.9	75	0.00
28	bis(2-Chloroethoxy)methane	0.443	0.405	8.6	70	0.00
29 C	2,4-Dichlorophenol	0.278	0.306	-10.1	82	0.00
30	1,2,4-Trichlorobenzene	0.284	0.315	-10.9	85	0.00
31	Naphthalene	1.064	1.092	-2.6	78	0.00
32	Benzoic acid	0.169	0.076	55.0#	34#	0.00
33	4-Chloroaniline	0.459	0.437	4.8	71	0.10
34 C	Hexachlorobutadiene	0.158	0.196	-24.1#	98	0.00
35	Caprolactam	0.140	0.124	11.4	67	0.02
36 C	4-Chloro-3-methylphenol	0.391	0.366	6.4	70	0.00
37	2-Methylnaphthalene	0.792	0.832	-5.1	80	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	78	0.00
39	1,2,4,5-Tetrachlorobenzene	0.526	0.615	-16.9	91	0.00
40 P	Hexachlorocyclopentadiene	0.248	0.310	-25.0	97	0.00
41 S	2,4,6-Tribromophenol	0.274	0.326	-19.0	90	0.00
42 C	2,4,6-Trichlorophenol	0.380	0.415	-9.2	84	0.00
43	2,4,5-Trichlorophenol	0.430	0.462	-7.4	81	0.00
44 S	2-Fluorobiphenyl	1.400	1.519	-8.5	83	0.00

Data Path : Z:\HPCHEM1\BNA_G\Data\BG070217\
 Data File : BG027812.D
 Acq On : 2 Jul 2017 10:37
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 03 08:18:07 2017
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG062817.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 28 19:26:37 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.603	1.671	-4.2	81	0.00
46	2-Chloronaphthalene	1.211	1.268	-4.7	81	0.00
47	2-Nitroaniline	0.485	0.468	3.5	73	0.00
48	Acenaphthylene	2.213	2.261	-2.2	78	0.00
49	Dimethylphthalate	1.706	1.692	0.8	77	0.00
50	2,6-Dinitrotoluene	0.371	0.387	-4.3	78	0.00
51 C	Acenaphthene	1.418	1.488	-4.9	80	0.00
52	3-Nitroaniline	0.436	0.394	9.6	68	0.02
53 P	2,4-Dinitrophenol	0.164	0.198	-20.7	93	0.00
54	Dibenzofuran	1.875	1.938	-3.4	80	0.00
55 P	4-Nitrophenol	0.333	0.224	32.7#	50	0.02
56	2,4-Dinitrotoluene	0.517	0.544	-5.2	79	0.00
57	Fluorene	1.818	1.856	-2.1	78	0.00
58	2,3,4,6-Tetrachlorophenol	0.371	0.364	1.9	76	0.00
59	Diethylphthalate	1.889	1.855	1.8	77	0.00
60	4-Chlorophenyl-phenylether	0.805	0.862	-7.1	82	0.00
61	4-Nitroaniline	0.455	0.396	13.0	67	0.00
62	Azobenzene	1.660	1.550	6.6	72	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	75	0.00
64	4,6-Dinitro-2-methylphenol	0.115	0.135	-17.4	87	0.00
65 c	n-Nitrosodiphenylamine	0.651	0.689	-5.8	78	0.00
66	4-Bromophenyl-phenylether	0.209	0.248	-18.7	87	0.00
67	Hexachlorobenzene	0.222	0.264	-18.9	88	0.00
68	Atrazine	0.205	0.195	4.9	69	0.01
69 C	Pentachlorophenol	0.123	0.133	-8.1	80	0.00
70	Phenanthrene	1.155	1.203	-4.2	77	0.00
71	Anthracene	1.166	1.222	-4.8	77	0.00
72	Carbazole	1.148	1.157	-0.8	74	0.00
73	Di-n-butylphthalate	1.262	1.323	-4.8	77	0.00
74 C	Fluoranthene	1.262	1.348	-6.8	80	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	81	0.00
76	Benzidine	0.447	0.025	94.4#	4#	0.05
77	Pyrene	1.255	1.274	-1.5	81	0.00
78 S	Terphenyl-d14	0.853	0.932	-9.3	84	0.00
79	Butylbenzylphthalate	0.565	0.548	3.0	77	0.00
80	Benzo(a)anthracene	1.209	1.238	-2.4	81	0.00
81	3,3'-Dichlorobenzidine	0.439	0.412	6.2	73	0.00
82	Chrysene	1.133	1.175	-3.7	82	0.00
83	Bis(2-ethylhexyl)phthalate	0.826	0.790	4.4	75	0.00
84 c	Di-n-octyl phthalate	1.365	1.314	3.7	75	0.00
85	Indeno(1,2,3-cd)pyrene	1.295	1.329	-2.6	80	0.00
86 I	Perylene-d12	1.000	1.000	0.0	85	0.00
87	Benzo(b)fluoranthene	1.289	1.305	-1.2	84	0.00

Data Path : Z:\HPCHEM1\BNA_G\Data\BG070217\
Data File : BG027812.D
Acq On : 2 Jul 2017 10:37
Operator : SJ/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jul 03 08:18:07 2017
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG062817.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 28 19:26:37 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.274	1.281	-0.5	83	0.00
89 C	Benzo(a)pyrene	1.209	1.229	-1.7	84	0.00
90	Dibenzo(a,h)anthracene	1.225	1.155	5.7	78	0.00
91	Benzo(g,h,i)perylene	1.177	1.132	3.8	80	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1