

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG070221\
 Data File : BG049163.D
 Acq On : 2 Jul 2021 11:51
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 02 13:41:26 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG063021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 30 15:39:12 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2	1,4-Dioxane	0.561	0.534	4.8	99	0.00
3	Pyridine	1.516	1.482	2.2	99	0.00
4	n-Nitrosodimethylamine	0.861	0.874	-1.5	103	0.00
5 S	2-Fluorophenol	1.277	1.277	0.0	102	0.00
6	Aniline	2.216	2.142	3.3	103	0.00
7 S	Phenol-d6	1.821	1.843	-1.2	104	0.00
8	2-Chlorophenol	1.401	1.375	1.9	103	0.00
9	Benzaldehyde	0.958	1.052	-9.8	103	0.00
10 C	Phenol	1.829	1.831	-0.1	102	0.00
11	bis(2-Chloroethyl)ether	1.334	1.310	1.8	101	0.00
12	1,3-Dichlorobenzene	1.567	1.534	2.1	104	0.00
13 C	1,4-Dichlorobenzene	1.576	1.550	1.6	103	0.00
14	1,2-Dichlorobenzene	1.517	1.475	2.8	103	0.00
15	Benzyl Alcohol	1.603	1.601	0.1	102	0.00
16	2,2'-oxybis(1-Chloropropane	2.264	2.238	1.1	103	0.00
17	2-Methylphenol	1.232	1.241	-0.7	104	0.00
18	Hexachloroethane	0.629	0.629	0.0	104	0.00
19 P	n-Nitroso-di-n-propylamine	1.308	1.292	1.2	103	0.00
20	3+4-Methylphenols	1.708	1.706	0.1	104	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
22	Acetophenone	0.597	0.613	-2.7	106	0.00
23 S	Nitrobenzene-d5	0.472	0.479	-1.5	105	0.00
24	Nitrobenzene	0.511	0.500	2.2	103	0.00
25	Isophorone	0.906	0.876	3.3	101	0.00
26 C	2-Nitrophenol	0.203	0.201	1.0	102	0.00
27	2,4-Dimethylphenol	0.305	0.299	2.0	100	0.00
28	bis(2-Chloroethoxy)methane	0.427	0.419	1.9	103	0.00
29 C	2,4-Dichlorophenol	0.366	0.368	-0.5	103	0.00
30	1,2,4-Trichlorobenzene	0.431	0.420	2.6	102	0.00
31	Naphthalene	1.164	1.131	2.8	102	0.00
32	Benzoic acid	0.222	0.230	-3.6	103	0.00
33	4-Chloroaniline	0.481	0.479	0.4	104	0.00
34 C	Hexachlorobutadiene	0.319	0.308	3.4	103	0.00
35	Caprolactam	0.116	0.121	-4.3	108	0.00
36 C	4-Chloro-3-methylphenol	0.421	0.415	1.4	102	0.00
37	2-Methylnaphthalene	0.877	0.865	1.4	103	0.00
38	1-Methylnaphthalene	0.830	0.818	1.4	104	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
40	1,2,4,5-Tetrachlorobenzene	0.698	0.698	0.0	101	0.00
41 P	Hexachlorocyclopentadiene	0.413	0.412	0.2	99	0.00
42 S	2,4,6-Tribromophenol	0.347	0.349	-0.6	103	0.00
43 C	2,4,6-Trichlorophenol	0.482	0.477	1.0	100	0.00
44	2,4,5-Trichlorophenol	0.527	0.503	4.6	97	0.00
45 S	2-Fluorobiphenyl	1.482	1.480	0.1	103	0.00
46	1,1'-Biphenyl	1.494	1.516	-1.5	103	0.00
47	2-Chloronaphthalene	1.229	1.208	1.7	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG070221\
 Data File : BG049163.D
 Acq On : 2 Jul 2021 11:51
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 02 13:41:26 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG063021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 30 15:39:12 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.448	0.461	-2.9	108	0.00
49	Acenaphthylene	1.938	1.902	1.9	104	0.00
50	Dimethylphthalate	1.605	1.556	3.1	100	0.00
51	2,6-Dinitrotoluene	0.368	0.356	3.3	100	0.00
52 C	Acenaphthene	1.208	1.190	1.5	102	0.00
53	3-Nitroaniline	0.353	0.350	0.8	107	0.00
54 P	2,4-Dinitrophenol	0.222	0.217	2.3	101	-0.01
55	Dibenzofuran	1.967	1.936	1.6	104	0.00
56 P	4-Nitrophenol	0.288	0.285	1.0	101	0.00
57	2,4-Dinitrotoluene	0.523	0.513	1.9	103	0.00
58	Fluorene	1.626	1.604	1.4	104	0.00
59	2,3,4,6-Tetrachlorophenol	0.490	0.486	0.8	100	0.00
60	Diethylphthalate	1.701	1.672	1.7	102	0.00
61	4-Chlorophenyl-phenylether	0.900	0.868	3.6	102	0.00
62	4-Nitroaniline	0.367	0.358	2.5	103	0.00
63	Azobenzene	1.715	1.682	1.9	104	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	0.143	0.138	3.5	99	0.00
66 c	n-Nitrosodiphenylamine	0.626	0.599	4.3	102	0.00
67	4-Bromophenyl-phenylether	0.270	0.261	3.3	102	0.00
68	Hexachlorobenzene	0.298	0.283	5.0	101	0.00
69	Atrazine	0.208	0.217	-4.3	102	0.00
70 C	Pentachlorophenol	0.174	0.177	-1.7	104	0.00
71	Phenanthrene	1.152	1.105	4.1	104	0.00
72	Anthracene	1.148	1.103	3.9	105	0.00
73	Carbazole	1.095	1.055	3.7	104	0.00
74	Di-n-butylphthalate	1.259	1.218	3.3	104	0.00
75 C	Fluoranthene	1.442	1.369	5.1	104	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
77	Benzydine	0.430	0.510	-18.6	103	0.00
78	Pyrene	1.485	1.436	3.3	103	0.00
79 S	Terphenyl-d14	1.180	1.150	2.5	101	0.00
80	Butylbenzylphthalate	0.563	0.536	4.8	99	0.00
81	Benzo(a)anthracene	1.492	1.411	5.4	101	0.00
82	3,3'-Dichlorobenzidine	0.507	0.479	5.5	102	0.00
83	Chrysene	1.417	1.344	5.2	101	0.00
84	Bis(2-ethylhexyl)phthalate	0.795	0.761	4.3	101	0.00
85 c	Di-n-octyl phthalate	1.338	1.288	3.7	102	-0.01
86 I	Perylene-d12	1.000	1.000	0.0	100	-0.02
87	Indeno(1,2,3-cd)pyrene	1.596	1.540	3.5	100	-0.02
88	Benzo(b)fluoranthene	1.443	1.393	3.5	99	-0.01
89	Benzo(k)fluoranthene	1.377	1.335	3.1	101	-0.01
90 C	Benzo(a)pyrene	1.331	1.296	2.6	101	-0.01
91	Dibenzo(a,h)anthracene	1.347	1.303	3.3	100	-0.01
92	Benzo(g,h,i)perylene	1.328	1.283	3.4	100	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG070221\
Data File : BG049163.D
Acq On : 2 Jul 2021 11:51
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jul 02 13:41:26 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG063021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 30 15:39:12 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0