

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG070224\
 Data File : BG061856.D
 Acq On : 2 Jul 2024 22:40
 Operator : MA/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SICV403

Quant Time: Jul 03 02:41:25 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070224.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 03 02:38:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.101	152	50792	20.000	ng/ul	0.00
20) Naphthalene-d8	10.909	136	247131	20.000	ng/ul	# 0.00
38) Acenaphthene-d10	14.722	164	173111	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.460	188	411271	20.000	ng/ul	0.00
79) Chrysene-d12	21.726	240	357896	20.000	ng/ul	0.00
88) Perylene-d12	24.963	264	415805	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.482	96	8682	6.423	ng/uL	0.00
4) Pyridine-d5	3.905	84	76667	18.447	ng/ul	0.00
7) Phenol-d5	7.249	99	102174	18.382	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.419	67	74683	21.217	ng/ul	0.00
11) 2-Chlorophenol-d4	7.625	132	76554	20.078	ng/ul	0.00
15) 4-Methylphenol-d8	8.800	113	83458	19.070	ng/ul	0.00
21) Nitrobenzene-d5	9.264	128	38929	20.712	ng/ul	0.00
24) 2-Nitrophenol-d4	9.987	143	45058	20.993	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.521	165	84920	20.512	ng/ul	-0.01
31) 4-Chloroaniline-d4	11.038	131	120603	20.065	ng/ul	0.00
46) Dimethylphthalate-d6	14.123	166	282219	19.830	ng/ul	0.00
49) Acenaphthylene-d8	14.417	160	308080	20.702	ng/ul	0.00
54) 4-Nitrophenol-d4	14.916	143	45075	19.832	ng/ul	0.00
60) Fluorene-d10	15.709	176	245166	20.463	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.821	200	46025	21.309	ng/ul	0.00
73) Anthracene-d10	17.560	188	382764	19.934	ng/ul	0.00
81) Pyrene-d10	19.840	212	433082	20.573	ng/ul	0.00
92) Benzo(a)pyrene-d12	24.746	264	431432	20.921	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.518	88	9546	6.399	ng/uL	87
5) Pyridine	3.923	79	68574	16.089	ng/ul	92
6) Benzaldehyde	7.231	77	58210	19.787	ng/ul	90
8) Phenol	7.278	94	102230	17.964	ng/ul	92
10) Bis(2-Chloroethyl)ether	7.513	93	81960	18.747	ng/ul	95
12) 2-Chlorophenol	7.660	128	75556	19.451	ng/ul	96
13) 2-Methylphenol	8.535	108	74334	18.301	ng/ul	91
14) 2,2'-oxybis(1-Chloropr...	8.612	45	171644	17.971	ng/ul#	97
16) Acetophenone	8.923	105	135618	18.760	ng/ul	93
17) N-Nitroso-di-n-propyla...	8.900	70	73283	18.107	ng/ul	95
18) 4-Methylphenol	8.864	108	81470	17.910	ng/ul	99
19) Hexachloroethane	9.182	117	30565	18.138	ng/ul	91
22) Nitrobenzene	9.305	77	103764	19.449	ng/ul	98
23) Isophorone	9.822	82	211153	17.518	ng/ul	100
25) 2-Nitrophenol	10.016	139	44545	20.351	ng/ul	88
26) 2,4-Dimethylphenol	10.069	107	74548	14.230	ng/ul	98
27) Bis(2-Chloroethoxy)met...	10.310	93	118743	17.889	ng/ul	97
29) 2,4-Dichlorophenol	10.551	162	75040	19.296	ng/ul	92
30) Naphthalene	10.962	128	264176	19.461	ng/ul#	94
32) 4-Chloroaniline	11.068	127	112808	19.127	ng/ul	98
33) Hexachlorobutadiene	11.232	225	54219	18.527	ng/ul#	77
34) Caprolactam	11.826	113	12099	7.941	ng/ul#	70
35) 4-Chloro-3-methylphenol	12.178	107	102997	19.886	ng/ul	99
36) 2-Methylnaphthalene	12.554	142	188047	19.121	ng/ul	95

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG070224\
 Data File : BG061856.D
 Acq On : 2 Jul 2024 22:40
 Operator : MA/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1


Instrument :
 BNA_G
 ClientSampleId :
 SICV403


Quant Time: Jul 03 02:41:25 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070224.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Jul 03 02:38:46 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.772	142	179995	17.612	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.918	216	100971	19.423	ng/ul	90
40) Hexachlorocyclopentadiene	12.889	237	66386	19.183	ng/ul	94
41) 2,4,6-Trichlorophenol	13.153	196	66853	19.362	ng/ul	91
42) 2,4,5-Trichlorophenol	13.230	196	73340	19.336	ng/ul	89
43) 1,1'-Biphenyl	13.553	154	261097	19.575	ng/ul	96
44) 2-Chloronaphthalene	13.600	162	193370	19.317	ng/ul	93
45) 2-Nitroaniline	13.806	65	83599	20.854	ng/ul	97
47) Dimethylphthalate	14.170	163	260973	19.064	ng/ul	100
48) 2,6-Dinitrotoluene	14.293	165	54886	21.135	ng/ul	93
50) Acenaphthylene	14.446	152	321437	19.680	ng/ul	97
51) 3-Nitroaniline	14.622	138	56768	22.457	ng/ul#	98
52) Acenaphthene	14.781	153	213029	18.889	ng/ul	98
53) 2,4-Dinitrophenol	14.828	184	25430	18.743	ng/ul	94
55) 4-Nitrophenol	14.928	109	50550	19.487	ng/ul#	91
56) Dibenzofuran	15.116	168	311224	19.525	ng/ul	100
57) 2,4-Dinitrotoluene	15.075	165	80632	21.172	ng/ul#	99
58) 2,3,4,6-Tetrachlorophenol	15.339	232	69163	20.193	ng/ul#	96
59) Diethylphthalate	15.527	149	270940	18.367	ng/ul	96
61) Fluorene	15.762	166	256712	18.937	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.750	204	128011	18.800	ng/ul	95
63) 4-Nitroaniline	15.780	138	58648	22.865	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.833	198	47542	20.452	ng/ul	92
67) N-Nitrosodiphenylamine	15.968	169	221453	19.717	ng/ul	97
68) 4-Bromophenyl-phenylether	16.644	248	88356	18.937	ng/ul	96
69) Hexachlorobenzene	16.755	284	103278	19.250	ng/ul	95
70) Atrazine	16.908	200	76331	17.216	ng/ul	98
71) Pentachlorophenol	17.102	266	54615	18.520	ng/ul	95
72) Phenanthrene	17.501	178	414287	19.104	ng/ul	99
74) Anthracene	17.595	178	422530	18.854	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.524	216	103305	20.224	ng/ul	96
76) Pentachlorobenzene	15.034	250	113309	19.449	ng/ul	96
77) Carbazole	17.860	167	380661	20.183	ng/ul	99
78) Di-n-butylphthalate	18.412	149	458797	18.947	ng/ul	99
80) Fluoranthene	19.505	202	467104	18.750	ng/ul	98
82) Pyrene	19.863	202	502161	19.513	ng/ul	99
83) Butylbenzylphthalate	20.745	149	204597	18.944	ng/ul	91
84) 3,3'-Dichlorobenzidine	21.614	252	195270	22.934	ng/ul	92
85) Benzo(a)anthracene	21.702	228	495444	19.120	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.602	149	287991	18.583	ng/ul	99
87) Chrysene	21.767	228	464057	19.130	ng/ul	99
89) Di-n-octyl phthalate	22.824	149	507822	19.991	ng/ul	100
90) Benzo(b)fluoranthene	23.929	252	496563	18.955	ng/ul	99
91) Benzo(k)fluoranthene	24.000	252	506249	19.379	ng/ul	95
93) Benzo(a)pyrene	24.810	252	473745	19.097	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	28.659	276	614776	19.393	ng/ul#	93
95) Dibenzo(a,h)anthracene	28.729	278	509378	19.455	ng/ul	99
96) Benzo(g,h,i)perylene	29.822	276	471230	18.685	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jul 03 02:41:25 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG070224.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Jul 03 02:38:46 2024
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 6.399 ng/uL
RT: 3.518 min Scan# 39
Delta R.T. -0.005 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40




Instrument :
BNA_G
ClientSampleId :
SICV403


Tgt Ion: 88 Resp: 9546
Ion Ratio Lower Upper
88 100
43 46.0 39.4 59.2
58 80.4 78.1 117.1



#5
Pyridine
Concen: 16.089 ng/uL
RT: 3.923 min Scan# 108
Delta R.T. -0.005 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40

Tgt Ion: 79 Resp: 68574
Ion Ratio Lower Upper
79 100
52 118.3 86.0 129.0
51 54.4 46.4 69.6









#6
Benzaldehyde
Concen: 19.787 ng/ul
RT: 7.231 min Scan# 61
Delta R.T. -0.005 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40




Instrument :
BNA_G
ClientSampleId :
SICV403


Tgt Ion: 77 Resp: 58210
Ion Ratio Lower Upper
77 100
105 90.9 79.6 119.4
106 88.5 62.6 93.8



#8
Phenol
Concen: 17.964 ng/ul
RT: 7.278 min Scan# 679
Delta R.T. 0.001 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40

Tgt Ion: 94 Resp: 102230
Ion Ratio Lower Upper
94 100
65 32.4 29.0 43.4
66 52.6 37.4 56.2





#10

Bis(2-Chloroethyl)ether

Concen: 18.747 ng/ul

RT: 7.513 min Scan# 719

Delta R.T. 0.001 min

Lab File: BG061856.D

Acq: 2 Jul 2024 22:40

Instrument :

BNA_G

ClientSampleId :

SICV403





Tgt Ion: 93 Resp: 81960

Ion Ratio Lower Upper

93 100

63 96.0 72.2 108.2

95 32.5 26.7 40.1



#12

2-Chlorophenol

Concen: 19.451 ng/ul

RT: 7.660 min Scan# 744

Delta R.T. 0.001 min

Lab File: BG061856.D

Acq: 2 Jul 2024 22:40




Tgt Ion:128 Resp: 75556


Ion Ratio Lower Upper

128 100

64 53.9 42.6 63.8

130 33.6 23.3 34.9





#13

2-Methylphenol

Concen: 18.301 ng/ul

RT: 8.535 min Scan# 894

Delta R.T. -0.005 min

Lab File: BG061856.D

Acq: 2 Jul 2024 22:40

Instrument :

BNA_G

ClientSampleId :





SICV403

Tgt Ion:108 Resp: 74334

Ion Ratio Lower Upper

108 100

107 85.3 74.9 112.3



#14

2,2'-oxybis(1-Chloropropane)

Concen: 17.971 ng/ul

RT: 8.612 min Scan# 906

Delta R.T. -0.011 min

Lab File: BG061856.D

Acq: 2 Jul 2024 22:40




Tgt Ion: 45 Resp: 171644


Ion Ratio Lower Upper

45 100

77 10.3 7.4 11.0

79 7.2 4.8 7.2#









#16
Acetophenone
Concen: 18.760 ng/ul
RT: 8.923 min Scan# 91
Delta R.T. 0.001 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40




Instrument :
BNA_G
ClientSampleId :
SICV403


Tgt Ion:105 Resp: 135618
Ion Ratio Lower Upper
105 100
77 86.9 64.1 96.1
51 45.5 33.8 50.8



#17
N-Nitroso-di-n-propylamine
Concen: 18.107 ng/ul
RT: 8.900 min Scan# 955
Delta R.T. 0.001 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40

Tgt Ion: 70 Resp: 73283
Ion Ratio Lower Upper
70 100
42 85.9 64.7 97.1
101 8.0 7.3 10.9
130 19.2 17.0 25.6









#18
4-Methylphenol
Concen: 17.910 ng/ul
RT: 8.864 min Scan# 949
Delta R.T. 0.001 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40




Instrument :
BNA_G
ClientSampleId :
SICV403


Tgt Ion:108 Resp: 81470
Ion Ratio Lower Upper
108 100
107 108.9 88.4 132.6



#19
Hexachloroethane
Concen: 18.138 ng/ul
RT: 9.182 min Scan# 1003
Delta R.T. 0.001 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40

Tgt Ion:117 Resp: 30565
Ion Ratio Lower Upper
117 100
201 92.6 69.3 103.9
199 70.4 48.9 73.3









#22
Nitrobenzene
Concen: 19.449 ng/ul
RT: 9.305 min Scan# 1024
Delta R.T. -0.005 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40




Instrument :
BNA_G
ClientSampleId :
SICV403


Tgt Ion: 77 Resp: 103764
Ion Ratio Lower Upper
77 100
123 40.7 31.0 46.6
65 15.7 12.7 19.1



#23
Isophorone
Concen: 17.518 ng/ul
RT: 9.822 min Scan# 1112
Delta R.T. -0.005 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40

Tgt Ion: 82 Resp: 211153
Ion Ratio Lower Upper
82 100
95 7.1 6.0 9.0
138 15.0 11.9 17.9









#25
2-Nitrophenol
Concen: 20.351 ng/ul
RT: 10.016 min Scan# 1145
Delta R.T. 0.001 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40




Instrument :
BNA_G
ClientSampleId :
SICV403


Tgt Ion:139 Resp: 44545
Ion Ratio Lower Upper
139 100
65 66.4 61.6 92.4
109 51.4 35.4 53.2



#26
2,4-Dimethylphenol
Concen: 14.230 ng/ul
RT: 10.069 min Scan# 1154
Delta R.T. -0.005 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40

Tgt Ion:107 Resp: 74548
Ion Ratio Lower Upper
107 100
121 46.1 35.6 53.4
122 76.3 62.2 93.2









#27
 Bis(2-Chloroethoxy)methane
 Concen: 17.889 ng/ul
 RT: 10.310 min Scan# 1195
 Delta R.T. 0.001 min
 Lab File: BG061856.D
 Acq: 2 Jul 2024 22:40




Instrument :
 BNA_G
 ClientSampleId :
 SICV403


Tgt Ion: 93 Resp: 118743
 Ion Ratio Lower Upper
 93 100
 95 30.8 26.2 39.2
 123 9.3 7.5 11.3



#29
 2,4-Dichlorophenol
 Concen: 19.296 ng/ul
 RT: 10.551 min Scan# 1236
 Delta R.T. -0.005 min
 Lab File: BG061856.D
 Acq: 2 Jul 2024 22:40

Tgt Ion: 162 Resp: 75040
 Ion Ratio Lower Upper
 162 100
 164 65.7 46.9 70.3
 98 45.2 33.4 50.2









#30
Naphthalene
Concen: 19.461 ng/ul
RT: 10.962 min Scan# 1306
Delta R.T. 0.001 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40




Instrument :
BNA_G
ClientSampleId :
SICV403


Tgt Ion:128 Resp: 264176
Ion Ratio Lower Upper
128 100
129 12.2 7.7 11.5#
127 13.7 9.6 14.4



#32
4-Chloroaniline
Concen: 19.127 ng/ul
RT: 11.068 min Scan# 1324
Delta R.T. 0.001 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40

Tgt Ion:127 Resp: 112808
Ion Ratio Lower Upper
127 100
129 29.0 24.2 36.4









#33
Hexachlorobutadiene
Concen: 18.527 ng/ul
RT: 11.232 min Scan# 11
Delta R.T. -0.005 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40




Instrument :
BNA_G
ClientSampleId :
SICV403


Tgt Ion:225 Resp: 54219
Ion Ratio Lower Upper
225 100
223 75.2 45.4 68.2#
227 70.2 44.0 66.0#



#34
Caprolactam
Concen: 7.941 ng/ul
RT: 11.826 min Scan# 1453
Delta R.T. -0.017 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40

Tgt Ion:113 Resp: 12099
Ion Ratio Lower Upper
113 100
55 279.3 184.6 276.8#
56 211.8 135.0 202.6#









#35
4-Chloro-3-methylphenol
Concen: 19.886 ng/ul
RT: 12.178 min Scan# 11
Delta R.T. -0.005 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40




Instrument :
BNA_G
ClientSampleId :
SICV403


Tgt Ion:107 Resp: 102997
Ion Ratio Lower Upper
107 100
144 24.1 19.6 29.4
142 74.2 60.3 90.5



#36
2-Methylnaphthalene
Concen: 19.121 ng/ul
RT: 12.554 min Scan# 1577
Delta R.T. -0.005 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40

Tgt Ion:142 Resp: 188047
Ion Ratio Lower Upper
142 100
141 82.0 69.4 104.2





#37

1-Methylnaphthalene

Concen: 17.612 ng/ul

RT: 12.772 min Scan# 1

Delta R.T. -0.005 min

Lab File: BG061856.D

Acq: 2 Jul 2024 22:40

Instrument :

BNA_G

ClientSampleId :

SICV403





Tgt Ion:142 Resp: 179995

Ion Ratio Lower Upper

142 100

141 87.8 71.4 107.2

116 4.7 3.8 5.8



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.423 ng/ul

RT: 12.918 min Scan# 1639

Delta R.T. 0.001 min

Lab File: BG061856.D

Acq: 2 Jul 2024 22:40

Tgt Ion:216 Resp: 100971




Ion Ratio Lower Upper


216 100

214 89.1 62.5 93.7

179 26.8 18.6 27.8

108 17.9 13.7 20.5









#40
Hexachlorocyclopentadiene
Concen: 19.183 ng/ul
RT: 12.889 min Scan# 1634
Delta R.T. -0.005 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40




Instrument :
BNA_G
ClientSampleId :
SICV403


Tgt Ion:237 Resp: 66386
Ion Ratio Lower Upper
237 100
235 60.4 51.7 77.5
272 12.8 12.4 18.6



#41
2,4,6-Trichlorophenol
Concen: 19.362 ng/ul
RT: 13.153 min Scan# 1679
Delta R.T. -0.005 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40

Tgt Ion:196 Resp: 66853
Ion Ratio Lower Upper
196 100
198 110.5 79.9 119.9
200 35.7 26.7 40.1





#42

2,4,5-Trichlorophenol

Concen: 19.336 ng/ul

RT: 13.230 min Scan# 1

Delta R.T. 0.001 min

Lab File: BG061856.D

Acq: 2 Jul 2024 22:40

Instrument :

BNA_G

ClientSampleId :

SICV403





Tgt Ion:196 Resp: 73340

Ion Ratio Lower Upper

196 100

198 103.2 72.5 108.7

200 32.8 23.7 35.5



#43

1,1'-Biphenyl

Concen: 19.575 ng/ul

RT: 13.553 min Scan# 1747

Delta R.T. -0.005 min

Lab File: BG061856.D

Acq: 2 Jul 2024 22:40




Tgt Ion:154 Resp: 261097


Ion Ratio Lower Upper

154 100

153 42.3 31.7 47.5

76 12.9 10.0 15.0









#44
2-Chloronaphthalene
Concen: 19.317 ng/ul
RT: 13.600 min Scan# 1755
Delta R.T. 0.001 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40




Instrument :
BNA_G
ClientSampleId :
SICV403


Tgt Ion:162 Resp: 193370
Ion Ratio Lower Upper
162 100
164 32.9 23.3 34.9
127 45.4 33.2 49.8



#45
2-Nitroaniline
Concen: 20.854 ng/ul
RT: 13.806 min Scan# 1790
Delta R.T. 0.001 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40

Tgt Ion: 65 Resp: 83599
Ion Ratio Lower Upper
65 100
92 57.3 48.2 72.4
138 69.2 56.8 85.2









#47
Dimethylphthalate
Concen: 19.064 ng/ul
RT: 14.170 min Scan# 1852
Delta R.T. 0.001 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40




Instrument :
BNA_G
ClientSampleId :
SICV403


Tgt Ion:163 Resp: 260973
Ion Ratio Lower Upper
163 100
194 3.4 2.6 4.0
164 10.2 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 21.135 ng/ul
RT: 14.293 min Scan# 1873
Delta R.T. 0.001 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40

Tgt Ion:165 Resp: 54886
Ion Ratio Lower Upper
165 100
89 71.0 62.6 93.8
121 21.9 18.4 27.6









#50
Acenaphthylene
Concen: 19.680 ng/ul
RT: 14.446 min Scan# 1899
Delta R.T. 0.001 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40




Instrument :
BNA_G
ClientSampleId :
SICV403


Tgt Ion:152 Resp: 321437
Ion Ratio Lower Upper
152 100
151 21.4 16.5 24.7
153 11.7 11.4 17.0



#51
3-Nitroaniline
Concen: 22.457 ng/ul
RT: 14.622 min Scan# 1929
Delta R.T. 0.001 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40

Tgt Ion:138 Resp: 56768
Ion Ratio Lower Upper
138 100
108 12.1 12.3 18.5#
92 147.0 116.3 174.5





#52

Acenaphthene

Concen: 18.889 ng/ul

RT: 14.781 min Scan# 1956

Delta R.T. -0.005 min

Lab File: BG061856.D

Acq: 2 Jul 2024 22:40

Instrument :

BNA_G

ClientSampleId :

SICV403





Tgt Ion:153 Resp: 213029

Ion Ratio Lower Upper

153 100

152 44.1 37.2 55.8

154 92.4 75.0 112.6



#53

2,4-Dinitrophenol

Concen: 18.743 ng/ul

RT: 14.828 min Scan# 1964

Delta R.T. 0.001 min

Lab File: BG061856.D

Acq: 2 Jul 2024 22:40




Tgt Ion:184 Resp: 25430


Ion Ratio Lower Upper

184 100

63 102.3 83.6 125.4

154 73.9 66.8 100.2









#55
4-Nitrophenol
Concen: 19.487 ng/ul
RT: 14.928 min Scan# 1981
Delta R.T. 0.001 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40




Instrument :
BNA_G
ClientSampleId :
SICV403


Tgt Ion:109 Resp: 50550
Ion Ratio Lower Upper
109 100
139 101.7 67.4 101.2#
65 130.7 102.1 153.1



#56
Dibenzofuran
Concen: 19.525 ng/ul
RT: 15.116 min Scan# 2013
Delta R.T. 0.001 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40

Tgt Ion:168 Resp: 311224
Ion Ratio Lower Upper
168 100
139 38.4 31.0 46.4









#57
2,4-Dinitrotoluene
Concen: 21.172 ng/ul
RT: 15.075 min Scan# 2006
Delta R.T. -0.005 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40




Instrument :
BNA_G
ClientSampleId :
SICV403


Tgt Ion:165 Resp: 80632
Ion Ratio Lower Upper
165 100
63 55.5 44.9 67.3
182 1.9 2.6 3.8#



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.193 ng/ul
RT: 15.339 min Scan# 2051
Delta R.T. 0.001 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40

Tgt Ion:232 Resp: 69163
Ion Ratio Lower Upper
232 100
131 49.6 35.8 53.8
130 1.8 2.0 3.0#
166 34.6 28.0 42.0









#59
Diethylphthalate
Concen: 18.367 ng/ul
RT: 15.527 min Scan# 2083
Delta R.T. 0.001 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40




Instrument :
BNA_G
ClientSampleId :
SICV403


Tgt Ion:149 Resp: 270940
Ion Ratio Lower Upper
149 100
177 22.4 16.3 24.5
150 12.8 9.6 14.4



#61
Fluorene
Concen: 18.937 ng/ul
RT: 15.762 min Scan# 2123
Delta R.T. -0.005 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40

Tgt Ion:166 Resp: 256712
Ion Ratio Lower Upper
166 100
165 96.2 78.4 117.6
167 11.5 10.4 15.6









#62
4-Chlorophenyl-phenylether
Concen: 18.800 ng/ul
RT: 15.750 min Scan# 2122
Delta R.T. -0.005 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40




Instrument :
BNA_G
ClientSampleId :
SICV403


Tgt Ion:204 Resp: 128011
Ion Ratio Lower Upper
204 100
206 28.4 24.5 36.7
141 61.9 46.6 69.8



#63
4-Nitroaniline
Concen: 22.865 ng/ul
RT: 15.780 min Scan# 2126
Delta R.T. 0.001 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40

Tgt Ion:138 Resp: 58648
Ion Ratio Lower Upper
138 100
92 66.4 52.3 78.5
108 122.6 94.1 141.1









#66
4,6-Dinitro-2-methylphenol
Concen: 20.452 ng/ul
RT: 15.833 min Scan# 2136
Delta R.T. -0.005 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40




Instrument :
BNA_G
ClientSampleId :
SICV403


Tgt Ion:198 Resp: 47542
Ion Ratio Lower Upper
198 100
182 4.3 3.6 5.4
77 31.9 21.8 32.8



#67
N-Nitrosodiphenylamine
Concen: 19.717 ng/ul
RT: 15.968 min Scan# 2158
Delta R.T. 0.001 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40

Tgt Ion:169 Resp: 221453
Ion Ratio Lower Upper
169 100
168 70.5 55.6 83.4
167 41.4 29.8 44.8









#68
4-Bromophenyl-phenylether
Concen: 18.937 ng/ul
RT: 16.644 min Scan# 21
Delta R.T. -0.005 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40




Instrument :
BNA_G
ClientSampleId :
SICV403


Tgt Ion	Ratio	Lower	Upper
248	100		
250	102.3	77.2	115.8
141	73.0	57.8	86.8



#69
Hexachlorobenzene
Concen: 19.250 ng/ul
RT: 16.755 min Scan# 2292
Delta R.T. -0.005 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40

Tgt Ion	Ratio	Lower	Upper
284	100		
142	36.4	26.4	39.6
249	31.5	27.2	40.8









#70
Atrazine
Concen: 17.216 ng/ul
RT: 16.908 min Scan# 21
Delta R.T. -0.005 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40




Instrument :
BNA_G
ClientSampleId :
SICV403


Tgt Ion:200 Resp: 76331
Ion Ratio Lower Upper
200 100
173 25.2 18.7 28.1
215 46.3 37.8 56.6



#71
Pentachlorophenol
Concen: 18.520 ng/ul
RT: 17.102 min Scan# 2351
Delta R.T. -0.005 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40

Tgt Ion:266 Resp: 54615
Ion Ratio Lower Upper
266 100
264 59.2 45.0 67.6
268 58.2 50.5 75.7





#72

Phenanthrene

Concen: 19.104 ng/ul

RT: 17.501 min Scan# 24

Delta R.T. -0.005 min

Lab File: BG061856.D

Acq: 2 Jul 2024 22:40

Instrument :

BNA_G

ClientSampleId :

SICV403





Tgt Ion:178 Resp: 414287

Ion Ratio Lower Upper

178 100

179 15.7 11.9 17.9

176 19.1 15.5 23.3



#74

Anthracene

Concen: 18.854 ng/ul

RT: 17.595 min Scan# 2435

Delta R.T. 0.001 min

Lab File: BG061856.D

Acq: 2 Jul 2024 22:40




Tgt Ion:178 Resp: 422530


Ion Ratio Lower Upper

178 100

179 14.7 11.9 17.9

176 19.0 14.7 22.1





#75

1,2,3,4-Tetrachlorobenzene

Concen: 20.224 ng/ul

RT: 13.524 min Scan# 1

Delta R.T. 0.001 min

Lab File: BG061856.D

Acq: 2 Jul 2024 22:40

Instrument :

BNA_G

ClientSampleId :

SICV403





Tgt Ion:216 Resp: 103305

Ion Ratio Lower Upper

216 100

214 79.8 61.7 92.5

218 48.3 35.7 53.5



#76

Pentachlorobenzene

Concen: 19.449 ng/ul

RT: 15.034 min Scan# 1999

Delta R.T. 0.001 min

Lab File: BG061856.D

Acq: 2 Jul 2024 22:40




Tgt Ion:250 Resp: 113309


Ion Ratio Lower Upper

250 100

248 62.3 47.5 71.3

252 65.6 50.0 75.0









#77
 Carbazole
 Concen: 20.183 ng/ul
 RT: 17.860 min Scan# 2481
 Delta R.T. -0.005 min
 Lab File: BG061856.D
 Acq: 2 Jul 2024 22:40




Instrument :
 BNA_G
 ClientSampleId :
 SICV403


Tgt Ion:167 Resp: 380661
 Ion Ratio Lower Upper
 167 100
 166 22.8 18.2 27.2
 139 13.0 9.6 14.4



#78
 Di-n-butylphthalate
 Concen: 18.947 ng/ul
 RT: 18.412 min Scan# 2574
 Delta R.T. 0.001 min
 Lab File: BG061856.D
 Acq: 2 Jul 2024 22:40

Tgt Ion:149 Resp: 458797
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.5 11.3
 104 6.2 5.0 7.4









#80
Fluoranthene
Concen: 18.750 ng/ul
RT: 19.505 min Scan# 21
Delta R.T. 0.001 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40




Instrument :
BNA_G
ClientSampleId :
SICV403


Tgt Ion:202 Resp: 467104
Ion Ratio Lower Upper
202 100
101 8.4 7.4 11.2
100 7.6 6.6 10.0



#82
Pyrene
Concen: 19.513 ng/ul
RT: 19.863 min Scan# 2821
Delta R.T. -0.005 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40

Tgt Ion:202 Resp: 502161
Ion Ratio Lower Upper
202 100
101 10.3 8.8 13.2
100 9.3 7.6 11.4





#83

Butylbenzylphthalate

Concen: 18.944 ng/ul

RT: 20.745 min Scan# 20

Delta R.T. 0.001 min

Lab File: BG061856.D

Acq: 2 Jul 2024 22:40

Instrument :

BNA_G

ClientSampleId :

SICV403





Tgt Ion:149 Resp: 204597

Ion Ratio Lower Upper

149 100

91 84.9 60.6 91.0

206 19.7 14.6 21.8



#84

3,3'-Dichlorobenzidine

Concen: 22.934 ng/ul

RT: 21.614 min Scan# 3119

Delta R.T. -0.005 min

Lab File: BG061856.D

Acq: 2 Jul 2024 22:40




Tgt Ion:252 Resp: 195270


Ion Ratio Lower Upper

252 100

254 58.7 52.8 79.2

126 8.7 6.6 10.0









#85
Benzo(a)anthracene
Concen: 19.120 ng/ul
RT: 21.702 min Scan# 3134
Delta R.T. 0.001 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40




Instrument :
BNA_G
ClientSampleId :
SICV403


Tgt Ion:228 Resp: 495444
Ion Ratio Lower Upper
228 100
229 20.0 15.7 23.5
226 26.0 21.7 32.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 18.583 ng/ul
RT: 21.602 min Scan# 3117
Delta R.T. 0.001 min
Lab File: BG061856.D
Acq: 2 Jul 2024 22:40

Tgt Ion:149 Resp: 287991
Ion Ratio Lower Upper
149 100
167 27.8 21.7 32.5
279 3.4 2.8 4.2





#87

Chrysene

Concen: 19.130 ng/ul

RT: 21.767 min Scan# 3146

Delta R.T. -0.005 min

Lab File: BG061856.D

Acq: 2 Jul 2024 22:40

Instrument :

BNA_G

ClientSampleId :

SICV403





Tgt Ion:228 Resp: 464057

Ion Ratio Lower Upper

228 100

226 30.2 23.3 34.9

229 19.3 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 19.991 ng/ul




RT: 22.824 min Scan# 3325


Delta R.T. 0.001 min

Lab File: BG061856.D

Acq: 2 Jul 2024 22:40

Tgt Ion:149 Resp: 507822





#90

Benzo(b)fluoranthene

Concen: 18.955 ng/ul

RT: 23.929 min Scan# 3513

Delta R.T. -0.005 min

Lab File: BG061856.D

Acq: 2 Jul 2024 22:40

Instrument :

BNA_G

ClientSampleId :

SICV403





Tgt Ion:252 Resp: 496563

Ion Ratio Lower Upper

252 100

253 21.9 17.2 25.8

125 8.1 6.1 9.1



#91

Benzo(k)fluoranthene

Concen: 19.379 ng/ul

RT: 24.000 min Scan# 3525

Delta R.T. -0.005 min

Lab File: BG061856.D

Acq: 2 Jul 2024 22:40




Tgt Ion:252 Resp: 506249


Ion Ratio Lower Upper

252 100

253 21.2 19.3 28.9

125 6.7 6.3 9.5





#93

Benzo(a)pyrene

Concen: 19.097 ng/ul

RT: 24.810 min Scan# 3

Delta R.T. -0.005 min

Lab File: BG061856.D

Acq: 2 Jul 2024 22:40

Instrument :

BNA_G

ClientSampleId :

SICV403





Tgt Ion:252 Resp: 473745

Ion Ratio Lower Upper

252 100

253 21.7 16.5 24.7

125 8.5 7.4 11.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.393 ng/ul

RT: 28.659 min Scan# 4318

Delta R.T. -0.005 min

Lab File: BG061856.D

Acq: 2 Jul 2024 22:40




Tgt Ion:276 Resp: 614776


Ion Ratio Lower Upper

276 100

138 11.4 11.9 17.9#

277 21.3 19.3 28.9





#95

Dibenzo(a,h)anthracene

Concen: 19.455 ng/ul

RT: 28.729 min Scan# 41

Delta R.T. -0.005 min

Lab File: BG061856.D

Acq: 2 Jul 2024 22:40

Instrument :

BNA_G

ClientSampleId :

SICV403





Tgt Ion:278 Resp: 509378

Ion Ratio Lower Upper

278 100

139 9.9 7.6 11.4

279 23.4 18.8 28.2



#96

Benzo(g,h,i)perylene

Concen: 18.685 ng/ul

RT: 29.822 min Scan# 4516

Delta R.T. -0.005 min

Lab File: BG061856.D

Acq: 2 Jul 2024 22:40

Tgt Ion:276 Resp: 471230

Ion Ratio Lower Upper

276 100

138 16.1 11.1 16.7

277 21.3 18.8 28.2

