


Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG070319\
 Data File : BG041590.D
 Acq On : 3 Jul 2019 14:38
 Operator : HP/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG062619.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jun 27 13:25:55 2019



AutoFind: Scans 2506, 2507, 2508; Background Corrected with Scan 2499

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	29.6	31277	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	31.5	33256	PASS
70	69	0.00	2	0.5	159	PASS
127	198	10	80	39.1	41269	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	105674	PASS
199	198	5	9	6.9	7334	PASS
275	198	10	60	31.4	33141	PASS
365	198	1	100	3.8	4009	PASS
441	198	0.01	100	10.0	10534	PASS
442	198	50	100	61.5	65040	PASS
443	442	15	24	19.9	12913	PASS