


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG070615\
Data File : BG017775.D
Acq On : 6 Jul 2015 18:22
Operator : TP/UM
Sample : DFTPP08
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP08

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG070615.M
Title : SVOA CALIBRATION
Last Update : Tue Jul 07 01:50:10 2015



AutoFind: Scans 2514, 2515, 2516; Background Corrected with Scan 2508

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	54.2	25922	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	54.7	26130	PASS
70	69	0.00	2	0.6	156	PASS
127	198	10	80	61.4	29365	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	47805	PASS
199	198	5	9	6.5	3110	PASS
275	198	10	60	24.9	11881	PASS
365	198	1	100	3.9	1885	PASS
441	442	0.01	24	15.0	5160	PASS
442	198	50	100	72.2	34512	PASS
443	442	15	24	19.8	6827	PASS

m/z	Abundance
35.90	319.0
38.00	626.0
39.00	2437.0
40.00	1403.0
43.90	502.0
49.00	522.0
50.00	7341.0
51.00	26648.0
52.00	1255.0
55.90	760.0
57.00	1555.0
57.90	153.0
60.20	198.0
61.00	316.0
61.90	430.0
62.90	915.0
64.10	155.0
64.30	154.0
65.10	825.0
69.00	28640.0
69.90	241.0
73.00	474.0
74.00	2887.0
75.00	4722.0
76.10	1620.0
77.00	29768.0
78.00	1803.0
78.90	2210.0
79.90	1773.0
80.90	2067.0
81.80	642.0
83.00	439.0
83.90	222.0
85.00	277.0
85.90	671.0
87.00	181.0
88.00	188.0
90.90	618.0
92.00	609.0
93.00	3169.0
94.00	295.0
98.00	2630.0
98.90	3151.0
100.00	190.0
101.00	982.0
103.00	487.0
104.00	1106.0
104.90	586.0
106.90	8716.0
107.90	1148.0
109.90	13840.0
111.00	2558.0
112.00	473.0
115.90	429.0
116.90	7747.0
118.00	709.0
119.80	154.0
121.80	626.0
123.00	1091.0
123.80	392.0
124.90	319.0
127.00	28368.0
127.90	2295.0
128.90	11044.0
129.90	1356.0
130.90	240.0
133.10	169.0
133.90	344.0
135.00	1146.0
135.90	276.0
136.90	778.0
137.60	184.0
140.00	153.0
140.90	1473.0
142.00	505.0
142.90	282.0
143.90	170.0
144.90	161.0
146.00	513.0
146.90	1077.0
147.90	2507.0
149.00	457.0
150.90	249.0
153.00	465.0
153.90	249.0
154.90	640.0
155.90	994.0
156.90	441.0
158.00	345.0
159.00	184.0
159.90	302.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

161.00	640.0
161.90	341.0
164.80	458.0
166.00	722.0
166.90	3346.0
167.90	1880.0
168.90	336.0
170.80	197.0
172.10	155.0
173.00	362.0
173.90	463.0
174.90	890.0
175.90	423.0
176.90	531.0
178.90	2170.0
179.90	1155.0
180.90	758.0
184.90	1055.0
186.00	6494.0
187.00	2013.0
187.90	273.0
188.80	352.0
190.70	333.0
191.90	672.0
193.00	594.0
194.80	296.0
195.90	1874.0
197.90	46800.0
198.90	2719.0
199.90	511.0
202.80	614.0
203.90	1937.0
204.90	3439.0
206.00	12520.0
207.00	2136.0
208.00	864.0
210.00	422.0
210.70	419.0
210.90	423.0
211.20	377.0
211.90	167.0
212.80	151.0
214.80	282.0
216.00	411.0
216.90	3478.0
217.90	440.0
219.80	226.0
221.00	2499.0
221.90	608.0
222.90	653.0
224.00	7272.0
224.90	1655.0
225.90	505.0
226.90	3255.0
227.90	375.0
228.90	613.0
229.80	178.0
230.80	171.0
233.80	187.0
235.90	168.0
236.90	326.0
238.80	174.0
240.80	283.0
241.90	459.0
243.00	484.0
244.00	4227.0
245.10	641.0
245.90	1550.0
247.00	520.0
251.90	154.0
252.90	294.0
253.90	338.0
254.90	26736.0
255.90	3495.0
257.00	493.0
257.90	2253.0
259.00	367.0
264.90	1023.0
265.80	172.0
273.00	736.0
273.90	2099.0
275.00	10046.0
276.00	1702.0
276.90	1008.0
277.90	236.0
280.80	151.0
282.20	161.0
293.00	217.0
294.10	229.0
295.90	3638.0
296.90	568.0
302.00	218.0
302.60	387.0
303.00	386.0
315.00	462.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

322.90	869.0
324.00	295.0
326.80	330.0
327.80	237.0
332.90	213.0
334.00	694.0
346.00	185.0
347.00	163.0
351.90	556.0
353.00	232.0
354.00	314.0
364.90	1071.0
365.70	416.0
370.90	186.0
371.80	571.0
372.90	161.0
382.90	168.0
384.00	151.0
396.90	156.0
402.00	169.0
421.00	209.0
422.00	198.0
423.00	1514.0
423.80	202.0
424.90	254.0
438.80	172.0
441.00	3561.0
442.00	26056.0
443.00	5114.0
443.90	765.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

m/z	Abundance
37.10	181.0
37.90	488.0
39.00	1806.0
40.00	1047.0
40.90	168.0
43.90	391.0
49.00	726.0
50.00	8925.0
51.00	31952.0
52.00	1505.0
53.10	222.0
54.90	180.0
55.90	1011.0
57.00	2047.0
60.90	379.0
62.00	539.0
63.00	1388.0
63.90	161.0
65.10	320.0
69.00	30024.0
69.80	228.0
70.90	200.0
72.90	482.0
74.00	3621.0
75.00	5269.0
76.00	2196.0
77.00	36816.0
78.00	1948.0
79.00	2724.0
80.00	2248.0
81.00	2474.0
81.90	857.0
82.90	597.0
83.90	737.0
84.90	346.0
86.00	858.0
87.00	372.0
91.00	704.0
92.00	716.0
92.90	4408.0
94.10	263.0
95.50	210.0
96.10	520.0
98.00	3784.0
98.90	2645.0
100.10	366.0
101.00	1558.0
102.90	423.0
103.90	653.0
105.00	1027.0
105.90	254.0
106.90	10363.0
108.00	1732.0
109.00	516.0
110.00	17008.0
110.90	3331.0
112.00	301.0
115.80	578.0
117.00	10496.0
117.90	543.0
122.00	696.0
122.90	1077.0
123.90	413.0
125.00	809.0
127.00	35320.0
127.90	2444.0
128.90	14594.0
130.00	1106.0
131.90	184.0
133.00	227.0
133.90	480.0
135.00	1155.0
135.90	354.0
137.00	833.0
140.80	1696.0
142.10	731.0
143.00	355.0
143.80	196.0
146.00	258.0
146.90	997.0
148.00	2873.0
148.90	395.0
150.80	306.0
151.90	249.0
153.00	679.0
153.90	446.0
154.90	969.0
156.00	1147.0
156.90	168.0
157.90	450.0
159.00	539.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

159.90	573.0
160.90	777.0
161.90	201.0
165.00	698.0
165.80	613.0
166.90	3587.0
168.00	2567.0
168.90	265.0
170.90	224.0
171.80	414.0
172.90	389.0
174.00	768.0
175.00	887.0
176.00	583.0
176.90	849.0
178.00	315.0
178.90	2735.0
179.90	1431.0
181.10	708.0
184.00	272.0
185.00	1124.0
186.00	9222.0
187.00	2793.0
187.90	332.0
188.90	827.0
189.80	180.0
190.80	355.0
192.00	1214.0
193.00	796.0
193.70	159.0
196.00	2386.0
197.90	56416.0
198.90	3496.0
200.00	159.0
201.00	340.0
201.40	522.0
203.00	709.0
204.00	2880.0
205.00	4385.0
206.00	17184.0
207.00	3351.0
208.00	1023.0
208.90	331.0
209.80	280.0
210.20	237.0
210.90	550.0
211.90	190.0
215.00	202.0
215.90	562.0
216.90	4854.0
217.80	196.0
219.90	182.0
221.00	2747.0
222.90	1105.0
224.00	10222.0
224.90	1951.0
226.00	456.0
226.90	5099.0
227.90	576.0
228.90	1101.0
229.70	257.0
230.80	195.0
233.80	324.0
235.80	342.0
236.80	357.0
240.80	247.0
242.00	686.0
243.00	364.0
244.00	7485.0
244.90	925.0
245.90	1596.0
246.90	262.0
248.80	229.0
250.80	182.0
252.20	159.0
253.20	308.0
254.90	36256.0
256.00	5523.0
257.00	554.0
257.90	2224.0
259.00	681.0
264.90	1638.0
265.90	319.0
270.10	223.0
272.10	220.0
272.90	1411.0
274.00	2890.0
274.90	14514.0
275.90	1800.0
276.80	1288.0
278.00	218.0
283.00	198.0
284.00	307.0
285.10	186.0
288.90	168.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

292.90	375.0
294.00	242.0
294.90	202.0
296.00	5345.0
296.90	539.0
302.90	608.0
314.00	355.0
314.90	529.0
316.00	185.0
322.90	1406.0
327.00	347.0
328.00	286.0
333.10	287.0
334.00	1105.0
335.00	210.0
341.00	303.0
345.80	330.0
352.00	766.0
352.90	525.0
353.90	448.0
365.00	2707.0
365.90	556.0
372.00	702.0
382.90	150.0
390.00	281.0
401.90	307.0
403.00	604.0
420.90	385.0
421.80	294.0
423.00	2612.0
424.00	410.0
434.40	176.0
441.00	6288.0
442.00	40136.0
443.00	7717.0
444.00	587.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

m/z	Abundance
38.00	412.0
39.00	1541.0
40.00	1205.0
44.00	566.0
46.90	155.0
48.90	438.0
50.00	6157.0
51.00	19168.0
52.10	1390.0
56.00	636.0
57.00	1265.0
59.20	163.0
60.90	182.0
62.00	446.0
63.00	489.0
64.10	180.0
65.10	372.0
69.00	19728.0
70.10	196.0
73.00	441.0
74.00	2625.0
75.00	3129.0
76.00	1133.0
77.00	24504.0
77.90	1277.0
79.00	1171.0
80.00	1200.0
81.00	1652.0
82.00	461.0
83.10	405.0
83.90	630.0
84.90	246.0
85.90	841.0
87.90	198.0
90.90	400.0
92.00	496.0
92.90	2411.0
94.00	159.0
98.00	1885.0
99.00	2135.0
99.90	205.0
100.80	683.0
101.80	251.0
103.00	460.0
103.90	612.0
105.00	604.0
107.00	7041.0
107.90	752.0
108.90	320.0
109.90	11458.0
110.90	1875.0
115.70	501.0
117.00	7568.0
118.00	431.0
119.00	250.0
121.90	544.0
123.00	686.0
124.00	380.0
125.00	615.0
127.00	24408.0
127.90	1692.0
128.90	10341.0
129.90	783.0
133.80	356.0
134.90	1193.0
135.90	551.0
137.00	595.0
137.80	196.0
139.90	309.0
140.80	1347.0
141.80	320.0
143.10	668.0
144.90	161.0
145.90	184.0
147.00	530.0
147.90	1227.0
148.90	275.0
151.00	235.0
153.00	449.0
154.00	398.0
154.90	491.0
156.00	1258.0
158.00	231.0
158.70	225.0
159.80	490.0
161.00	441.0
161.80	157.0
163.90	340.0
165.00	582.0
165.90	527.0
167.00	2441.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

168.00	1212.0
169.10	291.0
172.00	251.0
173.00	306.0
174.00	538.0
175.00	1100.0
175.80	296.0
176.80	492.0
178.80	1773.0
180.00	815.0
180.90	968.0
184.20	222.0
185.00	888.0
186.00	6708.0
187.00	1692.0
189.00	667.0
190.90	559.0
191.90	745.0
192.90	633.0
194.00	335.0
195.80	1118.0
197.90	40200.0
198.90	3115.0
199.80	239.0
202.90	241.0
203.90	1644.0
204.90	2805.0
206.00	12403.0
207.00	1973.0
207.90	571.0
208.90	382.0
209.80	324.0
210.40	253.0
211.10	466.0
215.10	273.0
215.90	290.0
216.90	3884.0
218.00	496.0
220.90	1993.0
221.50	726.0
223.00	770.0
224.00	6992.0
225.00	2009.0
225.80	157.0
226.90	3740.0
228.00	563.0
228.90	508.0
229.80	191.0
230.90	300.0
233.90	290.0
234.90	289.0
237.00	206.0
238.70	153.0
239.90	196.0
240.70	195.0
242.00	668.0
242.90	492.0
244.00	6170.0
245.00	713.0
245.90	1039.0
246.80	302.0
248.90	152.0
252.80	162.0
254.90	26104.0
256.00	4089.0
257.00	299.0
257.90	1923.0
258.90	332.0
264.90	828.0
272.20	185.0
272.90	1346.0
274.00	2400.0
275.00	11085.0
276.00	1202.0
276.90	1384.0
277.90	188.0
281.90	193.0
283.00	184.0
284.70	270.0
294.90	385.0
295.90	5312.0
296.90	971.0
303.00	464.0
303.80	161.0
314.00	240.0
314.90	613.0
321.90	250.0
323.00	1080.0
326.80	475.0
331.90	193.0
332.90	285.0
333.90	648.0
341.00	170.0
345.80	240.0
352.00	487.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

353.00	298.0
353.90	321.0
365.00	1879.0
365.90	373.0
372.00	853.0
372.80	348.0
402.10	247.0
402.90	263.0
421.00	296.0
422.00	199.0
423.00	1957.0
423.80	374.0
439.50	300.0
441.00	5633.0
442.00	37344.0
443.00	7650.0
443.90	663.0
459.00	181.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08

Scan 2508 (17.426 min): BG017775.D
DFTPP08

m/z	Abundance
39.90	1539.0
44.00	386.0
48.40	192.0
49.00	514.0
70.30	269.0
84.00	421.0
85.80	184.0
112.00	159.0
183.40	194.0
206.90	656.0
227.90	237.0
268.00	174.0
280.90	275.0
293.10	158.0

Instrument :
BNA_G
ClientSampleId :
DFTPP08