

Data Path : Z:\HPCHEM1\BNA G\DATA\BG070816\
 Data File : BG022918.D
 Acq On : 8 Jul 2016 6:03
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 08 07:12:03 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG070816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 07 16:51:07 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	100	0.00
2	1,4-Dioxane	40.000	40.284	-0.7	105	0.00
3	Pyridine	40.000	41.322	-3.3	104	0.00
4	n-Nitrosodimethylamine	40.000	41.261	-3.2	102	0.00
5 S	2-Fluorophenol	80.000	81.068	-1.3	104	0.00
6	Aniline	40.000	41.387	-3.5	103	0.00
7 S	Phenol-d6	80.000	82.754	-3.4	102	0.00
8	2-Chlorophenol	40.000	41.389	-3.5	104	0.00
9	Benzaldehyde	40.000	39.529	1.2	101	0.00
10 C	Phenol	40.000	41.305	-3.3	103	0.00
11	bis(2-Chloroethyl)ether	40.000	39.512	1.2	100	0.00
12	1,3-Dichlorobenzene	40.000	39.876	0.3	103	0.00
13 C	1,4-Dichlorobenzene	40.000	41.649	-4.1	106	0.00
14	1,2-Dichlorobenzene	40.000	39.696	0.8	104	0.00
15	Benzyl Alcohol	40.000	42.277	-5.7	103	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.188	-0.5	101	0.00
17	2-Methylphenol	40.000	40.485	-1.2	101	0.00
18	Hexachloroethane	40.000	39.713	0.7	107	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.045	-0.1	98	0.00
20	3+4-Methylphenols	40.000	41.948	-4.9	99	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	98	0.00
22	Acetophenone	40.000	39.217	2.0	97	0.00
23 S	Nitrobenzene-d5	80.000	79.556	0.6	99	0.00
24	Nitrobenzene	40.000	39.950	0.1	101	0.00
25	Isophorone	40.000	39.034	2.4	94	0.00
26 C	2-Nitrophenol	40.000	39.816	0.5	93	0.00
27	2,4-Dimethylphenol	40.000	39.568	1.1	98	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.484	1.3	96	0.00
29 C	2,4-Dichlorophenol	40.000	41.024	-2.6	99	0.00
30	1,2,4-Trichlorobenzene	40.000	40.093	-0.2	100	0.00
31	Naphthalene	40.000	40.299	-0.7	101	0.00
32	Benzoic acid	40.000	39.366	1.6	92	0.01
33	4-Chloroaniline	40.000	40.285	-0.7	99	0.00
34 C	Hexachlorobutadiene	40.000	40.576	-1.4	103	0.00
35	Caprolactam	40.000	35.058	12.4	86	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.234	1.9	95	0.00
37	2-Methylnaphthalene	40.000	40.313	-0.8	97	0.00
38 I	Acenaphthene-d10	20.000	20.000	0.0	91	0.00
39	1,2,4,5-Tetrachlorobenzene	40.000	40.828	-2.1	95	0.00
40 P	Hexachlorocyclopentadiene	40.000	40.924	-2.3	95	0.00
41 S	2,4,6-Tribromophenol	80.000	74.527	6.8	88	0.00
42 C	2,4,6-Trichlorophenol	40.000	40.462	-1.2	92	0.00
43	2,4,5-Trichlorophenol	40.000	39.970	0.1	94	0.00
44 S	2-Fluorobiphenyl	80.000	80.782	-1.0	96	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG070816\
 Data File : BG022918.D
 Acq On : 8 Jul 2016 6:03
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Jul 08 07:12:03 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG070816.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 07 16:51:07 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	40.000	41.229	-3.1	95	0.00
46	2-Chloronaphthalene	40.000	40.628	-1.6	94	0.00
47	2-Nitroaniline	40.000	40.446	-1.1	94	0.00
48	Acenaphthylene	40.000	39.852	0.4	93	0.00
49	Dimethylphthalate	40.000	38.259	4.4	90	0.00
50	2,6-Dinitrotoluene	40.000	38.700	3.2	90	0.00
51 C	Acenaphthene	40.000	45.434	-13.6	106	0.00
52	3-Nitroaniline	40.000	41.004	-2.5	94	0.00
53 P	2,4-Dinitrophenol	40.000	35.798	10.5	80	0.00
54	Dibenzofuran	40.000	39.292	1.8	93	0.00
55 P	4-Nitrophenol	40.000	37.221	6.9	86	0.00
56	2,4-Dinitrotoluene	40.000	38.852	2.9	91	0.00
57	Fluorene	40.000	39.163	2.1	93	0.00
58	2,3,4,6-Tetrachlorophenol	40.000	38.957	2.6	90	0.00
59	Diethylphthalate	40.000	37.809	5.5	89	0.00
60	4-Chlorophenyl-phenylether	40.000	38.545	3.6	89	0.00
61	4-Nitroaniline	40.000	37.788	5.5	89	0.00
62	Azobenzene	40.000	39.112	2.2	92	0.00
63 I	Phenanthrene-d10	20.000	20.000	0.0	90	0.00
64	4,6-Dinitro-2-methylphenol	40.000	39.232	1.9	84	0.00
65 c	n-Nitrosodiphenylamine	40.000	40.777	-1.9	90	0.00
66	4-Bromophenyl-phenylether	40.000	39.199	2.0	87	0.00
67	Hexachlorobenzene	40.000	40.411	-1.0	90	0.00
68	Atrazine	40.000	39.384	1.5	89	0.00
69 C	Pentachlorophenol	40.000	38.429	3.9	81	0.00
70	Phenanthrene	40.000	40.015	-0.0	91	0.00
71	Anthracene	40.000	40.224	-0.6	92	0.00
72	Carbazole	40.000	39.856	0.4	91	0.00
73	Di-n-butylphthalate	40.000	41.377	-3.4	91	0.00
74 C	Fluoranthene	40.000	38.261	4.3	90	0.00
75 I	Chrysene-d12	20.000	20.000	0.0	89	0.00
76	Benzidine	40.000	36.201	9.5	79	0.00
77	Pyrene	40.000	38.404	4.0	91	0.00
78 S	Terphenyl-d14	80.000	84.060	-5.1	93	0.00
79	Butylbenzylphthalate	40.000	39.695	0.8	89	0.00
80	Benzo(a)anthracene	40.000	39.950	0.1	92	0.00
81	3,3'-Dichlorobenzidine	40.000	39.773	0.6	88	0.00
82	Chrysene	40.000	39.965	0.1	92	0.00
83	Bis(2-ethylhexyl)phthalate	40.000	39.396	1.5	90	0.00
84 c	Di-n-octyl phthalate	40.000	39.979	0.1	90	0.00
85	Indeno(1,2,3-cd)pyrene	40.000	39.110	2.2	90	0.02
86 I	Perylene-d12	20.000	20.000	0.0	88	0.01
87	Benzo(b)fluoranthene	40.000	40.234	-0.6	91	0.01

Data Path : Z:\HPCHEM1\BNA G\DATA\BG070816\
Data File : BG022918.D
Acq On : 8 Jul 2016 6:03
Operator : UM/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Jul 08 07:12:03 2016
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG070816.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 07 16:51:07 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	40.000	41.235	-3.1	90	0.01
89 C	Benzo(a)pyrene	40.000	39.952	0.1	89	0.00
90	Dibenzo(a,h)anthracene	40.000	40.220	-0.5	90	0.03
91	Benzo(a,h,i)perylene	40.000	40.520	-1.3	90	0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0