


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG070821\
 Data File : BG049212.D
 Acq On : 8 Jul 2021 9:07
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG063021.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Jul 06 10:59:02 2021






AutoFind: Scans 2530, 2531, 2532; Background Corrected with Scan 2523

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.7	21307	PASS
68	69	0.00	2	1.3	266	PASS
69	198	0.00	100	38.1	21006	PASS
70	69	0.00	2	0.6	124	PASS
127	198	10	80	45.9	25275	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	55120	PASS
199	198	5	9	6.2	3443	PASS
275	198	10	60	33.7	18589	PASS
365	198	1	100	5.4	2963	PASS
441	198	0.01	100	17.0	9384	PASS
442	442	100	100	100.0	57651	PASS
443	442	15	24	19.2	11051	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG070821\
Data File : BG049212.D
Acq On : 8 Jul 2021 9:07
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Jul 08 11:55:04 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG063021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 06 10:59:02 2021
Response via : Initial Calibration



TIC: BG049212.D\data.ms

(77) Benzidine

19.706min (-0.000) 0.00 ng




response 355352

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.40	14.68
183.00	11.70	11.86
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG070821\
Data File : BG049212.D
Acq On : 8 Jul 2021 9:07
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: Jul 08 11:55:04 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_G\METHODS\8270-BG063021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 06 10:59:02 2021
Response via : Initial Calibration



TIC: BG049212.D\data.ms

(70) Pentachlorophenol (C)

17.127min (-0.000) 28914.57 ng

response 67525

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.20	58.50
264.00	61.30	59.69
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
8/7/2021	BNA_G	BG049212.D
Compound Name	Response	Retention Time
DDT	192562	20.94
DDD	5022	20.49
DDE	799	19.99
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
5821	198383	2.93